

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112423\
 Data File : BM042979.D
 Acq On : 24 Nov 2023 14:51
 Operator : MA/JU
 Sample : SSTD3.277
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2027

Quant Time: Nov 24 15:20:43 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 24 14:48:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.830	152	754	0.400	ng/ul	#-0.05
4) Naphthalene-d8	10.616	136	2367	0.400	ng/ul	#-0.03
9) Acenaphthene-d10	14.459	164	1380	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.216	188	2954	0.400	ng/ul	-0.02
17) Chrysene-d12	21.406	240	1831	0.400	ng/ul	-0.02
23) Perylene-d12	23.773	264	2103	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	3023	2.498	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.197	88	3142	2.403	ng/ul#	78
14) Pentachlorophenol	16.865	266	3049	4.005	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112423\
Data File : BM042979.D
Acq On : 24 Nov 2023 14:51
Operator : MA/JU
Sample : SST3.277
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2027

Quant Time: Nov 24 15:20:43 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 24 14:48:49 2023
Response via : Initial Calibration

