



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM112423\
Data File : BM042973.D
Acq On : 24 Nov 2023 11:14
Operator : MA/JU
Sample : DFTPP183
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP183

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Nov 27 02:36:19 2023



AutoFind: Scans 2480, 2481, 2482; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.3	28231	PASS
68	69	0.00	2	1.8	669	PASS
69	198	0.00	100	52.3	37565	PASS
70	69	0.00	2	0.8	314	PASS
127	198	10	80	53.6	38525	PASS
197	198	0.00	2	0.7	497	PASS
198	198	100	100	100.0	71891	PASS
199	198	5	9	7.6	5455	PASS
275	198	10	60	28.6	20560	PASS
365	198	1	100	6.0	4289	PASS
441	198	0.01	100	15.3	10971	PASS
442	198	50	100	93.0	66832	PASS
443	442	15	24	19.5	13064	PASS

m/z	Abundance
73.00	227.0
193.00	156.0
207.00	467.0
280.80	204.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

m/z	Abundance
37.10	279.0
38.00	558.0
39.10	2752.0
40.00	294.0
44.00	328.0
49.00	402.0
50.00	10985.0
51.10	35984.0
52.00	1944.0
55.00	263.0
56.00	914.0
57.00	2287.0
58.00	168.0
61.00	371.0
62.00	782.0
63.00	1505.0
63.90	279.0
65.00	615.0
68.00	751.0
69.00	47392.0
70.00	401.0
72.90	784.0
74.00	6770.0
75.00	9204.0
76.10	3244.0
77.10	63256.0
78.00	4322.0
79.00	3991.0
80.00	2892.0
81.00	4022.0
82.00	695.0
83.00	1009.0
84.90	447.0
86.00	1471.0
87.00	473.0
88.00	161.0
91.00	696.0
91.90	1120.0
93.00	9413.0
93.90	421.0
96.00	510.0
96.90	173.0
98.00	7513.0
99.00	4842.0
99.90	378.0
101.00	2051.0
103.10	526.0
104.00	1206.0
105.00	1474.0
106.00	470.0
107.00	16600.0
108.00	2230.0
109.00	620.0
110.00	21224.0
111.00	3915.0
112.00	704.0
112.80	169.0
116.00	748.0
117.00	27592.0
118.00	1932.0
118.90	189.0
119.90	265.0
122.00	1555.0
123.00	1861.0
124.00	1209.0
125.00	752.0
126.10	195.0
127.00	43584.0
128.00	3600.0
129.00	20728.0
130.00	1959.0
130.90	270.0
134.00	552.0
135.00	2033.0
135.90	569.0
137.00	1194.0
139.90	240.0
141.00	3002.0
141.90	699.0
143.00	641.0
143.80	179.0
144.90	178.0
146.00	664.0
146.90	1816.0
148.00	6530.0
149.00	954.0
149.90	318.0
151.00	432.0
152.90	947.0
153.90	585.0
155.00	1769.0
156.00	2402.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

156.90	319.0
158.00	379.0
159.00	537.0
159.90	1096.0
160.90	1491.0
161.90	406.0
165.00	1043.0
165.90	1010.0
167.00	10505.0
168.00	6637.0
168.90	938.0
169.90	159.0
170.90	346.0
171.90	591.0
172.90	876.0
174.00	1590.0
175.00	2211.0
176.00	591.0
177.00	914.0
177.90	335.0
178.90	5298.0
179.90	3116.0
181.00	1353.0
182.00	219.0
183.10	229.0
183.80	257.0
185.00	2138.0
186.00	15612.0
187.00	4546.0
188.00	653.0
188.90	1465.0
189.80	214.0
190.80	622.0
191.90	1818.0
192.90	1817.0
193.90	429.0
196.00	2377.0
197.00	540.0
197.90	79608.0
198.90	6180.0
199.80	526.0
200.80	164.0
201.50	352.0
202.90	992.0
204.00	5095.0
205.00	7440.0
206.00	26960.0
207.00	4579.0
208.00	1390.0
209.00	569.0
209.90	843.0
211.00	1223.0
211.90	267.0
214.80	289.0
216.00	1019.0
217.00	8835.0
218.00	1333.0
221.00	2803.0
221.80	1226.0
223.00	1785.0
224.00	17280.0
225.00	3951.0
225.90	408.0
227.00	9252.0
228.00	1575.0
228.90	1689.0
229.90	292.0
230.90	643.0
233.90	618.0
235.00	486.0
236.00	447.0
237.10	553.0
239.00	307.0
240.90	429.0
242.00	1233.0
243.00	1042.0
244.00	12676.0
245.00	1789.0
245.90	2738.0
246.90	745.0
248.90	316.0
252.90	200.0
254.10	413.0
255.00	58840.0
256.00	7932.0
257.00	1110.0
258.00	4558.0
259.00	694.0
259.90	197.0
265.00	1827.0
265.90	337.0
271.90	202.0
272.90	1841.0
274.00	4543.0
275.00	20128.0
276.00	2628.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

276.90	2495.0
277.90	577.0
280.90	275.0
283.00	369.0
284.00	301.0
285.00	306.0
286.10	159.0
291.80	176.0
293.00	531.0
293.80	282.0
294.80	251.0
296.00	10040.0
297.00	1370.0
301.00	155.0
302.00	215.0
303.00	835.0
304.00	160.0
314.00	484.0
314.90	1096.0
315.90	315.0
320.90	201.0
323.00	1895.0
324.10	409.0
326.90	486.0
327.90	272.0
331.90	189.0
332.90	188.0
334.00	1339.0
335.10	446.0
340.90	242.0
345.90	773.0
352.00	974.0
353.00	491.0
354.00	522.0
364.90	3558.0
366.00	489.0
371.10	170.0
372.10	855.0
373.00	353.0
383.00	354.0
402.00	454.0
403.00	726.0
404.10	163.0
421.00	357.0
421.80	331.0
423.00	2556.0
424.00	794.0
425.00	225.0
441.00	7920.0
442.00	51712.0
443.00	10393.0
444.10	823.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

m/z	Abundance
36.90	181.0
38.00	563.0
39.10	2305.0
39.90	186.0
44.00	246.0
48.90	436.0
50.00	10430.0
51.10	32936.0
52.00	1751.0
54.90	193.0
56.00	947.0
57.00	2102.0
60.90	346.0
62.00	609.0
63.00	1627.0
63.90	347.0
65.00	679.0
68.00	903.0
69.00	44568.0
70.00	380.0
72.90	836.0
74.10	6504.0
75.00	9242.0
76.00	3319.0
77.10	60520.0
78.10	3934.0
79.00	3687.0
80.00	3240.0
81.00	3777.0
82.10	1034.0
83.00	755.0
85.00	453.0
86.00	1382.0
86.90	336.0
87.90	154.0
91.00	828.0
92.00	869.0
93.00	8567.0
94.00	418.0
95.90	408.0
96.80	213.0
98.00	7350.0
98.90	3995.0
99.90	380.0
101.00	1906.0
103.00	485.0
103.90	1174.0
105.00	1476.0
106.00	440.0
107.00	16544.0
108.00	2169.0
108.90	406.0
110.00	21384.0
111.00	3978.0
112.00	605.0
113.00	197.0
116.00	794.0
117.00	30280.0
117.90	1922.0
118.90	157.0
119.90	203.0
122.00	994.0
123.00	2123.0
124.00	1012.0
124.90	740.0
126.10	209.0
127.00	47408.0
128.00	3431.0
129.00	22608.0
130.00	1990.0
131.10	312.0
132.90	171.0
134.00	696.0
134.90	1508.0
136.00	884.0
136.90	1149.0
137.90	293.0
139.90	303.0
141.00	2880.0
142.10	945.0
143.00	582.0
143.90	266.0
144.80	166.0
146.00	606.0
147.00	1749.0
148.00	6766.0
148.90	1187.0
150.00	191.0
151.10	418.0
151.80	155.0
152.90	1061.0
153.90	793.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

154.90	1892.0
156.00	2672.0
157.00	538.0
158.00	752.0
158.80	262.0
159.90	930.0
161.00	1529.0
161.90	464.0
164.10	152.0
164.90	1168.0
166.00	948.0
166.90	10775.0
168.00	6222.0
169.00	779.0
170.00	188.0
170.90	434.0
171.80	574.0
173.00	802.0
173.90	1220.0
175.00	2856.0
175.90	491.0
177.00	1162.0
177.80	481.0
178.90	6028.0
180.00	3034.0
181.00	1559.0
182.00	259.0
182.90	280.0
183.90	567.0
185.00	2356.0
186.00	17144.0
187.00	4974.0
188.10	442.0
188.90	1389.0
190.00	187.0
191.10	666.0
191.90	2176.0
193.00	2410.0
193.90	391.0
196.00	2586.0
197.00	704.0
197.90	88936.0
199.00	6369.0
199.90	586.0
201.40	364.0
203.00	835.0
204.00	5476.0
205.00	8939.0
206.00	30384.0
207.00	4787.0
208.00	1704.0
209.00	550.0
209.80	628.0
211.00	1368.0
211.70	231.0
212.20	164.0
214.80	421.0
216.00	772.0
217.00	10327.0
217.90	1544.0
221.00	3548.0
221.90	1441.0
223.00	2363.0
224.00	20152.0
225.00	5066.0
226.00	445.0
227.00	10614.0
228.00	1562.0
228.90	1747.0
230.90	788.0
232.90	209.0
234.00	646.0
234.90	570.0
236.00	408.0
237.00	686.0
239.00	418.0
240.00	367.0
240.90	540.0
242.00	1445.0
243.00	1296.0
244.00	15344.0
245.00	1739.0
246.00	3181.0
247.00	1019.0
248.90	277.0
250.80	153.0
253.00	378.0
253.90	474.0
255.00	73200.0
256.00	10985.0
257.00	830.0
258.00	5374.0
259.00	942.0
259.90	209.0
264.90	2504.0
265.90	481.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

270.80	304.0
272.10	309.0
273.00	2392.0
274.00	5378.0
275.00	25560.0
276.00	3226.0
276.90	3045.0
277.90	560.0
281.00	198.0
282.90	460.0
283.90	176.0
285.10	788.0
292.00	227.0
293.00	652.0
293.90	246.0
294.90	159.0
296.00	12390.0
297.00	1971.0
300.90	164.0
302.00	159.0
302.90	961.0
303.90	210.0
313.90	485.0
314.90	1259.0
315.90	659.0
321.10	199.0
323.00	2823.0
324.10	578.0
326.90	857.0
327.90	434.0
331.80	213.0
333.20	292.0
334.00	2190.0
335.00	559.0
340.90	376.0
346.00	852.0
352.00	997.0
353.00	543.0
354.00	840.0
355.00	158.0
365.00	5558.0
365.90	640.0
372.10	1084.0
373.00	322.0
382.90	399.0
383.90	156.0
390.00	245.0
390.80	159.0
402.00	514.0
402.90	659.0
404.00	285.0
421.00	724.0
422.00	578.0
423.00	3673.0
424.10	1176.0
424.90	222.0
441.10	14216.0
442.00	83832.0
443.10	16624.0
444.00	1650.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

m/z	Abundance
37.90	200.0
39.00	1317.0
40.10	197.0
43.90	273.0
49.10	287.0
50.10	4756.0
51.10	15774.0
52.00	738.0
55.00	157.0
56.00	430.0
57.00	1024.0
61.00	262.0
61.90	355.0
63.00	759.0
64.00	174.0
65.00	246.0
68.00	354.0
69.00	20736.0
69.90	162.0
73.00	517.0
74.00	3545.0
75.00	4213.0
76.10	1522.0
77.00	29128.0
78.00	1752.0
78.90	2066.0
80.00	1264.0
80.90	2143.0
82.00	479.0
83.00	391.0
85.00	243.0
85.90	858.0
87.00	323.0
90.90	589.0
92.00	438.0
92.90	4184.0
94.00	498.0
95.90	184.0
98.00	3911.0
99.00	2164.0
99.90	245.0
100.90	870.0
103.00	414.0
103.80	660.0
104.90	647.0
106.00	155.0
107.00	8497.0
108.00	1355.0
108.90	180.0
110.00	10952.0
111.00	2079.0
116.00	420.0
117.00	14679.0
117.90	1096.0
119.10	194.0
122.00	570.0
122.90	798.0
124.00	636.0
124.90	385.0
127.00	24584.0
128.00	1663.0
129.00	10830.0
130.00	842.0
131.00	165.0
133.90	362.0
134.90	1020.0
136.00	413.0
137.00	630.0
139.80	193.0
141.10	1873.0
142.00	375.0
143.00	390.0
145.80	333.0
146.90	957.0
148.00	3536.0
148.90	668.0
151.00	227.0
152.90	639.0
154.00	522.0
154.90	922.0
156.00	1545.0
157.00	216.0
157.90	393.0
159.00	377.0
159.90	656.0
160.90	992.0
162.00	228.0
165.00	691.0
166.00	496.0
167.00	6185.0
167.90	3719.0
168.90	286.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

170.90	153.0
171.80	211.0
172.90	579.0
173.90	767.0
175.00	1678.0
175.90	185.0
176.90	528.0
177.90	324.0
179.00	2971.0
180.00	1967.0
180.90	769.0
184.10	239.0
185.00	1407.0
186.00	9555.0
187.00	2761.0
187.90	421.0
189.00	676.0
190.90	494.0
192.00	1002.0
193.00	1407.0
193.90	319.0
196.00	1485.0
196.80	246.0
197.90	47128.0
198.90	3817.0
199.90	320.0
203.00	566.0
204.00	3213.0
205.00	4696.0
206.00	17840.0
207.00	3102.0
208.00	743.0
209.10	503.0
210.20	361.0
211.00	839.0
214.90	322.0
215.80	589.0
216.90	5512.0
217.90	681.0
220.90	2309.0
221.90	834.0
223.00	1167.0
224.00	11525.0
225.00	3069.0
225.90	211.0
227.00	5927.0
228.00	860.0
228.90	1219.0
230.80	375.0
233.90	452.0
234.90	480.0
236.00	426.0
236.90	526.0
239.90	224.0
240.90	343.0
241.90	670.0
243.00	1086.0
244.00	8877.0
245.00	1355.0
246.00	2017.0
246.90	390.0
248.80	205.0
250.80	166.0
252.90	248.0
253.90	352.0
255.00	44160.0
256.00	6005.0
256.90	515.0
257.90	2740.0
258.90	515.0
265.00	1680.0
266.00	201.0
273.00	1216.0
274.00	3519.0
275.00	15992.0
276.00	2217.0
277.00	1856.0
277.90	315.0
281.00	203.0
282.00	160.0
283.00	275.0
285.00	367.0
293.10	224.0
294.10	151.0
294.90	195.0
296.00	8118.0
297.00	1147.0
303.00	857.0
303.90	251.0
313.90	294.0
314.90	1105.0
316.00	448.0
323.00	1834.0
323.90	413.0
326.80	493.0
327.90	169.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183

331.80	240.0
332.90	310.0
334.00	1608.0
334.90	339.0
341.00	164.0
345.90	569.0
352.00	636.0
353.00	343.0
354.00	476.0
354.80	164.0
365.00	3750.0
365.90	515.0
372.00	924.0
372.90	302.0
383.10	195.0
402.00	644.0
403.00	548.0
404.00	242.0
421.00	259.0
422.00	365.0
423.00	2792.0
424.10	1083.0
441.00	10776.0
442.00	64952.0
443.10	12176.0
444.00	1055.0

Instrument :
BNA_M
ClientSampleId :
DFTPP183