


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM112719\
 Data File : BM023908.D
 Acq On : 27 Nov 2019 11:53
 Operator : JU
 Sample : DFTPP07
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP070

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM112719MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Nov 29 00:38:59 2019



AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.5	85294	PASS
68	69	0.00	2	1.9	2138	PASS
69	198	0.00	100	39.0	112746	PASS
70	69	0.00	2	0.7	734	PASS
127	198	10	80	47.0	136048	PASS
197	198	0.00	2	0.5	1348	PASS
198	198	100	100	100.0	289258	PASS
199	198	5	9	6.7	19374	PASS
275	198	10	60	32.5	93965	PASS
365	198	1	100	4.0	11459	PASS
441	198	0.01	100	10.9	31392	PASS
442	198	50	100	70.3	203370	PASS
443	442	15	24	18.9	38378	PASS

m/z	Abundance
39.00	560.0
40.00	1890.0
41.00	1022.0
42.10	671.0
43.10	2084.0
44.00	5863.0
45.10	1792.0
49.00	949.0
50.00	339.0
50.90	463.0
55.00	664.0
56.00	634.0
57.00	380.0
59.00	303.0
60.00	1266.0
73.00	1642.0
77.00	348.0
77.90	328.0
79.00	463.0
84.00	851.0
85.90	525.0
91.10	334.0
95.20	319.0
96.10	428.0
97.00	320.0
132.90	344.0
147.00	399.0
149.00	551.0
190.90	441.0
192.00	304.0
193.00	416.0
207.00	4083.0
208.00	1221.0
208.90	635.0
280.90	926.0
282.00	495.0
283.00	349.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

m/z	Abundance
37.00	403.0
38.10	1402.0
39.10	8498.0
40.00	1932.0
41.00	1036.0
42.10	706.0
43.10	2276.0
44.00	5684.0
45.00	1555.0
49.00	1689.0
50.10	24344.0
51.10	100072.0
52.00	5270.0
55.00	985.0
56.00	3600.0
57.00	7776.0
58.00	500.0
60.00	1448.0
61.00	1554.0
62.00	1583.0
63.10	4303.0
64.10	797.0
65.10	2704.0
67.00	429.0
68.00	2320.0
69.00	129848.0
70.00	923.0
73.10	2597.0
74.00	12267.0
75.00	19024.0
76.00	5960.0
77.10	144832.0
78.10	10290.0
79.00	9891.0
80.00	7670.0
81.00	10550.0
82.00	2326.0
83.00	2193.0
84.00	1158.0
85.00	1886.0
86.00	3611.0
87.00	1490.0
88.00	682.0
88.80	358.0
90.90	2473.0
92.00	3400.0
93.00	15910.0
94.10	1277.0
94.90	721.0
96.00	1031.0
97.10	550.0
98.00	12942.0
99.00	11507.0
99.90	613.0
101.00	5772.0
101.90	382.0
103.00	1883.0
104.00	3084.0
105.00	3315.0
106.00	1114.0
107.00	43136.0
108.00	6724.0
109.00	1100.0
110.00	73928.0
111.00	11877.0
112.00	1849.0
113.00	705.0
115.00	434.0
116.00	2452.0
117.00	32304.0
118.00	2824.0
118.90	764.0
120.00	515.0
121.00	349.0
122.00	2841.0
123.00	4921.0
124.00	2007.0
125.00	1427.0
126.10	925.0
127.00	145472.0
128.00	10957.0
129.00	57008.0
130.00	4847.0
131.00	1199.0
132.00	434.0
132.90	857.0
134.00	1644.0
135.00	4994.0
135.90	1859.0
137.00	2822.0
138.10	666.0
139.00	542.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

140.00	773.0
141.00	7012.0
142.00	2127.0
142.90	1343.0
144.10	690.0
144.90	633.0
146.00	964.0
147.00	3974.0
148.00	9102.0
149.00	2880.0
149.90	502.0
151.20	1013.0
152.00	667.0
153.00	2589.0
154.00	1760.0
155.00	4044.0
156.00	5977.0
157.00	1225.0
157.90	1433.0
158.90	1080.0
160.00	1923.0
161.00	3279.0
161.90	1263.0
163.10	326.0
164.00	433.0
165.00	3396.0
166.10	2271.0
167.00	14990.0
168.00	10622.0
169.00	1848.0
170.00	858.0
171.00	720.0
172.00	1553.0
173.00	1486.0
174.00	3240.0
175.00	6337.0
176.10	1816.0
177.00	3450.0
178.10	1028.0
179.00	10677.0
180.00	8802.0
181.00	3611.0
181.90	681.0
182.90	414.0
183.90	902.0
185.00	5429.0
186.00	43280.0
187.00	11526.0
188.00	1863.0
189.00	3192.0
189.90	467.0
191.00	1448.0
192.00	3982.0
193.00	4000.0
193.90	1172.0
195.00	540.0
196.00	8609.0
197.10	2217.0
198.00	292224.0
199.00	21168.0
200.00	1583.0
201.40	1709.0
203.00	2305.0
204.00	10179.0
205.00	18664.0
206.00	77504.0
207.00	15377.0
208.00	3500.0
209.00	1520.0
210.00	1115.0
210.20	1104.0
211.10	3161.0
211.80	702.0
213.00	364.0
215.00	1173.0
216.00	1456.0
217.00	19712.0
218.00	3140.0
218.90	310.0
221.00	13107.0
223.10	4424.0
224.00	42880.0
225.10	10752.0
225.90	1615.0
227.00	17456.0
228.00	2716.0
229.00	3811.0
230.00	428.0
231.00	1736.0
231.90	325.0
233.00	522.0
233.90	1164.0
235.00	1677.0
236.00	1116.0
237.00	2677.0
239.10	1231.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

240.00	654.0
241.00	747.0
242.00	1940.0
243.10	2837.0
244.00	33352.0
245.00	4407.0
246.00	6612.0
247.10	1195.0
247.90	389.0
248.90	1424.0
250.00	454.0
251.20	469.0
251.90	318.0
252.90	592.0
254.20	1892.0
255.00	190272.0
256.00	26824.0
257.00	2510.0
258.00	10919.0
259.00	1669.0
265.00	4677.0
265.90	821.0
267.10	327.0
270.10	439.0
271.00	731.0
272.10	451.0
273.00	5572.0
274.00	13912.0
275.00	89248.0
276.00	11388.0
277.00	7145.0
278.00	1679.0
279.00	436.0
281.00	1214.0
282.10	913.0
283.00	1053.0
284.00	546.0
285.00	1708.0
289.90	339.0
292.00	536.0
293.00	1541.0
293.90	649.0
295.10	616.0
296.00	24560.0
297.00	2947.0
300.90	482.0
303.10	3033.0
303.90	866.0
308.00	514.0
310.00	384.0
314.10	1229.0
315.00	2898.0
316.00	1720.0
320.90	619.0
323.10	7175.0
324.00	1413.0
326.90	1474.0
328.00	809.0
332.00	437.0
333.10	993.0
334.10	4441.0
335.00	1484.0
341.10	1064.0
346.10	1240.0
346.90	397.0
352.00	2117.0
353.00	1432.0
354.00	2214.0
355.20	375.0
365.00	9187.0
366.00	1773.0
371.10	587.0
372.10	3539.0
373.10	950.0
383.10	1311.0
390.00	732.0
402.10	1366.0
403.10	1608.0
404.10	474.0
421.10	1193.0
422.00	1296.0
423.10	9478.0
424.10	3995.0
425.00	548.0
441.10	23168.0
442.10	153792.0
443.10	29920.0
444.10	3170.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

m/z	Abundance
37.00	694.0
38.10	1382.0
39.10	8929.0
40.00	2069.0
41.10	1535.0
42.00	788.0
43.00	2506.0
44.00	5645.0
45.10	1926.0
49.00	1660.0
50.10	25648.0
51.10	98760.0
52.10	5029.0
53.00	312.0
55.00	1119.0
56.00	4147.0
57.00	7794.0
58.00	451.0
59.10	361.0
60.10	1112.0
61.00	2025.0
62.00	1280.0
63.00	4896.0
64.00	778.0
65.00	2190.0
65.90	408.0
66.90	406.0
68.00	2550.0
69.00	131520.0
70.10	708.0
73.00	2727.0
74.00	13073.0
75.00	20344.0
76.00	6727.0
77.10	150592.0
78.10	9745.0
79.00	9757.0
80.00	7492.0
81.00	11296.0
82.00	2256.0
83.00	2413.0
83.90	1548.0
85.00	1801.0
86.00	3467.0
87.00	1059.0
87.90	756.0
88.90	572.0
90.90	2404.0
92.00	2736.0
93.00	17232.0
93.90	1059.0
95.20	355.0
96.10	898.0
98.00	12763.0
99.00	11815.0
100.10	1045.0
101.00	5609.0
101.90	534.0
102.90	2137.0
104.00	3971.0
105.00	3689.0
106.00	1018.0
107.00	44480.0
108.00	6710.0
109.00	1586.0
110.00	83568.0
111.00	13291.0
112.00	1724.0
112.90	727.0
115.00	459.0
116.10	2385.0
117.00	34728.0
118.00	2609.0
118.80	396.0
120.00	781.0
121.00	455.0
122.00	3088.0
123.00	4383.0
123.90	2178.0
125.00	1912.0
126.10	1186.0
127.00	160192.0
128.00	12506.0
129.00	61704.0
130.00	4547.0
130.90	796.0
132.10	518.0
133.10	722.0
134.00	1701.0
135.00	5596.0
135.90	1735.0
137.00	3507.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

137.90	576.0
139.00	628.0
140.10	864.0
141.00	8478.0
142.00	2266.0
143.00	1627.0
144.00	627.0
145.00	505.0
145.90	1977.0
147.00	4165.0
148.00	9704.0
149.00	2628.0
150.00	610.0
151.20	1052.0
151.90	774.0
153.00	2712.0
154.00	2373.0
155.10	4373.0
156.10	6973.0
157.00	1383.0
158.00	1533.0
158.90	1284.0
160.00	3010.0
161.00	4408.0
161.90	1135.0
163.00	449.0
164.10	441.0
165.00	3496.0
166.00	2493.0
167.00	16448.0
168.00	11093.0
169.00	1794.0
169.90	665.0
171.00	666.0
172.00	1587.0
173.00	2045.0
174.00	3711.0
175.00	6669.0
175.90	1862.0
176.90	2847.0
178.00	1299.0
179.00	13630.0
180.00	10461.0
181.00	4300.0
182.00	643.0
183.10	322.0
184.10	737.0
185.00	6149.0
186.00	49296.0
187.00	14086.0
188.00	1380.0
189.00	3424.0
190.10	703.0
191.00	2067.0
192.00	3866.0
193.00	4676.0
194.10	708.0
196.00	10168.0
198.00	346880.0
199.00	21848.0
199.90	2396.0
201.20	982.0
201.70	1405.0
203.00	2511.0
204.00	12193.0
205.00	22152.0
206.00	94232.0
207.10	15778.0
208.00	4653.0
209.10	1673.0
210.30	1833.0
211.00	3379.0
211.90	695.0
215.00	928.0
216.00	2081.0
217.00	22744.0
218.00	2626.0
219.00	317.0
219.80	338.0
221.00	14482.0
223.00	5739.0
224.00	52112.0
225.10	12556.0
225.90	1628.0
227.00	22872.0
228.00	3297.0
229.00	4661.0
230.00	513.0
231.00	2140.0
231.90	447.0
232.90	409.0
234.00	1267.0
234.90	2112.0
236.00	1331.0
237.00	2449.0
238.00	348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

239.00	1248.0
239.90	956.0
240.90	1372.0
242.00	2887.0
243.00	3729.0
244.00	42224.0
245.00	5148.0
246.00	8312.0
247.00	1888.0
247.90	349.0
249.00	1809.0
249.90	367.0
251.10	491.0
252.00	482.0
253.10	791.0
255.00	232896.0
256.00	36264.0
256.90	3144.0
258.00	12892.0
259.00	2093.0
260.10	387.0
261.00	552.0
263.90	307.0
265.00	4980.0
266.00	743.0
270.00	329.0
271.00	683.0
271.90	766.0
273.00	7146.0
274.00	20024.0
275.00	112800.0
276.00	15247.0
277.00	9841.0
278.00	2278.0
279.00	551.0
281.10	1672.0
282.00	512.0
283.00	1065.0
283.90	483.0
285.00	1742.0
286.00	310.0
290.10	347.0
292.00	393.0
293.10	2427.0
293.90	725.0
295.00	808.0
296.00	32256.0
297.00	3834.0
298.10	315.0
301.00	326.0
302.00	714.0
303.00	3784.0
304.00	996.0
307.90	393.0
309.90	334.0
314.00	1905.0
315.00	3566.0
316.00	1634.0
317.00	455.0
321.10	1057.0
322.20	422.0
323.10	10320.0
324.10	1969.0
326.00	403.0
327.00	1820.0
328.00	1109.0
332.00	775.0
333.00	655.0
334.10	6894.0
335.00	1712.0
341.00	1021.0
341.90	447.0
346.00	2518.0
346.90	510.0
352.00	3072.0
353.00	1717.0
354.00	3597.0
355.10	771.0
365.00	14434.0
365.90	1949.0
371.20	772.0
372.10	5502.0
373.10	1261.0
383.10	1344.0
384.00	488.0
390.00	816.0
392.10	515.0
402.10	1899.0
403.00	2348.0
404.00	742.0
405.00	302.0
421.00	2103.0
422.00	2047.0
423.10	14562.0
424.10	4514.0
425.10	553.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

441.10	38808.0
442.10	249536.0
443.10	48536.0
444.10	4466.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

m/z	Abundance
37.00	441.0
37.90	863.0
39.10	4783.0
40.00	2082.0
41.00	1410.0
41.90	533.0
43.00	2375.0
44.00	6064.0
45.00	1785.0
49.00	1529.0
50.00	15433.0
51.10	58440.0
52.10	2959.0
53.10	336.0
55.10	876.0
56.00	2416.0
57.10	4900.0
59.10	317.0
60.00	1415.0
61.00	1115.0
61.90	1018.0
63.10	2540.0
63.90	485.0
65.00	1605.0
66.00	382.0
67.00	400.0
68.10	1546.0
69.00	76872.0
70.00	572.0
73.00	2015.0
74.00	7919.0
75.00	11739.0
76.10	4059.0
77.10	87360.0
78.00	6444.0
79.00	5793.0
80.00	4505.0
81.00	5921.0
82.00	1668.0
83.00	1785.0
84.00	735.0
85.00	1576.0
86.00	2554.0
87.00	1156.0
88.10	333.0
91.00	2177.0
92.00	2075.0
93.00	10969.0
94.00	671.0
95.00	440.0
96.00	817.0
97.00	685.0
98.00	7719.0
99.00	6977.0
99.90	399.0
101.00	3865.0
103.10	1188.0
104.00	2307.0
105.10	2492.0
106.10	886.0
107.00	27096.0
108.00	4402.0
109.10	963.0
110.00	50776.0
111.00	7617.0
111.90	895.0
113.00	340.0
115.00	476.0
116.00	1536.0
117.00	22728.0
118.00	1289.0
118.90	389.0
120.00	352.0
122.00	1493.0
123.00	3470.0
124.00	1178.0
125.00	985.0
126.20	865.0
127.00	102480.0
128.10	7624.0
129.00	39728.0
130.00	3432.0
131.00	461.0
132.00	369.0
132.80	385.0
134.00	1108.0
135.00	3531.0
136.00	1486.0
137.00	2494.0
137.90	467.0
140.00	468.0
141.00	6041.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

142.00	2174.0
142.90	1213.0
145.90	800.0
147.00	3209.0
148.00	6488.0
148.90	1869.0
149.90	497.0
151.00	824.0
151.70	570.0
153.00	2081.0
154.00	1349.0
155.10	3021.0
156.00	4148.0
157.10	546.0
158.00	881.0
159.00	741.0
160.00	1753.0
161.00	2963.0
162.00	887.0
162.80	332.0
165.00	2209.0
166.00	1718.0
167.00	10578.0
168.00	6746.0
169.00	1297.0
170.20	439.0
171.00	603.0
172.00	1254.0
172.90	1402.0
174.00	2499.0
175.10	4062.0
176.20	1284.0
177.00	2062.0
177.90	1039.0
179.00	8282.0
180.00	6930.0
181.00	2901.0
181.90	689.0
183.90	685.0
185.00	4448.0
186.00	33696.0
187.10	9417.0
188.00	973.0
188.90	1991.0
189.90	550.0
191.00	1799.0
191.90	2945.0
193.00	4299.0
194.10	919.0
195.10	459.0
196.00	7715.0
197.20	1829.0
197.90	228672.0
199.00	15108.0
200.00	1146.0
201.50	1555.0
202.10	445.0
203.00	1679.0
204.00	9039.0
205.00	14234.0
206.00	62208.0
207.00	11105.0
208.00	2476.0
209.00	1519.0
211.00	2188.0
215.00	762.0
216.00	1318.0
217.00	15446.0
218.00	2247.0
221.00	10709.0
221.80	2497.0
223.10	3984.0
224.00	34304.0
225.10	8770.0
226.00	1204.0
227.00	14225.0
228.00	2014.0
229.00	3437.0
229.90	343.0
231.00	1320.0
233.10	328.0
234.00	762.0
235.00	1071.0
235.90	623.0
237.00	1675.0
239.00	736.0
239.90	428.0
241.00	1087.0
242.00	1945.0
243.00	2409.0
244.00	29512.0
245.00	4124.0
246.00	6363.0
246.90	1064.0
248.00	471.0
249.00	1110.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070

250.90	433.0
253.10	642.0
254.20	1578.0
255.00	167040.0
256.00	24736.0
257.10	1805.0
258.00	9589.0
259.00	1824.0
261.00	334.0
264.00	412.0
265.10	4017.0
265.90	705.0
266.90	540.0
270.90	568.0
272.00	505.0
273.00	5791.0
274.00	14455.0
275.00	79848.0
276.00	9975.0
277.00	6270.0
278.10	1007.0
278.90	320.0
280.90	1460.0
282.00	568.0
283.00	959.0
284.00	438.0
285.10	1264.0
290.10	340.0
291.90	408.0
293.00	1597.0
293.90	343.0
294.90	591.0
296.00	23952.0
297.00	2799.0
301.00	370.0
302.10	448.0
303.00	2323.0
304.00	784.0
309.00	350.0
310.10	419.0
314.00	1141.0
315.00	2516.0
316.10	1512.0
321.00	698.0
322.00	439.0
323.10	7638.0
324.00	1581.0
327.00	1617.0
328.10	995.0
331.90	487.0
333.00	866.0
334.00	5559.0
335.00	1328.0
341.00	1005.0
342.10	401.0
346.00	2102.0
352.00	2985.0
353.00	1870.0
354.00	2436.0
355.00	688.0
365.00	10757.0
366.00	1532.0
371.10	487.0
372.10	4154.0
373.10	1094.0
383.10	1136.0
384.10	325.0
389.90	650.0
391.10	514.0
402.10	1443.0
403.00	2093.0
404.10	834.0
421.10	1509.0
422.10	1808.0
423.10	13141.0
424.10	4012.0
425.10	490.0
440.20	366.0
441.10	32200.0
442.10	206784.0
443.10	36680.0
444.10	3520.0

Instrument :
BNA_M
ClientSampleId :
DFTPP070