

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM113022\
 Data File : BM037797.D
 Acq On : 30 Nov 2022 10:13
 Operator : CG/JU
 Sample : SST0.191
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST0.1001

Quant Time: Nov 30 23:29:30 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM113022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 30 23:27:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.105	152	5499	0.400	ng/ul	0.00
4) Naphthalene-d8	10.925	136	16034	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.728	164	8225	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.466	188	17219	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.630	240	12357	0.400	ng/ul	# 0.00
23) Perylene-d12	24.118	264	11975	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	12.498	152	2251	0.097	ng/ul	0.00
18) Fluoranthene-d10	19.480	212	4206	0.100	ng/ul	0.00
Target Compounds				Qvalue		
5) Naphthalene	10.975	128	5279	0.105	ng/ul	99
7) 2-Methylnaphthalene	12.575	142	2981	0.098	ng/ul	99
8) 1-Methylnaphthalene	12.789	142	3039	0.097	ng/ul	100
10) Acenaphthylene	14.451	152	4073	0.099	ng/ul	94
11) Acenaphthene	14.789	153	3393	0.099	ng/ul	97
12) Fluorene	15.769	166	3775	0.100	ng/ul	97
15) Phenanthrene	17.508	178	6265	0.100	ng/ul	94
16) Anthracene	17.597	178	5352	0.096	ng/ul	93
19) Fluoranthene	19.508	202	6864	0.100	ng/ul#	80
20) Pyrene	19.871	202	7320	0.102	ng/ul#	82
21) Benzo(a)anthracene	21.613	228	6129	0.102	ng/ul	97
22) Chrysene	21.668	228	6199	0.103	ng/ul	97
24) Benzo(b)fluoranthene	23.352	252	6322	0.101	ng/ul#	45
25) Benzo(k)fluoranthene	23.404	252	6065	0.100	ng/ul#	48
26) Benzo(a)pyrene	24.004	252	5399	0.101	ng/ul#	24
27) Indeno(1,2,3-cd)pyrene	26.714	276	6947	0.099	ng/ul#	74
28) Dibenzo(a,h)anthracene	26.734	278	5444	0.099	ng/ul#	66
29) Benzo(g,h,i)perylene	27.516	276	5752	0.098	ng/ul#	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM113022\
Data File : BM037797.D
Acq On : 30 Nov 2022 10:13
Operator : CG/JU
Sample : SST00.191
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.1001

Quant Time: Nov 30 23:29:30 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM113022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 30 23:27:43 2022
Response via : Initial Calibration

