

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM113022\
 Data File : BM037806.D
 Acq On : 30 Nov 2022 17:10
 Operator : CG/JU
 Sample : N5811-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGCE0

Quant Time: Dec 01 00:20:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM113022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 01 00:14:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.105	152	4147	0.400	ng/ul	0.00
4) Naphthalene-d8	10.921	136	14038	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.730	164	8106	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.463	188	16011	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.631	240	11629	0.400	ng/ul	# 0.00
23) Perylene-d12	24.116	264	12208	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	24619	4.458	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.499	152	5768	0.285	ng/ul	0.00
18) Fluoranthene-d10	19.482	212	13352	0.338	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.472	88	1330	0.231	ng/ul#	73


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM113022\
Data File : BM037806.D
Acq On : 30 Nov 2022 17:10
Operator : CG/JU
Sample : N5811-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGCE0

Quant Time: Dec 01 00:20:22 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM113022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 01 00:14:21 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.231 ng/ul
 RT: 3.472 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BM037806.D
 Acq: 30 Nov 2022 17:10

Instrument :
 BNA_M
 ClientSampleId :
 BGCE0

Tgt Ion: 88 Resp: 1330
 Ion Ratio Lower Upper
 88 100
 43 95.2 102.8 154.2#
 58 61.4 67.3 100.9#

