

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120423\
 Data File : BM043174.D
 Acq On : 04 Dec 2023 13:36
 Operator : MA/JU
 Sample : 05473-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AE8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/05/2023
 Supervised By :mohammad ahmed 12/05/2023

Quant Time: Dec 04 14:17:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 02 23:54:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	956	0.400	ng/ul	#-0.03
4) Naphthalene-d8	10.590	136	2980	0.400	ng/ul	#-0.03
9) Acenaphthene-d10	14.442	164	1793m	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.196	188	4094m	0.400	ng/ul	-0.02
17) Chrysene-d12	21.406	240	2649m	0.400	ng/ul	0.00
23) Perylene-d12	23.738	264	2193m	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	3660	2.495	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.212	152	1162	0.291	ng/ul	-0.01
18) Fluoranthene-d10	19.230	212	3609	0.374	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120423\
Data File : BM043174.D
Acq On : 04 Dec 2023 13:36
Operator : MA/JU
Sample : 05473-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AE8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/05/2023
Supervised By :mohammad ahmed 12/05/2023

Quant Time: Dec 04 14:17:24 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM112423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 02 23:54:55 2023
Response via : Initial Calibration

