

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM120518\
 Data File : BM017914.D
 Acq On : 04 Dec 2018 22:11
 Operator : JU/SJ
 Sample : J5771-09 2X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EO-02-112918-B

Quant Time: Dec 05 00:27:15 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM111218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 12 18:37:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	267704	20.00	ng	-0.04
21) Naphthalene-d8	10.35	136	1002746	20.00	ng	-0.04
39) Acenaphthene-d10	14.23	164	471319	20.00	ng	-0.04
64) Phenanthrene-d10	16.99	188	877689	20.00	ng	-0.03
76) Chrysene-d12	21.20	240	925359	20.00	ng	-0.02
87) Perylene-d12	23.39	264	1018955	20.00	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.21	112	673257	42.44	ng	-0.03
7) Phenol-d6	6.77	99	909455	41.05	ng	-0.03
23) Nitrobenzene-d5	8.73	82	530799	27.33	ng	-0.04
42) 2,4,6-Tribromophenol	15.73	330	255646	37.75	ng	-0.03
45) 2-Fluorobiphenyl	12.84	172	1073010	29.54	ng	-0.04
79) Terphenyl-d14	19.64	244	1132781	27.74	ng	-0.02


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Phenol	6.80	94	51747	2.225	ng	96
31) Naphthalene	10.40	128	246314	4.996	ng	98
37) 2-Methylnaphthalene	12.02	142	241293	6.925	ng	98
38) 1-Methylnaphthalene	12.25	142	172453	5.048	ng	100
50) Dimethylphthalate	13.70	163	91680	2.392	ng	# 99
52) Acenaphthene	14.29	154	130463	4.506	ng	97
55) Dibenzofuran	14.63	168	170692	3.607	ng	# 96
56) 4-Nitrophenol	14.44	139	5167	5.576	ng	# 42
58) Fluorene	15.29	166	236433	6.527	ng	99
71) Phenanthrene	17.03	178	1360291	28.562	ng	100
72) Anthracene	17.12	178	460452	9.936	ng	99
75) Fluoranthene	19.06	202	1126833	20.911	ng	99
78) Pyrene	19.43	202	1045009	18.303	ng	99
81) Benzo(a)anthracene	21.18	228	542590	10.178	ng	98
83) Chrysene	21.23	228	463812	8.957	ng	96
86) Indeno(1,2,3-cd)pyrene	25.56	276	269569	6.515	ng	97
88) Benzo(b)fluoranthene	22.73	252	581565	9.746	ng	# 86
89) Benzo(k)fluoranthene	22.73	252	582761	9.662	ng	# 88
90) Benzo(a)pyrene	23.29	252	456075	8.382	ng	# 90
92) Benzo(g,h,i)perylene	26.23	276	258731	6.015	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM120518\
Data File : BM017914.D
Acq On : 04 Dec 2018 22:11
Operator : JU/SJ
Sample : J5771-09 2X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EO-02-112918-B

Quant Time: Dec 05 00:27:15 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM111218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 12 18:37:25 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.58 min Scan# 781
Delta R.T. -0.04 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Instrument :
BNA_M
ClientSampleId :
EO-02-112918-B

Tgt Ion:152 Resp: 267704
Ion Ratio Lower Upper
152 100
150 149.8 126.4 189.6
115 60.4 49.4 74.0




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 42.443 ng
RT: 5.21 min Scan# 378
Delta R.T. -0.03 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Tgt Ion:112 Resp: 673257
Ion Ratio Lower Upper
112 100
64 54.8 49.3 73.9
63 27.9 26.2 39.4



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 63.00 (62.70 to 63.70): BM0





#7

Phenol-d6

Concen: 41.051 ng

RT: 6.77 min Scan# 644

Delta R.T. -0.03 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion: 99 Resp: 909455

Ion Ratio Lower Upper

99 100

42 16.1 16.3 24.5#




71 33.8 29.8 44.8



Abundance Ion 99.00 (98.70 to 99.70): BM017914.D (-632) (-)

Ion 42.00 (41.70 to 42.70): BM017914.D (-632) (-)

Ion 71.00 (70.70 to 71.70): BM017914.D (-632) (-)



#10

Phenol

Concen: 2.225 ng

RT: 6.80 min Scan# 648

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion: 94 Resp: 51747

Ion Ratio Lower Upper

94 100

65 28.1 11.4 51.4



66 40.1 21.4 61.4



Abundance Ion 94.00 (93.70 to 94.70): BM017914.D (-637) (-)

Ion 65.00 (64.70 to 65.70): BM017914.D (-637) (-)

Ion 66.00 (65.70 to 66.70): BM017914.D (-637) (-)









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.35 min Scan# 1252
Delta R.T. -0.04 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Instrument :
BNA_M
ClientSampleId :
EO-02-112918-B

Tgt Ion:	136	Resp:	1002746
Ion	Ratio	Lower	Upper
136	100		
137	10.6	8.6	12.8
54	6.9	7.0	10.6#
68	5.5	4.9	7.3




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BM0
Ion 68.00 (67.70 to 68.70): BM0






#23
Nitrobenzene-d5
Concen: 27.328 ng
RT: 8.73 min Scan# 977
Delta R.T. -0.04 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Tgt Ion:	82	Resp:	530799
Ion	Ratio	Lower	Upper
82	100		
128	43.9	33.0	49.4
54	45.1	41.4	62.2



Abundance Ion 82.00 (81.70 to 82.70): BM0
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BM0





#31

Naphthalene

Concen: 4.996 ng

RT: 10.40 min Scan# 1260

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion:128 Resp: 246314

Ion Ratio Lower Upper

128 100

129 12.0 8.6 12.8

127 12.8 10.3 15.5



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E

150000



100000

50000

0

10.35 10.40 10.45

Time-->



#37

2-Methylnaphthalene

Concen: 6.925 ng

RT: 12.02 min Scan# 1536

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion:142 Resp: 241293

Ion Ratio Lower Upper

142 100

141 89.7 72.2 108.4

115 33.3 28.5 42.7



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E

150000


100000


50000

0

11.90 12.00 12.10

Time-->









#38
1-Methylnaphthalene
Concen: 5.048 ng
RT: 12.25 min Scan# 1574
Delta R.T. -0.04 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Instrument :
BNA_M
ClientSampleId :
EO-02-112918-B

Tgt Ion:142 Resp: 172453
Ion Ratio Lower Upper
142 100
141 93.0 74.2 111.4
116 3.9 3.1 4.7




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 116.00 (115.70 to 116.70): E






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.23 min Scan# 1911
Delta R.T. -0.04 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Tgt Ion:164 Resp: 471319
Ion Ratio Lower Upper
164 100
162 98.9 81.4 122.0
160 43.9 35.6 53.4



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E









#42
 2,4,6-Tribromophenol
 Concen: 37.749 ng
 RT: 15.73 min Scan# 2167
 Delta R.T. -0.03 min
 Lab File: BM017914.D
 Acq: 04 Dec 2018 22:11

Instrument :
 BNA_M
 ClientSampleId :
 EO-02-112918-B

Tgt Ion:330 Resp: 255646
 Ion Ratio Lower Upper
 330 100
 332 96.3 77.9 116.9
 141 32.5 27.5 41.3




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 29.535 ng
 RT: 12.84 min Scan# 1675
 Delta R.T. -0.04 min
 Lab File: BM017914.D
 Acq: 04 Dec 2018 22:11

Tgt Ion:172 Resp: 1073010
 Ion Ratio Lower Upper
 172 100
 171 35.3 28.7 43.1
 170 22.4 18.3 27.5



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#50

Dimethylphthalate

Concen: 2.392 ng

RT: 13.70 min Scan# 1821

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion:163 Resp: 91680

Ion Ratio Lower Upper

163 100

194 4.9 3.1 4.7#

164 10.0 8.2 12.4






Abundance

Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#52

Acenaphthene

Concen: 4.506 ng

RT: 14.29 min Scan# 1922

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion:154 Resp: 130463

Ion Ratio Lower Upper

154 100

153 109.3 90.2 135.4

152 51.2 41.6 62.4






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#55
Dibenzofuran
Concen: 3.607 ng
RT: 14.63 min Scan# 1980
Delta R.T. -0.03 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11




Instrument :
BNA_M
ClientSampleId :
EO-02-112918-B


Tgt Ion:168 Resp: 170692
Ion Ratio Lower Upper
168 100
139 38.5 31.0 46.6
169 19.3 10.6 16.0#



#56
4-Nitrophenol
Concen: 5.576 ng
RT: 14.44 min Scan# 1948
Delta R.T. -0.06 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Tgt Ion:139 Resp: 5167
Ion Ratio Lower Upper
139 100
109 133.1 58.6 98.6#
65 45.9 79.5 119.5#





#58

Fluorene

Concen: 6.527 ng

RT: 15.29 min Scan# 2091

Delta R.T. -0.03 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion:166 Resp: 236433

Ion Ratio Lower Upper

166 100

165 99.7 79.0 118.6




167 15.7 10.8 16.2



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 16.99 min Scan# 2380

Delta R.T. -0.03 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion:188 Resp: 877689

Ion Ratio Lower Upper

188 100

94 9.0 7.5 11.3



80 10.7 8.6 13.0




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM

Ion 80.00 (79.70 to 80.70): BM









#71
Phenanthrene
Concen: 28.562 ng
RT: 17.03 min Scan# 2388
Delta R.T. -0.02 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11




Instrument :
BNA_M
ClientSampleId :
EO-02-112918-B


Tgt Ion:178 Resp: 1360291
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.4
179 15.3 12.4 18.6



#72
Anthracene
Concen: 9.936 ng
RT: 17.12 min Scan# 2403
Delta R.T. -0.03 min
Lab File: BM017914.D
Acq: 04 Dec 2018 22:11

Tgt Ion:178 Resp: 460452
Ion Ratio Lower Upper
178 100
176 18.5 14.5 21.7
179 16.2 12.3 18.5









#75
 Fluoranthene
 Concen: 20.911 ng
 RT: 19.06 min Scan# 2732
 Delta R.T. -0.02 min
 Lab File: BM017914.D
 Acq: 04 Dec 2018 22:11




Instrument :
 BNA_M
 ClientSampleId :
 EO-02-112918-B


Tgt Ion: 202 Resp: 1126833
 Ion Ratio Lower Upper
 202 100
 101 11.5 0.0 31.5
 203 18.2 0.0 37.2



#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 21.20 min Scan# 3096
 Delta R.T. -0.02 min
 Lab File: BM017914.D
 Acq: 04 Dec 2018 22:11

Tgt Ion: 240 Resp: 925359
 Ion Ratio Lower Upper
 240 100
 120 10.4 8.5 12.7
 236 24.8 20.4 30.6









#78
 Pyrene
 Concen: 18.303 ng
 RT: 19.43 min Scan# 2795
 Delta R.T. -0.02 min
 Lab File: BM017914.D
 Acq: 04 Dec 2018 22:11




Instrument :
 BNA_M
 ClientSampleId :
 EO-02-112918-B


Tgt Ion: 202 Resp: 1045009
 Ion Ratio Lower Upper
 202 100
 200 19.9 16.4 24.6
 203 17.9 13.8 20.8



#79
 Terphenyl-d14
 Concen: 27.744 ng
 RT: 19.64 min Scan# 2831
 Delta R.T. -0.02 min
 Lab File: BM017914.D
 Acq: 04 Dec 2018 22:11

Tgt Ion: 244 Resp: 1132781
 Ion Ratio Lower Upper
 244 100
 212 6.6 5.8 8.6
 122 10.0 9.0 13.6





#81

Benzo(a)anthracene

Concen: 10.178 ng

RT: 21.18 min Scan# 3093

Delta R.T. -0.02 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11





Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B

Tgt Ion:	228	Resp:	542590
Ion	Ratio	Lower	Upper
228	100		
226	27.3	21.3	31.9
229	21.4	16.0	24.0



#83

Chrysene

Concen: 8.957 ng




RT: 21.23 min Scan# 3102


Delta R.T. -0.02 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Tgt Ion:	228	Resp:	463812
Ion	Ratio	Lower	Upper
228	100		
226	30.0	23.6	35.4
229	23.4	15.8	23.6





#86

Indeno(1,2,3-cd)pyrene

Concen: 6.515 ng

RT: 25.56 min Scan# 3838

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion:276 Resp: 269569

Ion Ratio Lower Upper

276 100

138 20.9 19.1 28.7




277 25.1 20.0 30.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 23.39 min Scan# 3468

Delta R.T. -0.02 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion:264 Resp: 1018955

Ion Ratio Lower Upper

264 100

260 23.5 18.2 27.2



265 22.2 17.5 26.3




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#88

Benzo(b)fluoranthene

Concen: 9.746 ng

RT: 22.73 min Scan# 3357

Delta R.T. -0.02 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion:252 Resp: 581565

Ion Ratio Lower Upper

252 100

253 27.3 17.3 25.9#




125 15.8 7.6 11.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(k)fluoranthene

Concen: 9.662 ng

RT: 22.73 min Scan# 3357

Delta R.T. -0.06 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion:252 Resp: 582761

Ion Ratio Lower Upper

252 100

253 27.3 17.7 26.5#



125 15.8 7.9 11.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#90

Benzo(a)pyrene

Concen: 8.382 ng

RT: 23.29 min Scan# 3451

Delta R.T. -0.03 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11

Instrument :

BNA_M

ClientSampleId :

EO-02-112918-B


Tgt Ion:252 Resp: 456075

Ion Ratio Lower Upper

252 100

253 25.8 17.4 26.2




125 15.0 8.2 12.2#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#92

Benzo(g,h,i)perylene

Concen: 6.015 ng

RT: 26.23 min Scan# 3951

Delta R.T. -0.04 min

Lab File: BM017914.D

Acq: 04 Dec 2018 22:11


Tgt Ion:276 Resp: 258731

Ion Ratio Lower Upper

276 100

277 25.3 19.0 28.4

138 22.1 15.9 23.9



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

