

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120721\
 Data File : BM033332.D
 Acq On : 07 Dec 2021 20:16
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM120721

Quant Time: Dec 08 02:49:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 02:33:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	131	0.00
2	1,4-Dioxane	40.000	32.970	17.6	94	0.00
3	Pyridine	40.000	38.844	2.9	119	0.00
4	n-Nitrosodimethylamine	40.000	38.811	3.0	120	0.00
5 S	2-Fluorophenol	80.000	78.945	1.3	128	0.00
6	Aniline	40.000	39.698	0.8	126	0.00
7 S	Phenol-d6	80.000	79.588	0.5	128	0.00
8	2-Chlorophenol	40.000	39.461	1.3	127	0.00
9	Benzaldehyde	40.000	38.484	3.8	130	0.00
10 C	Phenol	40.000	40.059	-0.1	129	0.00
11	bis(2-Chloroethyl)ether	40.000	39.766	0.6	131	0.00
12	1,3-Dichlorobenzene	40.000	39.234	1.9	129	0.00
13 C	1,4-Dichlorobenzene	40.000	39.966	0.1	131	0.00
14	1,2-Dichlorobenzene	40.000	39.407	1.5	129	0.00
15	Benzyl Alcohol	40.000	39.917	0.2	127	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.921	2.7	127	0.00
17	2-Methylphenol	40.000	39.968	0.1	126	0.00
18	Hexachloroethane	40.000	39.835	0.4	128	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.860	0.4	126	0.00
20	3+4-Methylphenols	40.000	40.315	-0.8	127	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	125	0.00
22	Acetophenone	40.000	40.248	-0.6	124	0.00
23 S	Nitrobenzene-d5	80.000	82.733	-3.4	128	0.00
24	Nitrobenzene	40.000	41.364	-3.4	129	0.00
25	Isophorone	40.000	40.520	-1.3	123	0.00
26 C	2-Nitrophenol	40.000	43.327	-8.3	128	0.00
27	2,4-Dimethylphenol	40.000	40.495	-1.2	125	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.158	-0.4	126	0.00
29 C	2,4-Dichlorophenol	40.000	41.447	-3.6	125	0.00
30	1,2,4-Trichlorobenzene	40.000	40.257	-0.6	127	0.00
31	Naphthalene	40.000	39.929	0.2	125	0.00
32	Benzoic acid	40.000	44.244	-10.6	130	0.01
33	4-Chloroaniline	40.000	40.859	-2.1	124	0.00
34 C	Hexachlorobutadiene	40.000	41.380	-3.5	131	0.00
35	Caprolactam	40.000	37.884	5.3	114	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.311	-0.8	120	0.00
37	2-Methylnaphthalene	40.000	39.471	1.3	122	0.00
38	1-Methylnaphthalene	40.000	39.146	2.1	122	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.023	-2.6	123	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.803	-12.0	127	0.00
42 S	2,4,6-Tribromophenol	80.000	78.996	1.3	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.674	-4.2	120	0.00
44	2,4,5-Trichlorophenol	40.000	40.978	-2.4	118	0.00
45 S	2-Fluorobiphenyl	80.000	80.145	-0.2	120	0.00
46	1,1'-Biphenyl	40.000	40.121	-0.3	120	0.00
47	2-Chloronaphthalene	40.000	40.373	-0.9	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120721\
 Data File : BM033332.D
 Acq On : 07 Dec 2021 20:16
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM120721

Quant Time: Dec 08 02:49:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM120721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 08 02:33:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.779	-4.4	115	0.00
49	Acenaphthylene	40.000	40.237	-0.6	118	0.00
50	Dimethylphthalate	40.000	38.734	3.2	113	0.00
51	2,6-Dinitrotoluene	40.000	40.134	-0.3	114	0.00
52 C	Acenaphthene	40.000	39.531	1.2	117	0.00
53	3-Nitroaniline	40.000	40.738	-1.8	112	0.00
54 P	2,4-Dinitrophenol	40.000	43.389	-8.5	112	0.00
55	Dibenzofuran	40.000	39.043	2.4	116	0.00
56 P	4-Nitrophenol	40.000	42.133	-5.3	111	0.00
57	2,4-Dinitrotoluene	40.000	40.569	-1.4	113	0.00
58	Fluorene	40.000	38.679	3.3	113	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.638	0.9	115	0.00
60	Diethylphthalate	40.000	38.559	3.6	111	0.00
61	4-Chlorophenyl-phenylether	40.000	38.795	3.0	115	0.00
62	4-Nitroaniline	40.000	40.041	-0.1	107	0.00
63	Azobenzene	40.000	39.196	2.0	113	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.835	-9.6	109	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.112	-2.8	111	0.00
67	4-Bromophenyl-phenylether	40.000	40.485	-1.2	109	0.00
68	Hexachlorobenzene	40.000	40.044	-0.1	111	0.00
69	Atrazine	40.000	41.994	-5.0	109	0.00
70 C	Pentachlorophenol	40.000	42.517	-6.3	107	0.00
71	Phenanthrene	40.000	39.333	1.7	108	0.00
72	Anthracene	40.000	39.890	0.3	107	0.00
73	Carbazole	40.000	39.044	2.4	104	0.00
74	Di-n-butylphthalate	40.000	39.291	1.8	103	0.00
75 C	Fluoranthene	40.000	37.830	5.4	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzidine	40.000	43.244	-8.1	94	0.00
78	Pyrene	40.000	42.133	-5.3	101	0.00
79 S	Terphenyl-d14	80.000	83.561	-4.5	101	0.00
80	Butylbenzylphthalate	40.000	42.213	-5.5	98	0.00
81	Benzo(a)anthracene	40.000	40.045	-0.1	97	0.00
82	3,3'-Dichlorobenzidine	40.000	39.886	0.3	95	0.00
83	Chrysene	40.000	39.947	0.1	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.532	-3.8	98	0.00
85 c	Di-n-octyl phthalate	40.000	40.624	-1.6	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	95	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.891	5.3	96	0.00
88	Benzo(b)fluoranthene	40.000	39.013	2.5	90	0.00
89	Benzo(k)fluoranthene	40.000	40.901	-2.3	100	0.00
90 C	Benzo(a)pyrene	40.000	40.000	0.0	95	0.00
91	Dibenzo(a,h)anthracene	40.000	38.000	5.0	96	0.00
92	Benzo(g,h,i)perylene	40.000	37.965	5.1	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120721\
Data File : BM033332.D
Acq On : 07 Dec 2021 20:16
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM120721

Quant Time: Dec 08 02:49:43 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM120721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 08 02:33:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0