

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120723\
 Data File : BM043185.D
 Acq On : 07 Dec 2023 17:30
 Operator : MA/JU
 Sample : SST04072
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040037

Quant Time: Dec 08 00:04:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 00:03:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	434513	20.000	ng/u1	0.00
20) Naphthalene-d8	10.822	136	1967811	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.633	164	1290613	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.374	188	2768193	20.000	ng/u1	0.00
79) Chrysene-d12	21.538	240	2660698	20.000	ng/u1	0.00
88) Perylene-d12	23.944	264	2684467	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.398	96	176557	15.949	ng/uL	0.00
4) Pyridine-d5	3.816	84	1404768	43.485	ng/u1	0.00
7) Phenol-d5	7.169	99	1824595	43.425	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	1139878	42.969	ng/u1	0.00
11) 2-Chlorophenol-d4	7.539	132	1395970	42.449	ng/u1	0.00
15) 4-Methylphenol-d8	8.728	113	1468138	43.804	ng/u1	0.00
21) Nitrobenzene-d5	9.186	128	728500	43.434	ng/u1	0.00
24) 2-Nitrophenol-d4	9.904	143	745083	46.068	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.439	165	1554105	43.110	ng/u1	0.00
31) 4-Chloroaniline-d4	10.963	131	2230039	42.952	ng/u1	0.00
46) Dimethylphthalate-d6	14.057	166	4747021	41.959	ng/u1	0.00
49) Acenaphthylene-d8	14.333	160	5521030	42.181	ng/u1	0.00
54) 4-Nitrophenol-d4	14.839	143	877837	44.399	ng/u1	0.01
60) Fluorene-d10	15.627	176	4134195	41.921	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.751	200	635818	46.418	ng/u1	0.00
73) Anthracene-d10	17.474	188	6362827	42.487	ng/u1	0.00
81) Pyrene-d10	19.756	212	7271892	42.714	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.791	264	6753293	43.113	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.440	88	195133	16.046	ng/uL	99
5) Pyridine	3.840	79	1439341	43.241	ng/u1	99
6) Benzaldehyde	7.157	77	1037209	43.415	ng/u1	100
8) Phenol	7.198	94	1804563	43.248	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.445	93	1460691	42.887	ng/u1	99
12) 2-Chlorophenol	7.575	128	1408260	42.211	ng/u1	99
13) 2-Methylphenol	8.457	108	1415004	43.843	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.545	45	2438099	42.498	ng/u1	100
16) Acetophenone	8.845	105	2364216	43.026	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.839	70	1271383	42.287	ng/u1	98
18) 4-Methylphenol	8.792	108	1542160	43.692	ng/u1	100
19) Hexachloroethane	9.092	117	586132	41.448	ng/u1	99
22) Nitrobenzene	9.227	77	1705968	42.304	ng/u1	99
23) Isophorone	9.757	82	3595922	41.848	ng/u1	99
25) 2-Nitrophenol	9.939	139	796743	44.948	ng/u1	99
26) 2,4-Dimethylphenol	9.992	107	1650684	47.496	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.245	93	2079007	41.151	ng/u1	99
29) 2,4-Dichlorophenol	10.469	162	1490611	42.602	ng/u1	98
30) Naphthalene	10.874	128	4967133	41.124	ng/u1	99
32) 4-Chloroaniline	10.986	127	2101261	42.890	ng/u1	99
33) Hexachlorobutadiene	11.145	225	895975	41.096	ng/u1	99
34) Caprolactam	11.769	113	235658	22.388	ng/u1	94
35) 4-Chloro-3-methylphenol	12.104	107	1553125	42.273	ng/u1	99
36) 2-Methylnaphthalene	12.474	142	3585043	41.256	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120723\
 Data File : BM043185.D
 Acq On : 07 Dec 2023 17:30
 Operator : MA/JU
 Sample : SST04072
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST040037

Quant Time: Dec 08 00:04:25 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 00:03:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.692	142	3635366	41.216	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.833	216	1853860	42.457	ng/ul	100
40) Hexachlorocyclopentadiene	12.804	237	1072090	45.372	ng/ul	100
41) 2,4,6-Trichlorophenol	13.074	196	1200662	44.189	ng/ul	100
42) 2,4,5-Trichlorophenol	13.145	196	1294185	43.763	ng/ul	98
43) 1,1'-Biphenyl	13.480	154	4742633	41.540	ng/ul	99
44) 2-Chloronaphthalene	13.521	162	3699676	41.804	ng/ul	100
45) 2-Nitroaniline	13.727	65	1047946	44.810	ng/ul	99
47) Dimethylphthalate	14.110	163	4647683	41.885	ng/ul	100
48) 2,6-Dinitrotoluene	14.227	165	911020	44.562	ng/ul	98
50) Acenaphthylene	14.362	152	6247252	42.231	ng/ul	100
51) 3-Nitroaniline	14.551	138	986790	43.840	ng/ul	100
52) Acenaphthene	14.704	153	4068222	41.692	ng/ul	100
53) 2,4-Dinitrophenol	14.757	184	383602	46.446	ng/ul	99
55) 4-Nitrophenol	14.851	109	633796	43.695	ng/ul	96
56) Dibenzofuran	15.033	168	5503260	41.288	ng/ul	99
57) 2,4-Dinitrotoluene	15.009	165	1297663	44.455	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.257	232	1076255	43.784	ng/ul	96
59) Diethylphthalate	15.468	149	4681753	41.946	ng/ul	99
61) Fluorene	15.680	166	4883744	42.327	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.680	204	2397166	42.053	ng/ul	99
63) 4-Nitroaniline	15.715	138	1017874	44.315	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.762	198	703884	46.047	ng/ul	99
67) N-Nitrosodiphenylamine	15.892	169	3898156	41.992	ng/ul	99
68) 4-Bromophenyl-phenylether	16.574	248	1407910	42.059	ng/ul	96
69) Hexachlorobenzene	16.668	284	1663648	41.885	ng/ul	98
70) Atrazine	16.851	200	1263152	46.150	ng/ul	99
71) Pentachlorophenol	17.015	266	1005524	45.253	ng/ul	98
72) Phenanthrene	17.421	178	7354378	42.605	ng/ul	99
74) Anthracene	17.509	178	7481934	43.028	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.439	216	1876478	42.848	ng/ul	99
76) Pentachlorobenzene	14.945	250	1866135	41.931	ng/ul	98
77) Carbazole	17.780	167	6686984	43.738	ng/ul	100
78) Di-n-butylphthalate	18.350	149	8479282	45.639	ng/ul	99
80) Fluoranthene	19.421	202	8587122	43.971	ng/ul	98
82) Pyrene	19.786	202	8848767	43.558	ng/ul	100
83) Butylbenzylphthalate	20.691	149	3670997	45.428	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.462	252	2993117	46.050	ng/ul	99
85) Benzo(a)anthracene	21.527	228	8658419	42.443	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.468	149	6004924	47.075	ng/ul	98
87) Chrysene	21.580	228	7784745	42.726	ng/ul	99
89) Di-n-octyl phthalate	22.409	149	9340887	48.043	ng/ul	100
90) Benzo(b)fluoranthene	23.215	252	8265945	43.156	ng/ul	99
91) Benzo(k)fluoranthene	23.262	252	8139926	42.670	ng/ul	100
93) Benzo(a)pyrene	23.844	252	7588490	42.864	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.450	276	8823612	43.753	ng/ul	98
95) Dibenzo(a,h)anthracene	26.473	278	7331865	43.484	ng/ul	99
96) Benzo(g,h,i)perylene	27.215	276	6539520	43.043	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 16.046 ng/uL
RT: 3.440 min Scan# 77
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion: 88 Resp: 195133
Ion Ratio Lower Upper
88 100
43 30.9 25.0 37.4
58 75.1 60.6 90.8



#5
Pyridine
Concen: 43.241 ng/uL
RT: 3.840 min Scan# 145
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion: 79 Resp: 1439341
Ion Ratio Lower Upper
79 100
52 80.7 64.9 97.3
51 37.7 31.0 46.4









#6
Benzaldehyde
Concen: 43.415 ng/ul
RT: 7.157 min Scan# 709
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion: 77 Resp: 1037209
Ion Ratio Lower Upper
77 100
105 96.0 76.3 114.5
106 93.4 74.6 112.0



#8
Phenol
Concen: 43.248 ng/ul
RT: 7.198 min Scan# 716
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion: 94 Resp: 1804563
Ion Ratio Lower Upper
94 100
65 29.3 23.5 35.3
66 36.8 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 42.887 ng/ul
RT: 7.445 min Scan# 758
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion: 93 Resp: 1460691
Ion Ratio Lower Upper
93 100
63 80.0 63.0 94.4
95 32.8 26.0 39.0



#12
2-Chlorophenol
Concen: 42.211 ng/ul
RT: 7.575 min Scan# 780
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:128 Resp: 1408260
Ion Ratio Lower Upper
128 100
64 47.4 38.3 57.5
130 32.5 26.6 40.0









#13
2-Methylphenol
Concen: 43.843 ng/ul
RT: 8.457 min Scan# 913
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:108 Resp: 1415004
Ion Ratio Lower Upper
108 100
107 86.3 71.8 107.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 42.498 ng/ul
RT: 8.545 min Scan# 945
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion: 45 Resp: 2438099
Ion Ratio Lower Upper
45 100
77 12.7 10.2 15.2
79 9.3 7.2 10.8









#16
Acetophenone
Concen: 43.026 ng/ul
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:105 Resp: 2364216
Ion Ratio Lower Upper
105 100
77 78.0 62.7 94.1
51 29.0 23.8 35.8



#17
N-Nitroso-di-n-propylamine
Concen: 42.287 ng/ul
RT: 8.839 min Scan# 995
Delta R.T. 0.012 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion: 70 Resp: 1271383
Ion Ratio Lower Upper
70 100
42 64.3 52.3 78.5
101 9.5 7.0 10.6
130 21.2 16.2 24.2









#18
4-Methylphenol
Concen: 43.692 ng/ul
RT: 8.792 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:108 Resp: 1542160
Ion Ratio Lower Upper
108 100
107 108.3 86.4 129.6



#19
Hexachloroethane
Concen: 41.448 ng/ul
RT: 9.092 min Scan# 1038
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:117 Resp: 586132
Ion Ratio Lower Upper
117 100
201 100.2 79.2 118.8
199 62.2 49.8 74.8









#22
Nitrobenzene
Concen: 42.304 ng/ul
RT: 9.227 min Scan# 1105
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion: 77 Resp: 1705968
Ion Ratio Lower Upper
77 100
123 44.9 35.6 53.4
65 13.5 11.0 16.4



#23
Isophorone
Concen: 41.848 ng/ul
RT: 9.757 min Scan# 1151
Delta R.T. 0.012 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion: 82 Resp: 3595922
Ion Ratio Lower Upper
82 100
95 6.7 5.2 7.8
138 18.3 14.0 21.0









#25
2-Nitrophenol
Concen: 44.948 ng/ul
RT: 9.939 min Scan# 1181
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:139 Resp: 796743
Ion Ratio Lower Upper
139 100
65 49.6 40.6 60.8
109 27.6 22.6 33.8



#26
2,4-Dimethylphenol
Concen: 47.496 ng/ul
RT: 9.992 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:107 Resp: 1650684
Ion Ratio Lower Upper
107 100
121 50.2 42.2 63.2
122 89.5 72.0 108.0





#27

Bis(2-Chloroethoxy)methane

Concen: 41.151 ng/ul

RT: 10.245 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion: 93 Resp: 2079007

Ion Ratio Lower Upper

93 100

95 33.2 26.2 39.2

123 11.9 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 42.602 ng/ul

RT: 10.469 min Scan# 1272

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:162 Resp: 1490611


Ion Ratio Lower Upper

162 100

164 65.7 51.4 77.0

98 33.0 26.9 40.3









#30
Naphthalene
Concen: 41.124 ng/ul
RT: 10.874 min Scan# 1341
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:128 Resp: 4967133
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 12.9 10.4 15.6



#32
4-Chloroaniline
Concen: 42.890 ng/ul
RT: 10.986 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:127 Resp: 2101261
Ion Ratio Lower Upper
127 100
129 32.2 26.0 39.0









#33
Hexachlorobutadiene
Concen: 41.096 ng/ul
RT: 11.145 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:225 Resp: 895975
Ion Ratio Lower Upper
225 100
223 63.5 50.2 75.4
227 64.2 52.0 78.0



#34
Caprolactam
Concen: 22.388 ng/ul
RT: 11.769 min Scan# 1493
Delta R.T. 0.018 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:113 Resp: 235658
Ion Ratio Lower Upper
113 100
55 183.6 139.5 209.3
56 135.8 103.5 155.3









#35
 4-Chloro-3-methylphenol
 Concen: 42.273 ng/ul
 RT: 12.104 min Scan# 1550
 Delta R.T. 0.006 min
 Lab File: BM043185.D
 Acq: 07 Dec 2023 17:30




Instrument :
 BNA_M
 ClientSampleId :
 SST040037


Tgt Ion:107 Resp: 1553125
 Ion Ratio Lower Upper
 107 100
 144 27.8 22.0 33.0
 142 84.9 67.3 100.9



#36
 2-Methylnaphthalene
 Concen: 41.256 ng/ul
 RT: 12.474 min Scan# 1613
 Delta R.T. -0.000 min
 Lab File: BM043185.D
 Acq: 07 Dec 2023 17:30

Tgt Ion:142 Resp: 3585043
 Ion Ratio Lower Upper
 142 100
 141 87.9 70.0 105.0





#37

1-Methylnaphthalene

Concen: 41.216 ng/ul

RT: 12.692 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:142 Resp: 3635366

Ion Ratio Lower Upper

142 100

141 90.9 71.9 107.9

116 3.6 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 42.457 ng/ul

RT: 12.833 min Scan# 1674

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Tgt Ion:216 Resp: 1853860




Ion Ratio Lower Upper


216 100

214 79.0 63.1 94.7

179 20.9 16.4 24.6

108 16.7 13.6 20.4









#40
Hexachlorocyclopentadiene
Concen: 45.372 ng/ul
RT: 12.804 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:237 Resp: 1072090
Ion Ratio Lower Upper
237 100
235 62.8 50.5 75.7
272 13.0 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 44.189 ng/ul
RT: 13.074 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:196 Resp: 1200662
Ion Ratio Lower Upper
196 100
198 96.2 77.0 115.4
200 30.9 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 43.763 ng/ul
RT: 13.145 min Scan# 1726
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:196 Resp: 1294185
Ion Ratio Lower Upper
196 100
198 97.4 76.1 114.1
200 31.7 24.6 36.8



#43
1,1'-Biphenyl
Concen: 41.540 ng/ul
RT: 13.480 min Scan# 1784
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:154 Resp: 4742633
Ion Ratio Lower Upper
154 100
153 39.9 31.8 47.6
76 12.3 9.3 13.9





#44

2-Chloronaphthalene

Concen: 41.804 ng/ul

RT: 13.521 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:162 Resp: 3699676

Ion Ratio Lower Upper

162 100

164 33.0 26.4 39.6

127 37.2 30.0 45.0



#45

2-Nitroaniline

Concen: 44.810 ng/ul

RT: 13.727 min Scan# 1826

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion: 65 Resp: 1047946


Ion Ratio Lower Upper

65 100

92 67.0 53.4 80.0

138 105.6 86.2 129.2









#47
Dimethylphthalate
Concen: 41.885 ng/ul
RT: 14.110 min Scan# 1891
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:163 Resp: 4647683
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 44.562 ng/ul
RT: 14.227 min Scan# 1911
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:165 Resp: 911020
Ion Ratio Lower Upper
165 100
89 58.0 47.5 71.3
121 21.1 17.1 25.7





#50

Acenaphthylene

Concen: 42.231 ng/ul

RT: 14.362 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:152 Resp: 6247252

Ion Ratio Lower Upper

152 100

151 20.0 16.0 24.0

153 13.2 10.3 15.5



#51

3-Nitroaniline

Concen: 43.840 ng/ul

RT: 14.551 min Scan# 1966

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:138 Resp: 986790


Ion Ratio Lower Upper

138 100

108 10.7 9.4 14.0

92 109.8 87.9 131.9









#52
Acenaphthene
Concen: 41.692 ng/ul
RT: 14.704 min Scan# 1991
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:153 Resp: 4068222
Ion Ratio Lower Upper
153 100
152 48.5 38.6 58.0
154 89.2 71.6 107.4



#53
2,4-Dinitrophenol
Concen: 46.446 ng/ul
RT: 14.757 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:184 Resp: 383602
Ion Ratio Lower Upper
184 100
63 63.9 52.2 78.4
154 60.5 47.8 71.8









#55
4-Nitrophenol
Concen: 43.695 ng/ul
RT: 14.851 min Scan# 2016
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:109 Resp: 633796
Ion Ratio Lower Upper
109 100
139 140.1 114.7 172.1
65 142.3 109.5 164.3



#56
Dibenzofuran
Concen: 41.288 ng/ul
RT: 15.033 min Scan# 2048
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:168 Resp: 5503260
Ion Ratio Lower Upper
168 100
139 36.5 28.6 43.0





#57

2,4-Dinitrotoluene

Concen: 44.455 ng/ul

RT: 15.009 min Scan# 2044

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:165 Resp: 1297663

Ion Ratio Lower Upper

165 100

63 43.9 34.6 52.0

182 2.6 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 43.784 ng/ul

RT: 15.257 min Scan# 2086

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Tgt Ion:232 Resp: 1076255




Ion Ratio Lower Upper


232 100

131 43.9 37.0 55.4

130 2.2 1.9 2.9

166 30.3 26.1 39.1









#59
Diethylphthalate
Concen: 41.946 ng/ul
RT: 15.468 min Scan# 2121
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:149 Resp: 4681753
Ion Ratio Lower Upper
149 100
177 23.0 18.2 27.2
150 12.3 10.0 15.0



#61
Fluorene
Concen: 42.327 ng/ul
RT: 15.680 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:166 Resp: 4883744
Ion Ratio Lower Upper
166 100
165 97.4 77.5 116.3
167 13.8 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 42.053 ng/ul
RT: 15.680 min Scan# 2158
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Instrument :
BNA_M
ClientSampleId :
SSTD040037

Tgt Ion:204 Resp: 2397166




Ion	Ratio	Lower	Upper
204	100		
206	32.5	26.2	39.2
141	56.3	44.2	66.4




#63
4-Nitroaniline
Concen: 44.315 ng/ul
RT: 15.715 min Scan# 2164
Delta R.T. 0.012 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:138 Resp: 1017874

Ion	Ratio	Lower	Upper
138	100		
92	49.0	40.1	60.1
108	76.7	63.0	94.6









#66
4,6-Dinitro-2-methylphenol
Concen: 46.047 ng/ul
RT: 15.762 min Scan# 2172
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:198 Resp: 703884
Ion Ratio Lower Upper
198 100
182 3.5 3.2 4.8
77 25.7 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 41.992 ng/ul
RT: 15.892 min Scan# 2194
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:169 Resp: 3898156
Ion Ratio Lower Upper
169 100
168 66.6 53.4 80.2
167 37.3 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 42.059 ng/ul
RT: 16.574 min Scan# 210
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:248 Resp: 1407910
Ion Ratio Lower Upper
248 100
250 97.0 78.9 118.3
141 68.2 60.2 90.2



#69
Hexachlorobenzene
Concen: 41.885 ng/ul
RT: 16.668 min Scan# 2326
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:284 Resp: 1663648
Ion Ratio Lower Upper
284 100
142 35.3 26.7 40.1
249 30.7 24.0 36.0





#70

Atrazine

Concen: 46.150 ng/ul

RT: 16.851 min Scan# 21

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:200 Resp: 1263152

Ion Ratio Lower Upper

200 100

173 26.7 20.6 30.8

215 49.5 39.4 59.2



#71

Pentachlorophenol

Concen: 45.253 ng/ul

RT: 17.015 min Scan# 2385

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:266 Resp: 1005524


Ion Ratio Lower Upper

266 100

264 61.8 50.5 75.7

268 64.1 50.4 75.6









#72
 Phenanthrene
 Concen: 42.605 ng/ul
 RT: 17.421 min Scan# 2453
 Delta R.T. 0.006 min
 Lab File: BM043185.D
 Acq: 07 Dec 2023 17:30




Instrument :
 BNA_M
 ClientSampleId :
 SSTD040037


Tgt Ion:178 Resp: 7354378
 Ion Ratio Lower Upper
 178 100
 179 15.5 12.2 18.2
 176 19.8 15.7 23.5



#74
 Anthracene
 Concen: 43.028 ng/ul
 RT: 17.509 min Scan# 2469
 Delta R.T. -0.000 min
 Lab File: BM043185.D
 Acq: 07 Dec 2023 17:30

Tgt Ion:178 Resp: 7481934
 Ion Ratio Lower Upper
 178 100
 179 15.7 12.1 18.1
 176 19.6 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 42.848 ng/uL
RT: 13.439 min Scan# 1776
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:216 Resp: 1876478
Ion Ratio Lower Upper
216 100
214 78.5 62.2 93.2
218 48.4 37.6 56.4



#76
Pentachlorobenzene
Concen: 41.931 ng/uL
RT: 14.945 min Scan# 2033
Delta R.T. -0.000 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:250 Resp: 1866135
Ion Ratio Lower Upper
250 100
248 63.6 51.8 77.8
252 63.6 51.9 77.9





#77

Carbazole

Concen: 43.738 ng/ul

RT: 17.780 min Scan# 2

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:167 Resp: 6686984

Ion Ratio Lower Upper

167 100

166 21.6 17.1 25.7

139 12.6 9.9 14.9



#78

Di-n-butylphthalate

Concen: 45.639 ng/ul

RT: 18.350 min Scan# 2612

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:149 Resp: 8479282


Ion Ratio Lower Upper

149 100

150 9.4 7.2 10.8

104 4.8 3.8 5.6





#80

Fluoranthene

Concen: 43.971 ng/ul

RT: 19.421 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:202 Resp: 8587122

Ion Ratio Lower Upper

202 100

101 11.1 8.2 12.2

100 8.2 6.1 9.1



#82

Pyrene

Concen: 43.558 ng/ul

RT: 19.786 min Scan# 2856

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:202 Resp: 8848767


Ion Ratio Lower Upper

202 100

101 12.5 10.2 15.2

100 10.0 8.2 12.2





#83

Butylbenzylphthalate

Concen: 45.428 ng/ul

RT: 20.691 min Scan# 3009

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:149 Resp: 3670997

Ion Ratio Lower Upper

149 100

91 66.5 56.9 85.3

206 21.5 16.1 24.1



#84

3,3'-Dichlorobenzidine

Concen: 46.050 ng/ul

RT: 21.462 min Scan# 3141

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:252 Resp: 2993117


Ion Ratio Lower Upper

252 100

254 64.5 51.3 76.9

126 10.3 8.8 13.2





#85

Benzo(a)anthracene

Concen: 42.443 ng/ul

RT: 21.527 min Scan# 3151

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:228 Resp: 8658419

Ion Ratio Lower Upper

228 100

229 20.3 15.8 23.8

226 28.5 22.2 33.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 47.075 ng/ul

RT: 21.468 min Scan# 3142

Delta R.T. -0.000 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:149 Resp: 6004924


Ion Ratio Lower Upper

149 100

167 29.4 22.6 34.0

279 4.6 3.7 5.5





#87

Chrysene

Concen: 42.726 ng/ul

RT: 21.580 min Scan# 31

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:228 Resp: 7784745

Ion Ratio Lower Upper

228 100

226 32.0 24.8 37.2

229 20.5 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 48.043 ng/ul




RT: 22.409 min Scan# 3302


Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Tgt Ion:149 Resp: 9340887









#90
Benzo(b)fluoranthene
Concen: 43.156 ng/ul
RT: 23.215 min Scan# 3438
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30




Instrument :
BNA_M
ClientSampleId :
SSTD040037


Tgt Ion:252 Resp: 8265945
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.2
125 9.4 7.2 10.8



#91
Benzo(k)fluoranthene
Concen: 42.670 ng/ul
RT: 23.262 min Scan# 3447
Delta R.T. 0.006 min
Lab File: BM043185.D
Acq: 07 Dec 2023 17:30

Tgt Ion:252 Resp: 8139926
Ion Ratio Lower Upper
252 100
253 22.5 17.9 26.9
125 9.9 7.7 11.5





#93

Benzo(a)pyrene

Concen: 42.864 ng/ul

RT: 23.844 min Scan# 3

Delta R.T. 0.012 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:252 Resp: 7588490

Ion Ratio Lower Upper

252 100

253 22.5 17.8 26.6

125 10.4 8.4 12.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.753 ng/ul

RT: 26.450 min Scan# 3989

Delta R.T. 0.012 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30




Tgt Ion:276 Resp: 8823612


Ion Ratio Lower Upper

276 100

138 22.2 16.9 25.3

277 24.8 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 43.484 ng/ul

RT: 26.473 min Scan# 3993

Delta R.T. 0.006 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Instrument :

BNA_M

ClientSampleId :

SSTD040037





Tgt Ion:278 Resp: 7331865

Ion Ratio Lower Upper

278 100

139 16.2 11.9 17.9

279 24.2 19.4 29.2



#96

Benzo(g,h,i)perylene

Concen: 43.043 ng/ul

RT: 27.215 min Scan# 4119

Delta R.T. 0.012 min

Lab File: BM043185.D

Acq: 07 Dec 2023 17:30

Tgt Ion:276 Resp: 6539520

Ion Ratio Lower Upper

276 100

138 20.6 16.2 24.2

277 24.0 18.7 28.1

