

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120723\
 Data File : BM043188.D
 Acq On : 07 Dec 2023 19:18
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV040

Quant Time: Dec 08 00:43:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 00:42:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.010	152	453266	20.000	ng/ul	0.00
20) Naphthalene-d8	10.822	136	2023676	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.633	164	1355221	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.374	188	2920491	20.000	ng/ul	0.00
79) Chrysene-d12	21.539	240	2724362	20.000	ng/ul	0.00
88) Perylene-d12	23.939	264	2905147	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.405	96	74690	6.467	ng/uL	0.00
4) Pyridine-d5	3.816	84	634459	18.515	ng/ul	0.00
7) Phenol-d5	7.169	99	820650	18.237	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.351	67	573266	20.579	ng/ul	0.00
11) 2-Chlorophenol-d4	7.540	132	634508	18.496	ng/ul	0.00
15) 4-Methylphenol-d8	8.722	113	669248	18.799	ng/ul	0.00
21) Nitrobenzene-d5	9.181	128	337803	19.584	ng/ul	0.00
24) 2-Nitrophenol-d4	9.904	143	338744	20.366	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.434	165	700321	18.890	ng/ul	0.00
31) 4-Chloroaniline-d4	10.957	131	1032920	19.364	ng/ul	0.00
46) Dimethylphthalate-d6	14.057	166	2262997	19.049	ng/ul	0.00
49) Acenaphthylene-d8	14.333	160	2606214	18.963	ng/ul	0.00
54) 4-Nitrophenol-d4	14.827	143	413107	19.406	ng/ul	0.00
60) Fluorene-d10	15.621	176	1913068	18.474	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	280863	17.473	ng/ul	0.00
73) Anthracene-d10	17.474	188	2940429	18.610	ng/ul	0.00
81) Pyrene-d10	19.751	212	3316299	19.024	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.780	264	3119310	18.401	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.440	88	82171	6.478	ng/uL	97
5) Pyridine	3.840	79	553596	15.718	ng/ul	99
6) Benzaldehyde	7.151	77	501384	21.648	ng/ul	100
8) Phenol	7.198	94	763892	17.165	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.445	93	621840	17.330	ng/ul	98
12) 2-Chlorophenol	7.569	128	611117	17.560	ng/ul	98
13) 2-Methylphenol	8.451	108	591445	17.187	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.540	45	518248	8.687	ng/ul	100
16) Acetophenone	8.845	105	1042043	18.211	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.828	70	551892	17.597	ng/ul	98
18) 4-Methylphenol	8.787	108	656434	17.580	ng/ul	98
19) Hexachloroethane	9.087	117	247578	16.783	ng/ul	97
22) Nitrobenzene	9.228	77	731061	17.628	ng/ul	97
23) Isophorone	9.751	82	1525090	17.258	ng/ul	98
25) 2-Nitrophenol	9.934	139	342128	18.768	ng/ul	99
26) 2,4-Dimethylphenol	9.992	107	569742	15.941	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.239	93	891478	17.158	ng/ul	99
29) 2,4-Dichlorophenol	10.463	162	640480	17.800	ng/ul	99
30) Naphthalene	10.869	128	2112350	17.006	ng/ul	100
32) 4-Chloroaniline	10.986	127	923817	18.361	ng/ul	99
33) Hexachlorobutadiene	11.145	225	367092	16.373	ng/ul	99
34) Caprolactam	11.751	113	221709	20.722	ng/ul	91
35) 4-Chloro-3-methylphenol	12.092	107	680570	18.012	ng/ul	99
36) 2-Methylnaphthalene	12.475	142	1512985	16.930	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM120723\
 Data File : BM043188.D
 Acq On : 07 Dec 2023 19:18
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV040


Quant Time: Dec 08 00:43:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120723.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 08 00:42:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.686	142	1484381	16.364	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.833	216	795936	17.359	ng/ul	100
40) Hexachlorocyclopentadiene	12.804	237	370887	14.256	ng/ul	100
41) 2,4,6-Trichlorophenol	13.075	196	508994	17.840	ng/ul	99
42) 2,4,5-Trichlorophenol	13.139	196	553868	17.836	ng/ul	99
43) 1,1'-Biphenyl	13.480	154	2134370	17.803	ng/ul	99
44) 2-Chloronaphthalene	13.516	162	1569681	16.891	ng/ul	100
45) 2-Nitroaniline	13.727	65	462170	18.820	ng/ul	98
47) Dimethylphthalate	14.104	163	1988291	17.064	ng/ul	99
48) 2,6-Dinitrotoluene	14.222	165	416450	19.399	ng/ul	98
50) Acenaphthylene	14.357	152	2685795	17.290	ng/ul	99
51) 3-Nitroaniline	14.551	138	472005	19.819	ng/ul	98
52) Acenaphthene	14.698	153	1712154	16.710	ng/ul	99
53) 2,4-Dinitrophenol	14.751	184	160372	16.273	ng/ul	99
55) 4-Nitrophenol	14.845	109	283699	18.431	ng/ul	100
56) Dibenzofuran	15.033	168	2449197	17.499	ng/ul	98
57) 2,4-Dinitrotoluene	15.004	165	553443	18.056	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.251	232	491936	19.059	ng/ul	99
59) Diethylphthalate	15.463	149	2002846	17.089	ng/ul	100
61) Fluorene	15.680	166	2075403	17.130	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.680	204	974570	16.282	ng/ul	97
63) 4-Nitroaniline	15.704	138	492140	20.566	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.757	198	272701	15.516	ng/ul	98
67) N-Nitrosodiphenylamine	15.892	169	1745810	17.826	ng/ul	100
68) 4-Bromophenyl-phenylether	16.568	248	584544	16.552	ng/ul	98
69) Hexachlorobenzene	16.668	284	704001	16.800	ng/ul	98
70) Atrazine	16.845	200	584193	21.152	ng/ul	97
71) Pentachlorophenol	17.015	266	391122	15.851	ng/ul	99
72) Phenanthrene	17.415	178	3149040	17.291	ng/ul	100
74) Anthracene	17.510	178	3182176	17.346	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.433	216	760335	16.456	ng/uL	98
76) Pentachlorobenzene	14.945	250	773565	16.475	ng/uL	99
77) Carbazole	17.780	167	2934667	19.475	ng/ul	100
78) Di-n-butylphthalate	18.351	149	3575746	18.243	ng/ul	100
80) Fluoranthene	19.421	202	3541479	17.711	ng/ul	100
82) Pyrene	19.780	202	3733166	17.947	ng/ul	100
83) Butylbenzylphthalate	20.686	149	1492852	18.042	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.456	252	1356507	20.951	ng/ul	100
85) Benzo(a)anthracene	21.521	228	3641482	17.433	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.462	149	2461813	18.848	ng/ul	99
87) Chrysene	21.574	228	3261239	17.481	ng/ul	99
89) Di-n-octyl phthalate	22.403	149	3877529	19.404	ng/ul	100
90) Benzo(b)fluoranthene	23.203	252	3374810	16.281	ng/ul	99
91) Benzo(k)fluoranthene	23.256	252	3365424	16.302	ng/ul	99
93) Benzo(a)pyrene	23.833	252	2989651	15.604	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.438	276	3609434	16.539	ng/ul	99
95) Dibenzo(a,h)anthracene	26.462	278	3027491	16.591	ng/ul	99
96) Benzo(g,h,i)perylene	27.197	276	2716986	16.524	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 08 00:43:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120723.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 08 00:42:52 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.478 ng/uL
RT: 3.440 min Scan# 71
Delta R.T. -0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion: 88 Resp: 82171
Ion Ratio Lower Upper
88 100
43 32.0 25.0 37.4
58 73.2 60.6 90.8



#5
Pyridine
Concen: 15.718 ng/uL
RT: 3.840 min Scan# 145
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion: 79 Resp: 553596
Ion Ratio Lower Upper
79 100
52 79.8 64.9 97.3
51 38.1 31.0 46.4









#6
Benzaldehyde
Concen: 21.648 ng/ul
RT: 7.151 min Scan# 709
Delta R.T. -0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion: 77 Resp: 501384
Ion Ratio Lower Upper
77 100
105 95.9 76.3 114.5
106 93.6 74.6 112.0



#8
Phenol
Concen: 17.165 ng/ul
RT: 7.198 min Scan# 716
Delta R.T. 0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion: 94 Resp: 763892
Ion Ratio Lower Upper
94 100
65 27.9 23.5 35.3
66 36.2 29.9 44.9









#10
Bis(2-Chloroethyl)ether
Concen: 17.330 ng/ul
RT: 7.445 min Scan# 758
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion: 93 Resp: 621840
Ion Ratio Lower Upper
93 100
63 76.3 63.0 94.4
95 32.1 26.0 39.0



#12
2-Chlorophenol
Concen: 17.560 ng/ul
RT: 7.569 min Scan# 779
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:128 Resp: 611117
Ion Ratio Lower Upper
128 100
64 46.7 38.3 57.5
130 32.2 26.6 40.0





#13

2-Methylphenol

Concen: 17.187 ng/ul

RT: 8.451 min Scan# 91

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :





SICV040

Tgt Ion:108 Resp: 591445

Ion Ratio Lower Upper

108 100

107 88.7 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 8.687 ng/ul

RT: 8.540 min Scan# 944

Delta R.T. -0.006 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion: 45 Resp: 518248


Ion Ratio Lower Upper

45 100

77 12.9 10.2 15.2

79 9.1 7.2 10.8









#16
Acetophenone
Concen: 18.211 ng/ul
RT: 8.845 min Scan# 996
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:105 Resp: 1042043
Ion Ratio Lower Upper
105 100
77 77.1 62.7 94.1
51 29.0 23.8 35.8



#17
N-Nitroso-di-n-propylamine
Concen: 17.597 ng/ul
RT: 8.828 min Scan# 993
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion: 70 Resp: 551892
Ion Ratio Lower Upper
70 100
42 64.2 52.3 78.5
101 10.0 7.0 10.6
130 21.0 16.2 24.2









#18
4-Methylphenol
Concen: 17.580 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:108 Resp: 656434
Ion Ratio Lower Upper
108 100
107 106.2 86.4 129.6



#19
Hexachloroethane
Concen: 16.783 ng/ul
RT: 9.087 min Scan# 1037
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:117 Resp: 247578
Ion Ratio Lower Upper
117 100
201 95.3 79.2 118.8
199 61.6 49.8 74.8









#22
Nitrobenzene
Concen: 17.628 ng/ul
RT: 9.228 min Scan# 1061
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion: 77 Resp: 731061
Ion Ratio Lower Upper
77 100
123 46.2 35.6 53.4
65 12.7 11.0 16.4



#23
Isophorone
Concen: 17.258 ng/ul
RT: 9.751 min Scan# 1150
Delta R.T. 0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion: 82 Resp: 1525090
Ion Ratio Lower Upper
82 100
95 6.5 5.2 7.8
138 18.5 14.0 21.0









#25
2-Nitrophenol
Concen: 18.768 ng/ul
RT: 9.934 min Scan# 1181
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:139 Resp: 342128
Ion Ratio Lower Upper
139 100
65 51.2 40.6 60.8
109 28.2 22.6 33.8



#26
2,4-Dimethylphenol
Concen: 15.941 ng/ul
RT: 9.992 min Scan# 1191
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:107 Resp: 569742
Ion Ratio Lower Upper
107 100
121 51.2 42.2 63.2
122 88.6 72.0 108.0





#27

Bis(2-Chloroethoxy)methane

Concen: 17.158 ng/ul

RT: 10.239 min Scan# 1233

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion: 93 Resp: 891478

Ion Ratio Lower Upper

93 100

95 32.9 26.2 39.2

123 11.9 9.2 13.8



#29

2,4-Dichlorophenol

Concen: 17.800 ng/ul

RT: 10.463 min Scan# 1271

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:162 Resp: 640480


Ion Ratio Lower Upper

162 100

164 64.5 51.4 77.0

98 34.6 26.9 40.3









#30
Naphthalene
Concen: 17.006 ng/ul
RT: 10.869 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:128 Resp: 2112350
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.4
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 18.361 ng/ul
RT: 10.986 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:127 Resp: 923817
Ion Ratio Lower Upper
127 100
129 33.1 26.0 39.0









#33
Hexachlorobutadiene
Concen: 16.373 ng/ul
RT: 11.145 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:225 Resp: 367092
Ion Ratio Lower Upper
225 100
223 62.4 50.2 75.4
227 64.5 52.0 78.0



#34
Caprolactam
Concen: 20.722 ng/ul
RT: 11.751 min Scan# 1490
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:113 Resp: 221709
Ion Ratio Lower Upper
113 100
55 187.2 139.5 209.3
56 140.1 103.5 155.3









#35
4-Chloro-3-methylphenol
Concen: 18.012 ng/ul
RT: 12.092 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:107 Resp: 680570
Ion Ratio Lower Upper
107 100
144 27.2 22.0 33.0
142 82.6 67.3 100.9



#36
2-Methylnaphthalene
Concen: 16.930 ng/ul
RT: 12.475 min Scan# 1613
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:142 Resp: 1512985
Ion Ratio Lower Upper
142 100
141 88.2 70.0 105.0









#37
1-Methylnaphthalene
Concen: 16.364 ng/ul
RT: 12.686 min Scan# 1649
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:142 Resp: 1484381
Ion Ratio Lower Upper
142 100
141 91.4 71.9 107.9
116 3.7 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.359 ng/ul
RT: 12.833 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:216 Resp: 795936
Ion Ratio Lower Upper
216 100
214 78.4 63.1 94.7
179 20.4 16.4 24.6
108 17.0 13.6 20.4





#40

Hexachlorocyclopentadiene

Concen: 14.256 ng/ul

RT: 12.804 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:237 Resp: 370887

Ion Ratio Lower Upper

237 100

235 63.3 50.5 75.7

272 13.3 10.6 16.0



#41

2,4,6-Trichlorophenol

Concen: 17.840 ng/ul

RT: 13.075 min Scan# 1715

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:196 Resp: 508994


Ion Ratio Lower Upper

196 100

198 95.8 77.0 115.4

200 30.5 25.0 37.4









#42
2,4,5-Trichlorophenol
Concen: 17.836 ng/ul
RT: 13.139 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:196 Resp: 553868
Ion Ratio Lower Upper
196 100
198 96.4 76.1 114.1
200 30.3 24.6 36.8



#43
1,1'-Biphenyl
Concen: 17.803 ng/ul
RT: 13.480 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:154 Resp: 2134370
Ion Ratio Lower Upper
154 100
153 39.1 31.8 47.6
76 12.0 9.3 13.9









#44
2-Chloronaphthalene
Concen: 16.891 ng/ul
RT: 13.516 min Scan# 1790
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:162 Resp: 1569681
Ion Ratio Lower Upper
162 100
164 33.0 26.4 39.6
127 37.5 30.0 45.0



#45
2-Nitroaniline
Concen: 18.820 ng/ul
RT: 13.727 min Scan# 1826
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion: 65 Resp: 462170
Ion Ratio Lower Upper
65 100
92 67.0 53.4 80.0
138 110.5 86.2 129.2









#47
Dimethylphthalate
Concen: 17.064 ng/ul
RT: 14.104 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:163 Resp: 1988291
Ion Ratio Lower Upper
163 100
194 4.2 3.3 4.9
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 19.399 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:165 Resp: 416450
Ion Ratio Lower Upper
165 100
89 58.0 47.5 71.3
121 20.7 17.1 25.7









#50
Acenaphthylene
Concen: 17.290 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:152 Resp: 2685795
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.3 10.3 15.5



#51
3-Nitroaniline
Concen: 19.819 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:138 Resp: 472005
Ion Ratio Lower Upper
138 100
108 10.3 9.4 14.0
92 107.5 87.9 131.9





#52

Acenaphthene

Concen: 16.710 ng/ul

RT: 14.698 min Scan# 1991

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:153 Resp: 1712154

Ion Ratio Lower Upper

153 100

152 47.7 38.6 58.0

154 89.0 71.6 107.4



#53

2,4-Dinitrophenol

Concen: 16.273 ng/ul

RT: 14.751 min Scan# 2000

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:184 Resp: 160372


Ion Ratio Lower Upper

184 100

63 65.1 52.2 78.4

154 61.1 47.8 71.8









#55
4-Nitrophenol
Concen: 18.431 ng/ul
RT: 14.845 min Scan# 2016
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:109 Resp: 283699
Ion Ratio Lower Upper
109 100
139 143.2 114.7 172.1
65 137.2 109.5 164.3



#56
Dibenzofuran
Concen: 17.499 ng/ul
RT: 15.033 min Scan# 2048
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:168 Resp: 2449197
Ion Ratio Lower Upper
168 100
139 36.8 28.6 43.0





#57

2,4-Dinitrotoluene

Concen: 18.056 ng/ul

RT: 15.004 min Scan# 2043

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:165 Resp: 553443

Ion Ratio Lower Upper

165 100

63 45.4 34.6 52.0

182 2.7 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.059 ng/ul

RT: 15.251 min Scan# 2085

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Tgt Ion:232 Resp: 491936




Ion Ratio Lower Upper


232 100

131 46.7 37.0 55.4

130 2.4 1.9 2.9

166 32.4 26.1 39.1





#59

Diethylphthalate

Concen: 17.089 ng/ul

RT: 15.463 min Scan# 2121

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:149 Resp: 2002846

Ion Ratio Lower Upper

149 100

177 22.9 18.2 27.2

150 12.5 10.0 15.0



#61

Fluorene

Concen: 17.130 ng/ul

RT: 15.680 min Scan# 2158

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:166 Resp: 2075403


Ion Ratio Lower Upper

166 100

165 97.3 77.5 116.3

167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 16.282 ng/ul
RT: 15.680 min Scan# 2158
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:204 Resp: 974570
Ion Ratio Lower Upper
204 100
206 32.9 26.2 39.2
141 58.1 44.2 66.4



#63
4-Nitroaniline
Concen: 20.566 ng/ul
RT: 15.704 min Scan# 2162
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:138 Resp: 492140
Ion Ratio Lower Upper
138 100
92 49.6 40.1 60.1
108 76.6 63.0 94.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 15.516 ng/ul
 RT: 15.757 min Scan# 2171
 Delta R.T. 0.000 min
 Lab File: BM043188.D
 Acq: 07 Dec 2023 19:18




Instrument :
 BNA_M
 ClientSampleId :
 SICV040


Tgt Ion:198 Resp: 272701
 Ion Ratio Lower Upper
 198 100
 182 4.3 3.2 4.8
 77 26.0 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 17.826 ng/ul
 RT: 15.892 min Scan# 2194
 Delta R.T. 0.000 min
 Lab File: BM043188.D
 Acq: 07 Dec 2023 19:18

Tgt Ion:169 Resp: 1745810
 Ion Ratio Lower Upper
 169 100
 168 66.9 53.4 80.2
 167 36.6 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 16.552 ng/ul
RT: 16.568 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:248 Resp: 584544
Ion Ratio Lower Upper
248 100
250 98.1 78.9 118.3
141 72.5 60.2 90.2



#69
Hexachlorobenzene
Concen: 16.800 ng/ul
RT: 16.668 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:284 Resp: 704001
Ion Ratio Lower Upper
284 100
142 34.1 26.7 40.1
249 29.0 24.0 36.0





#70

Atrazine

Concen: 21.152 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:200 Resp: 584193

Ion Ratio Lower Upper

200 100

173 27.5 20.6 30.8

215 47.5 39.4 59.2



#71

Pentachlorophenol

Concen: 15.851 ng/ul

RT: 17.015 min Scan# 2385

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:266 Resp: 391122


Ion Ratio Lower Upper

266 100

264 62.0 50.5 75.7

268 63.2 50.4 75.6









#72
 Phenanthrene
 Concen: 17.291 ng/ul
 RT: 17.415 min Scan# 2453
 Delta R.T. 0.000 min
 Lab File: BM043188.D
 Acq: 07 Dec 2023 19:18




Instrument :
 BNA_M
 ClientSampleId :
 SICV040


Tgt Ion:178 Resp: 3149040
 Ion Ratio Lower Upper
 178 100
 179 15.2 12.2 18.2
 176 19.4 15.7 23.5



#74
 Anthracene
 Concen: 17.346 ng/ul
 RT: 17.510 min Scan# 2469
 Delta R.T. 0.000 min
 Lab File: BM043188.D
 Acq: 07 Dec 2023 19:18

Tgt Ion:178 Resp: 3182176
 Ion Ratio Lower Upper
 178 100
 179 15.2 12.1 18.1
 176 18.9 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 16.456 ng/uL

RT: 13.433 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:216 Resp: 760335

Ion Ratio Lower Upper

216 100

214 79.3 62.2 93.2

218 48.0 37.6 56.4



#76

Pentachlorobenzene

Concen: 16.475 ng/uL

RT: 14.945 min Scan# 2033

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:250 Resp: 773565


Ion Ratio Lower Upper

250 100

248 63.0 51.8 77.8

252 64.5 51.9 77.9





#77

Carbazole

Concen: 19.475 ng/ul

RT: 17.780 min Scan# 2515

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:167 Resp: 2934667

Ion Ratio Lower Upper

167 100

166 21.4 17.1 25.7

139 12.4 9.9 14.9



#78

Di-n-butylphthalate

Concen: 18.243 ng/ul

RT: 18.351 min Scan# 2612

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:149 Resp: 3575746


Ion Ratio Lower Upper

149 100

150 9.1 7.2 10.8

104 4.8 3.8 5.6





#80

Fluoranthene

Concen: 17.711 ng/ul

RT: 19.421 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:202 Resp: 3541479

Ion Ratio Lower Upper

202 100

101 10.1 8.2 12.2

100 7.5 6.1 9.1



#82

Pyrene

Concen: 17.947 ng/ul

RT: 19.780 min Scan# 2855

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:202 Resp: 3733166


Ion Ratio Lower Upper

202 100

101 12.7 10.2 15.2

100 10.1 8.2 12.2









#83
Butylbenzylphthalate
Concen: 18.042 ng/ul
RT: 20.686 min Scan# 3009
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:149 Resp: 1492852
Ion Ratio Lower Upper
149 100
91 71.8 56.9 85.3
206 20.8 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 20.951 ng/ul
RT: 21.456 min Scan# 3140
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:252 Resp: 1356507
Ion Ratio Lower Upper
252 100
254 64.4 51.3 76.9
126 10.9 8.8 13.2





#85

Benzo(a)anthracene

Concen: 17.433 ng/ul

RT: 21.521 min Scan# 3151

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:228 Resp: 3641482

Ion Ratio Lower Upper

228 100

229 19.7 15.8 23.8

226 27.6 22.2 33.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.848 ng/ul

RT: 21.462 min Scan# 3141

Delta R.T. -0.006 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:149 Resp: 2461813


Ion Ratio Lower Upper

149 100

167 28.1 22.6 34.0

279 4.0 3.7 5.5





#87

Chrysene

Concen: 17.481 ng/ul

RT: 21.574 min Scan# 3160

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:228 Resp: 3261239

Ion Ratio Lower Upper

228 100

226 31.1 24.8 37.2

229 19.7 16.2 24.2



#89

Di-n-octyl phthalate

Concen: 19.404 ng/ul




RT: 22.403 min Scan# 3301


Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Tgt Ion:149 Resp: 3877529









#90
Benzo(b)fluoranthene
Concen: 16.281 ng/ul
RT: 23.203 min Scan# 3438
Delta R.T. -0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18




Instrument :
BNA_M
ClientSampleId :
SICV040


Tgt Ion:252 Resp: 3374810
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 9.5 7.2 10.8



#91
Benzo(k)fluoranthene
Concen: 16.302 ng/ul
RT: 23.256 min Scan# 3446
Delta R.T. 0.000 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:252 Resp: 3365424
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 9.5 7.7 11.5





#93

Benzo(a)pyrene

Concen: 15.604 ng/ul

RT: 23.833 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18

Instrument :

BNA_M

ClientSampleId :

SICV040





Tgt Ion:252 Resp: 2989651

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.6

125 10.0 8.4 12.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 16.539 ng/ul

RT: 26.438 min Scan# 3987

Delta R.T. 0.000 min

Lab File: BM043188.D

Acq: 07 Dec 2023 19:18




Tgt Ion:276 Resp: 3609434


Ion Ratio Lower Upper

276 100

138 22.0 16.9 25.3

277 24.3 19.5 29.3









#95
Dibenzo(a,h)anthracene
Concen: 16.591 ng/ul
RT: 26.462 min Scan# 3991
Delta R.T. -0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Instrument :
BNA_M
ClientSampleId :
SICV040

Tgt Ion:278 Resp: 3027491
Ion Ratio Lower Upper
278 100
139 15.3 11.9 17.9
279 24.5 19.4 29.2



#96
Benzo(g,h,i)perylene
Concen: 16.524 ng/ul
RT: 27.197 min Scan# 4116
Delta R.T. -0.006 min
Lab File: BM043188.D
Acq: 07 Dec 2023 19:18

Tgt Ion:276 Resp: 2716986
Ion Ratio Lower Upper
276 100
138 20.9 16.2 24.2
277 23.9 18.7 28.1

