

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121021\
 Data File : BM0333391.D
 Acq On : 10 Dec 2021 12:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 02:09:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM120821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 09 07:11:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	81	0.00
2	1,4-Dioxane	40.000	35.727	10.7	78	0.00
3	Pyridine	40.000	40.138	-0.3	79	0.00
4	n-Nitrosodimethylamine	40.000	45.221	-13.1	89	0.00
5 S	2-Fluorophenol	80.000	79.245	0.9	79	0.00
6	Aniline	40.000	40.897	-2.2	81	0.00
7 S	Phenol-d6	80.000	80.836	-1.0	80	0.00
8	2-Chlorophenol	40.000	41.546	-3.9	82	0.00
9	Benzaldehyde	40.000	40.432	-1.1	78	0.00
10 C	Phenol	40.000	40.294	-0.7	80	0.00
11	bis(2-Chloroethyl)ether	40.000	39.838	0.4	79	0.00
12	1,3-Dichlorobenzene	40.000	40.589	-1.5	81	0.00
13 C	1,4-Dichlorobenzene	40.000	40.491	-1.2	80	0.00
14	1,2-Dichlorobenzene	40.000	40.709	-1.8	82	0.00
15	Benzyl Alcohol	40.000	44.256	-10.6	86	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.693	-1.7	82	0.00
17	2-Methylphenol	40.000	42.415	-6.0	85	0.00
18	Hexachloroethane	40.000	42.140	-5.4	85	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.382	-8.5	86	0.00
20	3+4-Methylphenols	40.000	42.835	-7.1	84	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	40.014	-0.0	89	0.00
23 S	Nitrobenzene-d5	80.000	79.071	1.2	87	0.00
24	Nitrobenzene	40.000	40.015	-0.0	88	0.00
25	Isophorone	40.000	40.974	-2.4	90	0.00
26 C	2-Nitrophenol	40.000	37.285	6.8	80	0.00
27	2,4-Dimethylphenol	40.000	38.842	2.9	85	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.056	2.4	87	0.00
29 C	2,4-Dichlorophenol	40.000	41.574	-3.9	90	0.00
30	1,2,4-Trichlorobenzene	40.000	39.977	0.1	90	0.00
31	Naphthalene	40.000	39.424	1.4	88	0.00
32	Benzoic acid	40.000	38.488	3.8	82	0.00
33	4-Chloroaniline	40.000	41.221	-3.1	90	0.00
34 C	Hexachlorobutadiene	40.000	41.267	-3.2	93	0.00
35	Caprolactam	40.000	44.494	-11.2	94	0.01
36 C	4-Chloro-3-methylphenol	40.000	42.946	-7.4	93	0.00
37	2-Methylnaphthalene	40.000	41.486	-3.7	92	0.00
38	1-Methylnaphthalene	40.000	41.528	-3.8	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.892	5.3	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.627	5.9	90	0.00
42 S	2,4,6-Tribromophenol	80.000	87.921	-9.9	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.362	1.6	96	0.00
44	2,4,5-Trichlorophenol	40.000	40.136	-0.3	97	0.00
45 S	2-Fluorobiphenyl	80.000	75.954	5.1	96	0.00
46	1,1'-Biphenyl	40.000	38.025	4.9	96	0.00
47	2-Chloronaphthalene	40.000	38.217	4.5	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121021\
 Data File : BM033391.D
 Acq On : 10 Dec 2021 12:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 11 02:09:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM120821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 09 07:11:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.876	0.3	96	0.00
49	Acenaphthylene	40.000	39.864	0.3	99	0.00
50	Dimethylphthalate	40.000	40.558	-1.4	101	0.00
51	2,6-Dinitrotoluene	40.000	40.247	-0.6	99	0.00
52 C	Acenaphthene	40.000	39.453	1.4	100	0.00
53	3-Nitroaniline	40.000	41.432	-3.6	99	0.00
54 P	2,4-Dinitrophenol	40.000	39.472	1.3	94	0.00
55	Dibenzofuran	40.000	40.773	-1.9	102	0.00
56 P	4-Nitrophenol	40.000	45.389	-13.5	102	0.00
57	2,4-Dinitrotoluene	40.000	42.874	-7.2	101	0.00
58	Fluorene	40.000	41.435	-3.6	105	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.085	-7.7	107	0.00
60	Diethylphthalate	40.000	42.719	-6.8	108	0.00
61	4-Chlorophenyl-phenylether	40.000	41.424	-3.6	105	0.00
62	4-Nitroaniline	40.000	44.127	-10.3	102	0.00
63	Azobenzene	40.000	42.974	-7.4	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.278	-0.7	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.687	3.3	107	0.00
67	4-Bromophenyl-phenylether	40.000	38.695	3.3	107	0.00
68	Hexachlorobenzene	40.000	38.884	2.8	109	0.00
69	Atrazine	40.000	41.661	-4.2	109	0.00
70 C	Pentachlorophenol	40.000	43.861	-9.7	112	0.00
71	Phenanthrene	40.000	39.440	1.4	109	0.00
72	Anthracene	40.000	39.572	1.1	108	0.00
73	Carbazole	40.000	40.642	-1.6	111	0.00
74	Di-n-butylphthalate	40.000	41.344	-3.4	111	0.00
75 C	Fluoranthene	40.000	40.953	-2.4	112	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	33.913	15.2	88	0.00
78	Pyrene	40.000	41.799	-4.5	112	0.00
79 S	Terphenyl-d14	80.000	83.010	-3.8	113	0.00
80	Butylbenzylphthalate	40.000	41.975	-4.9	108	0.00
81	Benzo(a)anthracene	40.000	40.297	-0.7	108	0.00
82	3,3'-Dichlorobenzidine	40.000	38.264	4.3	100	0.00
83	Chrysene	40.000	39.607	1.0	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.486	-1.2	105	0.00
85 c	Di-n-octyl phthalate	40.000	37.589	6.0	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	34.923	12.7	81	0.01
88	Benzo(b)fluoranthene	40.000	42.755	-6.9	100	0.00
89	Benzo(k)fluoranthene	40.000	40.495	-1.2	92	0.00
90 C	Benzo(a)pyrene	40.000	40.134	-0.3	92	0.00
91	Dibenzo(a,h)anthracene	40.000	35.140	12.1	81	0.00
92	Benzo(g,h,i)perylene	40.000	34.474	13.8	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121021\
Data File : BM033391.D
Acq On : 10 Dec 2021 12:54
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Dec 11 02:09:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM120821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 09 07:11:07 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0