

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121115\
 Data File : BM003500.D
 Acq On : 12 Dec 2015 02:06
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Dec 12 03:03:33 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM120115.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 12 00:54:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.479	0.476	0.6	89	0.00
3 S	1,4-Dioxane-d8	0.402	0.398	1.0	89	0.00
4	Benzaldehyde	0.982	0.967	1.5	88	0.00
5 S	Phenol-d5	1.597	1.564	2.1	89	0.00
6	Phenol	1.686	1.662	1.4	89	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.881	0.878	0.3	89	0.00
8	Bis(2-Chloroethyl)ether	1.300	1.283	1.3	88	0.00
9 S	2-Chlorophenol-d4	1.320	1.341	-1.6	91	0.00
10	2-Chlorophenol	1.398	1.397	0.1	89	0.00
11	2-Methylphenol	1.312	1.273	3.0	88	0.00
12	2,2'-oxybis(1-Chloropropane	1.298	1.304	-0.5	89	0.00
13 S	4-Methylphenol-d8	1.339	1.276	4.7	86	0.00
14	Acetophenone	2.054	2.026	1.4	85	0.00
15 P	N-Nitroso-di-n-propylamine	0.969	0.944	2.6	84	0.00
16	4-Methylphenol	1.453	1.390	4.3	86	0.00
17	Hexachloroethane	0.514	0.531	-3.3	93	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
19 S	Nitrobenzene-d5	0.136	0.145	-6.6	88	0.00
20	Nitrobenzene	0.315	0.337	-7.0	88	0.00
21	Isophorone	0.589	0.590	-0.2	82	0.00
22 S	2-Nitrophenol-d4	0.143	0.153	-7.0	89	0.00
23 C	2-Nitrophenol	0.164	0.174	-6.1	87	0.00
24	2,4-Dimethylphenol	0.344	0.354	-2.9	85	0.00
25	Bis(2-Chloroethoxy)methane	0.394	0.390	1.0	84	0.00
26 S	2,4-Dichlorophenol-d3	0.283	0.289	-2.1	84	0.00
27 C	2,4-Dichlorophenol	0.298	0.301	-1.0	83	0.00
28	Naphthalene	1.010	1.042	-3.2	87	0.00
29 S	4-Chloroaniline-d4	0.320	0.361	-12.8	98	0.00
30	4-Chloroaniline	0.329	0.373	-13.4	98	0.00
31 C	Hexachlorobutadiene	0.180	0.192	-6.7	91	0.00
32	Caprolactam	0.087	0.074	14.9	74	0.00
33 C	4-Chloro-3-methylphenol	0.313	0.299	4.5	78	0.00
34	2-Methylnaphthalene	0.737	0.724	1.8	82	0.00
35	1-Methylnaphthalene	0.697	0.683	2.0	82	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
37	1,2,4,5-Tetrachlorobenzene	0.629	0.672	-6.8	82	0.00
38	Hexachlorocyclopentadiene	0.362	0.301	16.9	71	0.00
39 C	2,4,6-Trichlorophenol	0.392	0.402	-2.6	77	0.00
40	2,4,5-Trichlorophenol	0.434	0.432	0.5	75	0.00
41	1,1'-Biphenyl	1.667	1.758	-5.5	80	0.00
42	2-Chloronaphthalene	1.244	1.319	-6.0	81	0.00
43	2-Nitroaniline	0.278	0.291	-4.7	77	0.00
44 S	Dimethylphthalate-d6	1.504	1.511	-0.5	76	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121115\
 Data File : BM003500.D
 Acq On : 12 Dec 2015 02:06
 Operator : UM/IZ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Dec 12 03:03:33 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM120115.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 12 00:54:45 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.547	1.537	0.6	75	0.00
46	2,6-Dinitrotoluene	0.284	0.292	-2.8	75	0.00
47 S	Acenaphthylene-d8	1.848	1.914	-3.6	77	0.00
48	Acenaphthylene	2.019	2.122	-5.1	79	0.00
49	3-Nitroaniline	0.299	0.308	-3.0	80	0.00
50 C	Acenaphthene	1.349	1.380	-2.3	78	0.00
51	2,4-Dinitrophenol	0.148	0.085	42.6#	54	0.00
52 S	4-Nitrophenol-d4	0.270	0.234	13.3	69	0.00
53	4-Nitrophenol	0.211	0.192	9.0	72	0.00
54	Dibenzofuran	1.911	1.932	-1.1	77	0.00
55	2,4-Dinitrotoluene	0.413	0.421	-1.9	75	0.00
56	2,3,4,6-Tetrachlorophenol	0.339	0.335	1.2	74	0.00
57	Diethylphthalate	1.499	1.524	-1.7	77	0.00
58 S	Fluorene-d10	1.294	1.304	-0.8	77	0.00
59	Fluorene	1.481	1.517	-2.4	78	0.00
60	4-Chlorophenyl-phenylether	0.723	0.726	-0.4	77	0.00
61	4-Nitroaniline	0.335	0.320	4.5	75	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	77	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.102	0.089	12.7	72	0.00
64	4,6-Dinitro-2-methylphenol	0.111	0.099	10.8	72	0.00
65	N-Nitrosodiphenylamine	0.610	0.628	-3.0	76	0.00
66	4-Bromophenyl-phenylether	0.201	0.206	-2.5	75	0.00
67	Hexachlorobenzene	0.223	0.229	-2.7	76	0.00
68	Atrazine	0.203	0.204	-0.5	76	0.00
69 C	Pentachlorophenol	0.121	0.100	17.4	67	0.00
70	Phenanthrene	1.101	1.122	-1.9	76	0.00
71 S	Anthracene-d10	0.889	0.914	-2.8	76	0.00
72	Anthracene	1.110	1.148	-3.4	77	0.00
73	1,2,3,4-Tetrachlorobenzene	0.322	0.352	-9.3	82	0.00
74	Pentachlorobenzene	0.286	0.296	-3.5	77	0.00
75	Carbazole	0.959	1.018	-6.2	78	0.00
76	Di-n-butylphthalate	1.033	1.130	-9.4	80	0.00
77 C	Fluoranthene	1.112	1.212	-9.0	80	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
79 S	Pyrene-d10	0.979	0.946	3.4	80	0.00
80	Pyrene	1.294	1.254	3.1	80	0.00
81	Butylbenzylphthalate	0.422	0.464	-10.0	89	0.00
82	3,3'-Dichlorobenzidine	0.326	0.366	-12.3	96	0.00
83	Benzo(a)anthracene	1.143	1.197	-4.7	87	0.00
84	Bis(2-ethylhexyl)phthalate	0.587	0.680	-15.8	92	0.00
85	Chrysene	1.102	1.143	-3.7	88	0.00
86 I	Perylene-d12	1.000	1.000	0.0	95	0.00
87	Di-n-octyl phthalate	1.186	1.216	-2.5	98	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121115\
Data File : BM003500.D
Acq On : 12 Dec 2015 02:06
Operator : UM/IZ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02042

Quant Time: Dec 12 03:03:33 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM120115.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Dec 12 00:54:45 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.169	1.237	-5.8	96	0.00
89	Benzo(k)fluoranthene	1.147	1.136	1.0	90	0.00
90 S	Benzo(a)pyrene-d12	0.965	1.014	-5.1	95	0.00
91 C	Benzo(a)pyrene	1.134	1.174	-3.5	94	0.00
92	Indeno(1,2,3-cd)pyrene	1.271	1.345	-5.8	96	0.00
93	Dibenzo(a,h)anthracene	1.076	1.135	-5.5	95	0.00
94	Benzo(a,h,i)perylene	1.070	1.131	-5.7	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0