

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121118\
 Data File : BM018072.D
 Acq On : 12 Dec 2018 00:01
 Operator : JU/SJ
 Sample : PB115305BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK05

Manual Integrations
 APPROVED

Sohil
 12/12/2018 5:43:09 PM

Quant Time: Dec 12 01:51:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111918MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 12 00:58:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	168236	20.00	ng/ul	0.00
18) Naphthalene-d8	10.31	136	728915	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.20	164	402821	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.96	188	814458	20.00	ng/ul	0.00
77) Chrysene-d12	21.17	240	618284	20.00	ng/ul	0.00
85) Perylene-d12	23.34	264	615179	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	16843	4.58	ng/uL	0.00
5) Phenol-d5	6.74	99	348803	22.70	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.90	67	206468	22.28	ng/ul	0.00
9) 2-Chlorophenol-d4	7.09	132	312272	25.97	ng/ul	0.00
13) 4-Methylphenol-d8	8.26	113	287642	22.91	ng/ul	0.00
19) Nitrobenzene-d5	8.70	128	151929	27.14	ng/ul	0.00
22) 2-Nitrophenol-d4	9.42	143	178494	27.82	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.95	165	269721	22.09	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	405020	38.65	ng/ul	0.00
43) Dimethylphthalate-d6	13.62	166	1009369	28.88	ng/ul	0.00
46) Acenaphthylene-d8	13.89	160	1201421	28.40	ng/ul	0.00
51) 4-Nitrophenol-d4	14.45	143	120775m	19.01	ng/ul	0.00
57) Fluorene-d10	15.20	176	838199	27.77	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.35	200	118863	20.42	ng/ul	0.00
70) Anthracene-d10	17.06	188	1226291	30.24	ng/ul	0.00
78) Pyrene-d10	19.36	212	1178939	39.41	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.20	264	1014636	30.41	ng/ul	-0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121118\
Data File : BM018072.D
Acq On : 12 Dec 2018 00:01
Operator : JU/SJ
Sample : PB115305BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK05

Manual Integrations
APPROVED

Sohil
12/12/2018 5:43:09 PM

Quant Time: Dec 12 01:51:37 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111918MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 12 00:58:41 2018
Response via : Initial Calibration

