


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121118\
 Data File : BM018049.D
 Acq On : 11 Dec 2018 08:54
 Operator : JU/SJ
 Sample : DFTPP76
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP76

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111918MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Dec 12 06:24:35 2018



AutoFind: Scans 2451, 2452, 2453; Background Corrected with Scan 2446

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.1	106432	PASS
68	69	0.00	2	1.9	2356	PASS
69	198	0.00	100	35.9	127232	PASS
70	69	0.00	2	0.4	557	PASS
127	198	10	80	44.0	155586	PASS
197	198	0.00	2	0.3	1069	PASS
198	198	100	100	100.0	354005	PASS
199	198	5	9	7.0	24669	PASS
275	198	10	60	28.2	99749	PASS
365	198	1	100	3.8	13332	PASS
441	198	0.01	100	13.1	46522	PASS
442	198	50	100	88.6	313706	PASS
443	442	15	24	19.2	60330	PASS

Scan 2446 (17.374 min): BM018049.D
DFTPP76

m/z	Abundance
44.00	1248.0
73.10	603.0
207.10	1292.0
207.90	364.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

m/z	Abundance
37.10	605.0
38.10	1452.0
39.10	7594.0
40.00	713.0
41.00	309.0
44.00	1866.0
49.00	805.0
50.00	24744.0
51.00	90904.0
52.10	4852.0
52.90	355.0
54.90	534.0
56.00	2676.0
57.00	6588.0
58.00	314.0
61.10	1277.0
62.10	1368.0
63.00	4213.0
64.10	691.0
65.00	1693.0
68.10	2009.0
69.00	103944.0
69.90	407.0
73.00	1812.0
74.10	11823.0
75.00	17200.0
76.10	5500.0
77.10	115856.0
78.10	7657.0
79.00	9181.0
80.00	6476.0
81.00	8784.0
81.90	1909.0
83.00	2002.0
84.00	331.0
85.00	1409.0
86.00	2481.0
87.10	1055.0
87.90	392.0
91.00	2220.0
92.00	2299.0
93.00	14467.0
94.00	887.0
94.90	350.0
95.90	758.0
96.90	346.0
98.00	11243.0
99.00	8020.0
100.00	752.0
101.00	4023.0
102.10	399.0
103.00	1563.0
104.00	3688.0
105.00	2859.0
106.20	855.0
107.00	36392.0
108.00	6176.0
110.00	65024.0
111.00	9207.0
112.10	1264.0
112.90	321.0
116.00	2309.0
117.00	30448.0
118.00	2197.0
119.00	456.0
120.10	492.0
122.00	2922.0
123.00	3563.0
124.10	1446.0
125.00	1849.0
127.00	121352.0
128.00	9550.0
129.00	47464.0
129.90	3741.0
130.90	1055.0
132.10	473.0
133.00	437.0
134.10	1574.0
135.00	3554.0
136.00	2146.0
137.00	1944.0
138.10	327.0
138.90	389.0
140.00	958.0
141.00	6782.0
142.00	2176.0
143.00	1444.0
144.10	534.0
145.10	425.0
146.00	1509.0
147.00	3387.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

148.00	6508.0
149.00	1544.0
150.00	473.0
151.00	785.0
151.70	570.0
153.00	1975.0
154.00	1935.0
155.00	2983.0
156.00	5091.0
157.00	908.0
158.00	1215.0
159.00	1146.0
160.00	1825.0
161.00	3313.0
162.00	925.0
162.90	416.0
163.90	620.0
165.00	2387.0
166.00	1967.0
167.00	12070.0
168.00	7199.0
169.00	1035.0
170.10	340.0
170.80	608.0
171.90	1121.0
173.10	1594.0
174.00	3209.0
175.10	5235.0
176.00	1219.0
177.00	2624.0
178.00	1180.0
179.00	8850.0
180.00	6834.0
181.10	2958.0
182.00	476.0
184.00	546.0
185.00	3748.0
186.00	35504.0
187.00	9516.0
188.00	1066.0
189.10	2656.0
190.00	587.0
191.10	1571.0
192.00	2947.0
193.00	3536.0
194.10	652.0
195.00	421.0
196.00	7268.0
198.00	257088.0
198.90	18128.0
199.90	1271.0
201.60	1070.0
203.00	1471.0
204.00	9064.0
205.00	15646.0
206.10	66416.0
207.00	9335.0
208.00	2961.0
209.00	1007.0
210.00	1172.0
211.00	2363.0
212.00	479.0
215.10	691.0
216.10	1266.0
217.00	17104.0
218.00	2129.0
221.00	11187.0
221.80	2239.0
223.10	3277.0
224.10	34512.0
225.10	8948.0
226.00	1346.0
227.00	15795.0
228.00	2224.0
229.00	2973.0
231.00	1213.0
234.00	1206.0
235.00	1429.0
236.00	938.0
237.00	1567.0
239.00	838.0
240.00	453.0
241.10	1231.0
242.10	1779.0
243.10	2083.0
244.10	27624.0
245.10	3923.0
246.00	6075.0
247.00	1244.0
249.00	1165.0
251.00	412.0
253.10	572.0
255.00	151744.0
256.00	22160.0
257.10	2083.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

258.00	8398.0
258.90	1525.0
260.90	300.0
265.00	3336.0
266.00	595.0
273.10	4504.0
274.00	10608.0
275.10	67976.0
276.00	9679.0
277.00	6258.0
278.00	1112.0
281.00	508.0
283.10	809.0
284.00	385.0
285.10	1142.0
290.00	310.0
292.00	307.0
293.10	1393.0
295.10	382.0
296.00	19560.0
297.10	3543.0
303.00	2102.0
304.10	730.0
314.00	782.0
315.00	2275.0
316.00	1663.0
321.10	476.0
322.10	441.0
323.10	5829.0
324.10	947.0
327.00	1120.0
328.10	810.0
332.10	493.0
332.90	553.0
334.00	3609.0
335.00	920.0
341.00	511.0
346.00	1197.0
347.00	350.0
352.00	1726.0
353.00	1345.0
354.00	2106.0
355.00	617.0
365.00	8856.0
366.00	1447.0
370.90	432.0
372.10	3097.0
373.00	1230.0
373.90	334.0
383.00	745.0
384.00	332.0
389.90	309.0
391.00	459.0
402.10	1053.0
403.10	1451.0
404.00	625.0
421.00	1710.0
422.00	1565.0
423.10	10238.0
424.10	3184.0
425.00	430.0
441.10	24616.0
442.10	168128.0
443.10	34360.0
444.10	3256.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

m/z	Abundance
37.00	749.0
38.00	1837.0
39.10	9103.0
40.00	588.0
41.00	406.0
44.00	1237.0
50.10	33296.0
51.10	127408.0
52.10	6376.0
54.90	1053.0
56.00	3830.0
57.00	8634.0
61.00	2090.0
62.00	2364.0
63.00	5365.0
64.00	514.0
65.10	2566.0
67.10	419.0
68.10	2637.0
69.00	155648.0
70.00	743.0
73.10	1954.0
74.00	16848.0
75.00	24736.0
76.10	8077.0
77.10	166784.0
78.10	11556.0
79.00	14608.0
80.00	8979.0
81.00	11997.0
82.00	3382.0
83.00	2955.0
84.00	326.0
85.10	2548.0
86.00	2928.0
87.00	1565.0
87.80	475.0
89.00	331.0
91.00	2612.0
92.00	3642.0
93.00	21536.0
94.10	1347.0
95.20	316.0
96.10	1060.0
97.00	748.0
98.00	16088.0
99.00	11757.0
100.00	1083.0
101.00	7372.0
103.00	2266.0
104.10	3443.0
105.00	4073.0
106.00	1583.0
107.00	53776.0
108.00	8294.0
109.20	2021.0
110.00	101288.0
111.00	14636.0
112.10	1810.0
113.00	413.0
115.10	348.0
116.00	3414.0
117.00	44376.0
118.00	3585.0
118.90	573.0
120.00	624.0
122.00	4648.0
123.00	6759.0
123.90	2897.0
125.00	2327.0
126.20	1717.0
127.00	186432.0
128.10	14828.0
129.00	72816.0
130.00	5925.0
131.00	1002.0
131.90	784.0
133.00	358.0
134.00	2546.0
135.00	5429.0
136.00	2247.0
137.00	3832.0
138.00	648.0
140.10	928.0
141.00	10404.0
141.90	2729.0
143.00	2305.0
144.80	537.0
146.00	1708.0
147.00	5585.0
148.00	11069.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

149.00	2458.0
149.80	812.0
151.00	1281.0
151.70	880.0
152.00	914.0
153.00	3326.0
154.00	2601.0
155.00	5581.0
156.10	8506.0
157.10	1745.0
158.00	1700.0
159.00	1216.0
160.00	3514.0
161.00	4764.0
162.00	1167.0
163.10	466.0
164.00	505.0
165.00	3364.0
166.10	3100.0
167.00	18664.0
168.00	10216.0
169.00	1440.0
170.00	673.0
171.10	982.0
172.00	1635.0
173.00	2255.0
174.00	3991.0
175.00	7890.0
175.90	2334.0
177.10	3830.0
178.00	1326.0
179.00	13578.0
180.00	10194.0
181.00	4541.0
182.00	833.0
183.00	608.0
183.90	1409.0
185.00	6719.0
186.00	54368.0
187.00	15733.0
188.10	1718.0
189.00	4136.0
189.90	649.0
190.90	1324.0
192.00	5107.0
193.10	4911.0
194.00	1208.0
195.00	710.0
196.00	11033.0
197.10	3207.0
198.00	414464.0
199.00	28432.0
200.10	2448.0
201.60	1618.0
202.20	591.0
203.00	3319.0
204.00	15102.0
205.10	22688.0
206.10	111752.0
207.10	16032.0
208.00	3844.0
209.10	1390.0
210.00	1893.0
211.00	3856.0
211.90	761.0
215.00	1099.0
216.00	2474.0
217.00	29712.0
218.00	4034.0
219.00	322.0
219.80	374.0
221.00	15297.0
222.00	3200.0
223.00	6774.0
224.10	59080.0
225.00	14640.0
226.20	1944.0
227.00	25992.0
228.00	4603.0
229.00	5729.0
230.00	795.0
231.10	1865.0
232.00	317.0
233.00	363.0
233.90	1684.0
235.00	1981.0
236.00	1260.0
237.00	2560.0
238.00	362.0
239.00	705.0
239.90	680.0
241.00	1502.0
242.10	3561.0
243.10	3316.0
244.10	43704.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

245.10	6066.0
246.00	9383.0
247.00	1899.0
248.10	568.0
249.00	1540.0
251.10	503.0
252.00	751.0
252.90	905.0
254.20	2636.0
255.00	248128.0
256.00	38272.0
257.00	2841.0
258.00	14169.0
259.00	2435.0
260.00	316.0
261.10	415.0
264.10	354.0
265.00	6132.0
266.00	869.0
269.90	370.0
270.80	531.0
271.90	906.0
273.00	7319.0
274.00	21240.0
275.00	112088.0
276.00	14650.0
277.00	9942.0
278.00	2285.0
279.00	535.0
281.20	461.0
282.10	519.0
283.10	1180.0
284.00	714.0
284.90	1567.0
289.10	427.0
290.10	305.0
290.80	371.0
291.90	530.0
293.00	1999.0
294.10	586.0
295.10	644.0
296.00	31912.0
297.00	4414.0
301.00	460.0
302.00	724.0
303.10	3690.0
304.10	881.0
308.10	567.0
309.00	513.0
313.00	490.0
314.10	1623.0
315.00	3359.0
316.10	1699.0
321.10	1163.0
322.20	526.0
323.10	11232.0
324.00	1270.0
327.00	1679.0
328.00	1203.0
332.10	1057.0
332.90	980.0
334.00	6534.0
335.00	1655.0
341.00	1403.0
342.10	494.0
346.00	2072.0
346.80	717.0
352.00	3029.0
353.10	2594.0
354.10	3150.0
355.10	722.0
365.00	14510.0
366.10	1553.0
371.10	623.0
372.00	5318.0
373.00	1575.0
383.10	1697.0
384.20	333.0
390.00	794.0
391.00	370.0
392.10	489.0
401.00	382.0
402.00	2305.0
403.00	3167.0
404.10	1522.0
405.10	371.0
421.00	2616.0
422.10	2429.0
423.10	20672.0
424.10	4489.0
425.10	944.0
440.20	659.0
441.10	50656.0
442.10	353408.0
443.10	67400.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

444.106649.0

445.00567.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

m/z	Abundance
37.00	514.0
38.10	1245.0
39.10	8001.0
40.00	774.0
44.00	1217.0
49.10	921.0
50.10	27704.0
51.10	100984.0
52.00	5214.0
55.00	772.0
56.00	3131.0
57.00	6577.0
58.00	361.0
61.00	1301.0
62.00	1981.0
63.00	4366.0
63.90	736.0
65.00	2280.0
68.10	2423.0
69.00	122104.0
70.00	522.0
73.00	1687.0
74.10	12856.0
75.00	19768.0
76.10	6264.0
77.10	136320.0
78.00	9501.0
79.00	10728.0
80.00	7302.0
81.00	11094.0
82.00	2636.0
83.10	2177.0
85.00	1477.0
86.00	3293.0
87.00	1262.0
87.90	418.0
89.00	335.0
91.10	2718.0
92.00	3108.0
93.00	17792.0
93.90	1297.0
96.10	657.0
96.90	363.0
98.00	14426.0
99.00	9401.0
100.00	1074.0
101.00	5368.0
102.00	398.0
103.00	2180.0
104.10	3225.0
105.00	3706.0
105.90	1271.0
107.00	45128.0
108.00	6701.0
110.00	81704.0
110.90	11830.0
112.10	1247.0
113.00	463.0
114.90	305.0
116.00	3173.0
117.00	40208.0
118.00	3035.0
120.00	554.0
121.90	3416.0
123.00	5090.0
124.00	2533.0
125.00	2100.0
127.00	158976.0
128.00	12399.0
129.00	62952.0
130.00	5587.0
131.00	1273.0
132.00	547.0
133.00	350.0
134.00	2308.0
135.00	5859.0
135.90	2172.0
137.00	3277.0
138.10	427.0
139.00	440.0
139.90	1163.0
141.00	9572.0
142.10	2908.0
143.00	1810.0
143.90	504.0
145.00	427.0
146.00	1701.0
147.10	4854.0
148.00	9516.0
149.00	1969.0
149.90	429.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

151.00	1268.0
151.90	838.0
153.00	2949.0
154.10	2370.0
155.00	4973.0
156.10	7949.0
157.10	1383.0
158.00	1356.0
159.10	1719.0
160.10	3312.0
161.00	4082.0
162.10	1072.0
164.20	355.0
165.00	3096.0
166.00	2733.0
167.00	17992.0
168.00	9448.0
169.00	1507.0
170.00	576.0
171.00	1079.0
172.00	1815.0
173.00	2042.0
174.00	4199.0
175.00	7094.0
175.90	1891.0
177.00	3876.0
178.10	1342.0
178.90	12852.0
180.00	8634.0
181.00	4146.0
182.10	688.0
183.00	323.0
183.90	1281.0
185.00	6980.0
186.00	53480.0
187.00	14972.0
188.00	1502.0
189.10	3617.0
190.00	657.0
191.00	1653.0
192.00	4991.0
193.10	4664.0
194.00	886.0
194.90	496.0
196.00	10883.0
196.80	2385.0
198.00	390464.0
199.00	27448.0
199.90	1703.0
201.50	1788.0
203.00	2508.0
204.00	13587.0
205.10	24288.0
206.10	101312.0
207.10	14060.0
208.00	3712.0
209.00	1166.0
210.10	1497.0
210.90	3520.0
213.00	488.0
215.10	1060.0
216.00	2256.0
217.00	26960.0
218.00	3742.0
220.20	727.0
221.00	13933.0
222.00	3099.0
223.00	5966.0
224.10	54648.0
225.10	15867.0
226.00	1522.0
227.00	23824.0
228.00	3886.0
229.00	5613.0
229.90	628.0
231.00	1958.0
232.00	418.0
233.00	455.0
234.10	2005.0
235.10	1810.0
236.00	1203.0
237.00	1798.0
238.10	386.0
239.00	1384.0
240.10	763.0
241.00	1619.0
242.10	2970.0
243.10	2703.0
244.10	44896.0
245.10	6748.0
246.00	8953.0
247.00	1984.0
248.00	478.0
249.00	2218.0
251.00	476.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76

252.10	524.0
253.10	786.0
254.10	2538.0
255.00	255424.0
256.00	34152.0
257.00	3270.0
258.00	15897.0
259.00	2416.0
260.10	379.0
261.00	354.0
262.90	347.0
264.00	343.0
265.00	5448.0
266.00	590.0
270.00	367.0
271.00	596.0
272.10	548.0
273.00	6613.0
274.00	21752.0
275.00	119184.0
276.10	16090.0
277.00	10724.0
278.00	1742.0
279.00	351.0
280.90	493.0
282.00	411.0
283.00	1186.0
284.00	685.0
285.10	1563.0
286.00	414.0
289.00	438.0
290.00	360.0
291.10	334.0
292.00	512.0
293.00	2556.0
294.10	542.0
295.10	596.0
296.00	32520.0
297.00	4848.0
301.20	372.0
302.10	494.0
303.00	3240.0
304.00	1348.0
308.10	506.0
310.00	719.0
314.00	1774.0
315.00	3922.0
316.10	2427.0
316.90	308.0
321.00	912.0
322.20	616.0
323.10	12274.0
324.10	1822.0
326.00	329.0
327.00	2255.0
328.10	1492.0
329.00	382.0
332.00	1016.0
333.10	935.0
334.00	7852.0
335.00	2047.0
336.10	336.0
341.10	1432.0
342.10	334.0
346.00	2905.0
346.90	514.0
352.00	3583.0
353.10	2623.0
354.10	4573.0
355.00	1128.0
358.90	437.0
365.00	16632.0
366.00	2403.0
370.10	361.0
371.10	1263.0
372.10	6923.0
373.00	1556.0
383.00	1839.0
384.00	774.0
390.10	630.0
391.00	851.0
401.00	368.0
402.00	3073.0
403.10	4146.0
404.10	1242.0
405.10	304.0
421.00	3432.0
422.00	3511.0
423.10	26704.0
424.00	6215.0
425.10	447.0
441.10	64296.0
442.10	419584.0
443.10	79232.0
444.10	6978.0

Instrument :
BNA_M
ClientSampleId :
DFTPP76