

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121119\
 Data File : BM024034.D
 Acq On : 11 Dec 2019 15:29
 Operator : JU
 Sample : PB125195BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK95

Quant Time: Dec 12 00:43:54 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM112719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 07 04:26:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.47	152	498792	20.00	ng/ul	0.00
18) Naphthalene-d8	10.25	136	2190928	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.13	164	1413534	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.89	188	3009899	20.00	ng/ul	0.00
78) Chrysene-d12	21.10	240	2335199	20.00	ng/ul	0.00
86) Perylene-d12	23.24	264	2380457	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.03	96	78768	7.03	ng/uL	0.00
5) Phenol-d5	6.67	99	1467059	33.56	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.83	67	943249	37.10	ng/ul	0.00
9) 2-Chlorophenol-d4	7.01	132	1156396	34.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.19	113	1181017	34.24	ng/ul	0.00
19) Nitrobenzene-d5	8.63	128	579728	34.09	ng/ul	0.00
22) 2-Nitrophenol-d4	9.35	143	659636	34.13	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.88	165	1146315	31.32	ng/ul	0.00
29) 4-Chloroaniline-d4	10.39	131	1616856	40.25	ng/ul	0.00
44) Dimethylphthalate-d6	13.56	166	4077616	37.13	ng/ul	0.00
47) Acenaphthylene-d8	13.82	160	4473112	34.75	ng/ul	0.00
52) 4-Nitrophenol-d4	14.37	143	594803	32.21	ng/ul	0.00
58) Fluorene-d10	15.14	176	3358763	36.34	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.28	200	576215	30.50	ng/ul	0.00
71) Anthracene-d10	16.99	188	4674815	32.52	ng/ul	0.00
79) Pyrene-d10	19.30	212	5216737	43.23	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.10	264	4627276	36.77	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121119\
Data File : BM024034.D
Acq On : 11 Dec 2019 15:29
Operator : JU
Sample : PB125195BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK95

Quant Time: Dec 12 00:43:54 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM112719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Dec 07 04:26:57 2019
Response via : Initial Calibration

