


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121218\
 Data File : BM018092.D
 Acq On : 12 Dec 2018 14:08
 Operator : JU/SJ
 Sample : DFTPP77
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP77

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111918MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Dec 13 06:49:34 2018



AutoFind: Scans 2450, 2451, 2452; Background Corrected with Scan 2445

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.3	85725	PASS
68	69	0.00	2	1.7	1705	PASS
69	198	0.00	100	37.1	101720	PASS
70	69	0.00	2	0.6	575	PASS
127	198	10	80	43.6	119512	PASS
197	198	0.00	2	0.8	2137	PASS
198	198	100	100	100.0	274282	PASS
199	198	5	9	6.7	18302	PASS
275	198	10	60	28.1	77053	PASS
365	198	1	100	4.1	11380	PASS
441	198	0.01	100	13.0	35640	PASS
442	198	50	100	84.6	232072	PASS
443	442	15	24	19.1	44418	PASS

Scan 2445 (17.368 min): BM018092.D
DFTPP77

m/z	Abundance
39.90	371.0
41.00	317.0
42.90	347.0
44.00	1257.0
55.10	318.0
73.10	611.0
193.00	301.0
207.00	1638.0
208.00	416.0
281.20	471.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

m/z	Abundance
37.00	488.0
38.10	1092.0
39.10	5063.0
40.10	678.0
40.90	304.0
43.90	1374.0
49.00	802.0
50.10	19888.0
51.10	74872.0
52.10	4060.0
55.10	823.0
56.00	1956.0
57.00	4997.0
58.00	337.0
61.00	1286.0
62.00	1038.0
63.00	3509.0
63.90	489.0
65.00	1354.0
67.00	304.0
69.00	83392.0
70.00	466.0
73.00	1752.0
74.10	9041.0
75.00	14461.0
76.10	4680.0
77.10	94888.0
78.10	7547.0
79.00	7444.0
80.00	5142.0
81.00	6619.0
82.00	1910.0
83.10	1375.0
83.90	334.0
85.00	1317.0
86.00	1999.0
87.10	801.0
91.00	1721.0
92.00	1968.0
93.00	11403.0
93.90	632.0
96.00	494.0
98.00	7620.0
99.00	7310.0
100.00	564.0
101.00	3718.0
102.90	1283.0
103.90	2242.0
105.00	2326.0
105.90	807.0
107.00	27544.0
108.00	5238.0
109.10	1021.0
110.00	50392.0
111.00	7570.0
111.90	818.0
112.90	416.0
116.00	1798.0
117.00	25688.0
118.00	2021.0
122.00	2308.0
123.00	3161.0
124.00	1440.0
125.00	1291.0
126.20	957.0
127.00	94184.0
128.00	8377.0
129.00	38528.0
130.00	3220.0
131.00	669.0
132.20	375.0
134.00	1032.0
135.00	3395.0
136.10	1366.0
137.00	1523.0
139.00	317.0
140.10	528.0
141.00	5575.0
142.00	1247.0
142.90	1109.0
144.10	342.0
145.10	313.0
146.00	896.0
147.00	2476.0
148.00	6104.0
149.00	1168.0
150.00	438.0
151.20	390.0
152.10	636.0
153.00	1572.0
154.00	1408.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

155.00	2765.0
156.10	4182.0
157.10	964.0
158.00	991.0
159.00	661.0
159.90	1198.0
161.00	2272.0
161.70	412.0
165.00	1943.0
166.00	1251.0
167.00	9782.0
168.00	5964.0
169.00	788.0
170.20	478.0
171.00	652.0
172.00	798.0
173.00	1357.0
174.00	2577.0
175.00	4239.0
176.00	918.0
177.00	2434.0
178.00	869.0
179.00	8203.0
180.00	5251.0
181.00	2381.0
181.90	420.0
184.10	713.0
185.10	3298.0
186.00	26040.0
187.00	7787.0
188.00	995.0
188.90	1730.0
189.90	401.0
191.00	1158.0
192.00	1971.0
193.10	2885.0
194.10	492.0
195.10	387.0
196.00	5088.0
197.10	1733.0
198.00	195264.0
199.00	14163.0
200.10	854.0
201.40	739.0
203.00	1662.0
204.00	7334.0
205.10	12678.0
206.10	53584.0
207.00	7899.0
208.00	2502.0
209.00	776.0
210.40	1048.0
211.00	1843.0
215.10	640.0
216.00	1252.0
217.00	12985.0
218.00	1504.0
221.10	5945.0
221.80	2647.0
223.00	2978.0
224.00	26104.0
225.00	7523.0
226.00	629.0
227.00	13160.0
228.00	1890.0
229.00	2005.0
229.90	324.0
231.10	1156.0
232.90	370.0
234.10	791.0
235.00	841.0
235.90	362.0
237.00	827.0
238.00	316.0
239.00	651.0
240.10	463.0
240.90	876.0
241.90	1485.0
243.10	1207.0
244.10	21304.0
245.10	3083.0
246.00	4921.0
247.00	955.0
249.00	728.0
253.00	534.0
255.00	111368.0
256.00	18360.0
257.10	1388.0
258.00	6146.0
259.10	1018.0
265.00	2643.0
273.00	3664.0
274.00	9391.0
275.00	51312.0
276.10	7646.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

277.00	4766.0
278.00	1237.0
280.90	391.0
283.00	519.0
283.90	322.0
285.10	886.0
293.10	1120.0
294.00	424.0
295.00	385.0
296.00	14068.0
297.10	2054.0
303.10	1673.0
303.90	488.0
314.00	785.0
315.10	1140.0
316.00	1041.0
320.90	356.0
322.10	308.0
323.00	4698.0
324.00	994.0
327.00	888.0
328.00	523.0
331.90	303.0
333.00	510.0
334.10	3085.0
335.00	575.0
340.90	562.0
345.90	798.0
352.10	1324.0
353.00	1155.0
354.10	1341.0
355.10	447.0
365.00	6932.0
366.00	876.0
371.10	384.0
372.10	2095.0
373.10	489.0
383.00	537.0
389.90	419.0
402.00	1378.0
403.00	1468.0
403.90	548.0
421.10	951.0
422.10	1580.0
423.10	7443.0
424.10	2069.0
441.10	18392.0
442.10	121368.0
443.10	26032.0
444.00	2363.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

m/z	Abundance
37.10	694.0
38.00	1422.0
39.10	7823.0
39.90	591.0
41.00	329.0
44.00	1328.0
44.90	305.0
49.20	1111.0
50.10	27832.0
51.10	97872.0
52.10	5168.0
55.00	634.0
56.00	3965.0
57.00	7042.0
61.00	1459.0
62.00	1675.0
63.00	4053.0
64.00	717.0
65.10	1999.0
68.10	2697.0
69.00	120776.0
70.00	610.0
73.10	1912.0
74.00	12854.0
75.10	18008.0
76.10	7386.0
77.10	131200.0
78.10	8239.0
79.00	10180.0
80.00	7682.0
81.00	10397.0
82.00	2123.0
82.90	2015.0
84.90	1754.0
85.90	2880.0
87.00	1251.0
88.00	443.0
89.00	325.0
91.00	2125.0
92.00	3276.0
93.00	18008.0
94.10	971.0
95.20	513.0
96.10	828.0
97.00	362.0
98.00	13231.0
99.00	10240.0
100.10	964.0
101.00	5352.0
102.00	436.0
102.90	1655.0
104.00	3718.0
105.00	3174.0
106.00	924.0
107.00	39048.0
108.10	6687.0
109.00	1134.0
110.00	72576.0
111.00	11463.0
112.00	1237.0
112.90	358.0
116.00	2464.0
117.00	37368.0
118.00	2946.0
118.90	300.0
120.00	437.0
122.10	3086.0
123.00	3839.0
124.00	2171.0
125.00	1775.0
127.00	137920.0
128.00	10089.0
129.00	56296.0
130.00	5561.0
131.00	844.0
132.20	605.0
133.00	311.0
134.00	2070.0
135.00	5224.0
136.00	2332.0
137.00	2279.0
137.80	528.0
139.10	442.0
140.00	599.0
141.00	7222.0
142.00	2841.0
142.90	1391.0
144.00	458.0
144.80	554.0
146.10	1240.0
147.00	3883.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

148.00	9033.0
149.00	1754.0
150.10	514.0
151.10	1029.0
153.00	2610.0
153.90	1939.0
155.00	4339.0
156.10	6181.0
157.00	1142.0
158.00	1684.0
158.90	819.0
160.00	2413.0
161.00	3123.0
162.00	1223.0
163.10	581.0
165.00	3070.0
166.00	2003.0
167.00	14947.0
168.00	7456.0
169.10	1051.0
170.10	394.0
171.10	779.0
171.90	1220.0
173.00	1944.0
174.00	3438.0
175.00	6203.0
176.00	1361.0
177.00	2979.0
177.80	838.0
179.00	11179.0
180.00	8510.0
181.00	3817.0
181.80	463.0
183.00	462.0
184.00	1113.0
185.00	5334.0
186.00	43064.0
187.10	12260.0
187.90	1372.0
189.00	3199.0
189.90	541.0
191.00	1218.0
192.00	3785.0
193.10	4952.0
194.00	862.0
195.00	878.0
196.00	8408.0
197.10	2350.0
198.00	314752.0
199.00	21504.0
200.00	1773.0
201.40	1413.0
203.00	2276.0
204.10	10414.0
205.00	18720.0
206.10	81184.0
207.10	12481.0
208.00	3343.0
209.00	1199.0
210.00	1750.0
211.00	2806.0
211.70	972.0
214.90	1127.0
216.00	1495.0
217.00	23472.0
218.00	2948.0
221.00	13497.0
221.90	2713.0
223.00	4320.0
224.10	46696.0
225.10	12236.0
226.00	1619.0
227.00	19864.0
228.00	3236.0
229.00	4659.0
229.90	905.0
231.00	1710.0
233.00	402.0
234.00	1272.0
234.90	819.0
235.90	1063.0
237.00	2480.0
239.00	1042.0
240.00	790.0
240.90	1325.0
242.10	2373.0
243.10	2690.0
244.10	35592.0
245.10	4904.0
246.00	7863.0
246.90	1416.0
247.90	559.0
248.90	1240.0
250.80	403.0
252.10	435.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

253.10	747.0
254.20	1638.0
255.00	195072.0
256.00	28544.0
257.00	2253.0
258.00	11023.0
259.00	1837.0
260.00	324.0
260.90	355.0
263.20	324.0
265.00	5254.0
266.00	597.0
271.00	621.0
272.00	554.0
273.00	4888.0
274.00	15490.0
275.00	87104.0
276.00	12687.0
277.00	8864.0
278.00	1400.0
281.00	783.0
282.90	909.0
283.90	467.0
285.00	1213.0
293.00	1553.0
294.00	458.0
295.10	607.0
296.00	24480.0
297.00	4180.0
301.00	470.0
302.00	485.0
303.00	2760.0
304.00	782.0
308.00	474.0
308.90	394.0
310.10	390.0
314.10	1668.0
315.00	3273.0
316.00	1972.0
317.00	340.0
320.90	834.0
321.90	834.0
323.10	8275.0
324.10	1542.0
327.00	1777.0
328.00	927.0
332.00	469.0
333.00	784.0
334.10	5670.0
335.00	1582.0
341.20	880.0
346.00	1970.0
346.90	461.0
352.10	1910.0
353.10	1476.0
354.00	2527.0
355.10	736.0
365.00	12792.0
366.10	2019.0
370.00	335.0
371.00	900.0
372.00	4429.0
373.00	1179.0
383.10	1211.0
384.00	381.0
390.10	444.0
391.10	417.0
400.90	340.0
402.00	2317.0
403.00	2951.0
404.00	926.0
421.00	1929.0
422.00	2549.0
423.10	14918.0
424.00	3875.0
425.00	550.0
441.10	40592.0
442.10	256640.0
443.10	48776.0
444.10	4896.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

m/z	Abundance
37.10	565.0
38.10	1337.0
39.00	6199.0
40.00	442.0
41.00	361.0
43.10	467.0
43.90	1556.0
49.10	964.0
50.10	23288.0
51.10	84432.0
52.10	3568.0
55.00	811.0
55.90	2378.0
57.00	6676.0
58.00	363.0
61.00	1143.0
62.00	1276.0
63.00	3572.0
64.00	656.0
65.10	1368.0
68.10	2419.0
69.00	100992.0
70.00	650.0
73.10	1586.0
74.00	10824.0
75.00	16520.0
76.10	5432.0
77.10	108216.0
78.10	8003.0
79.00	8677.0
80.00	6406.0
81.00	8387.0
82.00	2537.0
83.10	2430.0
85.00	1418.0
86.00	2778.0
86.90	1111.0
87.90	402.0
91.00	1911.0
92.10	2696.0
93.00	14665.0
94.00	842.0
95.00	312.0
96.00	882.0
96.90	357.0
98.00	12261.0
99.00	8539.0
99.90	969.0
101.00	4269.0
101.90	308.0
103.00	1849.0
104.00	3367.0
105.00	2615.0
107.00	37168.0
108.00	5715.0
109.00	1483.0
110.00	70312.0
111.00	10740.0
112.00	1089.0
113.00	394.0
116.00	2476.0
117.00	34744.0
118.00	2255.0
119.00	431.0
120.10	472.0
122.00	2461.0
123.00	4097.0
124.00	1654.0
125.00	1781.0
127.00	126432.0
128.00	10019.0
129.00	51544.0
130.00	4827.0
130.90	1045.0
133.00	454.0
134.00	1623.0
135.00	4821.0
136.10	1737.0
137.00	2195.0
138.20	341.0
139.00	302.0
140.00	646.0
141.00	7897.0
142.00	1998.0
143.00	1949.0
144.10	650.0
144.90	369.0
146.00	1281.0
147.00	3875.0
148.00	9274.0
149.00	1459.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

149.90	611.0
150.90	724.0
152.00	608.0
153.00	2487.0
154.00	1664.0
155.00	3550.0
156.00	5658.0
157.00	1494.0
158.00	1266.0
159.00	921.0
160.10	2555.0
161.00	3575.0
161.90	878.0
163.10	344.0
164.00	553.0
165.00	3782.0
166.00	2345.0
167.00	12451.0
168.00	6659.0
169.10	987.0
169.90	314.0
170.90	419.0
171.90	1609.0
173.00	1403.0
174.00	3123.0
175.10	4557.0
176.10	2179.0
177.00	2552.0
178.10	1261.0
179.00	10669.0
180.00	7360.0
181.00	3353.0
182.00	593.0
184.00	1246.0
185.00	5513.0
186.00	41656.0
187.00	13102.0
188.00	1322.0
189.00	2844.0
191.00	1454.0
192.00	3591.0
193.10	4271.0
194.00	830.0
195.00	547.0
196.00	7367.0
197.10	2328.0
198.00	312832.0
199.00	19240.0
200.00	1536.0
201.50	1664.0
203.00	2698.0
204.00	11224.0
205.00	19200.0
206.10	81640.0
207.00	10836.0
208.00	3119.0
209.00	1329.0
210.40	1386.0
211.10	2923.0
213.10	303.0
215.00	1303.0
216.10	1954.0
217.00	22600.0
218.00	2576.0
221.00	12188.0
221.80	2849.0
223.00	4879.0
224.10	44752.0
225.10	10957.0
226.10	1240.0
227.00	20752.0
228.00	2649.0
229.00	3998.0
229.90	304.0
231.00	1936.0
231.90	319.0
233.90	1347.0
235.00	1214.0
235.90	1011.0
237.00	1702.0
239.00	929.0
239.90	1167.0
241.00	1000.0
242.00	2468.0
243.00	2480.0
244.10	36440.0
245.10	4280.0
246.00	7277.0
247.00	1621.0
247.90	369.0
249.00	1691.0
249.90	561.0
250.90	373.0
252.10	369.0
253.00	607.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77

254.20	2118.0
255.00	194624.0
256.00	29728.0
257.00	2484.0
258.00	11111.0
259.00	1812.0
264.10	400.0
264.90	4381.0
266.00	701.0
272.10	577.0
273.10	5604.0
274.00	16186.0
275.00	92744.0
276.10	12186.0
277.00	8099.0
278.00	1415.0
279.20	389.0
281.00	648.0
282.10	626.0
283.00	987.0
284.10	665.0
285.10	1504.0
286.00	425.0
289.10	323.0
292.10	453.0
293.00	2409.0
294.10	501.0
294.90	576.0
296.00	27864.0
297.00	3232.0
298.00	365.0
301.00	555.0
302.00	390.0
303.00	3639.0
304.00	889.0
309.90	558.0
314.10	1327.0
315.00	3454.0
316.00	1678.0
316.90	348.0
321.00	933.0
321.90	512.0
323.00	9293.0
324.10	1328.0
327.00	1808.0
327.90	779.0
331.90	488.0
332.80	598.0
334.10	6452.0
335.00	1937.0
341.00	1188.0
342.00	362.0
346.00	1950.0
346.90	403.0
352.10	3109.0
353.00	2368.0
354.00	2731.0
355.10	813.0
365.00	14416.0
366.00	2078.0
370.20	303.0
371.10	900.0
372.10	5591.0
373.10	1431.0
383.00	1219.0
383.90	333.0
390.00	597.0
390.90	495.0
391.90	418.0
401.10	361.0
402.10	1992.0
403.10	3103.0
404.00	1085.0
421.00	2265.0
422.00	2766.0
423.10	18064.0
424.10	4666.0
425.10	471.0
441.10	47936.0
442.10	318208.0
443.10	58448.0
444.10	5899.0
445.00	356.0

Instrument :
BNA_M
ClientSampleId :
DFTPP77