


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121518\
 Data File : BM018200.D
 Acq On : 15 Dec 2018 12:00
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM121518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Dec 15 21:39:37 2018



AutoFind: Scans 2462, 2463, 2464; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	34493	PASS
68	69	0.00	2	2.0	908	PASS
69	198	0.00	100	45.5	45717	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.1	49280	PASS
197	198	0.00	2	0.8	777	PASS
198	198	100	100	100.0	100416	PASS
199	198	5	9	7.2	7180	PASS
275	198	10	60	28.8	28922	PASS
365	198	1	100	4.4	4465	PASS
441	198	0.01	100	12.6	12674	PASS
442	198	50	100	82.4	82789	PASS
443	442	15	24	19.6	16253	PASS