


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121618\
 Data File : BM018219.D
 Acq On : 16 Dec 2018 10:36
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM121518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sat Dec 15 21:39:37 2018



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2455

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.3	37066	PASS
68	69	0.00	2	1.9	913	PASS
69	198	0.00	100	48.4	48096	PASS
70	69	0.00	2	0.3	143	PASS
127	198	10	80	49.3	48994	PASS
197	198	0.00	2	0.7	692	PASS
198	198	100	100	100.0	99458	PASS
199	198	5	9	7.3	7269	PASS
275	198	10	60	29.2	29040	PASS
365	198	1	100	5.5	5460	PASS
441	198	0.01	100	13.2	13102	PASS
442	198	50	100	84.3	83874	PASS
443	442	15	24	19.4	16257	PASS