

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM121718\
 Data File : BM018247.D
 Acq On : 17 Dec 2018 14:00
 Operator : JU/SJ
 Sample : J6378-10
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P001-SD012-0006-01

Quant Time: Dec 18 01:14:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM121518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 15 21:39:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	153236	20.00	ng	-0.01
21) Naphthalene-d8	10.26	136	599165	20.00	ng	0.00
39) Acenaphthene-d10	14.15	164	296471	20.00	ng	-0.01
64) Phenanthrene-d10	16.92	188	538598	20.00	ng	0.00
76) Chrysene-d12	21.14	240	539764	20.00	ng	0.00
87) Perylene-d12	23.30	264	613363	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.13	112	964772	105.17	ng	0.00
7) Phenol-d6	6.70	99	1290211	101.19	ng	0.00
23) Nitrobenzene-d5	8.65	82	804345	68.86	ng	-0.01
42) 2,4,6-Tribromophenol	15.66	330	312024	71.71	ng	0.00
45) 2-Fluorobiphenyl	12.76	172	1116865	48.79	ng	-0.01
79) Terphenyl-d14	19.57	244	1007776	42.23	ng	0.00
Target Compounds						
10) Phenol	6.72	94	75620	5.600	ng	96
13) 1,4-Dichlorobenzene	7.53	146	187023	15.218	ng	96
31) Naphthalene	10.31	128	172959	5.904	ng	98
37) 2-Methylnaphthalene	11.94	142	56244	2.722	ng	98
50) Dimethylphthalate	13.63	163	92159	3.739	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM121718\
Data File : BM018247.D
Acq On : 17 Dec 2018 14:00
Operator : JU/SJ
Sample : J6378-10
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P001-SD012-0006-01

Quant Time: Dec 18 01:14:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM121518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 15 21:39:37 2018
Response via : Initial Calibration

