

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM121718\
 Data File : BM018267.D
 Acq On : 18 Dec 2018 03:20
 Operator : JU/SJ
 Sample : J6385-03
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 121218-JETT-F-U-B

Quant Time: Dec 18 05:05:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM121518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 15 21:39:37 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	121180	20.00	ng	-0.02
21) Naphthalene-d8	10.26	136	432035	20.00	ng	-0.01
39) Acenaphthene-d10	14.15	164	194299	20.00	ng	-0.01
64) Phenanthrene-d10	16.91	188	380733	20.00	ng	-0.01
76) Chrysene-d12	21.13	240	387051	20.00	ng	-0.01
87) Perylene-d12	23.29	264	288971	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.12	112	346314	47.74	ng	0.00
7) Phenol-d6	6.69	99	252753	25.07	ng	-0.02
23) Nitrobenzene-d5	8.65	82	1072983	127.39	ng	-0.01
42) 2,4,6-Tribromophenol	15.66	330	287939	100.97	ng	-0.01
45) 2-Fluorobiphenyl	12.76	172	2036870	135.78	ng	-0.01
79) Terphenyl-d14	19.57	244	2002794	117.03	ng	0.00
Target Compounds						
10) Phenol	6.72	94	27108	2.539	ng	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM121718\
Data File : BM018267.D
Acq On : 18 Dec 2018 03:20
Operator : JU/SJ
Sample : J6385-03
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
121218-JETT-F-U-B

Quant Time: Dec 18 05:05:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM121518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 15 21:39:37 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.49 min Scan# 782
Delta R.T. -0.02 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20

Instrument :
BNA_M
ClientSampleId :
121218-JETT-F-U-B

Tgt Ion:152 Resp: 121180
Ion Ratio Lower Upper
152 100
150 151.5 116.1 174.1
115 62.3 44.6 67.0




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5



2-Fluorophenol
Concen: 47.740 ng
RT: 5.12 min Scan# 380
Delta R.T. -0.01 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20

Tgt Ion:112 Resp: 346314
Ion Ratio Lower Upper
112 100
64 60.3 45.4 68.2
63 31.3 24.0 36.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 63.00 (62.70 to 63.70): BM0





#7

Phenol-d6

Concen: 25.068 ng

RT: 6.69 min Scan# 646

Delta R.T. -0.02 min

Lab File: BM018267.D

Acq: 18 Dec 2018 03:20

Instrument :

BNA_M

ClientSampleId :

121218-JETT-F-U-B





Tgt Ion: 99 Resp: 252753

Ion Ratio Lower Upper

99 100

42 16.4 12.3 18.5

71 39.1 30.2 45.2



#10

Phenol

Concen: 2.539 ng

RT: 6.72 min Scan# 651

Delta R.T. -0.01 min

Lab File: BM018267.D

Acq: 18 Dec 2018 03:20


Tgt Ion: 94 Resp: 27108


Ion Ratio Lower Upper

94 100

65 29.5 9.5 49.5

66 46.2 21.3 61.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.26 min Scan# 1253
Delta R.T. -0.01 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20




Instrument :
BNA_M
ClientSampleId :
121218-JETT-F-U-B


Tgt Ion:	136	Resp:	432035
Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.2	13.8
54	6.3	5.2	7.8
68	6.3	5.0	7.6



#23
Nitrobenzene-d5
Concen: 127.390 ng
RT: 8.65 min Scan# 980
Delta R.T. -0.01 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20

Tgt Ion:	82	Resp:	1072983
Ion	Ratio	Lower	Upper
82	100		
128	37.7	33.4	50.2
54	41.5	33.4	50.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.15 min Scan# 1915

Delta R.T. -0.01 min

Lab File: BM018267.D

Acq: 18 Dec 2018 03:20

Instrument :

BNA_M

ClientSampleId :

121218-JETT-F-U-B


Tgt Ion:164 Resp: 194299

Ion Ratio Lower Upper

164 100

162 105.3 81.8 122.6




160 47.9 34.6 51.8



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42

2,4,6-Tribromophenol

Concen: 100.968 ng

RT: 15.66 min Scan# 2171

Delta R.T. -0.01 min

Lab File: BM018267.D

Acq: 18 Dec 2018 03:20


Tgt Ion:330 Resp: 287939

Ion Ratio Lower Upper

330 100

332 96.8 78.6 118.0



141 38.5 27.1 40.7




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#45
 2-Fluorobiphenyl
 Concen: 135.782 ng
 RT: 12.76 min Scan# 1678
 Delta R.T. -0.01 min
 Lab File: BM018267.D
 Acq: 18 Dec 2018 03:20




Instrument :
 BNA_M
 ClientSampleId :
 121218-JETT-F-U-B


Tgt Ion:172 Resp: 2036870
 Ion Ratio Lower Upper
 172 100
 171 34.4 28.3 42.5
 170 22.4 18.1 27.1



#64
 Phenanthrene-d10
 Concen: 20.000 ng
 RT: 16.91 min Scan# 2384
 Delta R.T. -0.01 min
 Lab File: BM018267.D
 Acq: 18 Dec 2018 03:20

Tgt Ion:188 Resp: 380733
 Ion Ratio Lower Upper
 188 100
 94 8.7 8.0 12.0
 80 10.1 9.0 13.6









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.13 min Scan# 3102
Delta R.T. -0.01 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20




Instrument :
BNA_M
ClientSampleId :
121218-JETT-F-U-B


Tgt Ion:240 Resp: 387051
Ion Ratio Lower Upper
240 100
120 8.7 8.2 12.2
236 25.2 20.3 30.5



#79
Terphenyl-d14
Concen: 117.029 ng
RT: 19.57 min Scan# 2836
Delta R.T. -0.01 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20

Tgt Ion:244 Resp: 2002794
Ion Ratio Lower Upper
244 100
212 7.1 5.8 8.8
122 9.7 9.0 13.6





#87
Perylene-d12
Concen: 20.000 ng
RT: 23.29 min Scan# 3469
Delta R.T. -0.01 min
Lab File: BM018267.D
Acq: 18 Dec 2018 03:20

Instrument :
BNA_M
ClientSampleId :
121218-JETT-F-U-B

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.1	17.5	26.3
265	21.6	17.3	25.9

