


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121721\
 Data File : BM033496.D
 Acq On : 17 Dec 2021 07:53
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM121621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Dec 16 17:10:15 2021






AutoFind: Scans 2469, 2470, 2471; Background Corrected with Scan 2462

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.1	58675	PASS
68	69	0.00	2	2.0	1055	PASS
69	198	0.00	100	47.0	52971	PASS
70	69	0.00	2	0.7	380	PASS
127	198	10	80	48.8	54992	PASS
197	198	0.00	2	0.7	830	PASS
198	198	100	100	100.0	112613	PASS
199	198	5	9	6.6	7431	PASS
275	198	10	60	26.2	29467	PASS
365	198	1	100	4.3	4833	PASS
441	198	0.01	100	11.0	12336	PASS
442	442	50	100	100.0	75176	PASS
443	442	15	24	19.4	14549	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121721\
Data File : BM033496.D
Acq On : 17 Dec 2021 07:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Dec 17 10:49:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM121621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 16 17:10:15 2021
Response via : Initial Calibration



TIC: BM033496.D\data.ms

(70) Pentachlorophenol (C)

16.977min (-0.006) 0.00 ng




response 109886

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	63.99
264.00	61.40	63.00
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121721\
Data File : BM033496.D
Acq On : 17 Dec 2021 07:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Dec 17 10:49:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM121621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 16 17:10:15 2021
Response via : Initial Calibration



TIC: BM033496.D\data.ms

(77) Benzidine

19.577min (-0.018) 767354.59 ng

response 401846

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.20	14.28
183.00	12.10	12.55
0.00	0.00	0.00

Instrument :
BNA_M
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/17/2021	BNA_M	<u>BM033496.D</u>
Compound Name	Response	Retention Time
DDT	319304	20.8
DDD	5705	20.36
DDE	698	19.85
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6403	325707	1.97