


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM121820\
 Data File : BM028156.D
 Acq On : 18 Dec 2020 10:55
 Operator : JU/CG
 Sample : DFTPP10
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP10

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121520MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 21 12:18:42 2020



AutoFind: Scans 2511, 2512, 2513; Background Corrected with Scan 2506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.0	115322	PASS
68	69	0.00	2	1.9	2154	PASS
69	198	0.00	100	44.5	111538	PASS
70	69	0.00	2	0.5	508	PASS
127	198	10	80	51.8	129712	PASS
197	198	0.00	2	0.7	1755	PASS
198	198	100	100	100.0	250432	PASS
199	198	5	9	6.8	16977	PASS
275	198	10	60	27.9	69781	PASS
365	198	1	100	3.4	8392	PASS
441	198	0.01	100	13.5	33741	PASS
442	198	50	100	87.0	217834	PASS
443	442	15	24	19.6	42784	PASS

Scan 2506 (17.927 min): BM028156.D
DFTPP10

m/z	Abundance
40.00	664.0
40.90	155.0
43.00	193.0
44.00	1316.0
49.00	239.0
54.90	224.0
73.10	914.0
95.90	195.0
117.00	210.0
133.00	221.0
146.90	178.0
191.00	255.0
192.90	195.0
207.00	2000.0
207.90	337.0
209.10	303.0
281.00	529.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
37.00	484.0
38.00	1267.0
39.10	7583.0
40.00	863.0
41.00	347.0
43.10	302.0
44.00	1273.0
45.00	300.0
49.00	909.0
50.10	24760.0
51.10	96312.0
52.10	4847.0
52.90	229.0
55.00	538.0
56.00	2815.0
57.00	7155.0
57.90	357.0
58.90	170.0
61.00	1410.0
62.00	1372.0
63.00	3658.0
64.00	608.0
65.00	1639.0
66.10	256.0
67.00	193.0
68.10	1595.0
69.00	89536.0
70.00	386.0
73.00	1438.0
74.00	9273.0
75.00	13858.0
76.10	4575.0
77.10	92304.0
78.10	7048.0
79.00	7045.0
80.00	4596.0
81.00	7077.0
82.00	1555.0
83.00	1471.0
83.90	252.0
85.00	1216.0
86.00	2304.0
87.00	1031.0
88.00	433.0
88.90	233.0
91.00	1625.0
92.00	1739.0
93.00	11634.0
94.00	780.0
94.90	374.0
96.00	692.0
97.20	306.0
98.00	8781.0
99.00	6599.0
100.00	579.0
101.00	3537.0
102.00	383.0
103.00	1511.0
104.00	2300.0
105.00	2381.0
106.00	830.0
107.00	26984.0
108.00	4500.0
109.10	893.0
110.00	49568.0
111.00	7838.0
112.00	954.0
112.90	222.0
115.00	275.0
116.10	1615.0
117.00	23816.0
118.00	1603.0
119.00	324.0
119.90	356.0
122.00	2380.0
123.00	2830.0
124.00	1186.0
125.00	1103.0
127.00	96336.0
128.00	7965.0
129.00	37960.0
130.00	3148.0
130.90	684.0
132.10	248.0
133.00	455.0
134.00	997.0
135.00	3043.0
136.00	1203.0
137.00	1881.0
138.00	479.0
139.00	157.0
139.90	471.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

141.00	4284.0
142.00	1238.0
143.00	974.0
144.00	224.0
145.10	303.0
146.00	1091.0
147.00	2364.0
148.00	5556.0
149.00	1315.0
150.00	372.0
151.10	512.0
152.00	517.0
153.00	1772.0
154.00	1048.0
155.00	2747.0
156.00	3538.0
157.00	775.0
158.00	810.0
159.00	731.0
159.90	1568.0
161.00	1937.0
162.00	690.0
163.00	259.0
163.90	224.0
165.00	1680.0
166.00	1499.0
167.00	8967.0
168.00	4981.0
169.00	793.0
170.10	291.0
170.90	397.0
172.00	938.0
173.00	1209.0
174.00	2176.0
175.10	3430.0
176.00	888.0
177.00	1925.0
178.00	615.0
179.00	6816.0
180.00	4728.0
181.00	2150.0
181.90	549.0
183.10	205.0
184.00	545.0
185.00	2850.0
186.00	22672.0
187.00	6879.0
188.00	703.0
189.00	1905.0
189.90	310.0
191.00	798.0
192.00	2055.0
193.00	2184.0
194.00	427.0
196.00	6135.0
197.10	1206.0
198.00	169472.0
199.00	12109.0
200.00	941.0
201.50	935.0
203.00	1229.0
204.00	5697.0
205.00	10442.0
206.10	44544.0
207.00	8170.0
208.00	2073.0
209.00	924.0
210.20	1015.0
211.00	1825.0
215.00	533.0
216.00	1228.0
217.00	12131.0
218.00	1711.0
219.10	155.0
219.90	205.0
221.00	7459.0
221.90	1591.0
223.00	3077.0
224.10	23304.0
225.00	6316.0
226.10	810.0
227.00	9755.0
228.00	1425.0
229.00	2222.0
230.00	293.0
231.00	1025.0
232.10	271.0
232.90	193.0
234.00	624.0
234.90	684.0
235.90	538.0
237.00	1013.0
238.90	328.0
240.10	385.0
241.00	634.0
242.00	1164.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

243.10	1449.0
244.10	18896.0
245.00	2836.0
246.00	3701.0
247.00	546.0
248.10	206.0
249.00	797.0
250.00	173.0
252.00	212.0
253.00	389.0
255.00	92416.0
256.00	15227.0
257.00	1030.0
258.00	5054.0
259.00	932.0
259.90	195.0
261.00	190.0
264.10	163.0
265.00	2282.0
265.90	320.0
266.90	186.0
271.10	181.0
271.90	236.0
273.00	2796.0
274.00	7437.0
275.00	44160.0
276.00	5643.0
277.00	3626.0
278.00	403.0
281.00	634.0
282.00	211.0
283.00	424.0
283.80	353.0
285.00	636.0
289.00	185.0
292.90	744.0
293.90	269.0
295.00	334.0
296.00	11805.0
297.00	1704.0
301.90	211.0
303.00	1387.0
304.00	412.0
307.90	185.0
313.90	485.0
314.90	1054.0
316.00	714.0
321.00	362.0
323.00	3726.0
324.00	686.0
327.00	934.0
328.00	421.0
331.90	155.0
333.00	283.0
334.00	2018.0
335.00	665.0
336.00	169.0
341.00	525.0
341.90	161.0
346.00	786.0
346.90	151.0
352.00	1101.0
353.00	761.0
354.00	1354.0
355.00	358.0
365.00	4842.0
366.00	675.0
371.00	179.0
372.00	2142.0
373.00	536.0
383.00	483.0
389.90	159.0
391.00	243.0
392.00	246.0
402.00	596.0
402.90	862.0
403.90	465.0
421.00	1102.0
422.00	990.0
423.00	6312.0
424.00	2190.0
424.90	355.0
441.00	17720.0
442.00	116672.0
443.00	23576.0
444.00	2212.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
35.90	150.0
37.00	622.0
38.10	1664.0
39.10	10877.0
40.00	1125.0
41.10	468.0
43.10	255.0
44.00	1309.0
45.00	310.0
49.00	1297.0
50.10	36352.0
51.10	135616.0
52.10	6980.0
53.00	287.0
55.10	700.0
56.00	4545.0
57.00	8786.0
58.00	468.0
61.00	1710.0
62.00	1923.0
63.00	5135.0
64.00	725.0
65.10	2261.0
66.10	340.0
66.90	302.0
68.20	2571.0
69.00	132224.0
70.00	550.0
73.10	1798.0
74.00	12730.0
75.00	21112.0
76.10	6956.0
77.10	144000.0
78.10	9480.0
79.00	9841.0
80.00	7858.0
81.00	10222.0
82.00	2570.0
83.00	2243.0
84.00	370.0
85.00	1826.0
86.00	3098.0
87.10	1305.0
88.00	557.0
89.00	271.0
91.00	2710.0
92.00	2510.0
93.00	16616.0
94.00	1180.0
95.00	322.0
96.00	1049.0
97.10	338.0
98.00	14044.0
99.00	9575.0
100.00	1104.0
101.00	6304.0
102.10	327.0
103.00	2039.0
104.00	3845.0
105.00	3557.0
106.00	969.0
107.00	41920.0
108.00	6877.0
109.10	1180.0
110.00	77488.0
111.00	11919.0
112.00	1702.0
113.10	515.0
115.00	310.0
116.00	2564.0
117.00	35896.0
118.00	2968.0
118.90	586.0
120.00	684.0
121.00	267.0
122.00	2745.0
123.00	4485.0
124.00	2309.0
125.00	1821.0
127.00	153792.0
128.00	11323.0
129.00	59912.0
130.00	5505.0
131.00	1000.0
132.00	530.0
133.00	494.0
133.90	1537.0
135.00	4802.0
136.00	1853.0
137.00	2397.0
138.00	461.0
138.90	291.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

140.00	797.0
141.00	7869.0
142.00	2443.0
142.90	1673.0
143.90	316.0
145.10	503.0
146.00	1354.0
147.00	3898.0
148.00	8798.0
149.00	1884.0
150.00	452.0
151.00	920.0
152.00	673.0
153.00	2380.0
154.00	1778.0
155.00	4476.0
156.00	6083.0
157.00	1345.0
158.00	1424.0
158.90	908.0
160.00	2387.0
161.00	3557.0
162.00	875.0
162.80	333.0
164.00	381.0
165.00	2622.0
166.00	2210.0
167.00	15936.0
168.00	7831.0
169.00	1196.0
170.00	491.0
171.00	668.0
172.00	1374.0
173.00	1759.0
174.00	2971.0
175.00	5399.0
176.00	1707.0
177.00	2898.0
178.00	836.0
179.00	11294.0
180.00	7759.0
181.00	3612.0
182.00	707.0
182.90	293.0
184.00	961.0
185.00	5584.0
186.00	38728.0
187.00	11278.0
188.10	1416.0
189.00	2363.0
190.00	444.0
191.00	1543.0
192.00	3417.0
193.00	3516.0
194.00	958.0
194.90	758.0
196.00	9219.0
197.10	2125.0
198.00	294720.0
199.00	20144.0
199.90	1754.0
201.50	1556.0
203.00	2261.0
204.00	9928.0
205.00	17376.0
206.10	73760.0
207.00	12342.0
208.00	2424.0
209.00	1093.0
210.10	1444.0
211.00	3123.0
212.10	422.0
213.10	246.0
215.00	865.0
216.00	1766.0
217.00	21024.0
218.00	2396.0
219.00	203.0
221.10	12764.0
222.00	2624.0
223.00	4257.0
224.00	41096.0
225.00	10448.0
226.00	1403.0
227.00	17968.0
228.00	2583.0
229.00	3794.0
230.00	708.0
231.00	1652.0
232.00	222.0
233.00	251.0
234.00	955.0
235.00	1278.0
236.00	761.0
237.00	1546.0
238.10	273.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

239.00	781.0
240.10	567.0
241.00	1025.0
242.00	2168.0
243.00	2690.0
244.10	33984.0
245.00	4789.0
246.00	6744.0
247.00	1164.0
247.90	377.0
249.00	961.0
250.00	265.0
251.00	259.0
252.00	314.0
253.10	622.0
254.10	1329.0
255.00	169984.0
256.00	25624.0
257.00	1962.0
258.00	10248.0
259.00	1555.0
260.00	314.0
261.00	296.0
264.00	311.0
265.00	3922.0
266.00	548.0
266.80	181.0
270.90	254.0
272.10	585.0
273.00	4768.0
274.00	12758.0
275.00	78904.0
276.00	11025.0
277.00	6259.0
278.00	995.0
278.90	300.0
280.90	463.0
281.90	364.0
282.90	750.0
284.00	516.0
285.10	975.0
286.10	204.0
289.00	295.0
290.00	250.0
291.00	226.0
291.90	267.0
293.00	1713.0
294.00	276.0
295.10	609.0
296.00	20912.0
297.00	3277.0
297.90	194.0
301.00	186.0
302.00	326.0
303.00	2121.0
304.00	798.0
305.00	163.0
308.00	370.0
308.80	194.0
310.00	243.0
312.90	264.0
314.00	833.0
315.00	2422.0
316.00	1314.0
316.80	323.0
320.90	730.0
322.00	354.0
323.10	6884.0
324.00	1351.0
325.10	192.0
327.00	1276.0
328.00	738.0
332.00	561.0
333.00	600.0
334.00	4370.0
335.10	985.0
341.10	728.0
342.10	190.0
346.00	1426.0
346.90	306.0
351.00	170.0
352.10	1933.0
353.10	1362.0
354.00	2186.0
354.90	499.0
358.90	206.0
364.10	171.0
365.00	9386.0
366.00	1515.0
367.00	168.0
369.90	304.0
371.00	681.0
372.00	3512.0
373.00	849.0
383.00	864.0
384.00	232.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

389.90	369.0
391.00	367.0
392.00	274.0
401.00	312.0
402.00	1639.0
403.00	1917.0
403.90	687.0
421.00	1507.0
422.00	1879.0
423.00	13405.0
424.00	3392.0
425.00	332.0
438.90	171.0
440.10	312.0
441.00	36520.0
442.00	241152.0
443.00	47016.0
444.00	4327.0
444.90	269.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

m/z	Abundance
37.00	585.0
38.10	1440.0
39.10	8771.0
40.00	822.0
41.00	350.0
43.10	296.0
44.00	1057.0
44.90	366.0
49.10	1101.0
50.10	29712.0
51.10	114040.0
52.00	6029.0
53.00	282.0
55.00	783.0
56.00	3613.0
57.00	7338.0
57.90	464.0
59.00	171.0
59.90	173.0
61.00	1590.0
62.00	1412.0
63.10	4184.0
64.00	839.0
65.00	1800.0
65.80	309.0
67.00	203.0
68.10	2297.0
69.00	112856.0
70.10	589.0
73.10	1557.0
74.00	11975.0
75.00	17880.0
76.10	6179.0
77.10	120872.0
78.10	8306.0
79.00	8766.0
80.00	6873.0
81.00	8827.0
82.00	2119.0
83.00	1853.0
84.00	243.0
85.00	1551.0
86.00	2565.0
87.10	1074.0
87.90	454.0
88.90	328.0
91.00	1938.0
92.00	2420.0
93.00	15646.0
94.00	1078.0
95.20	307.0
96.00	960.0
97.00	373.0
98.00	12058.0
99.00	8734.0
100.00	806.0
101.00	4660.0
101.90	291.0
103.00	1835.0
104.00	3013.0
105.00	3140.0
106.10	1039.0
107.00	37360.0
108.00	5713.0
109.00	1069.0
110.00	68384.0
111.00	10435.0
112.00	1248.0
113.00	519.0
113.90	153.0
115.10	337.0
116.00	2211.0
117.00	33336.0
118.00	2017.0
118.90	474.0
120.00	442.0
121.10	196.0
122.00	2995.0
123.00	4015.0
124.00	1752.0
125.00	1656.0
127.00	139008.0
128.00	10949.0
129.00	53752.0
130.00	4680.0
131.00	836.0
132.00	450.0
132.80	352.0
134.00	1594.0
135.00	4625.0
136.00	1470.0
137.00	2545.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

138.00	420.0
138.80	177.0
140.00	766.0
141.00	7283.0
142.00	2145.0
143.00	1436.0
144.00	469.0
145.00	385.0
146.00	1411.0
147.00	4085.0
148.00	8453.0
149.00	1567.0
149.90	636.0
151.00	1070.0
151.90	693.0
153.00	1995.0
154.00	1355.0
155.00	3925.0
156.10	5926.0
157.10	1032.0
157.90	1402.0
159.00	955.0
160.00	2117.0
161.00	3525.0
162.00	883.0
162.90	341.0
163.90	608.0
165.00	2537.0
166.00	2066.0
167.00	15207.0
168.00	7096.0
169.00	1102.0
169.90	393.0
170.90	753.0
172.00	1039.0
173.00	1759.0
174.00	2974.0
175.10	5186.0
176.00	1454.0
177.00	2452.0
178.00	831.0
179.00	11144.0
180.00	7448.0
181.00	2996.0
182.10	512.0
182.90	285.0
184.00	943.0
185.00	4872.0
186.00	36808.0
187.00	10269.0
188.00	1156.0
189.00	2630.0
189.90	351.0
191.00	1731.0
192.00	3668.0
193.00	3963.0
194.00	913.0
196.00	8401.0
197.10	1934.0
198.00	287104.0
199.00	18680.0
200.00	1386.0
201.60	1568.0
203.00	2055.0
204.00	10784.0
205.00	16576.0
206.10	72280.0
207.00	10786.0
208.00	2734.0
209.00	1013.0
210.00	1041.0
211.00	2695.0
212.00	399.0
213.00	203.0
215.00	960.0
216.00	1599.0
217.00	20464.0
218.00	2413.0
218.90	205.0
219.80	255.0
221.00	13189.0
221.90	2557.0
223.00	4408.0
224.10	41768.0
225.00	10476.0
226.00	1115.0
227.00	17536.0
228.00	2570.0
229.00	4039.0
230.00	390.0
231.10	1863.0
232.00	303.0
233.00	222.0
234.00	1203.0
235.00	1593.0
236.00	1042.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

237.00	1435.0
237.90	157.0
239.00	777.0
240.00	621.0
241.00	1054.0
242.00	2514.0
243.10	2342.0
244.00	34144.0
245.00	4674.0
246.00	6421.0
247.00	1447.0
248.00	252.0
249.00	1384.0
249.90	387.0
251.00	495.0
252.00	309.0
253.00	907.0
254.20	1464.0
255.00	173824.0
256.00	25160.0
257.00	2128.0
258.00	10156.0
259.00	1608.0
260.00	270.0
260.90	378.0
263.00	186.0
264.00	229.0
265.00	4173.0
265.90	572.0
267.10	221.0
270.00	261.0
271.00	324.0
272.00	431.0
273.00	5499.0
274.00	14860.0
275.00	86280.0
276.00	10100.0
277.00	6815.0
278.00	1270.0
280.90	621.0
281.90	320.0
283.10	744.0
284.00	485.0
285.00	1263.0
286.10	234.0
289.00	272.0
290.00	217.0
291.90	284.0
293.00	1216.0
294.00	442.0
295.10	574.0
296.00	22192.0
297.00	3077.0
298.00	209.0
301.00	392.0
302.10	433.0
303.00	2400.0
304.00	764.0
305.00	188.0
307.90	267.0
308.90	261.0
309.90	350.0
313.10	213.0
314.00	1225.0
315.00	2824.0
316.10	1416.0
317.00	265.0
321.00	822.0
322.10	441.0
323.10	8179.0
324.00	1199.0
326.00	168.0
327.00	1419.0
328.00	679.0
332.00	740.0
333.00	835.0
334.00	4743.0
335.00	1331.0
335.90	153.0
340.10	187.0
340.90	682.0
341.90	275.0
346.00	1747.0
347.00	234.0
352.00	2443.0
353.00	1915.0
354.10	2491.0
355.10	487.0
359.00	261.0
365.00	10949.0
365.90	1410.0
370.00	416.0
371.00	569.0
372.00	4582.0
373.00	1053.0
383.00	992.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10

384.10	358.0
390.10	424.0
390.90	375.0
391.90	223.0
400.90	203.0
402.00	2097.0
403.00	2581.0
404.00	953.0
421.00	2278.0
422.00	2249.0
423.00	15984.0
424.00	3817.0
424.90	479.0
439.60	153.0
440.20	294.0
441.00	46984.0
442.00	295680.0
443.00	57760.0
444.00	5227.0
445.00	256.0

Instrument :
BNA_M
ClientSampleId :
DFTPP10