

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121918\
 Data File : BM018311.D
 Acq On : 20 Dec 2018 00:01
 Operator : JU/SJ
 Sample : J6445-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EB06_1-3

Manual Integrations
 APPROVED

Sohil
 12/20/2018 2:55:12 PM

Quant Time: Dec 20 02:53:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM121518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 15 21:39:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	121574	20.00	ng	-0.02
21) Naphthalene-d8	10.25	136	442242	20.00	ng	-0.02
39) Acenaphthene-d10	14.14	164	225458	20.00	ng	-0.02
64) Phenanthrene-d10	16.90	188	468406	20.00	ng	-0.02
76) Chrysene-d12	21.13	240	509540	20.00	ng	-0.02
87) Perylene-d12	23.29	264	567874	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.12	112	401959	55.23	ng	0.00
7) Phenol-d6	6.69	99	514538	50.87	ng	-0.02
23) Nitrobenzene-d5	8.64	82	318920	36.99	ng	-0.02
42) 2,4,6-Tribromophenol	15.65	330	158324	47.85	ng	-0.02
45) 2-Fluorobiphenyl	12.75	172	631717	36.29	ng	-0.02
79) Terphenyl-d14	19.56	244	840177	37.29	ng	-0.01
Target Compounds						Qvalue
10) Phenol	6.72	94	33390	3.117	ng	95
31) Naphthalene	10.30	128	108721	5.028	ng	99
50) Dimethylphthalate	13.62	163	51726	2.760	ng	98
52) Acenaphthene	14.21	154	48540	3.531	ng	96
55) Dibenzofuran	14.55	168	54998	2.490	ng	96
58) Fluorene	15.20	166	62322	3.653	ng	96
71) Phenanthrene	16.95	178	949361	37.314	ng	99
72) Anthracene	17.04	178	171713	6.803	ng	96
73) Carbazole	17.33	167	76861	2.967	ng	97
75) Fluoranthene	18.98	202	1236972	41.817	ng	99
78) Pyrene	19.34	202	1037687	34.178	ng	99
81) Benzo(a)anthracene	21.11	228	572920	19.248	ng	99
83) Chrysene	21.16	228	536527	18.740	ng	98
86) Indeno(1,2,3-cd)pyrene	25.41	276	366307	12.845	ng	# 95
88) Benzo(b)fluoranthene	22.65	252	694220m	21.763	ng	
89) Benzo(k)fluoranthene	22.69	252	204488m	6.437	ng	
90) Benzo(a)pyrene	23.19	252	517936	17.072	ng	98
91) Dibenzo(a,h)anthracene	25.41	278	100198	3.565	ng	# 89
92) Benzo(a,h,i)perylene	26.06	276	350010	13.175	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121918\
Data File : BM018311.D
Acq On : 20 Dec 2018 00:01
Operator : JU/SJ
Sample : J6445-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EB06_1-3

Manual Integrations
APPROVED

Sohil
12/20/2018 2:55:12 PM

Quant Time: Dec 20 02:53:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM121518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 15 21:39:37 2018
Response via : Initial Calibration

