

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM122018\
 Data File : BM018349.D
 Acq On : 21 Dec 2018 02:37
 Operator : JU/SJ
 Sample : J6347-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P001-SD002-0612-01

Manual Integrations
 APPROVED

Sohil
 12/21/2018 2:49:46 PM

Quant Time: Dec 21 03:57:35 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM121518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 15 21:39:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	172898	20.00	ng	-0.02
21) Naphthalene-d8	10.26	136	711983	20.00	ng	-0.01
39) Acenaphthene-d10	14.15	164	365739	20.00	ng	-0.01
64) Phenanthrene-d10	16.92	188	645422	20.00	ng	0.00
76) Chrysene-d12	21.14	240	647378	20.00	ng	0.00
87) Perylene-d12	23.31	264	690070	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.18	112	1909058m	184.45	ng	0.05
7) Phenol-d6	6.72	99	2283644	158.74	ng	0.01
23) Nitrobenzene-d5	8.65	82	1441187	103.83	ng	-0.01
42) 2,4,6-Tribromophenol	15.67	330	618684	115.25	ng	0.00
45) 2-Fluorobiphenyl	12.76	172	2710912	96.00	ng	-0.01
79) Terphenyl-d14	19.57	244	2708546	94.62	ng	0.00

Target Compounds

						Qvalue
10) Phenol	6.74	94	237040	15.559	ng	96
12) 1,3-Dichlorobenzene	7.37	146	90075	6.596	ng	# 92
13) 1,4-Dichlorobenzene	7.53	146	1058237	76.314	ng	100
14) 1,2-Dichlorobenzene	7.83	146	335692	25.112	ng	97
20) 3+4-Methylphenols	8.29	107	93074	6.529	ng	# 65
30) 1,2,4-Trichlorobenzene	10.13	180	987725	79.885	ng	93
31) Naphthalene	10.32	128	4523100	129.921	ng	99
37) 2-Methylnaphthalene	11.94	142	3031083	123.457	ng	98
38) 1-Methylnaphthalene	12.16	142	1585615	66.185	ng	99
40) 1,2,4,5-Tetrachlorobenzene	12.31	216	57865	4.601	ng	# 93
46) 1,1'-Biphenyl	12.97	154	1095027	33.280	ng	99
50) Dimethylphthalate	13.62	163	215862	7.099	ng	99
52) Acenaphthene	14.22	154	61018	2.736	ng	# 90
58) Fluorene	15.22	166	203156	7.340	ng	# 91
71) Phenanthrene	16.96	178	1008440	28.765	ng	97
78) Pyrene	19.36	202	104470	2.708	ng	# 94
84) Bis(2-ethylhexyl)phthalate	21.06	149	85567	3.129	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM122018\
Data File : BM018349.D
Acq On : 21 Dec 2018 02:37
Operator : JU/SJ
Sample : J6347-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P001-SD002-0612-01

Manual Integrations
APPROVED

Sohil
12/21/2018 2:49:46 PM

Quant Time: Dec 21 03:57:35 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM121518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 15 21:39:37 2018
Response via : Initial Calibration

