




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122121\
Data File : BM033548.D
Acq On : 21 Dec 2021 10:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 21 13:56:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM122121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 21 13:51:44 2021
Response via : Initial Calibration



TIC: BM033548.D\data.ms

(70) Pentachlorophenol (C)




16.954min (-0.012) 0.00 ng

response 53830

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	61.63
264.00	61.40	60.51
0.00	0.00	0.00

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122121\  
Data File : BM033548.D  
Acq On    : 21 Dec 2021  10:34  
Operator  : CG/JU  
Sample    : DFTPP  
Misc      :  
ALS Vial  : 1    Sample Multiplier: 1
```

Quant Time: Dec 21 13:56:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM122121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 21 13:51:44 2021
Response via : Initial Calibration



TIC: BM033548.D\data.ms

(77) Benzidine

19.553min (-0.012) 0.00 ng

response **284591**

Ion	Exp%	Act%
-----	------	------

184.00	100.00	100.00
--------	--------	--------

185.00	14.20	14.37
--------	-------	-------

183.00	12.10	12.43
--------	-------	-------

0.00 0.00 0.00