


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM123019\
Data File : BM024365.D
Acq On : 30 Dec 2019 10:02
Operator : JU
Sample : DFTPP07
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP07

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121719MA.M
Title : SVOA CALIBRATION
Last Update : Tue Dec 31 03:12:42 2019



AutoFind: Scans 2450, 2451, 2452; Background Corrected with Scan 2444

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.1	51656	PASS
68	69	0.00	2	1.6	1013	PASS
69	198	0.00	100	47.9	64882	PASS
70	69	0.00	2	0.6	391	PASS
127	198	10	80	52.2	70786	PASS
197	198	0.00	2	0.7	906	PASS
198	198	100	100	100.0	135522	PASS
199	198	5	9	6.5	8770	PASS
275	198	10	60	32.2	43701	PASS
365	198	1	100	4.7	6385	PASS
441	198	0.01	100	9.9	13383	PASS
442	198	50	100	63.1	85549	PASS
443	442	15	24	19.4	16576	PASS

Scan 2444 (17.262 min): BM024365.D
DFTPP07

m/z	Abundance
40.00	1373.0
41.10	455.0
44.00	3643.0
49.00	732.0
73.00	947.0
84.00	360.0
85.90	396.0
96.10	351.0
132.80	334.0
190.90	396.0
207.00	2280.0
208.00	788.0
281.10	611.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

m/z	Abundance
36.90	473.0
38.10	1259.0
39.00	5659.0
40.00	1046.0
41.00	429.0
44.00	3672.0
45.00	466.0
49.00	854.0
50.00	19352.0
51.00	68784.0
52.00	3949.0
54.90	601.0
56.00	2309.0
57.00	4812.0
58.00	373.0
60.90	1211.0
62.00	1393.0
63.00	2961.0
63.90	509.0
65.10	1732.0
67.10	392.0
68.00	1544.0
69.00	83520.0
70.10	528.0
73.10	1244.0
74.00	9253.0
75.00	15251.0
76.00	4630.0
77.00	94208.0
78.00	6778.0
79.00	6198.0
80.00	4614.0
81.00	6220.0
82.00	1916.0
83.00	1480.0
83.90	432.0
85.00	1132.0
86.00	2084.0
87.00	788.0
87.80	371.0
91.00	1486.0
92.00	2044.0
93.00	10271.0
93.90	981.0
96.00	798.0
97.10	424.0
98.00	8329.0
99.00	7612.0
99.90	704.0
101.00	3734.0
102.00	384.0
103.00	1201.0
104.00	1997.0
105.00	2027.0
106.00	905.0
107.00	27832.0
108.00	4596.0
109.10	809.0
110.00	45072.0
111.00	7626.0
112.00	1007.0
115.90	1319.0
117.00	23848.0
118.00	1931.0
119.00	352.0
120.10	394.0
122.00	1873.0
123.00	2848.0
124.00	1372.0
125.00	1078.0
126.20	386.0
127.00	87168.0
128.00	6689.0
129.00	36880.0
130.00	2669.0
130.90	576.0
131.90	481.0
133.00	374.0
134.00	1046.0
135.00	3190.0
136.00	1104.0
137.00	1964.0
140.00	597.0
140.90	5357.0
142.00	1508.0
143.00	1002.0
144.90	479.0
146.00	809.0
147.00	2584.0
148.00	5868.0
149.00	1594.0
150.00	398.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

151.20	633.0
152.00	585.0
153.00	1428.0
154.10	956.0
155.00	2786.0
156.00	3966.0
157.00	637.0
157.90	1119.0
158.90	623.0
160.00	1280.0
160.90	1966.0
161.90	759.0
162.90	461.0
165.00	1613.0
166.00	1662.0
167.00	9227.0
168.00	7782.0
168.90	1111.0
170.20	335.0
171.00	436.0
171.90	1046.0
172.90	954.0
173.90	2290.0
175.00	4036.0
176.10	476.0
176.90	2125.0
178.00	633.0
179.00	6732.0
180.00	6121.0
180.90	2009.0
182.00	307.0
183.90	516.0
185.00	3256.0
186.00	23792.0
187.00	7368.0
188.00	980.0
189.00	1719.0
190.90	1058.0
191.90	2611.0
193.00	3047.0
194.00	577.0
196.00	5018.0
197.10	1486.0
197.90	158784.0
199.00	10468.0
199.90	520.0
201.30	663.0
203.00	1330.0
204.00	6218.0
205.00	10429.0
206.00	44648.0
207.00	8535.0
208.00	2283.0
208.90	797.0
210.00	617.0
211.00	2219.0
211.80	658.0
214.80	648.0
215.90	887.0
217.00	11483.0
217.90	1521.0
221.00	5952.0
221.80	2016.0
222.90	2802.0
224.00	24288.0
225.00	6744.0
226.00	1321.0
227.00	10089.0
228.00	1715.0
229.00	2493.0
230.00	318.0
230.90	793.0
234.00	802.0
234.90	868.0
235.80	518.0
237.00	1706.0
239.00	340.0
240.00	618.0
240.90	715.0
241.90	1731.0
243.00	1491.0
244.00	18512.0
245.00	1977.0
246.00	4433.0
247.00	1019.0
249.00	726.0
252.00	307.0
252.90	431.0
254.10	877.0
255.00	104664.0
256.00	14709.0
257.00	1701.0
258.00	7091.0
258.90	1076.0
265.00	3149.0
265.90	491.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

273.00	2814.0
274.00	8589.0
275.00	47360.0
276.00	6537.0
277.00	4825.0
278.00	1006.0
281.00	591.0
282.00	412.0
282.90	1083.0
284.00	411.0
285.00	879.0
292.00	407.0
293.00	1121.0
294.00	401.0
295.00	365.0
296.00	15700.0
297.00	2284.0
303.00	1837.0
304.00	377.0
314.00	626.0
315.00	2221.0
316.00	1040.0
321.10	504.0
323.00	4466.0
324.00	869.0
326.90	998.0
328.00	611.0
333.00	341.0
334.00	2615.0
334.90	710.0
341.10	719.0
346.00	961.0
352.00	1338.0
353.10	741.0
354.00	1406.0
354.90	451.0
365.00	6415.0
365.90	1049.0
372.00	1819.0
373.00	432.0
382.90	527.0
401.90	683.0
403.10	551.0
404.10	369.0
421.00	891.0
422.00	923.0
423.10	4721.0
424.00	2116.0
425.00	347.0
441.10	12784.0
442.10	78744.0
443.10	15817.0
444.10	1267.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

m/z	Abundance
36.90	373.0
38.10	605.0
39.00	4564.0
40.10	1684.0
41.00	480.0
44.00	3332.0
49.10	756.0
50.00	14457.0
51.00	57944.0
52.00	3287.0
55.00	442.0
56.00	2069.0
57.00	4577.0
57.90	311.0
61.00	930.0
62.00	961.0
63.00	2810.0
64.00	405.0
65.10	1322.0
67.00	323.0
68.10	1051.0
69.00	72416.0
69.90	321.0
73.00	1361.0
74.00	6481.0
75.00	12162.0
76.10	3964.0
77.00	81904.0
78.10	6310.0
79.00	5564.0
80.00	3739.0
81.00	6051.0
82.00	1430.0
83.00	1555.0
83.90	507.0
85.00	1194.0
86.00	1957.0
86.90	609.0
87.90	461.0
91.00	1056.0
92.00	1521.0
93.00	9520.0
94.00	790.0
96.00	606.0
98.00	7669.0
99.00	6499.0
99.90	446.0
101.00	2984.0
103.00	883.0
103.90	2071.0
105.00	1697.0
106.00	571.0
107.00	23600.0
108.00	3874.0
109.10	524.0
110.00	40008.0
111.00	7160.0
112.00	822.0
115.90	1533.0
117.00	21648.0
118.10	1272.0
122.00	1503.0
123.00	2680.0
124.00	1547.0
124.90	1085.0
126.20	512.0
127.00	81440.0
128.00	5915.0
129.00	31384.0
129.90	2732.0
131.00	555.0
132.90	533.0
133.90	1078.0
135.00	2825.0
135.90	881.0
136.90	1919.0
138.00	335.0
140.00	417.0
141.00	4472.0
142.00	969.0
143.10	925.0
144.00	346.0
145.00	324.0
145.90	940.0
147.00	2587.0
147.90	5281.0
149.00	1299.0
149.90	421.0
150.90	432.0
151.30	399.0
151.90	437.0
153.00	1402.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

153.90	977.0
155.00	2335.0
156.00	3958.0
157.00	679.0
157.90	651.0
159.00	530.0
160.00	1143.0
161.00	2246.0
161.80	576.0
163.00	304.0
164.90	1860.0
165.90	992.0
167.00	9210.0
168.00	6514.0
169.00	742.0
170.90	373.0
172.00	1047.0
173.00	1010.0
174.00	1382.0
175.00	3557.0
176.00	871.0
176.90	1377.0
178.10	681.0
178.90	7030.0
180.00	4428.0
181.00	2211.0
183.90	605.0
185.00	3077.0
186.00	25536.0
187.00	7185.0
188.00	839.0
189.00	1503.0
190.90	741.0
191.90	2676.0
193.00	2713.0
193.90	469.0
196.00	4081.0
197.10	815.0
197.90	159296.0
199.00	10062.0
200.00	923.0
201.50	545.0
203.00	1421.0
204.00	6651.0
205.00	11656.0
206.00	44480.0
207.00	8357.0
208.00	2033.0
208.90	923.0
209.90	937.0
211.00	1849.0
211.80	481.0
215.00	439.0
215.90	1017.0
217.00	13178.0
217.90	1514.0
219.80	323.0
221.00	6874.0
221.90	1405.0
223.00	2949.0
224.00	25160.0
225.00	6945.0
226.00	762.0
227.00	11540.0
228.00	1741.0
229.00	2113.0
230.90	1070.0
234.00	463.0
235.00	728.0
235.80	393.0
237.00	1546.0
239.10	814.0
239.80	495.0
241.10	694.0
242.00	1685.0
243.00	2030.0
244.00	19536.0
245.00	2420.0
246.00	4437.0
247.00	815.0
248.90	920.0
253.00	392.0
254.10	867.0
255.00	113536.0
256.00	17216.0
256.90	1794.0
258.00	6622.0
259.00	1325.0
265.00	3172.0
266.00	380.0
270.90	324.0
272.90	3206.0
274.00	9653.0
275.00	52784.0
276.00	7096.0
277.00	4881.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

277.90	1008.0
278.90	316.0
281.00	594.0
282.00	325.0
283.00	677.0
283.90	323.0
285.00	759.0
289.00	341.0
293.00	1227.0
294.10	522.0
295.00	446.0
296.00	17064.0
297.00	2325.0
303.00	1794.0
303.90	406.0
308.10	325.0
314.00	816.0
314.90	1529.0
316.00	994.0
321.00	436.0
323.00	4607.0
324.00	903.0
327.00	920.0
328.00	635.0
332.90	493.0
334.00	3110.0
335.00	812.0
341.00	727.0
345.90	942.0
347.00	379.0
352.10	1414.0
353.00	1054.0
354.00	1303.0
355.00	525.0
365.10	7352.0
366.00	999.0
371.00	430.0
372.00	2798.0
373.00	635.0
383.00	696.0
384.00	325.0
402.10	735.0
403.00	1075.0
404.00	614.0
421.00	879.0
422.10	898.0
423.10	6237.0
424.00	1996.0
441.10	16480.0
442.10	102896.0
443.10	19936.0
444.00	2055.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

m/z	Abundance
37.90	411.0
39.00	2248.0
39.90	1085.0
41.00	308.0
44.00	3923.0
49.00	573.0
50.00	7721.0
51.00	28240.0
52.00	1526.0
55.00	390.0
55.90	718.0
57.00	2282.0
61.00	576.0
62.00	872.0
63.00	1670.0
65.10	698.0
68.10	445.0
69.00	38712.0
70.00	324.0
73.00	1520.0
74.00	4336.0
75.00	5691.0
76.00	2181.0
77.00	42024.0
78.00	3482.0
79.00	3166.0
80.00	1823.0
81.00	3082.0
81.90	722.0
83.00	663.0
83.90	351.0
84.90	582.0
85.90	1078.0
87.10	475.0
87.90	456.0
91.00	1067.0
91.90	778.0
93.00	5203.0
94.00	378.0
95.90	433.0
98.00	3312.0
98.90	3449.0
99.90	322.0
101.00	1670.0
103.10	570.0
103.90	1205.0
105.00	1390.0
107.00	12299.0
108.00	1689.0
109.10	461.0
110.00	22816.0
110.90	3574.0
112.10	424.0
116.10	715.0
117.00	12657.0
118.00	1039.0
119.80	310.0
122.00	852.0
123.00	1643.0
123.90	934.0
124.90	698.0
127.00	43752.0
128.00	2826.0
129.00	17184.0
129.90	1542.0
133.90	812.0
135.00	1625.0
135.90	714.0
137.00	1125.0
140.90	2454.0
141.90	969.0
143.00	520.0
146.10	697.0
147.00	1549.0
148.00	3231.0
149.00	1081.0
151.10	381.0
152.90	683.0
153.90	736.0
155.00	1377.0
156.00	1993.0
156.90	510.0
158.00	547.0
160.00	892.0
161.00	1258.0
161.90	358.0
164.90	933.0
166.00	699.0
167.00	5551.0
168.00	4235.0
168.90	575.0
170.90	303.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

171.90	398.0
173.00	781.0
174.00	1002.0
174.90	2406.0
176.00	370.0
176.90	1073.0
178.00	423.0
178.90	4190.0
180.00	2836.0
181.00	1186.0
184.00	435.0
185.00	2152.0
186.00	14106.0
187.00	3617.0
188.10	474.0
189.00	870.0
190.90	584.0
192.00	1463.0
193.00	1549.0
195.90	2556.0
197.20	417.0
197.90	88488.0
199.00	5780.0
199.90	371.0
201.40	385.0
203.00	901.0
204.00	3731.0
205.00	6871.0
206.00	25872.0
207.00	5444.0
208.00	1606.0
209.00	647.0
210.10	482.0
211.00	1266.0
215.00	372.0
215.90	884.0
217.00	7033.0
217.90	869.0
221.00	4008.0
221.90	1104.0
222.90	1522.0
224.00	14790.0
225.00	3426.0
226.00	354.0
227.00	7594.0
228.00	1154.0
229.00	1504.0
231.00	416.0
233.90	469.0
234.80	401.0
235.90	444.0
236.90	802.0
238.90	318.0
241.00	505.0
241.90	717.0
243.00	845.0
244.00	11414.0
245.00	1685.0
246.00	2749.0
247.00	488.0
249.00	717.0
252.90	374.0
255.00	69704.0
256.00	9710.0
256.90	884.0
258.00	3779.0
259.00	521.0
265.00	2100.0
265.90	469.0
272.90	1954.0
274.00	5105.0
275.00	30960.0
276.00	3977.0
277.00	3118.0
277.90	759.0
281.10	809.0
283.00	653.0
284.00	339.0
285.00	685.0
293.00	742.0
295.20	453.0
296.00	11543.0
296.90	1495.0
303.00	1199.0
314.00	661.0
315.00	1283.0
316.00	566.0
323.00	3358.0
324.00	600.0
327.00	674.0
332.90	404.0
334.00	2644.0
334.90	767.0
341.00	491.0
346.00	822.0
351.90	860.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07

352.90	404.0
354.00	882.0
365.00	5390.0
366.00	658.0
372.10	1965.0
382.80	356.0
402.00	726.0
403.10	510.0
404.10	435.0
421.10	825.0
422.10	502.0
423.00	4417.0
424.00	1234.0
441.10	10885.0
442.10	75008.0
443.10	13975.0
444.00	1238.0

Instrument :
BNA_M
ClientSampleId :
DFTPP07