


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM010421\
 Data File : BM028331.D
 Acq On : 04 Jan 2021 11:42
 Operator : JU/CG
 Sample : DFTPP07
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP007

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM010421.M
 Title : SVOA CALIBRATION
 Last Update : Mon Jan 04 17:00:49 2021



AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2495

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.6	19212	PASS
68	69	0.00	2	2.0	361	PASS
69	198	0.00	100	46.3	18307	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.4	20713	PASS
197	198	0.00	2	0.3	124	PASS
198	198	100	100	100.0	39520	PASS
199	198	5	9	6.9	2712	PASS
275	198	10	60	29.3	11598	PASS
365	198	1	100	4.1	1609	PASS
441	198	0.01	100	14.2	5597	PASS
442	198	50	100	92.6	36581	PASS
443	442	15	24	19.2	7023	PASS

Scan 2495 (17.863 min): BM028331.D
DFTPP07

m/z	Abundance
39.90	151.0
44.00	417.0
207.00	289.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

m/z	Abundance
38.00	291.0
39.00	1767.0
40.00	204.0
44.00	450.0
49.00	198.0
50.00	6302.0
51.00	21840.0
52.00	1092.0
55.00	175.0
56.00	707.0
57.00	1570.0
61.00	274.0
61.90	331.0
63.00	718.0
65.00	388.0
68.10	410.0
69.00	20120.0
73.00	348.0
74.00	2092.0
75.00	3267.0
76.10	1091.0
77.10	21264.0
78.00	1576.0
79.00	1663.0
80.00	1145.0
81.00	1676.0
81.90	376.0
83.00	392.0
85.00	243.0
85.90	453.0
86.90	277.0
90.90	369.0
92.00	389.0
93.00	2831.0
93.90	186.0
98.00	2069.0
99.00	1900.0
100.00	161.0
101.00	803.0
103.00	322.0
103.90	611.0
105.00	585.0
106.00	182.0
107.00	6275.0
108.00	987.0
109.10	188.0
110.00	11006.0
111.00	1888.0
112.00	257.0
116.00	396.0
117.00	6424.0
118.00	469.0
122.00	394.0
123.00	675.0
124.00	331.0
124.90	285.0
127.00	21824.0
128.00	1693.0
129.00	8846.0
130.00	796.0
130.90	161.0
133.90	274.0
135.00	706.0
136.00	303.0
137.10	404.0
141.00	1239.0
141.90	345.0
143.00	268.0
145.90	299.0
147.00	658.0
148.00	1496.0
149.00	357.0
152.90	344.0
153.90	265.0
155.00	634.0
155.90	840.0
157.00	152.0
157.90	245.0
160.00	388.0
160.90	516.0
164.90	518.0
166.00	338.0
167.00	2306.0
168.00	1397.0
168.90	178.0
171.90	179.0
173.00	299.0
173.90	435.0
175.00	909.0
176.00	263.0
177.00	430.0
178.90	1618.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

180.00	1124.0
181.00	491.0
185.00	793.0
186.00	5317.0
187.00	1594.0
187.90	210.0
188.90	451.0
191.00	177.0
191.90	585.0
193.00	583.0
196.00	1215.0
198.00	37928.0
198.90	2718.0
200.00	221.0
201.50	250.0
203.00	277.0
204.00	1543.0
205.00	2406.0
206.00	10640.0
207.00	1765.0
208.00	483.0
208.90	188.0
210.00	167.0
211.00	436.0
215.90	300.0
217.00	2815.0
218.00	411.0
221.00	1646.0
222.00	450.0
222.90	655.0
224.00	6090.0
225.00	1484.0
226.00	178.0
227.00	2605.0
227.90	392.0
229.00	517.0
231.00	186.0
234.00	205.0
234.90	191.0
237.00	305.0
241.00	166.0
242.00	308.0
243.00	338.0
244.00	4570.0
245.00	633.0
246.00	926.0
246.90	200.0
248.90	191.0
255.00	23280.0
256.00	3473.0
256.90	313.0
258.00	1339.0
259.00	289.0
265.00	576.0
273.00	638.0
274.00	1657.0
275.00	10044.0
276.00	1349.0
276.90	777.0
277.90	167.0
285.00	154.0
293.00	194.0
296.00	2870.0
297.00	337.0
303.00	294.0
314.90	353.0
316.00	179.0
323.00	843.0
326.90	191.0
334.00	664.0
345.90	206.0
352.00	205.0
352.90	183.0
353.90	305.0
364.90	1266.0
366.00	181.0
372.00	422.0
383.00	158.0
401.90	155.0
402.90	242.0
421.00	217.0
421.90	180.0
423.00	1581.0
423.90	597.0
441.00	3965.0
442.00	26632.0
443.00	5162.0
444.00	440.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

m/z	Abundance
38.00	336.0
39.00	1804.0
40.00	198.0
44.00	367.0
49.00	272.0
50.00	6459.0
51.00	22792.0
52.00	1174.0
55.00	152.0
56.00	696.0
57.00	1521.0
60.90	307.0
62.00	291.0
63.00	867.0
65.00	393.0
68.00	425.0
69.00	22032.0
73.00	316.0
74.00	2328.0
75.00	3437.0
76.10	1048.0
77.10	23208.0
78.00	1616.0
79.00	1770.0
80.00	1241.0
81.00	1726.0
82.00	449.0
83.00	384.0
85.00	375.0
86.00	661.0
86.90	180.0
91.00	362.0
92.00	423.0
93.00	3080.0
94.00	156.0
96.00	165.0
98.00	2416.0
99.00	1972.0
100.10	155.0
101.00	983.0
103.00	315.0
103.90	654.0
105.00	613.0
106.10	198.0
107.00	7056.0
108.00	1122.0
108.90	193.0
110.00	12274.0
111.00	2001.0
112.00	282.0
115.90	421.0
117.00	6952.0
117.90	494.0
122.00	442.0
123.00	753.0
123.90	366.0
125.00	346.0
127.00	24888.0
128.00	1997.0
129.00	9840.0
130.00	887.0
130.90	155.0
134.00	294.0
135.00	912.0
136.00	264.0
137.00	504.0
141.00	1336.0
141.90	393.0
142.90	229.0
146.00	290.0
147.00	675.0
148.00	1698.0
148.90	338.0
151.00	151.0
152.90	441.0
153.90	287.0
155.00	752.0
156.00	1127.0
157.00	214.0
157.90	224.0
158.90	162.0
160.00	422.0
161.00	562.0
164.90	521.0
165.90	364.0
167.00	2725.0
168.00	1447.0
169.00	254.0
171.90	209.0
173.00	259.0
173.90	528.0
175.00	959.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

175.90	306.0
177.00	533.0
178.00	195.0
179.00	1898.0
180.00	1476.0
181.00	774.0
185.00	1002.0
186.00	6793.0
187.00	2099.0
188.00	247.0
188.90	464.0
190.90	224.0
192.00	579.0
193.00	729.0
194.00	160.0
196.00	1366.0
196.70	325.0
197.10	373.0
198.00	48192.0
199.00	3310.0
200.00	215.0
201.50	210.0
202.90	384.0
204.00	1849.0
205.00	2996.0
206.00	13468.0
207.00	2138.0
208.00	493.0
209.00	201.0
210.00	212.0
210.90	571.0
214.90	163.0
216.00	299.0
217.00	3639.0
218.00	463.0
220.90	1934.0
221.80	420.0
223.00	688.0
224.00	7217.0
225.00	1697.0
226.00	234.0
227.00	3358.0
227.90	430.0
228.90	632.0
230.90	234.0
233.90	217.0
234.90	251.0
235.90	183.0
237.00	341.0
240.90	181.0
242.00	412.0
243.00	437.0
244.00	6019.0
245.00	728.0
246.00	1222.0
246.90	258.0
248.90	181.0
254.20	274.0
255.00	30688.0
256.00	4532.0
256.90	414.0
258.00	1688.0
258.90	298.0
265.00	731.0
273.00	968.0
274.00	2488.0
275.00	14294.0
276.00	1921.0
277.00	1180.0
278.00	246.0
282.90	166.0
292.90	265.0
296.00	3768.0
297.00	584.0
302.90	483.0
314.90	394.0
316.00	199.0
323.00	1200.0
323.90	202.0
326.90	301.0
334.00	871.0
334.90	179.0
341.00	164.0
345.90	338.0
352.00	454.0
352.90	326.0
354.00	407.0
365.00	2009.0
365.90	345.0
372.00	684.0
373.00	177.0
382.90	150.0
401.90	277.0
403.00	462.0
420.90	438.0
421.90	370.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

423.00	2276.0
423.90	792.0
441.00	6794.0
442.00	43888.0
443.00	8464.0
443.90	768.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

m/z	Abundance
38.00	201.0
39.00	1044.0
39.90	192.0
44.00	441.0
49.00	157.0
50.10	3729.0
51.10	13004.0
52.00	634.0
56.00	396.0
57.00	899.0
60.90	178.0
62.00	208.0
63.00	573.0
65.00	215.0
68.00	250.0
69.00	12769.0
72.90	319.0
74.00	1268.0
75.00	2153.0
76.00	651.0
77.10	13975.0
78.00	1011.0
79.00	979.0
80.00	799.0
81.00	1063.0
82.00	286.0
83.00	246.0
85.90	358.0
90.90	267.0
91.90	247.0
93.00	1877.0
98.00	1557.0
99.00	1101.0
100.00	167.0
101.00	566.0
102.90	172.0
104.00	486.0
104.90	449.0
107.00	4246.0
107.90	772.0
110.00	7738.0
111.00	1216.0
116.00	258.0
117.00	4294.0
118.00	226.0
121.90	373.0
123.00	475.0
123.90	248.0
124.90	246.0
127.00	15429.0
128.00	1177.0
129.00	6156.0
129.90	465.0
133.90	210.0
135.00	561.0
136.00	198.0
137.00	309.0
141.00	785.0
141.90	312.0
142.90	155.0
147.00	476.0
148.00	1134.0
149.00	241.0
153.00	249.0
153.90	188.0
154.90	462.0
156.00	749.0
157.80	199.0
158.90	157.0
160.00	307.0
161.00	322.0
164.90	304.0
166.00	302.0
167.00	1787.0
168.00	877.0
169.00	159.0
172.00	180.0
173.00	168.0
174.00	286.0
175.00	574.0
175.90	229.0
177.00	281.0
178.90	1366.0
180.00	969.0
180.90	425.0
185.00	661.0
186.00	4982.0
187.00	1225.0
187.90	184.0
188.90	371.0
190.90	153.0
192.00	468.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007

193.00	511.0
196.00	1098.0
198.00	32440.0
198.90	2109.0
199.90	153.0
202.90	271.0
204.00	1314.0
205.00	2104.0
206.00	9073.0
207.00	1494.0
208.00	313.0
209.00	195.0
210.90	341.0
215.90	215.0
217.00	2737.0
218.00	309.0
221.00	1390.0
222.90	586.0
224.00	4934.0
225.00	1471.0
227.00	2090.0
227.90	287.0
229.00	448.0
231.00	189.0
234.00	161.0
237.00	228.0
242.00	344.0
243.10	341.0
244.00	4127.0
245.10	536.0
246.00	862.0
248.90	170.0
254.20	232.0
255.00	21864.0
256.00	3057.0
257.00	218.0
258.00	1283.0
259.00	245.0
265.00	513.0
273.00	697.0
274.00	1891.0
275.00	10457.0
276.00	1377.0
277.00	829.0
277.90	182.0
292.90	171.0
296.00	2962.0
296.90	390.0
303.00	325.0
314.00	151.0
314.90	319.0
316.00	188.0
323.00	1036.0
324.00	196.0
326.90	163.0
334.00	606.0
345.90	260.0
352.00	290.0
352.90	204.0
353.90	361.0
365.00	1553.0
365.90	214.0
372.00	514.0
382.90	153.0
401.90	209.0
402.90	281.0
404.00	172.0
420.90	333.0
422.00	284.0
423.00	2049.0
424.00	665.0
441.00	6033.0
442.00	39224.0
443.00	7444.0
444.00	680.0

Instrument :
BNA_M
ClientSampleId :
DFTPP007