

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010623\
 Data File : BM038398.D
 Acq On : 06 Jan 2023 22:26
 Operator : CG/JU
 Sample : 01036-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0LD7

Quant Time: Jan 07 00:40:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:11:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	78379	20.000	ng/ul	0.00
20) Naphthalene-d8	10.810	136	297704	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.627	164	178792	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.368	188	391047	20.000	ng/ul	0.00
79) Chrysene-d12	21.539	240	381109	20.000	ng/ul	0.00
88) Perylene-d12	23.968	264	455333	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	9908	4.099	ng/uL	0.00
4) Pyridine-d5	3.805	84	34822	4.980	ng/ul	0.00
7) Phenol-d5	7.157	99	36750	5.147	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.322	67	131134	25.530	ng/ul	0.00
11) 2-Chlorophenol-d4	7.522	132	100351	19.615	ng/ul	0.00
15) 4-Methylphenol-d8	8.710	113	64348	11.675	ng/ul	0.00
21) Nitrobenzene-d5	9.169	128	59516	25.481	ng/ul	0.00
24) 2-Nitrophenol-d4	9.892	143	65922	24.886	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.428	165	118119	23.091	ng/ul	0.00
31) 4-Chloroaniline-d4	10.951	131	99843	14.145	ng/ul	0.00
46) Dimethylphthalate-d6	14.039	166	377244	27.861	ng/ul	0.00
49) Acenaphthylene-d8	14.321	160	409542	25.954	ng/ul	0.00
54) 4-Nitrophenol-d4	14.833	143	8696	3.568	ng/ul	0.00
60) Fluorene-d10	15.616	176	335972	27.284	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.739	200	66640	23.513	ng/ul	0.00
73) Anthracene-d10	17.468	188	594464	31.513	ng/ul	0.00
81) Pyrene-d10	19.756	212	715625	33.744	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.809	264	715862	31.654	ng/ul	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010623\
 Data File : BM038398.D
 Acq On : 06 Jan 2023 22:26
 Operator : CG/JU
 Sample : 01036-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COLD7

Quant Time: Jan 07 00:40:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:11:50 2023
 Response via : Initial Calibration

