

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010623\
 Data File : BM038391.D
 Acq On : 06 Jan 2023 18:07
 Operator : CG/JU
 Sample : PB150075BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK075

Quant Time: Jan 07 00:38:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Jan 05 02:11:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.998	152	56970	20.000	ng/u1	0.00
20) Naphthalene-d8	10.810	136	214395	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.633	164	126720	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.374	188	278522	20.000	ng/u1	0.00
79) Chrysene-d12	21.539	240	264579	20.000	ng/u1	0.00
88) Perylene-d12	23.968	264	337566	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.363	96	9745	5.547	ng/uL	0.00
4) Pyridine-d5	3.799	84	139711	27.489	ng/u1	0.00
7) Phenol-d5	7.157	99	149067	28.723	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.328	67	115129	30.838	ng/u1	0.00
11) 2-Chlorophenol-d4	7.528	132	114456	30.780	ng/u1	0.00
15) 4-Methylphenol-d8	8.710	113	103913	25.938	ng/u1	0.00
21) Nitrobenzene-d5	9.175	128	47551	28.270	ng/u1	0.00
24) 2-Nitrophenol-d4	9.892	143	52397	27.467	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.433	165	95368	25.888	ng/u1	0.00
31) 4-Chloroaniline-d4	10.957	131	124219	24.436	ng/u1	0.00
46) Dimethylphthalate-d6	14.039	166	275871	28.746	ng/u1	0.00
49) Acenaphthylene-d8	14.327	160	309778	27.699	ng/u1	0.00
54) 4-Nitrophenol-d4	14.833	143	36343	21.036	ng/u1	0.00
60) Fluorene-d10	15.621	176	237769	27.244	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.745	200	46437	23.005	ng/u1	0.00
73) Anthracene-d10	17.468	188	383795	28.565	ng/u1	0.00
81) Pyrene-d10	19.756	212	459570	31.215	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	474859	28.323	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM010623\
Data File : BM038391.D
Acq On : 06 Jan 2023 18:07
Operator : CG/JU
Sample : PB150075BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK075

Quant Time: Jan 07 00:38:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM010423.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Jan 05 02:11:50 2023
Response via : Initial Calibration

