

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM010920\
 Data File : BM024447.D
 Acq On : 10 Jan 2020 05:13
 Operator : JU
 Sample : K6466-09
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 H0BP7

Quant Time: Jan 10 07:55:12 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM010820MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jan 10 07:52:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.49	152	301198	20.00	ng/ul	0.00
18) Naphthalene-d8	10.25	136	1134375	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.13	164	681580	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.89	188	1472362	20.00	ng/ul	0.00
78) Chrysene-d12	21.09	240	1484682	20.00	ng/ul	0.00
86) Perylene-d12	23.23	264	1756709	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.10	96	40087	5.74	ng/uL	0.00
5) Phenol-d5	6.68	99	115935	4.93	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.84	67	342551	24.83	ng/ul	0.00
9) 2-Chlorophenol-d4	7.03	132	388509	19.54	ng/ul	0.00
13) 4-Methylphenol-d8	8.20	113	220823	11.69	ng/ul	0.00
19) Nitrobenzene-d5	8.63	128	242987	28.56	ng/ul	0.00
22) 2-Nitrophenol-d4	9.35	143	194558	23.86	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.88	165	411353	22.99	ng/ul	0.00
29) 4-Chloroaniline-d4	10.39	131	45591	2.26	ng/ul	0.00
44) Dimethylphthalate-d6	13.56	166	1626172	33.22	ng/ul	0.00
47) Acenaphthylene-d8	13.82	160	1855016	29.35	ng/ul	0.00
52) 4-Nitrophenol-d4	14.34	143	39170	4.54	ng/ul	0.00
58) Fluorene-d10	15.14	176	1402743	32.33	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.26	200	154577	23.43	ng/ul	0.00
71) Anthracene-d10	16.98	188	2268381	32.89	ng/ul	0.00
79) Pyrene-d10	19.29	212	2617649	33.00	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.09	264	2989125	33.32	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM010920\
Data File : BM024447.D
Acq On : 10 Jan 2020 05:13
Operator : JU
Sample : K6466-09
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
H0BP7

Quant Time: Jan 10 07:55:12 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM010820MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jan 10 07:52:51 2020
Response via : Initial Calibration

