

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011923\
 Data File : BM038485.D
 Acq On : 19 Jan 2023 09:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 20 02:45:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 19 06:09:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.599	0.606	-1.2	85	0.00
3	Pyridine	1.733	1.810	-4.4	81	0.00
4	n-Nitrosodimethylamine	1.026	1.143	-11.4	89	0.00
5 S	2-Fluorophenol	1.311	1.321	-0.8	84	0.00
6	Aniline	2.121	2.208	-4.1	84	0.00
7 S	Phenol-d6	1.680	1.695	-0.9	80	0.00
8	2-Chlorophenol	1.298	1.333	-2.7	82	0.00
9	Benzaldehyde	1.123	1.105	1.6	78	0.00
10 C	Phenol	1.750	1.793	-2.5	81	0.00
11	bis(2-Chloroethyl)ether	1.418	1.437	-1.3	82	0.00
12	1,3-Dichlorobenzene	1.391	1.386	0.4	80	-0.01
13 C	1,4-Dichlorobenzene	1.399	1.406	-0.5	80	0.00
14	1,2-Dichlorobenzene	1.368	1.404	-2.6	82	0.00
15	Benzyl Alcohol	1.345	1.503	-11.7	84	0.00
16	2,2'-oxybis(1-Chloropropane	1.977	2.206	-11.6	87	0.00
17	2-Methylphenol	1.182	1.278	-8.1	83	0.00
18	Hexachloroethane	0.573	0.640	-11.7	87	-0.01
19 P	n-Nitroso-di-n-propylamine	1.219	1.424	-16.8	88	-0.01
20	3+4-Methylphenols	1.587	1.726	-8.8	84	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
22	Acetophenone	0.591	0.602	-1.9	86	0.00
23 S	Nitrobenzene-d5	0.491	0.516	-5.1	88	0.00
24	Nitrobenzene	0.519	0.534	-2.9	87	0.00
25	Isophorone	0.869	0.892	-2.6	86	0.00
26 C	2-Nitrophenol	0.182	0.183	-0.5	83	0.00
27	2,4-Dimethylphenol	0.310	0.302	2.6	82	0.00
28	bis(2-Chloroethoxy)methane	0.489	0.517	-5.7	88	0.00
29 C	2,4-Dichlorophenol	0.343	0.342	0.3	83	0.00
30	1,2,4-Trichlorobenzene	0.396	0.382	3.5	82	0.00
31	Naphthalene	1.060	1.070	-0.9	86	0.00
32	Benzoic acid	0.236	0.242	-2.5	87	0.00
33	4-Chloroaniline	0.465	0.475	-2.2	83	0.00
34 C	Hexachlorobutadiene	0.278	0.269	3.2	84	0.00
35	Caprolactam	0.097	0.099	-2.1	82	0.00
36 C	4-Chloro-3-methylphenol	0.373	0.394	-5.6	86	0.00
37	2-Methylnaphthalene	0.775	0.784	-1.2	85	0.00
38	1-Methylnaphthalene	0.728	0.740	-1.6	84	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	0.749	0.701	6.4	82	0.00
41 P	Hexachlorocyclopentadiene	0.426	0.416	2.3	82	0.00
42 S	2,4,6-Tribromophenol	0.338	0.329	2.7	81	0.00
43 C	2,4,6-Trichlorophenol	0.473	0.461	2.5	84	0.00
44	2,4,5-Trichlorophenol	0.558	0.533	4.5	81	0.00
45 S	2-Fluorobiphenyl	1.604	1.563	2.6	85	0.00
46	1,1'-Biphenyl	1.552	1.503	3.2	85	0.00
47	2-Chloronaphthalene	1.221	1.172	4.0	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011923\
 Data File : BM038485.D
 Acq On : 19 Jan 2023 09:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 20 02:45:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 19 06:09:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.435	0.482	-10.8	90	0.00
49	Acenaphthylene	1.925	1.893	1.7	85	0.00
50	Dimethylphthalate	1.621	1.554	4.1	83	0.00
51	2,6-Dinitrotoluene	0.330	0.331	-0.3	84	0.00
52 C	Acenaphthene	1.170	1.165	0.4	86	0.00
53	3-Nitroaniline	0.329	0.338	-2.7	84	0.00
54 P	2,4-Dinitrophenol	0.230	0.226	1.7	83	0.00
55	Dibenzofuran	1.992	1.988	0.2	86	0.00
56 P	4-Nitrophenol	0.273	0.273	0.0	84	0.00
57	2,4-Dinitrotoluene	0.458	0.471	-2.8	85	0.00
58	Fluorene	1.665	1.679	-0.8	87	0.00
59	2,3,4,6-Tetrachlorophenol	0.495	0.475	4.0	82	0.00
60	Diethylphthalate	1.663	1.692	-1.7	88	0.00
61	4-Chlorophenyl-phenylether	0.914	0.929	-1.6	86	0.00
62	4-Nitroaniline	0.342	0.372	-8.8	87	0.00
63	Azobenzene	1.830	2.102	-14.9	97	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
65	4,6-Dinitro-2-methylphenol	0.136	0.142	-4.4	86	0.00
66 c	n-Nitrosodiphenylamine	0.587	0.614	-4.6	87	0.00
67	4-Bromophenyl-phenylether	0.231	0.237	-2.6	84	0.00
68	Hexachlorobenzene	0.259	0.263	-1.5	84	0.00
69	Atrazine	0.208	0.215	-3.4	85	0.00
70 C	Pentachlorophenol	0.189	0.201	-6.3	85	0.00
71	Phenanthrene	1.106	1.098	0.7	83	0.00
72	Anthracene	1.137	1.164	-2.4	85	0.00
73	Carbazole	0.994	1.069	-7.5	89	0.00
74	Di-n-butylphthalate	1.181	1.322	-11.9	90	0.00
75 C	Fluoranthene	1.391	1.502	-8.0	88	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
77	Benzidine	0.438	0.421	3.9	88	0.00
78	Pyrene	1.331	1.302	2.2	90	0.00
79 S	Terphenyl-d14	1.068	1.129	-5.7	93	0.00
80	Butylbenzylphthalate	0.487	0.497	-2.1	92	0.00
81	Benzo(a)anthracene	1.438	1.455	-1.2	93	0.00
82	3,3'-Dichlorobenzidine	0.463	0.464	-0.2	91	0.00
83	Chrysene	1.398	1.410	-0.9	93	0.00
84	Bis(2-ethylhexyl)phthalate	0.699	0.747	-6.9	98	0.00
85 c	Di-n-octyl phthalate	1.231	1.311	-6.5	99	0.00
86 I	Perylene-d12	1.000	1.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	1.480	1.543	-4.3	96	0.00
88	Benzo(b)fluoranthene	1.240	1.242	-0.2	91	0.00
89	Benzo(k)fluoranthene	1.282	1.326	-3.4	95	0.00
90 C	Benzo(a)pyrene	1.217	1.255	-3.1	94	0.00
91	Dibenzo(a,h)anthracene	1.242	1.304	-5.0	97	0.00
92	Benzo(g,h,i)perylene	1.232	1.263	-2.5	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM011923\
Data File : BM038485.D
Acq On : 19 Jan 2023 09:22
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Jan 20 02:45:15 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM011623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 19 06:09:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0