

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012523\
 Data File : BM038526.D
 Acq On : 25 Jan 2023 10:59
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 01/26/2023
 Supervised By : Jagrut Upadhyay 01/26/2023

Quant Time: Jan 25 23:12:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 21 01:21:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	67675	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	247214	20.000	ng	0.00
39) Acenaphthene-d10	14.598	164	168776	20.000	ng	0.00
64) Phenanthrene-d10	17.339	188	414630	20.000	ng	0.00
76) Chrysene-d12	21.515	240	475086	20.000	ng	0.00
86) Perylene-d12	23.927	264	535335	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	339417	82.911	ng	0.00
7) Phenol-d6	7.128	99	435791	83.435	ng	0.00
23) Nitrobenzene-d5	9.139	82	469820	82.528	ng	0.00
42) 2,4,6-Tribromophenol	16.086	330	230157	94.698	ng	0.00
45) 2-Fluorobiphenyl	13.227	172	1142768	83.405	ng	0.00
79) Terphenyl-d14	19.956	244	1941121	79.654	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.375	88	77435	40.328	ng	99
3) Pyridine	3.787	79	216421	42.915	ng	99
4) n-Nitrosodimethylamine	3.704	42	99801	41.689	ng	100
6) Aniline	7.298	93	272498	42.031	ng	99
8) 2-Chlorophenol	7.528	128	178456	42.673	ng	99
9) Benzaldehyde	7.110	77	123303m	39.799	ng	
10) Phenol	7.157	94	215915	41.368	ng	95
11) bis(2-Chloroethyl)ether	7.392	93	173238	42.068	ng	99
12) 1,3-Dichlorobenzene	7.851	146	198816	41.818	ng	98
13) 1,4-Dichlorobenzene	8.004	146	202488	42.151	ng	98
14) 1,2-Dichlorobenzene	8.322	146	198820	42.729	ng	97
15) Benzyl Alcohol	8.216	79	170869	43.754	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.498	45	179393	41.483	ng	99
17) 2-Methylphenol	8.410	107	156303	42.312	ng	97
18) Hexachloroethane	9.045	117	76613	42.899	ng	97
19) n-Nitroso-di-n-propyla...	8.781	70	155802	43.699	ng	99
20) 3+4-Methylphenols	8.745	107	211461	42.711	ng	99
22) Acetophenone	8.798	105	278547	40.901	ng	# 99
24) Nitrobenzene	9.181	77	236740	41.100	ng	99
25) Isophorone	9.704	82	393133	41.846	ng	99
26) 2-Nitrophenol	9.892	139	94468	41.317	ng	99
27) 2,4-Dimethylphenol	9.951	122	153940	40.455	ng	97
28) bis(2-Chloroethoxy)met...	10.186	93	230741	41.984	ng	99
29) 2,4-Dichlorophenol	10.428	162	189323	42.410	ng	97
30) 1,2,4-Trichlorobenzene	10.639	180	216606	42.428	ng	97
31) Naphthalene	10.828	128	548416	40.835	ng	100
32) Benzoic acid	10.092	122	110220	38.273	ng	96
33) 4-Chloroaniline	10.945	127	234857	41.897	ng	99
34) Hexachlorobutadiene	11.098	225	144593	43.058	ng	97
35) Caprolactam	11.745	113	46945	43.375	ng	96
36) 4-Chloro-3-methylphenol	12.063	107	177352	43.688	ng	96
37) 2-Methylnaphthalene	12.433	142	409148	42.391	ng	97
38) 1-Methylnaphthalene	12.651	142	386756	42.488	ng	99
40) 1,2,4,5-Tetrachloroben...	12.798	216	264373	41.642	ng	99
41) Hexachlorocyclopentadiene	12.769	237	149953	42.974	ng	99
43) 2,4,6-Trichlorophenol	13.039	196	172586	42.466	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012523\
 Data File : BM038526.D
 Acq On : 25 Jan 2023 10:59
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 01/26/2023
 Supervised By : Jagrut Upadhyay 01/26/2023


Quant Time: Jan 25 23:12:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 21 01:21:49 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.110	196	200450	42.111	ng	97
46) 1,1'-Biphenyl	13.439	154	554469	41.021	ng	99
47) 2-Chloronaphthalene	13.480	162	433237	41.059	ng	99
48) 2-Nitroaniline	13.692	65	131658	42.488	ng	98
49) Acenaphthylene	14.327	152	692947	41.976	ng	99
50) Dimethylphthalate	14.063	163	552286	42.510	ng	100
51) 2,6-Dinitrotoluene	14.186	165	116528	42.619	ng	96
52) Acenaphthene	14.663	154	421899	41.905	ng	98
53) 3-Nitroaniline	14.516	138	118726	43.697	ng	99
54) 2,4-Dinitrophenol	14.727	184	80879	40.337	ng	96
55) Dibenzofuran	14.998	168	722901	42.947	ng	99
56) 4-Nitrophenol	14.821	139	96341	43.057	ng	96
57) 2,4-Dinitrotoluene	14.974	165	163938	44.232	ng	97
58) Fluorene	15.645	166	610310	44.713	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.221	232	175706	45.159	ng	99
60) Diethylphthalate	15.415	149	563177	44.264	ng	100
61) 4-Chlorophenyl-phenyle...	15.639	204	314702	43.915	ng	99
62) 4-Nitroaniline	15.680	138	130238	42.004	ng	98
63) Azobenzene	15.927	77	589472	43.729	ng	100
65) 4,6-Dinitro-2-methylph...	15.733	198	116749	43.379	ng	97
66) n-Nitrosodiphenylamine	15.857	169	516266	41.476	ng	98
67) 4-Bromophenyl-phenylether	16.527	248	197104	42.856	ng	96
68) Hexachlorobenzene	16.645	284	210362	42.387	ng	100
69) Atrazine	16.804	200	180723	43.518	ng	97
70) Pentachlorophenol	16.992	266	157879	44.679	ng	95
71) Phenanthrene	17.386	178	976180	42.159	ng	100
72) Anthracene	17.474	178	999120	42.405	ng	100
73) Carbazole	17.751	167	889528	42.871	ng	100
74) Di-n-butylphthalate	18.298	149	1019469	44.722	ng	100
75) Fluoranthene	19.392	202	1233358	44.320	ng	99
77) Benzidine	19.580	184	469054	43.326	ng	99
78) Pyrene	19.756	202	1318495	39.857	ng	100
80) Butylbenzylphthalate	20.645	149	479552	41.922	ng	98
81) Benzo(a)anthracene	21.497	228	1373708	40.807	ng	99
82) 3,3'-Dichlorobenzidine	21.433	252	462821	42.801	ng	100
83) Chrysene	21.556	228	1355064	41.161	ng	99
84) Bis(2-ethylhexyl)phtha...	21.409	149	753477	45.450	ng	100
85) Di-n-octyl phthalate	22.339	149	1291798	43.822	ng	99
87) Indeno(1,2,3-cd)pyrene	26.444	276	1656301	41.812	ng	99
88) Benzo(b)fluoranthene	23.191	252	1385310	42.898	ng	99
89) Benzo(k)fluoranthene	23.238	252	1406130	42.080	ng	99
90) Benzo(a)pyrene	23.821	252	1358175	42.421	ng	99
91) Dibenzo(a,h)anthracene	26.456	278	1382235	41.920	ng	99
92) Benzo(g,h,i)perylene	27.221	276	1388702	41.116	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations APPROVED


Quant Time: Jan 25 23:12:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM012023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 21 01:21:49 2023
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.963 min Scan# 81
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:152 Resp: 67675
Ion Ratio Lower Upper
152 100
150 152.0 124.2 186.4
115 58.2 44.8 67.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#2
1,4-Dioxane
Concen: 40.328 ng
RT: 3.375 min Scan# 32
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion: 88 Resp: 77435
Ion Ratio Lower Upper
88 100
58 64.0 50.5 75.7
43 30.8 23.9 35.9





#3
Pyridine
Concen: 42.915 ng
RT: 3.787 min Scan# 103
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 21642
Ion Ratio Lower Upper
79 100
52 57.1 45.0 67.4
51 34.5 27.1 40.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#4
n-Nitrosodimethylamine
Concen: 41.689 ng
RT: 3.704 min Scan# 88
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion: 42 Resp: 99801
Ion Ratio Lower Upper
42 100
74 131.8 105.0 157.4
44 4.9 4.1 6.1









#5
2-Fluorophenol
Concen: 82.911 ng
RT: 5.516 min Scan# 396
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 112 Resp: 339417
Ion Ratio Lower Upper
112 100
64 57.9 47.3 70.9
63 36.4 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#6
Aniline
Concen: 42.031 ng
RT: 7.298 min Scan# 699
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion: 93 Resp: 272498
Ion Ratio Lower Upper
93 100
66 40.2 31.6 47.4
65 21.0 16.6 24.8









#7
Phenol-d6
Concen: 83.435 ng
RT: 7.128 min Scan# 671
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 435791
Ion Ratio Lower Upper
99 100
42 20.4 16.0 24.0
71 43.7 33.8 50.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#8
2-Chlorophenol
Concen: 42.673 ng
RT: 7.528 min Scan# 738
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59


Tgt Ion:128 Resp: 178456
Ion Ratio Lower Upper
128 100
130 32.0 13.2 53.2
64 47.9 27.7 67.7





#9
Benzaldehyde
Concen: 39.799 ng m
RT: 7.110 min Scan# 667
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 123303
Ion Ratio Lower Upper
77 100
105 91.7 72.9 112.9
106 83.3 63.8 103.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#10
Phenol
Concen: 41.368 ng
RT: 7.157 min Scan# 675
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion: 94 Resp: 215915
Ion Ratio Lower Upper
94 100
65 35.1 12.7 52.7
66 49.5 26.0 66.0





#11
bis(2-Chloroethyl)ether
Concen: 42.068 ng
RT: 7.392 min Scan# 716
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion: 93 Resp: 173238
Ion Ratio Lower Upper
93 100
63 73.9 54.4 94.4
95 31.6 11.4 51.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#12
1,3-Dichlorobenzene
Concen: 41.818 ng
RT: 7.851 min Scan# 793
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:146 Resp: 198816
Ion Ratio Lower Upper
146 100
148 66.8 52.1 78.1
75 37.6 28.7 43.1









#13
1,4-Dichlorobenzene
Concen: 42.151 ng
RT: 8.004 min Scan# 819
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:146 Resp: 202488
Ion Ratio Lower Upper
146 100
148 63.6 51.3 76.9
111 37.2 31.4 47.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#14
1,2-Dichlorobenzene
Concen: 42.729 ng
RT: 8.322 min Scan# 873
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59


Tgt Ion:146 Resp: 198820
Ion Ratio Lower Upper
146 100
148 63.7 52.7 79.1
111 39.4 33.2 49.8





#15
Benzyl Alcohol
Concen: 43.754 ng
RT: 8.216 min Scan# 855
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 79 Resp: 170869
Ion Ratio Lower Upper
79 100
108 66.6 55.0 82.6
77 69.1 55.0 82.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#16
2,2'-oxybis(1-Chloropropane)
Concen: 41.483 ng
RT: 8.498 min Scan# 903
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion: 45 Resp: 179393
Ion Ratio Lower Upper
45 100
77 20.1 0.0 39.2
79 15.0 0.0 34.9





#17
2-Methylphenol
Concen: 42.312 ng
RT: 8.410 min Scan# 889
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:	107	Resp:	156303
Ion Ratio	Lower	Upper	
107	100		
108	109.0	90.8	136.2
77	59.7	46.1	69.1
79	54.9	43.0	64.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#18
Hexachloroethane
Concen: 42.899 ng
RT: 9.045 min Scan# 996
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:	117	Resp:	76613
Ion Ratio	Lower	Upper	
117	100		
119	96.9	77.9	116.9
201	132.1	100.6	151.0






#19
n-Nitroso-di-n-propylamine
Concen: 43.699 ng
RT: 8.781 min Scan# 951
Delta R.T. -0.005 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :




SSTDCCC040




Tgt Ion: 70 Resp: 155802
Ion Ratio Lower Upper
70 100
42 52.6 41.4 62.2
101 8.4 6.5 9.7
130 17.2 14.3 21.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#20
3+4-Methylphenols
Concen: 42.711 ng
RT: 8.745 min Scan# 945
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:107 Resp: 211461
Ion Ratio Lower Upper
107 100
108 89.7 70.7 110.7
77 42.4 22.0 62.0
79 35.3 15.8 55.8





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.775 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:136 Resp: 247214
Ion Ratio Lower Upper
136 100
137 11.4 8.4 12.6
54 6.9 6.0 9.0
68 4.8 4.5 6.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#22
Acetophenone
Concen: 40.901 ng
RT: 8.798 min Scan# 954
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:105 Resp: 278547
Ion Ratio Lower Upper
105 100
71 2.0 0.9 1.3#
51 27.3 21.8 32.8
120 21.1 17.3 25.9





#23

Nitrobenzene-d5

Concen: 82.528 ng

RT: 9.139 min Scan# 1012

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 469820





Ion	Ratio	Lower	Upper
82	100		
128	37.5	29.3	43.9
54	46.3	36.9	55.3

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#24

Nitrobenzene

Concen: 41.100 ng

RT: 9.181 min Scan# 1019




Delta R.T. -0.006 min


Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Tgt Ion: 77 Resp: 236740


Ion	Ratio	Lower	Upper
77	100		
123	38.8	31.5	47.3
65	13.8	11.2	16.8





#25
Isophorone
Concen: 41.846 ng
RT: 9.704 min Scan# 1108
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 82 Resp: 393133
Ion Ratio Lower Upper
82 100
95 7.6 5.6 8.4
138 16.2 13.2 19.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#26
2-Nitrophenol
Concen: 41.317 ng
RT: 9.892 min Scan# 1140
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59



Tgt Ion:139 Resp: 94468
Ion Ratio Lower Upper
139 100
109 38.2 29.3 43.9
65 61.0 48.6 72.8





#27

2,4-Dimethylphenol

Concen: 40.455 ng

RT: 9.951 min Scan# 1151

Delta R.T. -0.006 min

Lab File: BM038526.D


Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040



Tgt Ion:122 Resp: 153940

Ion Ratio Lower Upper

122 100

107 127.2 98.2 147.2




121 58.9 47.1 70.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#28

bis(2-Chloroethoxy)methane

Concen: 41.984 ng

RT: 10.186 min Scan# 1190

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59




Tgt Ion: 93 Resp: 230741


Ion Ratio Lower Upper

93 100

95 31.7 25.3 37.9


123 10.4 8.8 13.2





#29
2,4-Dichlorophenol
Concen: 42.410 ng
RT: 10.428 min Scan# 1231
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 189323
Ion Ratio Lower Upper
162 100
164 66.4 43.7 83.7
98 31.9 12.9 52.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#30
1,2,4-Trichlorobenzene
Concen: 42.428 ng
RT: 10.639 min Scan# 1267
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:180 Resp: 216606
Ion Ratio Lower Upper
180 100
182 94.4 78.3 117.5
145 24.8 21.0 31.6





#31
Naphthalene
Concen: 40.835 ng
RT: 10.828 min Scan# 1299
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 548416
Ion Ratio Lower Upper
128 100
129 10.7 8.8 13.2
127 13.0 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#32
Benzoic acid
Concen: 38.273 ng
RT: 10.092 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59


Tgt Ion:122 Resp: 110220
Ion Ratio Lower Upper
122 100
105 134.2 107.9 147.9
77 112.1 88.6 128.6





#33
4-Chloroaniline
Concen: 41.897 ng
RT: 10.945 min Scan# 1319
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:127 Resp: 23485
Ion Ratio Lower Upper
127 100
129 32.7 25.8 38.6
65 34.9 28.2 42.2
92 20.9 16.6 24.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#34
Hexachlorobutadiene
Concen: 43.058 ng
RT: 11.098 min Scan# 1345
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:225 Resp: 144593
Ion Ratio Lower Upper
225 100
223 64.6 48.8 73.2
227 62.4 49.4 74.0





#35
 Caprolactam
 Concen: 43.375 ng
 RT: 11.745 min Scan# 1455
 Delta R.T. 0.000 min
 Lab File: BM038526.D
 Acq: 25 Jan 2023 10:59




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion:113 Resp: 46945
 Ion Ratio Lower Upper
 113 100
 55 171.3 145.0 185.0
 56 130.2 105.7 145.7

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 01/26/2023
 Supervised By :Jagrut Upadhyay 01/26/2023




#36
 4-Chloro-3-methylphenol
 Concen: 43.688 ng
 RT: 12.063 min Scan# 1509
 Delta R.T. -0.006 min
 Lab File: BM038526.D
 Acq: 25 Jan 2023 10:59



Tgt Ion:107 Resp: 177352
 Ion Ratio Lower Upper
 107 100
 144 25.6 22.0 33.0
 142 80.1 67.5 101.3









#37
2-Methylnaphthalene
Concen: 42.391 ng
RT: 12.433 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 409148
Ion Ratio Lower Upper
142 100
141 91.6 70.7 106.1
115 40.0 30.4 45.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#38
1-Methylnaphthalene
Concen: 42.488 ng
RT: 12.651 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59


Tgt Ion:142 Resp: 386756
Ion Ratio Lower Upper
142 100
141 93.1 73.4 110.0
116 4.1 3.3 4.9





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.598 min Scan# 1941
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:164 Resp: 168776
Ion Ratio Lower Upper
164 100
162 105.3 82.0 123.0
160 48.0 35.8 53.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.642 ng
RT: 12.798 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:216 Resp: 264373
Ion Ratio Lower Upper
216 100
214 78.0 61.4 92.0
179 19.0 15.7 23.5
108 17.4 14.0 21.0





#41
Hexachlorocyclopentadiene
Concen: 42.974 ng
RT: 12.769 min Scan# 1629
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:237 Resp: 149953
Ion Ratio Lower Upper
237 100
235 62.6 42.3 82.3
272 11.7 0.0 32.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#42
2,4,6-Tribromophenol
Concen: 94.698 ng
RT: 16.086 min Scan# 2193
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59



Tgt Ion:330 Resp: 230157
Ion Ratio Lower Upper
330 100
332 96.6 77.5 116.3
141 34.9 29.0 43.4





#43

2,4,6-Trichlorophenol

Concen: 42.466 ng

RT: 13.039 min Scan# 1675

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:196 Resp: 172586

Ion Ratio Lower Upper

196 100

198 96.4 77.9 116.9





200 31.9 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#44

2,4,5-Trichlorophenol

Concen: 42.111 ng

RT: 13.110 min Scan# 1687

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Tgt Ion:196 Resp: 200450




Ion Ratio Lower Upper


196 100

198 98.6 76.0 114.0

97 48.1 36.9 55.3

132 20.5 17.4 26.2






#45
2-Fluorobiphenyl
Concen: 83.405 ng
RT: 13.227 min Scan# 1707
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:172 Resp: 1142768
Ion Ratio Lower Upper
172 100
171 35.2 28.0 42.0
170 23.9 19.5 29.3


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#46
1,1'-Biphenyl
Concen: 41.021 ng
RT: 13.439 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:154 Resp: 554469
Ion Ratio Lower Upper
154 100
153 40.4 20.1 60.1
76 11.0 0.0 31.3





#47
2-Chloronaphthalene
Concen: 41.059 ng
RT: 13.480 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 433237
Ion Ratio Lower Upper
162 100
127 34.2 27.4 41.2
164 32.9 26.7 40.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#48
2-Nitroaniline
Concen: 42.488 ng
RT: 13.692 min Scan# 1786
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion: 65 Resp: 131658
Ion Ratio Lower Upper
65 100
92 63.8 50.0 75.0
138 95.7 78.2 117.2





#49
Acenaphthylene
Concen: 41.976 ng
RT: 14.327 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:152 Resp: 69294
Ion Ratio Lower Upper
152 100
151 19.7 16.0 24.0
153 12.6 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#50
Dimethylphthalate
Concen: 42.510 ng
RT: 14.063 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:163 Resp: 552286
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 9.9 8.0 12.0





#51
2,6-Dinitrotoluene
Concen: 42.619 ng
RT: 14.186 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:165 Resp: 116528
Ion Ratio Lower Upper
165 100
63 64.9 51.0 76.4
89 68.0 50.2 75.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#52
Acenaphthene
Concen: 41.905 ng
RT: 14.663 min Scan# 1951
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:154 Resp: 421899
Ion Ratio Lower Upper
154 100
153 115.6 90.7 136.1
152 54.1 43.8 65.8





#53
3-Nitroaniline
Concen: 43.697 ng
RT: 14.516 min Scan# 1926
Delta R.T. -0.005 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 118726
Ion Ratio Lower Upper
138 100
108 9.3 7.0 10.4
92 120.9 95.7 143.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#54
2,4-Dinitrophenol
Concen: 40.337 ng
RT: 14.727 min Scan# 1962
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59


Tgt Ion:184 Resp: 80879
Ion Ratio Lower Upper
184 100
63 67.4 52.2 78.4
154 66.0 48.9 73.3





#55
Dibenzenofuran
Concen: 42.947 ng
RT: 14.998 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 722901
Ion Ratio Lower Upper
168 100
139 35.7 27.9 41.9
169 13.4 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#56
4-Nitrophenol
Concen: 43.057 ng
RT: 14.821 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:139 Resp: 96341
Ion Ratio Lower Upper
139 100
109 101.7 76.8 116.8
65 109.0 91.7 131.7






#57
2,4-Dinitrotoluene
Concen: 44.232 ng
RT: 14.974 min Scan# 2004
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:165 Resp: 163938
Ion Ratio Lower Upper
165 100
63 48.6 37.8 56.8
89 80.0 60.9 91.3
182 2.5 1.7 2.5


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#58
Fluorene
Concen: 44.713 ng
RT: 15.645 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:166 Resp: 610310
Ion Ratio Lower Upper
166 100
165 97.8 79.6 119.4
167 13.4 10.5 15.7





#59
2,3,4,6-Tetrachlorophenol
Concen: 45.159 ng
RT: 15.221 min Scan# 2046
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:232 Resp: 175706
Ion Ratio Lower Upper
232 100
131 38.9 31.0 46.4
130 2.2 1.8 2.6
166 28.0 22.7 34.1

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#60
Diethylphthalate
Concen: 44.264 ng
RT: 15.415 min Scan# 2079
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59



Tgt Ion:149 Resp: 563177
Ion Ratio Lower Upper
149 100
177 23.3 18.6 27.8
150 12.1 9.5 14.3









#61
4-Chlorophenyl-phenylether
Concen: 43.915 ng
RT: 15.639 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 314702
Ion Ratio Lower Upper
204 100
206 33.3 26.2 39.4
141 50.8 41.5 62.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#62
4-Nitroaniline
Concen: 42.004 ng
RT: 15.680 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59


Tgt Ion:138 Resp: 130238
Ion Ratio Lower Upper
138 100
92 53.3 35.2 75.2
108 105.2 87.3 127.3





#63
Azobenzene
Concen: 43.729 ng
RT: 15.927 min Scan# 2166
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 589472
Ion Ratio Lower Upper
77 100
182 24.7 4.7 44.7
105 16.5 0.0 36.7
51 29.6 9.4 49.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.339 min Scan# 2406
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:188 Resp: 414630
Ion Ratio Lower Upper
188 100
94 7.4 6.0 9.0
80 7.9 6.9 10.3





#65
4,6-Dinitro-2-methylphenol
Concen: 43.379 ng
RT: 15.733 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:198 Resp: 116749
Ion Ratio Lower Upper
198 100
51 37.7 18.8 58.8
105 37.7 19.7 59.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#66
n-Nitrosodiphenylamine
Concen: 41.476 ng
RT: 15.857 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:169 Resp: 516266
Ion Ratio Lower Upper
169 100
168 67.8 55.8 83.8
167 37.7 30.3 45.5





#67
4-Bromophenyl-phenylether
Concen: 42.856 ng
RT: 16.527 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:248 Resp: 197104
Ion Ratio Lower Upper
248 100
250 93.3 77.6 116.4
141 63.1 52.5 78.7

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#68
Hexachlorobenzene
Concen: 42.387 ng
RT: 16.645 min Scan# 2288
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:284 Resp: 210362
Ion Ratio Lower Upper
284 100
142 31.2 25.2 37.8
249 30.5 24.4 36.6





#69
 Atrazine
 Concen: 43.518 ng
 RT: 16.804 min Scan# 2316
 Delta R.T. -0.006 min
 Lab File: BM038526.D
 Acq: 25 Jan 2023 10:59




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040




Tgt Ion:200 Resp: 180723
 Ion Ratio Lower Upper
 200 100
 173 24.3 3.5 43.5
 215 47.2 30.0 70.0

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 01/26/2023
 Supervised By :Jagrut Upadhyay 01/26/2023




#70
 Pentachlorophenol
 Concen: 44.679 ng
 RT: 16.992 min Scan# 2347
 Delta R.T. -0.006 min
 Lab File: BM038526.D
 Acq: 25 Jan 2023 10:59



Tgt Ion:266 Resp: 157879
 Ion Ratio Lower Upper
 266 100
 268 68.9 51.7 77.5
 264 65.5 49.4 74.0









#71
Phenanthrene
Concen: 42.159 ng
RT: 17.386 min Scan# 2415
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 976180
Ion Ratio Lower Upper
178 100
176 19.6 15.4 23.2
179 15.2 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#72
Anthracene
Concen: 42.405 ng
RT: 17.474 min Scan# 2429
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:178 Resp: 999120
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.6
179 15.4 12.2 18.4





#73

Carbazole

Concen: 42.871 ng

RT: 17.751 min Scan# 24

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:167 Resp: 889528





Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.7	26.5
139	11.7	9.4	14.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#74

Di-n-butylphthalate

Concen: 44.722 ng

RT: 18.298 min Scan# 2569




Delta R.T. -0.006 min


Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Tgt Ion:149 Resp: 1019469


Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.0
104	5.9	4.6	6.8





#75
Fluoranthene
Concen: 44.320 ng
RT: 19.392 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:202 Resp: 1233358
Ion Ratio Lower Upper
202 100
101 8.3 0.0 28.7
203 17.7 0.0 37.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.515 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:240 Resp: 475086
Ion Ratio Lower Upper
240 100
120 8.3 6.2 9.2
236 25.8 20.2 30.2





#77
Benzidine
Concen: 43.326 ng
RT: 19.580 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 469054
Ion Ratio Lower Upper
184 100
185 13.9 10.9 16.3
183 11.0 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#78
Pyrene
Concen: 39.857 ng
RT: 19.756 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:202 Resp: 1318495
Ion Ratio Lower Upper
202 100
200 21.0 17.0 25.4
203 17.7 14.2 21.2






#79
 Terphenyl-d14
 Concen: 79.654 ng
 RT: 19.956 min Scan# 2851
 Delta R.T. -0.006 min
 Lab File: BM038526.D
 Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040






Tgt Ion:244 Resp: 194112
 Ion Ratio Lower Upper
 244 100
 212 7.6 6.0 9.0
 122 8.4 6.7 10.1


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 01/26/2023
 Supervised By :Jagrut Upadhyay 01/26/2023




#80
 Butylbenzylphthalate
 Concen: 41.922 ng
 RT: 20.645 min Scan# 2968
 Delta R.T. -0.006 min
 Lab File: BM038526.D
 Acq: 25 Jan 2023 10:59



Tgt Ion:149 Resp: 479552
 Ion Ratio Lower Upper
 149 100
 91 73.6 57.5 86.3
 206 22.0 18.2 27.2





#81

Benzo(a)anthracene

Concen: 40.807 ng

RT: 21.497 min Scan# 3114

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 1373708

Ion Ratio Lower Upper

228 100

226 27.1 22.2 33.2





229 19.8 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#82

3,3'-Dichlorobenzidine

Concen: 42.801 ng

RT: 21.433 min Scan# 3102

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59




Tgt Ion:252 Resp: 462821


Ion Ratio Lower Upper

252 100

254 64.9 51.9 77.9


126 8.9 7.7 11.5





#83
Chrysene
Concen: 41.161 ng
RT: 21.556 min Scan# 3123
Delta R.T. 0.000 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040




Tgt Ion:228 Resp: 1355064
Ion Ratio Lower Upper
228 100
226 30.4 24.4 36.6
229 19.1 15.7 23.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023




#84
Bis(2-ethylhexyl)phthalate
Concen: 45.450 ng
RT: 21.409 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Tgt Ion:149 Resp: 753477
Ion Ratio Lower Upper
149 100
167 26.8 21.3 31.9
279 3.6 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 43.822 ng
RT: 22.339 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 1291798
Ion Ratio Lower Upper
149 100
167 1.6 1.3 1.9
43 8.1 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.927 min Scan# 3526
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:264 Resp: 535335
Ion Ratio Lower Upper
264 100
260 23.7 18.5 27.7
265 22.2 17.8 26.6





#87

Indeno(1,2,3-cd)pyrene

Concen: 41.812 ng

RT: 26.444 min Scan# 3956

Delta R.T. -0.012 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 165630

Ion Ratio Lower Upper

276 100

138 19.6 15.8 23.8





277 24.8 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#88

Benzo(b)fluoranthene

Concen: 42.898 ng

RT: 23.191 min Scan# 3401

Delta R.T. -0.006 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59




Tgt Ion:252 Resp: 1385310


Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0


125 7.7 6.4 9.6





#89
Benzo(k)fluoranthene
Concen: 42.080 ng
RT: 23.238 min Scan# 3410
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040



Tgt Ion:252 Resp: 1406130
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 8.0 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 01/26/2023
Supervised By :Jagrut Upadhyay 01/26/2023



#90
Benzo(a)pyrene
Concen: 42.421 ng
RT: 23.821 min Scan# 3508
Delta R.T. -0.006 min
Lab File: BM038526.D
Acq: 25 Jan 2023 10:59

Tgt Ion:252 Resp: 1358175
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.0
125 8.6 7.3 10.9





#91

Dibenzo(a,h)anthracene

Concen: 41.920 ng

RT: 26.456 min Scan# 3958

Delta R.T. -0.012 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Instrument :

BNA_M

ClientSampleId :

SSTDCCC040

Tgt Ion:278 Resp: 1382235

Ion Ratio Lower Upper

278 100

139 12.4 10.3 15.5





279 24.0 19.3 28.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 01/26/2023

Supervised By :Jagrut Upadhyay 01/26/2023



#92

Benzo(g,h,i)perylene

Concen: 41.116 ng

RT: 27.221 min Scan# 4086

Delta R.T. -0.012 min

Lab File: BM038526.D

Acq: 25 Jan 2023 10:59

Tgt Ion:276 Resp: 1388702

Ion Ratio Lower Upper

276 100

277 24.0 18.8 28.2

138 16.4 13.8 20.8

