

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012622\
 Data File : BM033899.D
 Acq On : 27 Jan 2022 13:09
 Operator : CG/JU
 Sample : N1279-06
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0VD6

Quant Time: Jan 28 01:50:25 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 24 15:20:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.939	152	1634	0.400	ng/u1	0.00
4) Naphthalene-d8	10.747	136	5642	0.400	ng/u1 #	0.00
9) Acenaphthene-d10	14.572	164	3205	0.400	ng/u1	-0.01
13) Phenanthrene-d10	17.309	188	6107	0.400	ng/u1	-0.01
17) Chrysene-d12	21.475	240	4664	0.400	ng/u1 #	0.00
23) Perylene-d12	23.863	264	5459	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.292	96	7428	4.348	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.339	152	2728	0.351	ng/u1	0.00
18) Fluoranthene-d10	19.326	212	6163	0.429	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM012622\
Data File : BM033899.D
Acq On : 27 Jan 2022 13:09
Operator : CG/JU
Sample : N1279-06
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0VD6

Quant Time: Jan 28 01:50:25 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM012422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 24 15:20:03 2022
Response via : Initial Calibration

