

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020123\
 Data File : BM038643.D
 Acq On : 01 Feb 2023 15:20
 Operator : CG/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC2.5

Quant Time: Feb 02 02:37:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 02 02:35:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	50036	20.000	ng	0.00
21) Naphthalene-d8	10.746	136	154942	20.000	ng	0.00
39) Acenaphthene-d10	14.575	164	102421	20.000	ng	0.00
64) Phenanthrene-d10	17.316	188	208728	20.000	ng	0.00
76) Chrysene-d12	21.492	240	198991	20.000	ng	0.00
86) Perylene-d12	23.892	264	305031	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						Qvalue
19) n-Nitroso-di-n-propyla...	8.746	70	3485	1.405	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020123\
Data File : BM038643.D
Acq On : 01 Feb 2023 15:20
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Feb 02 02:37:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 02 02:35:13 2023
Response via : Initial Calibration

