

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020123\
 Data File : BM038652.D
 Acq On : 01 Feb 2023 21:31
 Operator : CG/JU
 Sample : PB150609TB
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB150609TB

Quant Time: Feb 02 04:02:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 02 03:31:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.939	152	64579	20.000	ng	0.00
21) Naphthalene-d8	10.745	136	198639	20.000	ng	0.00
39) Acenaphthene-d10	14.574	164	125376	20.000	ng	0.00
64) Phenanthrene-d10	17.315	188	227953	20.000	ng	0.00
76) Chrysene-d12	21.491	240	156002	20.000	ng	0.00
86) Perylene-d12	23.891	264	197909	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	481688	146.113	ng	0.00
7) Phenol-d6	7.110	99	523742	139.295	ng	0.00
23) Nitrobenzene-d5	9.116	82	323371	92.467	ng	0.00
42) 2,4,6-Tribromophenol	16.062	330	224217	131.572	ng	0.00
45) 2-Fluorobiphenyl	13.198	172	919793	93.398	ng	0.00
79) Terphenyl-d14	19.933	244	901952	102.027	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM020123\
Data File : BM038652.D
Acq On : 01 Feb 2023 21:31
Operator : CG/JU
Sample : PB150609TB
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB150609TB

Quant Time: Feb 02 04:02:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 02 03:31:45 2023
Response via : Initial Calibration

