

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020921\
 Data File : BM028680.D
 Acq On : 09 Feb 2021 10:33
 Operator : CG/JU
 Sample : PB134489BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK89

Quant Time: Feb 09 11:05:37 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Feb 09 10:35:42 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	20971	20.00	ng/ul	0.00
18) Naphthalene-d8	10.70	136	84100	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.54	164	47622	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.28	188	96169	20.00	ng/ul	0.00
78) Chrysene-d12	21.43	240	96211	20.00	ng/ul	0.00
86) Perylene-d12	23.67	264	96413	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.20	96	3686	7.40	ng/uL	0.00
5) Phenol-d5	7.06	99	64791	36.19	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.23	67	38273	34.16	ng/ul	0.00
9) 2-Chlorophenol-d4	7.42	132	56755	36.56	ng/ul	0.00
13) 4-Methylphenol-d8	8.62	113	51949	35.93	ng/ul	0.00
19) Nitrobenzene-d5	9.07	128	25928	36.03	ng/ul	0.00
22) 2-Nitrophenol-d4	9.79	143	28943	36.83	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.33	165	50574	35.31	ng/ul	0.00
29) 4-Chloroaniline-d4	10.85	131	71868	43.11	ng/ul	0.00
44) Dimethylphthalate-d6	13.96	166	147223	38.93	ng/ul	0.00
47) Acenaphthylene-d8	14.24	160	187051	38.72	ng/ul	0.00
52) 4-Nitrophenol-d4	14.76	143	24511	34.60	ng/ul	0.00
58) Fluorene-d10	15.54	176	122098	37.36	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.67	200	21110	31.90	ng/ul	0.00
71) Anthracene-d10	17.38	188	188759	38.12	ng/ul	0.00
79) Pyrene-d10	19.66	212	203539	36.22	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.53	264	217365	39.73	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM020921\
Data File : BM028680.D
Acq On : 09 Feb 2021 10:33
Operator : CG/JU
Sample : PB134489BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK89

Quant Time: Feb 09 11:05:37 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011621MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Feb 09 10:35:42 2021
Response via : Initial Calibration

