

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021121\
 Data File : BM028717.D
 Acq On : 11 Feb 2021 18:37
 Operator : CG/JU
 Sample : PB134593BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Feb 11 23:47:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011621MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 11 16:46:31 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.89	152	29506	20.00	ng/ul	0.00
18) Naphthalene-d8	10.70	136	116764	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.54	164	63631	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.28	188	120653	20.00	ng/ul	0.00
78) Chrysene-d12	21.43	240	111367	20.00	ng/ul	0.00
86) Perylene-d12	23.66	264	114161	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.20	96	3689	5.27	ng/uL	0.00
5) Phenol-d5	7.06	99	67945	26.97	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.22	67	40221	25.51	ng/ul	0.00
9) 2-Chlorophenol-d4	7.42	132	58897	26.97	ng/ul	0.00
13) 4-Methylphenol-d8	8.61	113	53716	26.40	ng/ul	0.00
19) Nitrobenzene-d5	9.06	128	26693	26.72	ng/ul	0.00
22) 2-Nitrophenol-d4	9.79	143	29884	27.39	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.32	165	51716	26.01	ng/ul	0.00
29) 4-Chloroaniline-d4	10.84	131	74082	32.01	ng/ul	0.00
44) Dimethylphthalate-d6	13.96	166	142173	28.14	ng/ul	0.00
47) Acenaphthylene-d8	14.24	160	185678	28.77	ng/ul	0.00
52) 4-Nitrophenol-d4	14.76	143	22243	23.50	ng/ul	0.00
58) Fluorene-d10	15.53	176	119774	27.43	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.67	200	17849	21.50	ng/ul	0.00
71) Anthracene-d10	17.37	188	178538	28.74	ng/ul	0.00
79) Pyrene-d10	19.66	212	182201	28.01	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.52	264	193225	29.82	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM021121\
Data File : BM028717.D
Acq On : 11 Feb 2021 18:37
Operator : CG/JU
Sample : PB134593BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Feb 11 23:47:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM011621MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 11 16:46:31 2021
Response via : Initial Calibration

