

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021523\
 Data File : BM038752.D
 Acq On : 15 Feb 2023 17:37
 Operator : CG/JU
 Sample : 01539-09
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COMP-8

Quant Time: Feb 16 02:17:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 15 02:07:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.892	152	72630	20.000	ng	0.00
21) Naphthalene-d8	10.704	136	265327	20.000	ng	0.00
39) Acenaphthene-d10	14.533	164	178704	20.000	ng	-0.01
64) Phenanthrene-d10	17.280	188	354991	20.000	ng	# 0.00
76) Chrysene-d12	21.456	240	274471	20.000	ng	0.00
86) Perylene-d12	23.838	264	263257	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	321861	86.809	ng	0.00
7) Phenol-d6	7.069	99	394494	93.290	ng	-0.01
23) Nitrobenzene-d5	9.069	82	224165	47.989	ng	-0.01
42) 2,4,6-Tribromophenol	16.027	330	191866	78.990	ng	0.00
45) 2-Fluorobiphenyl	13.163	172	647455	46.125	ng	0.00
79) Terphenyl-d14	19.898	244	888508	57.125	ng	0.00
Target Compounds						
						Qvalue
75) Fluoranthene	19.333	202	56372	2.547	ng	98
78) Pyrene	19.698	202	50533	2.391	ng	99
84) Bis(2-ethylhexyl)phtha...	21.356	149	8329	2.023	ng	# 96
88) Benzo(b)fluoranthene	23.109	252	37143	2.356	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM021523\
Data File : BM038752.D
Acq On : 15 Feb 2023 17:37
Operator : CG/JU
Sample : 01539-09
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COMP-8

Quant Time: Feb 16 02:17:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM020123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 15 02:07:17 2023
Response via : Initial Calibration

