

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022621\
 Data File : BM028898.D
 Acq On : 25 Feb 2021 15:57
 Operator : CG/JU
 Sample : PB134805BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK05

Quant Time: Feb 25 16:39:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM022121MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 25 15:24:23 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	20732	20.00	ng/ul	0.00
18) Naphthalene-d8	10.60	136	85932	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.46	164	54284	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.20	188	110133	20.00	ng/ul	0.00
78) Chrysene-d12	21.36	240	94334	20.00	ng/ul	0.00
86) Perylene-d12	23.58	264	84392	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	2374	5.45	ng/uL	0.00
5) Phenol-d5	6.99	99	43267	27.40	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.14	67	25774	28.97	ng/ul	0.00
9) 2-Chlorophenol-d4	7.33	132	38543	28.42	ng/ul	0.00
13) 4-Methylphenol-d8	8.53	113	36662	27.67	ng/ul	0.00
19) Nitrobenzene-d5	8.98	128	18365	30.20	ng/ul	0.00
22) 2-Nitrophenol-d4	9.70	143	21313	30.88	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	36206	27.58	ng/ul	0.00
29) 4-Chloroaniline-d4	10.76	131	53793	34.58	ng/ul	0.00
44) Dimethylphthalate-d6	13.88	166	122039	31.18	ng/ul	0.00
47) Acenaphthylene-d8	14.16	160	150747	30.25	ng/ul	0.00
52) 4-Nitrophenol-d4	14.70	143	19005	28.05	ng/ul	0.00
58) Fluorene-d10	15.46	176	101290	30.79	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.60	200	16772	24.90	ng/ul	0.00
71) Anthracene-d10	17.30	188	157387	31.38	ng/ul	0.00
79) Pyrene-d10	19.59	212	158498	33.71	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.44	264	137115	31.52	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM022621\
Data File : BM028898.D
Acq On : 25 Feb 2021 15:57
Operator : CG/JU
Sample : PB134805BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK05

Quant Time: Feb 25 16:39:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM022121MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 25 15:24:23 2021
Response via : Initial Calibration

