

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031622\
 Data File : BM034264.D
 Acq On : 16 Mar 2022 19:13
 Operator : CG/JU
 Sample : PB143171BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB143171BL

Quant Time: Mar 17 04:55:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 16 02:38:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.948	152	134516	20.000	ng	0.00
21) Naphthalene-d8	10.760	136	542975	20.000	ng	0.00
39) Acenaphthene-d10	14.589	164	362491	20.000	ng	0.00
64) Phenanthrene-d10	17.324	188	779960	20.000	ng	0.00
76) Chrysene-d12	21.494	240	830136	20.000	ng	0.00
86) Perylene-d12	23.912	264	858776	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.501	112	1086403	144.887	ng	0.00
7) Phenol-d6	7.107	99	1418183	132.099	ng	0.00
23) Nitrobenzene-d5	9.107	82	922904	82.754	ng	0.00
42) 2,4,6-Tribromophenol	16.071	330	629757	128.892	ng	0.00
45) 2-Fluorobiphenyl	13.219	172	2142412	80.763	ng	0.00
79) Terphenyl-d14	19.942	244	3981022	83.329	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM031622\
Data File : BM034264.D
Acq On : 16 Mar 2022 19:13
Operator : CG/JU
Sample : PB143171BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB143171BL

Quant Time: Mar 17 04:55:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 16 02:38:26 2022
Response via : Initial Calibration

