

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041423\
 Data File : BM039430.D
 Acq On : 14 Apr 2023 12:26
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 14 14:49:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 14 14:49:28 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	84	0.00
2	1,4-Dioxane	40.000	39.694	0.8	85	0.00
3	Pyridine	40.000	38.989	2.5	81	0.00
4	n-Nitrosodimethylamine	40.000	38.892	2.8	83	0.00
5 S	2-Fluorophenol	80.000	78.035	2.5	82	0.00
6	Aniline	40.000	40.109	-0.3	84	0.00
7 S	Phenol-d6	80.000	80.935	-1.2	85	0.00
8	2-Chlorophenol	40.000	40.414	-1.0	85	0.00
9	Benzaldehyde	40.000	34.734	13.2	73	0.00
10 C	Phenol	40.000	39.427	1.4	83	0.00
11	bis(2-Chloroethyl)ether	40.000	41.504	-3.8	88	0.00
12	1,3-Dichlorobenzene	40.000	41.201	-3.0	87	0.00
13 C	1,4-Dichlorobenzene	40.000	40.995	-2.5	87	0.00
14	1,2-Dichlorobenzene	40.000	41.513	-3.8	88	0.00
15	Benzyl Alcohol	40.000	40.004	-0.0	83	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	45.079	-12.7	96	0.00
17	2-Methylphenol	40.000	40.928	-2.3	87	0.00
18	Hexachloroethane	40.000	42.237	-5.6	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.207	-8.0	92	0.00
20	3+4-Methylphenols	40.000	41.336	-3.3	86	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	40.705	-1.8	88	0.00
23 S	Nitrobenzene-d5	80.000	82.284	-2.9	89	0.00
24	Nitrobenzene	40.000	40.768	-1.9	88	0.00
25	Isophorone	40.000	43.730	-9.3	94	0.00
26 C	2-Nitrophenol	40.000	41.447	-3.6	88	0.00
27	2,4-Dimethylphenol	40.000	41.530	-3.8	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.204	-5.5	92	0.00
29 C	2,4-Dichlorophenol	40.000	42.340	-5.9	90	0.00
30	1,2,4-Trichlorobenzene	40.000	42.478	-6.2	93	0.00
31	Naphthalene	40.000	41.437	-3.6	90	0.00
32	Benzoic acid	40.000	38.206	4.5	79	0.00
33	4-Chloroaniline	40.000	39.840	0.4	86	0.00
34 C	Hexachlorobutadiene	40.000	44.199	-10.5	96	0.00
35	Caprolactam	40.000	42.235	-5.6	89	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.975	-7.4	92	0.00
37	2-Methylnaphthalene	40.000	42.423	-6.1	92	0.00
38	1-Methylnaphthalene	40.000	42.524	-6.3	92	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.706	-4.3	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.266	6.8	82	0.00
42 S	2,4,6-Tribromophenol	80.000	85.436	-6.8	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.718	-4.3	93	0.00
44	2,4,5-Trichlorophenol	40.000	40.935	-2.3	92	0.00
45 S	2-Fluorobiphenyl	80.000	83.424	-4.3	96	0.00
46	1,1'-Biphenyl	40.000	41.150	-2.9	95	0.00
47	2-Chloronaphthalene	40.000	40.574	-1.4	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041423\
 Data File : BM039430.D
 Acq On : 14 Apr 2023 12:26
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 14 14:49:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 14 14:49:28 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.359	1.6	89	0.00
49	Acenaphthylene	40.000	41.357	-3.4	95	0.00
50	Dimethylphthalate	40.000	42.982	-7.5	99	0.00
51	2,6-Dinitrotoluene	40.000	42.989	-7.5	97	0.00
52 C	Acenaphthene	40.000	41.543	-3.9	96	0.00
53	3-Nitroaniline	40.000	37.763	5.6	84	0.00
54 P	2,4-Dinitrophenol	40.000	37.668	5.8	84	0.00
55	Dibenzofuran	40.000	41.401	-3.5	96	0.00
56 P	4-Nitrophenol	40.000	32.061	19.8	68	0.00
57	2,4-Dinitrotoluene	40.000	42.956	-7.4	95	0.00
58	Fluorene	40.000	41.781	-4.5	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.662	-4.2	95	0.00
60	Diethylphthalate	40.000	44.515	-11.3	103	0.00
61	4-Chlorophenyl-phenylether	40.000	42.931	-7.3	100	0.00
62	4-Nitroaniline	40.000	36.147	9.6	79	0.00
63	Azobenzene	40.000	43.406	-8.5	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.868	-4.7	94	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.471	-3.7	96	0.00
67	4-Bromophenyl-phenylether	40.000	43.562	-8.9	101	0.00
68	Hexachlorobenzene	40.000	43.247	-8.1	101	0.00
69	Atrazine	40.000	44.375	-10.9	103	0.00
70 C	Pentachlorophenol	40.000	40.685	-1.7	90	0.00
71	Phenanthrene	40.000	41.394	-3.5	97	0.00
72	Anthracene	40.000	41.898	-4.7	96	0.00
73	Carbazole	40.000	40.447	-1.1	93	0.00
74	Di-n-butylphthalate	40.000	47.946	-19.9	109	0.00
75 C	Fluoranthene	40.000	42.501	-6.3	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	84	0.00
77	Benzydine	40.000	34.910	12.7	69	0.00
78	Pyrene	40.000	46.297	-15.7	97	0.00
79 S	Terphenyl-d14	80.000	93.682	-17.1	98	0.00
80	Butylbenzylphthalate	40.000	50.973	-27.4#	105	0.00
81	Benzo(a)anthracene	40.000	41.917	-4.8	88	0.00
82	3,3'-Dichlorobenzidine	40.000	42.119	-5.3	87	0.00
83	Chrysene	40.000	41.797	-4.5	87	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	51.890	-29.7#	108	0.00
85 c	Di-n-octyl phthalate	40.000	47.334	-18.3	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	77	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.678	0.8	74	0.00
88	Benzo(b)fluoranthene	40.000	41.720	-4.3	80	0.00
89	Benzo(k)fluoranthene	40.000	42.684	-6.7	81	0.00
90 C	Benzo(a)pyrene	40.000	41.648	-4.1	79	0.00
91	Dibenzo(a,h)anthracene	40.000	39.830	0.4	75	0.00
92	Benzo(g,h,i)perylene	40.000	39.472	1.3	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041423\
Data File : BM039430.D
Acq On : 14 Apr 2023 12:26
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Apr 14 14:49:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM032923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 14 14:49:28 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0