

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042423\
 Data File : BM039628.D
 Acq On : 24 Apr 2023 22:56
 Operator : CG/JU
 Sample : PB152350BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS350

Quant Time: Apr 25 00:45:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.807	152	212125	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.613	136	964438	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.465	164	599964	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.224	188	1247816	20.000	ng/u1	0.00
79) Chrysene-d12	21.418	240	1166685	20.000	ng/u1	0.00
88) Perylene-d12	23.777	264	1164496	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.207	96	32202	5.827	ng/uL	0.00
4) Pyridine-d5	3.631	84	408079	26.280	ng/u1	0.00
7) Phenol-d5	6.995	99	625716	32.229	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.148	67	417933	32.568	ng/u1	0.00
11) 2-Chlorophenol-d4	7.342	132	490515	32.208	ng/u1	0.00
15) 4-Methylphenol-d8	8.542	113	497500	31.811	ng/u1	0.00
21) Nitrobenzene-d5	8.983	128	246075	31.714	ng/u1	0.00
24) 2-Nitrophenol-d4	9.707	143	277030	31.796	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.248	165	480588	32.057	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.766	131	603639	26.960	ng/u1	-0.01
46) Dimethylphthalate-d6	13.883	166	1492333	30.842	ng/u1	0.00
49) Acenaphthylene-d8	14.160	160	1735031	31.918	ng/u1	0.00
54) 4-Nitrophenol-d4	14.712	143	200930	27.438	ng/u1	-0.02
60) Fluorene-d10	15.465	176	1236327	31.187	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.607	200	250986	30.268	ng/u1	0.00
73) Anthracene-d10	17.324	188	1919334	31.984	ng/u1	0.00
81) Pyrene-d10	19.624	212	2233635	28.812	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.630	264	2023443	32.185	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.237	88	71198	11.773	ng/uL	94
5) Pyridine	3.648	79	444106	27.124	ng/u1	99
6) Benzaldehyde	6.954	77	333130	40.162	ng/u1	94
8) Phenol	7.019	94	669145	33.103	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.237	93	540815	32.090	ng/u1	97
12) 2-Chlorophenol	7.378	128	494366	32.259	ng/u1	97
13) 2-Methylphenol	8.272	108	506278	32.246	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.342	45	786385	33.595	ng/u1	98
16) Acetophenone	8.642	105	845392	32.910	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.631	70	464208	32.650	ng/u1	95
18) 4-Methylphenol	8.607	108	549515	32.418	ng/u1	99
19) Hexachloroethane	8.883	117	217591	31.262	ng/u1	97
22) Nitrobenzene	9.025	77	643438	33.586	ng/u1	95
23) Isophorone	9.548	82	1309849	32.062	ng/u1	98
25) 2-Nitrophenol	9.736	139	286661	31.533	ng/u1	92
26) 2,4-Dimethylphenol	9.801	107	591674	31.850	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.036	93	758124	31.559	ng/u1	100
29) 2,4-Dichlorophenol	10.277	162	473512	32.065	ng/u1	100
30) Naphthalene	10.666	128	1687989	31.633	ng/u1	100
32) 4-Chloroaniline	10.789	127	566556	25.680	ng/u1	99
33) Hexachlorobutadiene	10.942	225	292664	30.076	ng/u1	99
34) Caprolactam	11.583	113	170069	30.925	ng/u1	97
35) 4-Chloro-3-methylphenol	11.930	107	555439	32.214	ng/u1	98
36) 2-Methylnaphthalene	12.283	142	1105659	30.908	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042423\
 Data File : BM039628.D
 Acq On : 24 Apr 2023 22:56
 Operator : CG/JU
 Sample : PB152350BS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS350

Quant Time: Apr 25 00:45:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.501	142	1138234	31.663	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.654	216	563705	31.302	ng/ul	98
40) Hexachlorocyclopentadiene	12.624	237	219288	20.438	ng/ul	99
41) 2,4,6-Trichlorophenol	12.907	196	375774	31.234	ng/ul	99
42) 2,4,5-Trichlorophenol	12.989	196	395970	31.667	ng/ul	98
43) 1,1'-Biphenyl	13.301	154	1486934	32.069	ng/ul	100
44) 2-Chloronaphthalene	13.342	162	1148811	31.585	ng/ul	99
45) 2-Nitroaniline	13.566	65	374671	35.696	ng/ul	95
47) Dimethylphthalate	13.930	163	1498557	31.281	ng/ul	100
48) 2,6-Dinitrotoluene	14.060	165	304521	31.430	ng/ul	94
50) Acenaphthylene	14.189	152	1865504	32.083	ng/ul	100
51) 3-Nitroaniline	14.395	138	282742	31.134	ng/ul	97
52) Acenaphthene	14.530	153	1275771	31.868	ng/ul	100
53) 2,4-Dinitrophenol	14.613	184	132859	25.760	ng/ul	96
55) 4-Nitrophenol	14.724	109	161772	29.084	ng/ul	91
56) Dibenzofuran	14.871	168	1740571	31.988	ng/ul	100
57) 2,4-Dinitrotoluene	14.854	165	439151	32.620	ng/ul	87
58) 2,3,4,6-Tetrachlorophenol	15.101	232	348261	31.116	ng/ul	96
59) Diethylphthalate	15.295	149	1542984	31.070	ng/ul	100
61) Fluorene	15.518	166	1476990	32.813	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.512	204	707459	31.710	ng/ul	99
63) 4-Nitroaniline	15.565	138	260647	37.427	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.618	198	246351	30.901	ng/ul	94
67) N-Nitrosodiphenylamine	15.736	169	1194115	32.119	ng/ul	99
68) 4-Bromophenyl-phenylether	16.412	248	417411	30.972	ng/ul	98
69) Hexachlorobenzene	16.524	284	483687	31.240	ng/ul	99
70) Atrazine	16.689	200	350528	25.409	ng/ul	98
71) Pentachlorophenol	16.877	266	259828	28.707	ng/ul	97
72) Phenanthrene	17.265	178	2229951	32.190	ng/ul	99
74) Anthracene	17.359	178	2279139	32.795	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.266	216	574505	30.914	ng/ul	99
76) Pentachlorobenzene	14.789	250	576535	31.519	ng/ul	98
77) Carbazole	17.636	167	2044335	33.103	ng/ul	100
78) Di-n-butylphthalate	18.189	149	2706120	30.981	ng/ul	99
80) Fluoranthene	19.289	202	2644891	29.107	ng/ul	98
82) Pyrene	19.653	202	2765708	29.401	ng/ul	99
83) Butylbenzylphthalate	20.547	149	1235223	27.818	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.341	252	854552	33.030	ng/ul	98
85) Benzo(a)anthracene	21.400	228	2650463	31.409	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.324	149	1836948	27.818	ng/ul	100
87) Chrysene	21.453	228	2523203	31.347	ng/ul	100
89) Di-n-octyl phthalate	22.236	149	3113165	26.608	ng/ul	100
90) Benzo(b)fluoranthene	23.059	252	2532988	30.914	ng/ul	99
91) Benzo(k)fluoranthene	23.106	252	2600725	32.320	ng/ul	99
93) Benzo(a)pyrene	23.677	252	2225941	32.230	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.224	276	2572496	33.863	ng/ul	100
95) Dibenzo(a,h)anthracene	26.241	278	2163500	34.349	ng/ul	99
96) Benzo(g,h,i)perylene	26.976	276	2008933	33.196	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.773 ng/uL
RT: 3.237 min Scan# 34
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion: 88 Resp: 71198
Ion Ratio Lower Upper
88 100
43 33.8 24.6 36.8
58 76.5 56.7 85.1



#5
Pyridine
Concen: 27.124 ng/uL
RT: 3.648 min Scan# 104
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion: 79 Resp: 444106
Ion Ratio Lower Upper
79 100
52 76.5 61.9 92.9
51 38.0 29.7 44.5









#6
Benzaldehyde
Concen: 40.162 ng/ul
RT: 6.954 min Scan# 666
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion: 77 Resp: 333130
Ion Ratio Lower Upper
77 100
105 90.7 76.3 114.5
106 86.8 75.4 113.0



#8
Phenol
Concen: 33.103 ng/ul
RT: 7.019 min Scan# 677
Delta R.T. -0.012 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion: 94 Resp: 669145
Ion Ratio Lower Upper
94 100
65 29.6 22.2 33.4
66 37.3 29.6 44.4





#10

Bis(2-Chloroethyl)ether

Concen: 32.090 ng/ul

RT: 7.237 min Scan# 715

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion: 93 Resp: 540815

Ion Ratio Lower Upper

93 100

63 81.6 61.9 92.9

95 32.2 25.8 38.8



#12

2-Chlorophenol

Concen: 32.259 ng/ul

RT: 7.378 min Scan# 738

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion:128 Resp: 494366


Ion Ratio Lower Upper

128 100

64 51.1 38.8 58.2

130 32.1 25.4 38.0





#13

2-Methylphenol

Concen: 32.246 ng/ul

RT: 8.272 min Scan# 890

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :





SLCS350

Tgt Ion:108 Resp: 506278

Ion Ratio Lower Upper

108 100

107 89.3 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 33.595 ng/ul

RT: 8.342 min Scan# 902

Delta R.T. -0.012 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion: 45 Resp: 786385


Ion Ratio Lower Upper

45 100

77 14.6 12.3 18.5

79 11.3 9.8 14.6









#16
Acetophenone
Concen: 32.910 ng/ul
RT: 8.642 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:105 Resp: 845392
Ion Ratio Lower Upper
105 100
77 83.0 64.1 96.1
51 30.1 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 32.650 ng/ul
RT: 8.631 min Scan# 951
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion: 70 Resp: 464208
Ion Ratio Lower Upper
70 100
42 58.5 43.0 64.4
101 9.0 7.4 11.2
130 18.7 16.2 24.2









#18
4-Methylphenol
Concen: 32.418 ng/ul
RT: 8.607 min Scan# 947
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:108 Resp: 549515
Ion Ratio Lower Upper
108 100
107 110.2 88.6 133.0



#19
Hexachloroethane
Concen: 31.262 ng/ul
RT: 8.883 min Scan# 994
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:117 Resp: 217591
Ion Ratio Lower Upper
117 100
201 93.7 78.3 117.5
199 58.9 47.4 71.0









#22
Nitrobenzene
Concen: 33.586 ng/ul
RT: 9.025 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion: 77 Resp: 643438
Ion Ratio Lower Upper
77 100
123 41.1 36.1 54.1
65 13.4 10.2 15.2



#23
Isophorone
Concen: 32.062 ng/ul
RT: 9.548 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion: 82 Resp: 1309849
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.2
138 15.6 13.2 19.8





#25

2-Nitrophenol

Concen: 31.533 ng/ul

RT: 9.736 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:139 Resp: 286661

Ion Ratio Lower Upper

139 100

65 58.4 40.7 61.1

109 31.6 23.9 35.9



#26

2,4-Dimethylphenol

Concen: 31.850 ng/ul

RT: 9.801 min Scan# 1150

Delta R.T. -0.012 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion:107 Resp: 591674


Ion Ratio Lower Upper

107 100

121 49.1 39.1 58.7

122 83.1 68.1 102.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 31.559 ng/ul
 RT: 10.036 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BM039628.D
 Acq: 24 Apr 2023 22:56




Instrument :
 BNA_M
 ClientSampleId :
 SLCS350


Tgt Ion: 93 Resp: 758124
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.8 38.6
 123 11.1 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 32.065 ng/ul
 RT: 10.277 min Scan# 1231
 Delta R.T. -0.012 min
 Lab File: BM039628.D
 Acq: 24 Apr 2023 22:56

Tgt Ion: 162 Resp: 473512
 Ion Ratio Lower Upper
 162 100
 164 65.6 52.4 78.6
 98 37.3 29.4 44.0





#30

Naphthalene

Concen: 31.633 ng/ul

RT: 10.666 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:128 Resp: 1687989

Ion Ratio Lower Upper

128 100

129 10.9 8.9 13.3

127 12.9 10.5 15.7



#32

4-Chloroaniline

Concen: 25.680 ng/ul

RT: 10.789 min Scan# 1318

Delta R.T. -0.012 min

Lab File: BM039628.D




Acq: 24 Apr 2023 22:56


Tgt Ion:127 Resp: 566556

Ion Ratio Lower Upper

127 100

129 31.7 25.8 38.8









#33
Hexachlorobutadiene
Concen: 30.076 ng/ul
RT: 10.942 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:225 Resp: 292664
Ion Ratio Lower Upper
225 100
223 62.2 50.1 75.1
227 63.7 50.4 75.6



#34
Caprolactam
Concen: 30.925 ng/ul
RT: 11.583 min Scan# 1453
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:113 Resp: 170069
Ion Ratio Lower Upper
113 100
55 185.1 145.5 218.3
56 134.0 103.9 155.9









#35
4-Chloro-3-methylphenol
Concen: 32.214 ng/ul
RT: 11.930 min Scan# 1513
Delta R.T. -0.012 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:107 Resp: 555439
Ion Ratio Lower Upper
107 100
144 24.9 21.0 31.4
142 76.7 62.9 94.3



#36
2-Methylnaphthalene
Concen: 30.908 ng/ul
RT: 12.283 min Scan# 1572
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:142 Resp: 1105659
Ion Ratio Lower Upper
142 100
141 88.4 71.1 106.7





#37

1-Methylnaphthalene

Concen: 31.663 ng/ul

RT: 12.501 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:142 Resp: 1138234

Ion Ratio Lower Upper

142 100

141 91.3 72.0 108.0

116 4.1 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 31.302 ng/ul

RT: 12.654 min Scan# 1635

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Tgt Ion:216 Resp: 563705




Ion Ratio Lower Upper


216 100

214 79.1 62.2 93.2

179 20.5 16.5 24.7

108 20.0 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 20.438 ng/ul
RT: 12.624 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:237 Resp: 219288
Ion Ratio Lower Upper
237 100
235 62.5 50.9 76.3
272 13.5 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 31.234 ng/ul
RT: 12.907 min Scan# 1678
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56


Tgt Ion:196 Resp: 375774
Ion Ratio Lower Upper
196 100
198 96.4 76.7 115.1
200 31.0 23.8 35.6








#42
2,4,5-Trichlorophenol
Concen: 31.667 ng/ul
RT: 12.989 min Scan# 1692
Delta R.T. -0.012 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56


Instrument :
BNA_M
ClientSampleId :
SLCS350





Tgt Ion:196 Resp: 395970
Ion Ratio Lower Upper
196 100
198 96.7 75.7 113.5
200 31.8 24.6 36.8




#43
1,1'-Biphenyl
Concen: 32.069 ng/ul
RT: 13.301 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56



Tgt Ion:154 Resp: 1486934
Ion Ratio Lower Upper
154 100
153 40.1 31.9 47.9
76 14.7 11.5 17.3









#44
2-Chloronaphthalene
Concen: 31.585 ng/ul
RT: 13.342 min Scan# 1752
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:162 Resp: 1148811
Ion Ratio Lower Upper
162 100
164 32.8 26.1 39.1
127 39.8 31.4 47.2



#45
2-Nitroaniline
Concen: 35.696 ng/ul
RT: 13.566 min Scan# 1790
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion: 65 Resp: 374671
Ion Ratio Lower Upper
65 100
92 67.8 53.4 80.2
138 96.5 83.2 124.8









#47
Dimethylphthalate
Concen: 31.281 ng/ul
RT: 13.930 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:163 Resp: 1498557
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 31.430 ng/ul
RT: 14.060 min Scan# 1874
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:165 Resp: 304521
Ion Ratio Lower Upper
165 100
89 68.3 50.6 76.0
121 23.0 17.3 25.9









#50
Acenaphthylene
Concen: 32.083 ng/ul
RT: 14.189 min Scan# 1896
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:152 Resp: 1865504
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 31.134 ng/ul
RT: 14.395 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:138 Resp: 282742
Ion Ratio Lower Upper
138 100
108 10.2 8.8 13.2
92 123.3 95.5 143.3





#52

Acenaphthene

Concen: 31.868 ng/ul

RT: 14.530 min Scan# 1955

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:153 Resp: 1275771

Ion Ratio Lower Upper

153 100

152 47.7 38.2 57.4

154 88.8 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 25.760 ng/ul

RT: 14.613 min Scan# 1968

Delta R.T. -0.012 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion:184 Resp: 132859


Ion Ratio Lower Upper

184 100

63 80.1 60.0 90.0

154 65.7 51.8 77.6









#55
4-Nitrophenol
Concen: 29.084 ng/ul
RT: 14.724 min Scan# 1989
Delta R.T. -0.018 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:109 Resp: 161772
Ion Ratio Lower Upper
109 100
139 126.7 106.4 159.6
65 146.5 104.8 157.2



#56
Dibenzofuran
Concen: 31.988 ng/ul
RT: 14.871 min Scan# 2012
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:168 Resp: 1740571
Ion Ratio Lower Upper
168 100
139 38.4 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 32.620 ng/ul
RT: 14.854 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:165 Resp: 439151
Ion Ratio Lower Upper
165 100
63 58.9 39.4 59.2
182 2.5 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 31.116 ng/ul
RT: 15.101 min Scan# 2051
Delta R.T. -0.012 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:232 Resp: 348261
Ion Ratio Lower Upper
232 100
131 51.6 38.0 57.0
130 2.6 2.0 3.0
166 32.1 24.8 37.2









#59
Diethylphthalate
Concen: 31.070 ng/ul
RT: 15.295 min Scan# 2084
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:149 Resp: 1542984
Ion Ratio Lower Upper
149 100
177 21.5 17.3 25.9
150 12.5 9.8 14.6



#61
Fluorene
Concen: 32.813 ng/ul
RT: 15.518 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:166 Resp: 1476990
Ion Ratio Lower Upper
166 100
165 97.2 78.2 117.2
167 13.3 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 31.710 ng/ul
RT: 15.512 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:204 Resp: 707459
Ion Ratio Lower Upper
204 100
206 32.2 26.8 40.2
141 61.4 48.5 72.7



#63
4-Nitroaniline
Concen: 37.427 ng/ul
RT: 15.565 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:138 Resp: 260647
Ion Ratio Lower Upper
138 100
92 54.5 44.1 66.1
108 83.7 64.6 97.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 30.901 ng/ul
 RT: 15.618 min Scan# 2139
 Delta R.T. -0.006 min
 Lab File: BM039628.D
 Acq: 24 Apr 2023 22:56




Instrument :
 BNA_M
 ClientSampleId :
 SLCS350


Tgt Ion:198 Resp: 246351
 Ion Ratio Lower Upper
 198 100
 182 3.8 3.1 4.7
 77 29.5 20.6 31.0



#67
 N-Nitrosodiphenylamine
 Concen: 32.119 ng/ul
 RT: 15.736 min Scan# 2159
 Delta R.T. -0.006 min
 Lab File: BM039628.D
 Acq: 24 Apr 2023 22:56

Tgt Ion:169 Resp: 1194115
 Ion Ratio Lower Upper
 169 100
 168 67.1 53.2 79.8
 167 36.7 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 30.972 ng/ul
RT: 16.412 min Scan# 211
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:248 Resp: 417411
Ion Ratio Lower Upper
248 100
250 95.7 77.7 116.5
141 73.4 61.0 91.4



#69
Hexachlorobenzene
Concen: 31.240 ng/ul
RT: 16.524 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:284 Resp: 483687
Ion Ratio Lower Upper
284 100
142 36.9 28.9 43.3
249 30.5 23.9 35.9





#70

Atrazine

Concen: 25.409 ng/ul

RT: 16.689 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:200 Resp: 350528

Ion Ratio Lower Upper

200 100

173 26.8 22.2 33.2

215 47.4 38.9 58.3



#71

Pentachlorophenol

Concen: 28.707 ng/ul

RT: 16.877 min Scan# 2353

Delta R.T. -0.012 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion:266 Resp: 259828


Ion Ratio Lower Upper

266 100

264 64.0 50.2 75.4

268 62.2 51.9 77.9









#72
Phenanthrene
Concen: 32.190 ng/ul
RT: 17.265 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:178 Resp: 2229951
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 19.4 15.8 23.8



#74
Anthracene
Concen: 32.795 ng/ul
RT: 17.359 min Scan# 2435
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:178 Resp: 2279139
Ion Ratio Lower Upper
178 100
179 15.3 12.1 18.1
176 18.9 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 30.914 ng/uL

RT: 13.266 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:216 Resp: 574505

Ion Ratio Lower Upper

216 100

214 77.7 63.0 94.4

218 47.6 38.1 57.1



#76

Pentachlorobenzene

Concen: 31.519 ng/uL

RT: 14.789 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion:250 Resp: 576535


Ion Ratio Lower Upper

250 100

248 63.1 50.4 75.6

252 65.4 50.2 75.2









#77
 Carbazole
 Concen: 33.103 ng/ul
 RT: 17.636 min Scan# 2482
 Delta R.T. -0.006 min
 Lab File: BM039628.D
 Acq: 24 Apr 2023 22:56




Instrument :
 BNA_M
 ClientSampleId :
 SLCS350


Tgt Ion:167 Resp: 2044335
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.0 25.4
 139 13.1 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 30.981 ng/ul
 RT: 18.189 min Scan# 2576
 Delta R.T. -0.006 min
 Lab File: BM039628.D
 Acq: 24 Apr 2023 22:56

Tgt Ion:149 Resp: 2706120
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.0 10.6
 104 5.4 4.1 6.1









#80
Fluoranthene
Concen: 29.107 ng/u1
RT: 19.289 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:202 Resp: 2644891
Ion Ratio Lower Upper
202 100
101 12.0 10.4 15.6
100 9.3 7.7 11.5



#82
Pyrene
Concen: 29.401 ng/u1
RT: 19.653 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:202 Resp: 2765708
Ion Ratio Lower Upper
202 100
101 14.4 12.1 18.1
100 11.6 9.6 14.4





#83

Butylbenzylphthalate

Concen: 27.818 ng/ul

RT: 20.547 min Scan# 2977

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:149 Resp: 1235223

Ion Ratio Lower Upper

149 100

91 79.2 59.4 89.0

206 18.9 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 33.030 ng/ul

RT: 21.341 min Scan# 3112

Delta R.T. -0.000 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56




Tgt Ion:252 Resp: 854552


Ion Ratio Lower Upper

252 100

254 64.8 50.9 76.3

126 11.7 10.4 15.6









#85
Benzo(a)anthracene
Concen: 31.409 ng/ul
RT: 21.400 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:228 Resp: 2650463
Ion Ratio Lower Upper
228 100
229 19.7 15.4 23.2
226 27.3 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 27.818 ng/ul
RT: 21.324 min Scan# 3109
Delta R.T. -0.000 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:149 Resp: 1836948
Ion Ratio Lower Upper
149 100
167 27.1 21.8 32.6
279 3.8 3.1 4.7









#87
Chrysene
Concen: 31.347 ng/ul
RT: 21.453 min Scan# 3132
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:228 Resp: 2523203
Ion Ratio Lower Upper
228 100
226 30.0 23.9 35.9
229 19.7 15.8 23.6



#89
Di-n-octyl phthalate
Concen: 26.608 ng/ul
RT: 22.236 min Scan# 3264
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:149 Resp: 3113165









#90
Benzo(b)fluoranthene
Concen: 30.914 ng/ul
RT: 23.059 min Scan# 3404
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:252 Resp: 2532988
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 11.3 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 32.320 ng/ul
RT: 23.106 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:252 Resp: 2600725
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 11.3 8.6 12.8









#93
Benzo(a)pyrene
Concen: 32.230 ng/ul
RT: 23.677 min Scan# 3510
Delta R.T. -0.006 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56




Instrument :
BNA_M
ClientSampleId :
SLCS350


Tgt Ion:252 Resp: 2225941
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 11.9 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 33.863 ng/ul
RT: 26.224 min Scan# 3942
Delta R.T. -0.012 min
Lab File: BM039628.D
Acq: 24 Apr 2023 22:56

Tgt Ion:276 Resp: 2572496
Ion Ratio Lower Upper
276 100
138 26.7 21.2 31.8
277 24.9 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 34.349 ng/ul

RT: 26.241 min Scan# 3947

Delta R.T. -0.006 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Instrument :

BNA_M

ClientSampleId :

SLCS350





Tgt Ion:278 Resp: 2163500

Ion Ratio Lower Upper

278 100

139 17.4 13.5 20.3

279 23.7 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 33.196 ng/ul

RT: 26.976 min Scan# 4070

Delta R.T. -0.012 min

Lab File: BM039628.D

Acq: 24 Apr 2023 22:56

Tgt Ion:276 Resp: 2008933

Ion Ratio Lower Upper

276 100

138 23.3 18.7 28.1

277 23.9 18.7 28.1

