

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042423\
 Data File : BM039629.D
 Acq On : 24 Apr 2023 23:32
 Operator : CG/JU
 Sample : PB152352BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS352

Quant Time: Apr 25 00:45:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.807	152	201886	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.613	136	998404	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.466	164	612267	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.224	188	1276564	20.000	ng/u1	0.00
79) Chrysene-d12	21.418	240	1212263	20.000	ng/u1	0.00
88) Perylene-d12	23.777	264	1240709	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.202	96	35674	6.783	ng/uL	0.00
4) Pyridine-d5	3.625	84	429882	29.088	ng/u1	0.00
7) Phenol-d5	6.990	99	609864	33.005	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.143	67	404907	33.154	ng/u1	0.00
11) 2-Chlorophenol-d4	7.343	132	475582	32.811	ng/u1	0.00
15) 4-Methylphenol-d8	8.542	113	493955	33.186	ng/u1	0.00
21) Nitrobenzene-d5	8.984	128	263402	32.793	ng/u1	0.00
24) 2-Nitrophenol-d4	9.707	143	298027	33.043	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.248	165	510054	32.865	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.766	131	637482	27.503	ng/u1	-0.01
46) Dimethylphthalate-d6	13.883	166	1585712	32.113	ng/u1	0.00
49) Acenaphthylene-d8	14.160	160	1847770	33.309	ng/u1	0.00
54) 4-Nitrophenol-d4	14.713	143	223784	29.945	ng/u1	-0.02
60) Fluorene-d10	15.466	176	1318406	32.589	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.601	200	272561	32.130	ng/u1	0.00
73) Anthracene-d10	17.324	188	2055536	33.482	ng/u1	0.00
81) Pyrene-d10	19.624	212	2413190	29.957	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.624	264	2248142	33.562	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.237	88	76962	13.372	ng/uL	94
5) Pyridine	3.649	79	476720	30.593	ng/u1	99
6) Benzaldehyde	6.954	77	317031	40.159	ng/u1	98
8) Phenol	7.019	94	658239	34.215	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.237	93	517798	32.282	ng/u1	96
12) 2-Chlorophenol	7.372	128	485178	33.265	ng/u1	96
13) 2-Methylphenol	8.272	108	490455	32.822	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.348	45	748935	33.617	ng/u1	99
16) Acetophenone	8.642	105	894233	36.577	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.631	70	497495	36.766	ng/u1	96
18) 4-Methylphenol	8.601	108	589947	36.569	ng/u1	97
19) Hexachloroethane	8.884	117	228627	34.513	ng/u1	97
22) Nitrobenzene	9.025	77	696473	35.117	ng/u1	95
23) Isophorone	9.548	82	1396766	33.027	ng/u1	98
25) 2-Nitrophenol	9.736	139	314181	33.385	ng/u1	94
26) 2,4-Dimethylphenol	9.801	107	637655	33.157	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.036	93	804215	32.338	ng/u1	99
29) 2,4-Dichlorophenol	10.278	162	510549	33.396	ng/u1	99
30) Naphthalene	10.666	128	1802365	32.628	ng/u1	100
32) 4-Chloroaniline	10.789	127	605886	26.528	ng/u1	98
33) Hexachlorobutadiene	10.942	225	314801	31.251	ng/u1	98
34) Caprolactam	11.583	113	185561	32.595	ng/u1	98
35) 4-Chloro-3-methylphenol	11.930	107	588169	32.952	ng/u1	98
36) 2-Methylnaphthalene	12.283	142	1180991	31.891	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042423\
 Data File : BM039629.D
 Acq On : 24 Apr 2023 23:32
 Operator : CG/JU
 Sample : PB152352BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS352

Quant Time: Apr 25 00:45:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:18:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.501	142	1213144	32.599	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.654	216	599021	32.594	ng/ul	99
40) Hexachlorocyclopentadiene	12.625	237	237926	21.729	ng/ul	98
41) 2,4,6-Trichlorophenol	12.907	196	402711	32.800	ng/ul	98
42) 2,4,5-Trichlorophenol	12.983	196	420117	32.923	ng/ul	99
43) 1,1'-Biphenyl	13.301	154	1576877	33.326	ng/ul	99
44) 2-Chloronaphthalene	13.342	162	1235778	33.294	ng/ul	100
45) 2-Nitroaniline	13.566	65	402052	37.535	ng/ul	97
47) Dimethylphthalate	13.930	163	1598594	32.698	ng/ul	100
48) 2,6-Dinitrotoluene	14.060	165	329335	33.308	ng/ul	97
50) Acenaphthylene	14.189	152	1999652	33.699	ng/ul	100
51) 3-Nitroaniline	14.395	138	306916	33.117	ng/ul	97
52) Acenaphthene	14.530	153	1361488	33.325	ng/ul	100
53) 2,4-Dinitrophenol	14.613	184	146422	27.820	ng/ul	96
55) 4-Nitrophenol	14.724	109	178791	31.498	ng/ul	92
56) Dibenzofuran	14.866	168	1863408	33.557	ng/ul	99
57) 2,4-Dinitrotoluene	14.854	165	475712	34.625	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.101	232	376170	32.934	ng/ul	96
59) Diethylphthalate	15.295	149	1645139	32.462	ng/ul	100
61) Fluorene	15.518	166	1557814	33.914	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.513	204	754433	33.136	ng/ul	99
63) 4-Nitroaniline	15.566	138	269074	37.861	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.618	198	264721	32.458	ng/ul#	91
67) N-Nitrosodiphenylamine	15.730	169	1262064	33.183	ng/ul	99
68) 4-Bromophenyl-phenylether	16.413	248	444319	32.226	ng/ul	98
69) Hexachlorobenzene	16.524	284	516542	32.611	ng/ul	97
70) Atrazine	16.689	200	372420	26.388	ng/ul	98
71) Pentachlorophenol	16.877	266	286418	30.932	ng/ul	100
72) Phenanthrene	17.265	178	2393232	33.769	ng/ul	99
74) Anthracene	17.354	178	2445191	34.392	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.266	216	610794	32.127	ng/ul	99
76) Pentachlorobenzene	14.789	250	615771	32.906	ng/ul	99
77) Carbazole	17.636	167	2199720	34.817	ng/ul	99
78) Di-n-butylphthalate	18.189	149	2886638	32.303	ng/ul	99
80) Fluoranthene	19.289	202	2826341	29.934	ng/ul	98
82) Pyrene	19.654	202	2968093	30.366	ng/ul	99
83) Butylbenzylphthalate	20.548	149	1345608	29.165	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.342	252	930946	34.630	ng/ul	98
85) Benzo(a)anthracene	21.400	228	2888529	32.944	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.324	149	2003943	29.206	ng/ul	99
87) Chrysene	21.453	228	2731117	32.654	ng/ul	100
89) Di-n-octyl phthalate	22.236	149	3414414	27.390	ng/ul	100
90) Benzo(b)fluoranthene	23.059	252	2826382	32.376	ng/ul	100
91) Benzo(k)fluoranthene	23.106	252	2806152	32.731	ng/ul	99
93) Benzo(a)pyrene	23.677	252	2452785	33.333	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.230	276	2922271	36.104	ng/ul	99
95) Dibenzo(a,h)anthracene	26.241	278	2449872	36.506	ng/ul	100
96) Benzo(g,h,i)perylene	26.977	276	2268795	35.187	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.372 ng/uL
RT: 3.237 min Scan# 34
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion: 88 Resp: 76962
Ion Ratio Lower Upper
88 100
43 33.4 24.6 36.8
58 76.2 56.7 85.1



#5
Pyridine
Concen: 30.593 ng/uL
RT: 3.649 min Scan# 104
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion: 79 Resp: 476720
Ion Ratio Lower Upper
79 100
52 76.7 61.9 92.9
51 38.0 29.7 44.5









#6
Benzaldehyde
Concen: 40.159 ng/ul
RT: 6.954 min Scan# 666
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion: 77 Resp: 317031
Ion Ratio Lower Upper
77 100
105 94.7 76.3 114.5
106 90.5 75.4 113.0



#8
Phenol
Concen: 34.215 ng/ul
RT: 7.019 min Scan# 677
Delta R.T. -0.012 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion: 94 Resp: 658239
Ion Ratio Lower Upper
94 100
65 29.6 22.2 33.4
66 38.7 29.6 44.4





#10

Bis(2-Chloroethyl)ether

Concen: 32.282 ng/ul

RT: 7.237 min Scan# 714

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion: 93 Resp: 517798

Ion Ratio Lower Upper

93 100

63 81.6 61.9 92.9

95 32.6 25.8 38.8



#12

2-Chlorophenol

Concen: 33.265 ng/ul

RT: 7.372 min Scan# 737

Delta R.T. -0.012 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:128 Resp: 485178


Ion Ratio Lower Upper

128 100

64 52.8 38.8 58.2

130 31.9 25.4 38.0





#13

2-Methylphenol

Concen: 32.822 ng/ul

RT: 8.272 min Scan# 890

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :





SLCS352

Tgt Ion:108 Resp: 490455

Ion Ratio Lower Upper

108 100

107 90.0 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 33.617 ng/ul

RT: 8.348 min Scan# 903

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion: 45 Resp: 748935


Ion Ratio Lower Upper

45 100

77 15.0 12.3 18.5

79 11.6 9.8 14.6









#16
Acetophenone
Concen: 36.577 ng/ul
RT: 8.642 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:105 Resp: 894233
Ion Ratio Lower Upper
105 100
77 83.3 64.1 96.1
51 30.3 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 36.766 ng/ul
RT: 8.631 min Scan# 951
Delta R.T. -0.000 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion: 70 Resp: 497495
Ion Ratio Lower Upper
70 100
42 58.1 43.0 64.4
101 9.3 7.4 11.2
130 19.5 16.2 24.2









#18
4-Methylphenol
Concen: 36.569 ng/ul
RT: 8.601 min Scan# 947
Delta R.T. -0.012 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:108 Resp: 589947
Ion Ratio Lower Upper
108 100
107 114.1 88.6 133.0



#19
Hexachloroethane
Concen: 34.513 ng/ul
RT: 8.884 min Scan# 994
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:117 Resp: 228627
Ion Ratio Lower Upper
117 100
201 94.4 78.3 117.5
199 60.0 47.4 71.0









#22
Nitrobenzene
Concen: 35.117 ng/ul
RT: 9.025 min Scan# 1018
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion: 77 Resp: 696473
Ion Ratio Lower Upper
77 100
123 40.8 36.1 54.1
65 12.8 10.2 15.2



#23
Isophorone
Concen: 33.027 ng/ul
RT: 9.548 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion: 82 Resp: 1396766
Ion Ratio Lower Upper
82 100
95 6.4 5.4 8.2
138 15.6 13.2 19.8





#25

2-Nitrophenol

Concen: 33.385 ng/ul

RT: 9.736 min Scan# 1139

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:139 Resp: 314181

Ion Ratio Lower Upper

139 100

65 56.7 40.7 61.1

109 31.0 23.9 35.9



#26

2,4-Dimethylphenol

Concen: 33.157 ng/ul

RT: 9.801 min Scan# 1150

Delta R.T. -0.012 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:107 Resp: 637655


Ion Ratio Lower Upper

107 100

121 49.0 39.1 58.7

122 82.8 68.1 102.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.338 ng/ul
 RT: 10.036 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32




Instrument :
 BNA_M
 ClientSampleId :
 SLCS352


Tgt Ion: 93 Resp: 804215
 Ion Ratio Lower Upper
 93 100
 95 31.6 25.8 38.6
 123 10.6 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 33.396 ng/ul
 RT: 10.278 min Scan# 1231
 Delta R.T. -0.012 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32

Tgt Ion: 162 Resp: 510549
 Ion Ratio Lower Upper
 162 100
 164 65.2 52.4 78.6
 98 37.8 29.4 44.0





#30

Naphthalene

Concen: 32.628 ng/ul

RT: 10.666 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:128 Resp: 1802365

Ion Ratio Lower Upper

128 100

129 10.9 8.9 13.3

127 12.9 10.5 15.7



#32

4-Chloroaniline

Concen: 26.528 ng/ul

RT: 10.789 min Scan# 1318

Delta R.T. -0.012 min

Lab File: BM039629.D




Acq: 24 Apr 2023 23:32


Tgt Ion:127 Resp: 605886

Ion Ratio Lower Upper

127 100

129 31.3 25.8 38.8





#33

Hexachlorobutadiene

Concen: 31.251 ng/ul

RT: 10.942 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:225 Resp: 314801

Ion Ratio Lower Upper

225 100

223 63.8 50.1 75.1

227 65.4 50.4 75.6



#34

Caprolactam

Concen: 32.595 ng/ul

RT: 11.583 min Scan# 1453

Delta R.T. -0.000 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:113 Resp: 185561


Ion Ratio Lower Upper

113 100

55 184.4 145.5 218.3

56 132.7 103.9 155.9









#35
 4-Chloro-3-methylphenol
 Concen: 32.952 ng/ul
 RT: 11.930 min Scan# 1513
 Delta R.T. -0.012 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32




Instrument :
 BNA_M
 ClientSampleId :
 SLCS352


Tgt Ion:107 Resp: 588169
 Ion Ratio Lower Upper
 107 100
 144 25.0 21.0 31.4
 142 77.5 62.9 94.3



#36
 2-Methylnaphthalene
 Concen: 31.891 ng/ul
 RT: 12.283 min Scan# 1572
 Delta R.T. -0.006 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32

Tgt Ion:142 Resp: 1180991
 Ion Ratio Lower Upper
 142 100
 141 88.3 71.1 106.7





#37

1-Methylnaphthalene

Concen: 32.599 ng/ul

RT: 12.501 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:142 Resp: 1213144

Ion Ratio Lower Upper

142 100

141 92.8 72.0 108.0

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 32.594 ng/ul

RT: 12.654 min Scan# 1635

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Tgt Ion:216 Resp: 599021




Ion Ratio Lower Upper


216 100

214 78.4 62.2 93.2

179 20.6 16.5 24.7

108 18.8 14.0 21.0





#40

Hexachlorocyclopentadiene

Concen: 21.729 ng/ul

RT: 12.625 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:237 Resp: 237926

Ion Ratio Lower Upper

237 100

235 61.8 50.9 76.3

272 13.1 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 32.800 ng/ul

RT: 12.907 min Scan# 1678

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:196 Resp: 402711


Ion Ratio Lower Upper

196 100

198 97.5 76.7 115.1

200 31.2 23.8 35.6









#42
2,4,5-Trichlorophenol
Concen: 32.923 ng/ul
RT: 12.983 min Scan# 1691
Delta R.T. -0.018 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:196 Resp: 420117
Ion Ratio Lower Upper
196 100
198 96.1 75.7 113.5
200 30.9 24.6 36.8



#43
1,1'-Biphenyl
Concen: 33.326 ng/ul
RT: 13.301 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:154 Resp: 1576877
Ion Ratio Lower Upper
154 100
153 40.0 31.9 47.9
76 15.1 11.5 17.3









#44
2-Chloronaphthalene
Concen: 33.294 ng/ul
RT: 13.342 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:162 Resp: 1235778
Ion Ratio Lower Upper
162 100
164 32.7 26.1 39.1
127 39.6 31.4 47.2



#45
2-Nitroaniline
Concen: 37.535 ng/ul
RT: 13.566 min Scan# 1790
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion: 65 Resp: 402052
Ion Ratio Lower Upper
65 100
92 68.1 53.4 80.2
138 99.2 83.2 124.8









#47
Dimethylphthalate
Concen: 32.698 ng/ul
RT: 13.930 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:163 Resp: 1598594
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.2 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 33.308 ng/ul
RT: 14.060 min Scan# 1874
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:165 Resp: 329335
Ion Ratio Lower Upper
165 100
89 65.6 50.6 76.0
121 22.6 17.3 25.9





#50

Acenaphthylene

Concen: 33.699 ng/ul

RT: 14.189 min Scan# 1896

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:152 Resp: 1999652

Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.8

153 13.2 10.4 15.6



#51

3-Nitroaniline

Concen: 33.117 ng/ul

RT: 14.395 min Scan# 1931

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:138 Resp: 306916


Ion Ratio Lower Upper

138 100

108 9.7 8.8 13.2

92 123.3 95.5 143.3









#52
Acenaphthene
Concen: 33.325 ng/ul
RT: 14.530 min Scan# 1955
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:153 Resp: 1361488
Ion Ratio Lower Upper
153 100
152 48.1 38.2 57.4
154 89.0 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 27.820 ng/ul
RT: 14.613 min Scan# 1968
Delta R.T. -0.012 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:184 Resp: 146422
Ion Ratio Lower Upper
184 100
63 79.3 60.0 90.0
154 66.4 51.8 77.6









#55
4-Nitrophenol
Concen: 31.498 ng/ul
RT: 14.724 min Scan# 1987
Delta R.T. -0.018 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:109 Resp: 178791
Ion Ratio Lower Upper
109 100
139 127.0 106.4 159.6
65 142.9 104.8 157.2



#56
Dibenzofuran
Concen: 33.557 ng/ul
RT: 14.866 min Scan# 2011
Delta R.T. -0.012 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:168 Resp: 1863408
Ion Ratio Lower Upper
168 100
139 38.8 30.7 46.1





#57

2,4-Dinitrotoluene

Concen: 34.625 ng/ul

RT: 14.854 min Scan# 2010

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:165 Resp: 475712

Ion Ratio Lower Upper

165 100

63 59.6 39.4 59.2#

182 2.6 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 32.934 ng/ul

RT: 15.101 min Scan# 2051

Delta R.T. -0.012 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Tgt Ion:232 Resp: 376170




Ion Ratio Lower Upper


232 100

131 50.7 38.0 57.0

130 2.8 2.0 3.0

166 32.1 24.8 37.2





#59

Diethylphthalate

Concen: 32.462 ng/ul

RT: 15.295 min Scan# 2084

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:149 Resp: 1645139

Ion Ratio Lower Upper

149 100

177 21.5 17.3 25.9

150 12.3 9.8 14.6



#61

Fluorene

Concen: 33.914 ng/ul

RT: 15.518 min Scan# 2122

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:166 Resp: 1557814


Ion Ratio Lower Upper

166 100

165 96.9 78.2 117.2

167 13.6 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 33.136 ng/ul
RT: 15.513 min Scan# 2121
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:204 Resp: 754433
Ion Ratio Lower Upper
204 100
206 32.4 26.8 40.2
141 59.8 48.5 72.7



#63
4-Nitroaniline
Concen: 37.861 ng/ul
RT: 15.566 min Scan# 2130
Delta R.T. -0.000 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:138 Resp: 269074
Ion Ratio Lower Upper
138 100
92 55.3 44.1 66.1
108 83.6 64.6 97.0









#66
4,6-Dinitro-2-methylphenol
Concen: 32.458 ng/ul
RT: 15.618 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:198 Resp: 264721
Ion Ratio Lower Upper
198 100
182 3.9 3.1 4.7
77 31.1 20.6 31.0#



#67
N-Nitrosodiphenylamine
Concen: 33.183 ng/ul
RT: 15.730 min Scan# 2158
Delta R.T. -0.012 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:169 Resp: 1262064
Ion Ratio Lower Upper
169 100
168 67.3 53.2 79.8
167 37.0 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 32.226 ng/ul
RT: 16.413 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:248 Resp: 444319
Ion Ratio Lower Upper
248 100
250 97.7 77.7 116.5
141 72.9 61.0 91.4



#69
Hexachlorobenzene
Concen: 32.611 ng/ul
RT: 16.524 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:284 Resp: 516542
Ion Ratio Lower Upper
284 100
142 37.4 28.9 43.3
249 31.9 23.9 35.9





#70

Atrazine

Concen: 26.388 ng/ul

RT: 16.689 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:200 Resp: 372420

Ion Ratio Lower Upper

200 100

173 26.7 22.2 33.2

215 47.1 38.9 58.3



#71

Pentachlorophenol

Concen: 30.932 ng/ul

RT: 16.877 min Scan# 2353

Delta R.T. -0.012 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:266 Resp: 286418


Ion Ratio Lower Upper

266 100

264 62.8 50.2 75.4

268 64.3 51.9 77.9









#72
Phenanthrene
Concen: 33.769 ng/ul
RT: 17.265 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:178 Resp: 2393232
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.4
176 19.4 15.8 23.8



#74
Anthracene
Concen: 34.392 ng/ul
RT: 17.354 min Scan# 2434
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:178 Resp: 2445191
Ion Ratio Lower Upper
178 100
179 15.6 12.1 18.1
176 19.0 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 32.127 ng/uL

RT: 13.266 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:216 Resp: 610794

Ion Ratio Lower Upper

216 100

214 77.4 63.0 94.4

218 47.4 38.1 57.1



#76

Pentachlorobenzene

Concen: 32.906 ng/uL

RT: 14.789 min Scan# 1998

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:250 Resp: 615771


Ion Ratio Lower Upper

250 100

248 62.9 50.4 75.6

252 64.1 50.2 75.2









#77
 Carbazole
 Concen: 34.817 ng/ul
 RT: 17.636 min Scan# 2482
 Delta R.T. -0.006 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32




Instrument :
 BNA_M
 ClientSampleId :
 SLCS352


Tgt Ion:167 Resp: 2199720
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.0 25.4
 139 13.1 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 32.303 ng/ul
 RT: 18.189 min Scan# 2576
 Delta R.T. -0.006 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32

Tgt Ion:149 Resp: 2886638
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.0 10.6
 104 5.4 4.1 6.1









#80
Fluoranthene
Concen: 29.934 ng/ul
RT: 19.289 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:202 Resp: 2826341
Ion Ratio Lower Upper
202 100
101 11.9 10.4 15.6
100 9.2 7.7 11.5



#82
Pyrene
Concen: 30.366 ng/ul
RT: 19.654 min Scan# 2825
Delta R.T. -0.000 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:202 Resp: 2968093
Ion Ratio Lower Upper
202 100
101 14.4 12.1 18.1
100 11.7 9.6 14.4









#83
 Butylbenzylphthalate
 Concen: 29.165 ng/ul
 RT: 20.548 min Scan# 2977
 Delta R.T. -0.006 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32




Instrument :
 BNA_M
 ClientSampleId :
 SLCS352


Tgt Ion:149 Resp: 1345608
 Ion Ratio Lower Upper
 149 100
 91 77.5 59.4 89.0
 206 19.1 15.8 23.8



#84
 3,3'-Dichlorobenzidine
 Concen: 34.630 ng/ul
 RT: 21.342 min Scan# 3112
 Delta R.T. -0.000 min
 Lab File: BM039629.D
 Acq: 24 Apr 2023 23:32

Tgt Ion:252 Resp: 930946
 Ion Ratio Lower Upper
 252 100
 254 65.1 50.9 76.3
 126 11.3 10.4 15.6





#85

Benzo(a)anthracene

Concen: 32.944 ng/ul

RT: 21.400 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:228 Resp: 2888529

Ion Ratio Lower Upper

228 100

229 19.7 15.4 23.2

226 27.0 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 29.206 ng/ul

RT: 21.324 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:149 Resp: 2003943


Ion Ratio Lower Upper

149 100

167 26.9 21.8 32.6

279 4.0 3.1 4.7





#87

Chrysene

Concen: 32.654 ng/ul

RT: 21.453 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:228 Resp: 2731117

Ion Ratio Lower Upper

228 100

226 29.9 23.9 35.9

229 19.7 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 27.390 ng/ul




RT: 22.236 min Scan# 3264


Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Tgt Ion:149 Resp: 3414414









#90
Benzo(b)fluoranthene
Concen: 32.376 ng/ul
RT: 23.059 min Scan# 3404
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32




Instrument :
BNA_M
ClientSampleId :
SLCS352


Tgt Ion:252 Resp: 2826382
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 11.1 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 32.731 ng/ul
RT: 23.106 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BM039629.D
Acq: 24 Apr 2023 23:32

Tgt Ion:252 Resp: 2806152
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 11.1 8.6 12.8





#93

Benzo(a)pyrene

Concen: 33.333 ng/ul

RT: 23.677 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:252 Resp: 2452785

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 12.0 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 36.104 ng/ul

RT: 26.230 min Scan# 3943

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32




Tgt Ion:276 Resp: 2922271


Ion Ratio Lower Upper

276 100

138 27.5 21.2 31.8

277 25.2 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 36.506 ng/ul

RT: 26.241 min Scan# 3947

Delta R.T. -0.006 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Instrument :

BNA_M

ClientSampleId :

SLCS352





Tgt Ion:278 Resp: 2449872

Ion Ratio Lower Upper

278 100

139 17.0 13.5 20.3

279 23.6 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 35.187 ng/ul

RT: 26.977 min Scan# 4070

Delta R.T. -0.012 min

Lab File: BM039629.D

Acq: 24 Apr 2023 23:32

Tgt Ion:276 Resp: 2268795

Ion Ratio Lower Upper

276 100

138 24.0 18.7 28.1

277 24.1 18.7 28.1

