


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042623\
 Data File : BM039670.D
 Acq On : 26 Apr 2023 10:03
 Operator : CG/JU
 Sample : DFTPP113
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP113

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM040623.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 27 01:14:42 2023



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.6	124794	PASS
68	69	0.00	2	1.5	1915	PASS
69	198	0.00	100	44.7	127893	PASS
70	69	0.00	2	0.5	682	PASS
127	198	10	80	50.2	143416	PASS
197	198	0.00	2	0.4	1058	PASS
198	198	100	100	100.0	285952	PASS
199	198	5	9	6.7	19091	PASS
275	198	10	60	25.2	72032	PASS
365	198	1	100	3.4	9802	PASS
441	198	0.01	100	12.7	36416	PASS
442	198	50	100	78.0	223019	PASS
443	442	15	24	19.5	43525	PASS

m/z	Abundance
36.10	150.0
39.90	1133.0
41.00	267.0
44.00	3538.0
45.10	155.0
46.90	280.0
49.00	671.0
50.90	433.0
56.90	156.0
72.90	1026.0
74.90	158.0
76.90	230.0
82.90	151.0
83.90	558.0
85.80	322.0
95.90	297.0
114.90	158.0
118.90	160.0
132.90	381.0
146.90	337.0
190.90	415.0
192.90	340.0
207.00	3368.0
207.90	794.0
208.90	647.0
220.90	156.0
266.90	165.0
281.00	646.0
282.00	210.0
355.00	214.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

m/z	Abundance
36.00	239.0
36.90	438.0
38.00	1393.0
39.00	10167.0
40.00	1566.0
41.00	510.0
43.00	309.0
44.00	2759.0
44.90	339.0
47.00	278.0
47.90	175.0
49.00	1500.0
50.00	33696.0
51.00	131648.0
52.00	6311.0
53.00	307.0
54.90	955.0
56.00	3603.0
57.00	8676.0
57.90	476.0
59.00	182.0
61.00	1676.0
62.00	2033.0
63.00	5235.0
64.00	746.0
65.00	2673.0
66.10	279.0
66.90	307.0
67.90	1915.0
69.00	134464.0
70.00	716.0
73.00	1901.0
74.00	13525.0
75.00	20808.0
76.00	7293.0
77.00	150848.0
78.00	11366.0
79.00	9618.0
80.00	7583.0
81.00	11729.0
82.00	2560.0
83.00	2306.0
83.90	698.0
85.00	1961.0
85.90	3276.0
87.00	1422.0
88.00	617.0
88.90	254.0
91.00	2278.0
92.00	2965.0
93.00	17272.0
93.90	1148.0
94.90	303.0
96.00	1162.0
96.90	520.0
97.90	12373.0
98.90	11085.0
99.90	1005.0
100.90	6018.0
102.00	334.0
102.90	1548.0
104.00	3340.0
104.90	3355.0
105.90	967.0
107.00	42744.0
108.00	7404.0
109.00	1045.0
110.00	73192.0
111.00	11743.0
111.90	1226.0
113.00	403.0
115.10	301.0
116.00	2186.0
117.00	36448.0
117.90	2972.0
119.10	410.0
119.80	470.0
121.10	189.0
122.00	2512.0
122.90	4068.0
123.90	1765.0
124.90	1872.0
126.00	384.0
127.00	143360.0
128.00	11049.0
129.00	53040.0
129.90	4920.0
130.90	914.0
131.80	403.0
132.90	490.0
133.90	1353.0
134.90	3954.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

136.00	1913.0
137.00	2570.0
137.80	471.0
138.90	223.0
139.90	536.0
141.00	7111.0
141.90	2228.0
142.90	1446.0
144.00	499.0
145.10	385.0
146.00	1104.0
146.90	3994.0
147.90	9011.0
148.90	1910.0
149.90	477.0
151.00	929.0
151.60	598.0
152.90	2208.0
153.90	1636.0
155.00	3842.0
156.00	5291.0
157.00	1062.0
158.00	1397.0
158.90	1036.0
159.90	2001.0
161.00	3019.0
161.90	1120.0
162.90	458.0
163.70	299.0
164.90	2639.0
165.90	1802.0
167.00	14995.0
168.00	9669.0
168.90	1506.0
169.90	563.0
170.80	707.0
171.90	1257.0
173.00	1548.0
174.00	2944.0
175.00	5119.0
176.00	1586.0
176.90	2270.0
177.90	667.0
178.90	9852.0
179.90	7997.0
180.90	3286.0
182.00	564.0
182.80	214.0
183.90	842.0
184.90	4435.0
186.00	36352.0
187.00	10850.0
187.90	1537.0
188.90	2408.0
189.80	392.0
190.90	1410.0
191.90	3427.0
193.00	3624.0
194.00	884.0
194.90	344.0
195.90	6680.0
196.80	935.0
197.90	263872.0
198.90	17968.0
199.90	1277.0
201.30	1132.0
202.90	1876.0
204.00	9588.0
205.00	16488.0
206.00	67672.0
207.00	13460.0
208.00	2874.0
209.00	1283.0
210.10	1257.0
210.90	2756.0
211.80	544.0
212.80	194.0
215.00	713.0
216.00	1551.0
216.90	19280.0
218.00	2700.0
218.90	355.0
220.10	286.0
221.00	9093.0
221.80	2795.0
223.00	3753.0
224.00	37944.0
225.00	10318.0
225.90	1262.0
226.90	15713.0
227.90	2540.0
228.90	3201.0
229.90	448.0
230.90	1451.0
231.90	208.0
232.80	292.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

234.00	1042.0
235.00	1205.0
235.90	751.0
237.00	1934.0
237.90	212.0
238.90	611.0
239.90	621.0
240.90	921.0
242.00	2168.0
243.00	1918.0
244.00	28856.0
245.00	3464.0
245.90	5397.0
247.00	1083.0
247.90	252.0
249.00	1161.0
249.90	246.0
250.90	286.0
252.00	258.0
253.00	733.0
254.00	734.0
255.00	145408.0
256.00	22776.0
257.00	2042.0
257.90	8429.0
258.90	1331.0
259.90	264.0
260.80	273.0
264.90	3461.0
265.90	662.0
266.90	208.0
271.00	310.0
271.90	404.0
273.00	4192.0
274.00	11629.0
275.00	61496.0
276.00	8424.0
277.00	5370.0
277.90	1185.0
278.80	224.0
281.00	601.0
282.10	329.0
282.90	813.0
283.90	463.0
284.90	850.0
288.90	330.0
290.00	212.0
291.80	370.0
292.90	1302.0
293.90	436.0
294.80	269.0
296.00	17736.0
296.90	2692.0
297.80	216.0
301.00	276.0
302.00	330.0
303.00	2006.0
304.00	684.0
304.90	154.0
308.00	268.0
308.90	169.0
309.80	303.0
314.00	615.0
314.90	1849.0
315.90	1126.0
320.90	542.0
322.00	197.0
323.00	5503.0
324.00	1069.0
326.90	1291.0
327.90	599.0
332.00	359.0
333.00	463.0
334.00	3439.0
335.00	1203.0
340.90	567.0
342.00	300.0
345.90	1251.0
346.90	284.0
352.00	1708.0
353.00	1201.0
354.00	1466.0
354.90	383.0
358.90	175.0
364.90	7408.0
365.90	1482.0
369.90	221.0
371.00	384.0
372.00	2768.0
372.90	814.0
382.90	564.0
383.90	207.0
389.90	289.0
391.00	206.0
391.80	192.0
400.70	174.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

402.00	1227.0
403.00	1622.0
403.90	486.0
404.90	223.0
420.90	1218.0
421.90	1329.0
423.00	8872.0
423.90	2719.0
425.00	635.0
441.00	24320.0
442.00	154496.0
443.00	31136.0
444.00	2887.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

m/z	Abundance
36.00	269.0
37.00	502.0
38.10	1903.0
39.00	10667.0
40.00	1279.0
41.00	553.0
43.00	408.0
44.00	2905.0
44.90	339.0
46.90	267.0
48.00	196.0
49.00	1618.0
50.00	37552.0
51.00	147136.0
52.00	7029.0
53.00	416.0
54.90	901.0
56.00	4273.0
57.00	9685.0
57.90	503.0
58.90	280.0
59.80	150.0
61.00	2003.0
62.00	2126.0
63.00	5717.0
64.00	819.0
65.00	3149.0
65.90	248.0
66.90	311.0
68.00	2308.0
69.00	151616.0
70.00	826.0
73.00	1999.0
74.00	15864.0
75.00	23648.0
76.00	8724.0
77.00	180032.0
78.00	12829.0
79.00	10529.0
79.90	7985.0
81.00	12360.0
81.90	3531.0
83.00	2843.0
83.90	693.0
84.90	2324.0
86.00	3336.0
87.00	1301.0
88.00	799.0
88.90	309.0
91.00	2500.0
91.90	2786.0
93.00	21376.0
93.90	1428.0
94.90	374.0
95.90	934.0
97.00	322.0
98.00	14625.0
99.00	12078.0
100.00	940.0
100.90	6730.0
101.90	413.0
102.90	1977.0
103.90	3859.0
105.00	3910.0
106.00	848.0
107.00	51192.0
108.00	7261.0
109.00	1250.0
110.00	89528.0
110.90	13608.0
112.00	1761.0
112.90	579.0
114.90	150.0
116.00	2688.0
117.00	45672.0
117.90	3388.0
118.90	454.0
119.90	450.0
121.00	329.0
121.90	3420.0
122.90	4928.0
123.90	2270.0
124.90	2225.0
126.00	470.0
127.00	171776.0
128.00	12857.0
128.90	64504.0
129.90	5239.0
131.00	1126.0
131.80	463.0
132.90	504.0
134.00	1715.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

134.90	5442.0
135.90	2211.0
136.90	3076.0
137.90	941.0
138.80	425.0
139.90	679.0
140.90	8290.0
141.90	2615.0
143.00	1732.0
143.90	561.0
145.00	379.0
145.90	1569.0
146.90	3817.0
147.90	11639.0
149.00	2374.0
149.90	694.0
150.90	1329.0
152.00	544.0
152.90	2787.0
153.90	1786.0
155.00	4808.0
156.00	6613.0
157.00	1458.0
157.90	1621.0
158.90	1327.0
160.00	2304.0
161.00	4213.0
161.90	1132.0
162.90	457.0
163.90	371.0
164.90	3283.0
165.90	2644.0
167.00	17752.0
168.00	12107.0
169.00	1522.0
170.00	587.0
170.90	809.0
171.90	1683.0
173.00	2247.0
173.90	3934.0
175.00	7095.0
175.90	1833.0
176.90	3396.0
177.90	1052.0
178.90	13424.0
179.90	9350.0
181.00	4450.0
181.90	652.0
182.90	498.0
183.90	938.0
185.00	5485.0
186.00	47080.0
187.00	13264.0
188.00	2001.0
188.90	3161.0
189.90	374.0
191.00	1629.0
191.90	4267.0
193.00	4711.0
193.90	1081.0
195.00	592.0
196.00	9376.0
196.80	1424.0
197.90	347648.0
198.90	23456.0
200.00	1859.0
201.30	1226.0
202.90	2454.0
204.00	13675.0
205.00	21456.0
206.00	92136.0
207.00	15141.0
207.90	3673.0
208.90	1456.0
210.00	1231.0
211.00	3685.0
211.70	795.0
212.90	275.0
213.90	169.0
214.90	858.0
215.90	2064.0
216.90	24088.0
218.00	3160.0
218.80	238.0
219.70	235.0
221.00	13737.0
221.70	3080.0
223.00	5430.0
224.00	50592.0
225.00	13280.0
226.00	1381.0
226.90	21160.0
227.90	3029.0
228.90	4374.0
230.00	631.0
230.90	2047.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

232.00	341.0
233.00	387.0
233.90	1256.0
234.90	1475.0
235.90	901.0
237.00	2366.0
237.90	326.0
238.90	918.0
240.00	722.0
240.90	1254.0
242.00	2770.0
243.00	2793.0
244.00	41056.0
245.00	5460.0
245.90	7892.0
247.00	1535.0
248.00	239.0
249.00	1629.0
250.00	287.0
251.00	280.0
251.80	359.0
253.00	859.0
254.00	1125.0
255.00	203392.0
256.00	29680.0
257.00	2337.0
257.90	11400.0
258.90	1908.0
259.90	383.0
261.00	452.0
263.90	450.0
264.90	5062.0
266.00	889.0
269.80	186.0
270.90	451.0
272.00	516.0
272.90	6372.0
274.00	16448.0
275.00	87520.0
276.00	11388.0
276.90	7917.0
277.90	1370.0
278.90	256.0
281.00	758.0
281.90	418.0
282.90	895.0
283.90	481.0
285.10	1362.0
286.00	354.0
289.00	338.0
290.00	259.0
291.00	261.0
292.00	476.0
292.90	1465.0
294.00	474.0
294.90	668.0
295.90	27312.0
296.90	3650.0
298.00	262.0
300.90	333.0
301.90	494.0
303.00	2838.0
303.90	813.0
307.90	388.0
308.90	360.0
309.90	363.0
314.00	1302.0
314.90	2833.0
315.90	1543.0
316.80	195.0
320.90	887.0
322.00	414.0
323.00	8371.0
324.00	1427.0
326.90	1624.0
328.00	644.0
328.90	185.0
331.90	669.0
332.90	942.0
334.00	5026.0
335.00	1235.0
341.00	876.0
342.10	253.0
346.00	1628.0
346.90	501.0
351.00	204.0
352.00	2589.0
353.00	1544.0
354.00	2147.0
355.00	656.0
359.10	202.0
364.90	11581.0
366.00	1614.0
370.00	310.0
371.00	539.0
372.00	3994.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

372.90	1065.0
383.00	1234.0
384.00	286.0
389.90	675.0
391.00	336.0
391.90	240.0
401.00	290.0
401.90	1630.0
403.00	2189.0
404.00	872.0
420.90	1873.0
422.00	2074.0
423.00	14590.0
424.00	4048.0
424.90	539.0
441.00	43008.0
442.00	264256.0
443.00	52392.0
444.00	4866.0
444.90	263.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

m/z	Abundance
37.00	415.0
38.00	944.0
39.00	6866.0
40.00	1182.0
41.00	474.0
42.90	244.0
44.00	3293.0
44.90	287.0
47.00	243.0
49.00	1402.0
50.00	23072.0
51.00	96896.0
52.00	4833.0
53.00	223.0
55.00	637.0
55.90	3038.0
57.00	6668.0
57.90	235.0
58.90	202.0
61.00	1010.0
61.90	1500.0
63.00	3557.0
64.00	491.0
65.00	1957.0
66.00	214.0
66.90	204.0
68.00	1522.0
69.00	97600.0
70.00	503.0
73.00	1775.0
74.00	10019.0
75.00	16712.0
76.00	5519.0
77.00	113664.0
78.00	7719.0
79.00	7313.0
79.90	5806.0
81.00	7865.0
82.00	2031.0
82.90	2034.0
84.00	781.0
84.90	1815.0
86.00	2548.0
87.00	1141.0
88.00	505.0
89.00	191.0
91.00	2035.0
91.90	1845.0
93.00	12708.0
94.00	881.0
94.80	198.0
96.00	819.0
97.00	332.0
98.00	9608.0
99.00	8043.0
99.90	745.0
101.00	4546.0
101.80	249.0
102.90	1145.0
103.90	2356.0
105.00	2609.0
106.00	1060.0
107.00	32872.0
108.00	5058.0
109.00	906.0
110.00	59976.0
111.00	9064.0
111.90	1332.0
112.90	391.0
114.90	272.0
116.00	1557.0
116.90	30368.0
118.00	2089.0
118.90	425.0
120.00	449.0
120.90	182.0
121.90	1825.0
122.90	3384.0
124.00	1407.0
124.90	1303.0
126.00	525.0
127.00	115112.0
128.00	8685.0
129.00	44472.0
130.00	3815.0
131.00	787.0
132.00	315.0
132.90	563.0
134.00	1510.0
135.00	3773.0
136.00	1400.0
137.00	1971.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

137.90	445.0
138.90	164.0
139.90	692.0
140.90	5373.0
142.00	1857.0
142.90	1285.0
144.00	422.0
144.90	407.0
146.00	1150.0
147.00	3102.0
148.00	7452.0
148.90	1652.0
150.00	343.0
151.10	880.0
151.80	349.0
152.90	1674.0
153.90	1416.0
155.00	3039.0
156.00	4498.0
157.10	789.0
157.90	972.0
158.90	792.0
160.00	1734.0
161.00	3116.0
162.00	821.0
162.90	309.0
163.80	257.0
165.00	1961.0
165.90	1904.0
167.00	12871.0
167.90	6963.0
168.90	1086.0
169.90	327.0
171.00	397.0
172.00	1092.0
172.90	1441.0
173.90	2568.0
175.00	4889.0
176.00	1352.0
176.90	2137.0
177.90	888.0
178.90	9528.0
180.00	6375.0
180.90	3511.0
182.00	426.0
183.90	659.0
185.00	4178.0
186.00	33632.0
187.00	10123.0
188.00	1111.0
188.90	2130.0
189.90	361.0
191.00	1541.0
191.90	3182.0
193.00	3639.0
193.90	750.0
195.00	473.0
196.00	6890.0
196.90	816.0
197.90	246336.0
198.90	15850.0
199.90	1249.0
201.50	1057.0
203.00	1753.0
204.00	9506.0
205.00	16184.0
206.00	68144.0
207.00	12956.0
208.00	2864.0
209.00	1429.0
210.10	1067.0
211.00	2950.0
211.90	557.0
213.00	230.0
214.90	679.0
216.00	1597.0
216.90	18808.0
218.00	2267.0
219.00	223.0
220.10	368.0
221.00	10072.0
221.90	2368.0
223.00	3943.0
224.00	39016.0
225.00	9412.0
225.90	1019.0
226.90	16696.0
228.00	2234.0
228.90	3462.0
229.90	384.0
230.90	1312.0
232.00	241.0
232.90	225.0
234.00	788.0
235.00	1128.0
235.90	717.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

236.90	1739.0
237.90	278.0
238.90	599.0
240.00	472.0
241.00	940.0
241.90	1985.0
243.00	2061.0
244.00	31104.0
245.00	4318.0
246.00	5974.0
247.00	1282.0
247.90	251.0
248.90	1270.0
249.90	200.0
251.00	430.0
251.90	280.0
253.00	641.0
253.90	790.0
255.00	154816.0
256.00	21680.0
256.90	1952.0
258.00	9645.0
258.90	1353.0
259.80	251.0
260.90	335.0
263.80	240.0
264.90	4005.0
265.90	453.0
266.80	209.0
271.00	332.0
271.90	438.0
272.90	4682.0
274.00	12074.0
275.00	67080.0
276.00	8315.0
276.90	5505.0
277.90	1229.0
280.90	606.0
281.90	253.0
282.90	861.0
284.00	550.0
284.90	1029.0
286.00	247.0
289.00	175.0
290.00	311.0
290.90	186.0
292.00	166.0
292.90	1659.0
294.00	373.0
294.90	296.0
296.00	22008.0
296.90	3109.0
297.90	259.0
300.80	312.0
301.90	376.0
303.00	2349.0
303.90	606.0
307.90	433.0
308.90	207.0
309.80	200.0
313.00	237.0
314.00	945.0
315.00	2467.0
316.00	1371.0
316.90	184.0
320.90	714.0
322.00	307.0
323.00	7353.0
324.00	1167.0
325.90	225.0
326.90	1424.0
327.90	573.0
332.00	301.0
332.90	539.0
334.00	4369.0
334.90	833.0
341.00	766.0
342.00	323.0
346.00	1662.0
347.00	371.0
350.90	172.0
351.90	1725.0
353.00	1373.0
354.00	2068.0
354.90	418.0
359.00	209.0
365.00	10416.0
365.90	1509.0
369.90	189.0
370.90	616.0
372.00	3308.0
373.00	925.0
383.00	1044.0
384.00	279.0
389.90	399.0
390.90	362.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113

392.00	172.0
400.90	298.0
402.00	1810.0
403.00	1910.0
404.00	884.0
405.00	228.0
421.00	1778.0
422.00	1860.0
423.00	12084.0
424.00	3666.0
424.90	517.0
441.00	41920.0
442.00	250304.0
443.00	47048.0
444.00	4244.0
444.90	378.0

Instrument :
BNA_M
ClientSampleId :
DFTPP113