



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM052721\
 Data File : BM030191.D
 Acq On : 27 May 2021 08:48
 Operator : CG/JU
 Sample : DFTPP42
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP042

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM052621.M
 Title : SVOA CALIBRATION
 Last Update : Thu May 27 10:27:00 2021



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.8	256807	PASS
68	69	0.00	2	1.7	4317	PASS
69	198	0.00	100	39.9	251306	PASS
70	69	0.00	2	0.6	1470	PASS
127	198	10	80	47.8	300672	PASS
197	198	0.00	2	0.5	3172	PASS
198	198	100	100	100.0	629333	PASS
199	198	5	9	6.9	43200	PASS
275	198	10	60	27.9	175467	PASS
365	198	1	100	3.8	23800	PASS
441	198	0.01	100	10.5	66104	PASS
442	198	50	100	68.7	432320	PASS
443	442	15	24	19.6	84731	PASS

m/z	Abundance
35.90	234.0
38.90	364.0
40.00	1053.0
41.00	402.0
42.00	183.0
43.00	441.0
44.00	2848.0
48.90	391.0
50.90	260.0
55.10	351.0
56.00	249.0
57.00	247.0
69.10	299.0
73.10	1441.0
74.10	203.0
74.90	177.0
77.00	194.0
78.00	240.0
79.10	182.0
81.00	269.0
84.00	329.0
86.10	155.0
91.10	217.0
95.70	460.0
118.90	215.0
133.00	590.0
134.90	202.0
147.00	609.0
148.10	201.0
148.90	173.0
166.10	180.0
167.00	389.0
176.80	258.0
191.00	601.0
192.90	568.0
207.00	5489.0
208.00	976.0
209.00	788.0
221.00	322.0
266.90	272.0
281.00	1133.0
282.00	439.0
283.00	255.0
341.90	185.0
354.90	232.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

m/z	Abundance
36.00	353.0
37.00	1052.0
38.00	3248.0
39.10	18000.0
40.00	2038.0
41.00	732.0
42.00	265.0
43.10	457.0
44.00	3186.0
45.00	784.0
47.00	199.0
48.00	201.0
49.10	2291.0
50.10	63784.0
51.10	241536.0
52.10	13250.0
53.00	560.0
55.00	1489.0
56.00	6736.0
57.00	15816.0
58.00	695.0
59.10	382.0
60.10	250.0
61.00	3011.0
62.00	3168.0
63.00	9112.0
64.00	1224.0
65.10	4912.0
66.00	315.0
66.90	289.0
68.00	3964.0
69.00	231680.0
70.00	1307.0
70.90	361.0
72.10	183.0
73.00	3297.0
74.00	23808.0
75.00	37120.0
76.10	13498.0
77.10	267264.0
78.10	18328.0
79.00	15807.0
80.00	12497.0
81.00	18664.0
82.00	4671.0
83.00	4383.0
83.90	809.0
85.00	3482.0
86.00	5278.0
87.10	2015.0
88.00	882.0
89.10	480.0
90.00	226.0
91.00	3978.0
92.00	4765.0
93.00	29096.0
94.00	2178.0
94.90	519.0
96.10	1772.0
96.90	437.0
98.00	21520.0
99.00	19984.0
100.00	1720.0
101.00	10277.0
102.00	633.0
103.00	3139.0
104.00	6824.0
105.00	5717.0
106.00	2067.0
107.00	74232.0
108.00	11166.0
109.00	2108.0
110.00	128776.0
111.00	20952.0
112.00	2364.0
113.00	835.0
113.80	221.0
114.90	483.0
116.00	3821.0
117.00	68088.0
118.00	5122.0
119.00	811.0
120.10	959.0
121.10	440.0
122.00	4633.0
123.00	7946.0
124.00	3701.0
125.00	2931.0
126.10	1180.0
127.00	261440.0
128.00	21232.0
129.00	102584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

130.00	9211.0
131.10	1899.0
132.10	775.0
132.90	781.0
134.00	2921.0
135.00	8933.0
136.00	3352.0
137.00	5092.0
137.90	815.0
139.00	565.0
139.90	1093.0
141.00	14264.0
142.00	3900.0
143.00	3113.0
144.00	894.0
145.00	570.0
146.00	2358.0
147.00	7864.0
148.00	17328.0
149.00	3954.0
150.00	954.0
151.00	1662.0
151.80	807.0
153.00	4593.0
154.00	3265.0
155.00	7875.0
156.00	11175.0
157.00	2293.0
158.00	2248.0
159.00	2116.0
160.00	4448.0
161.00	5589.0
162.00	1532.0
163.00	723.0
164.00	896.0
165.00	5122.0
166.00	4126.0
167.00	29152.0
168.00	19120.0
169.10	2939.0
170.10	631.0
171.00	1340.0
172.00	2445.0
173.00	3472.0
174.00	5502.0
175.00	11251.0
176.00	2835.0
177.00	5097.0
178.00	1737.0
179.00	21744.0
180.00	16424.0
181.00	7209.0
182.00	1370.0
183.00	612.0
184.00	1593.0
185.00	9216.0
186.00	70632.0
187.00	20440.0
188.00	2727.0
189.00	5343.0
190.00	894.0
191.00	2672.0
192.00	6669.0
193.10	8040.0
194.00	1525.0
195.00	862.0
196.00	15040.0
197.00	2163.0
198.00	525504.0
199.00	37256.0
200.00	2920.0
201.50	2565.0
203.00	3740.0
204.00	20216.0
205.00	34152.0
206.10	139968.0
207.10	25568.0
208.00	6336.0
209.00	2612.0
210.20	2501.0
211.00	5367.0
212.10	607.0
213.00	571.0
215.00	1332.0
216.00	3133.0
217.00	38712.0
218.00	5229.0
219.00	621.0
220.20	845.0
221.00	21208.0
221.80	4628.0
223.00	8616.0
224.00	79568.0
225.00	21280.0
226.00	2721.0
227.00	31416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

228.00	5152.0
229.00	6626.0
230.00	1011.0
231.00	2813.0
231.90	630.0
232.90	513.0
234.00	2021.0
234.90	2074.0
236.00	1550.0
237.00	4989.0
238.10	673.0
238.90	1256.0
240.00	1256.0
241.00	1683.0
242.00	4075.0
243.10	4602.0
244.10	61496.0
245.10	7860.0
246.00	12213.0
247.00	2602.0
247.90	559.0
249.00	2211.0
250.00	370.0
251.00	336.0
252.00	677.0
253.00	1061.0
254.10	1945.0
255.00	308800.0
256.00	44912.0
257.10	4124.0
258.00	17704.0
259.00	3251.0
260.00	448.0
261.00	506.0
264.00	494.0
265.00	7358.0
266.00	973.0
267.00	333.0
267.90	211.0
269.90	465.0
271.00	562.0
272.00	695.0
273.00	9121.0
274.00	23120.0
275.00	134784.0
276.00	17136.0
277.00	11103.0
278.00	2303.0
278.90	397.0
281.00	1342.0
282.10	832.0
283.00	1316.0
284.00	929.0
285.00	2239.0
285.90	356.0
289.00	423.0
290.00	436.0
290.90	298.0
292.00	574.0
293.00	2625.0
293.90	871.0
295.00	1012.0
296.00	39080.0
297.00	5563.0
297.90	256.0
301.00	564.0
301.90	945.0
303.00	4763.0
304.00	1078.0
304.90	236.0
308.00	556.0
309.00	516.0
310.10	445.0
310.90	175.0
312.00	162.0
313.00	381.0
314.00	1709.0
315.00	4426.0
316.00	1878.0
316.80	250.0
320.00	151.0
321.00	1222.0
322.10	470.0
323.10	11467.0
324.00	2252.0
324.90	308.0
326.00	243.0
327.00	2614.0
328.00	1236.0
329.00	269.0
332.00	862.0
333.00	1014.0
334.00	7413.0
335.00	1904.0
336.10	274.0
339.00	152.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

339.90	165.0
341.00	1565.0
342.00	318.0
343.00	166.0
346.00	2282.0
347.00	620.0
352.00	2772.0
353.00	2852.0
354.00	2800.0
355.10	849.0
356.90	154.0
359.00	326.0
365.00	16488.0
366.00	2785.0
366.90	206.0
370.00	419.0
370.90	787.0
372.00	5476.0
373.00	1281.0
376.90	194.0
383.00	1264.0
383.90	379.0
384.90	243.0
390.00	757.0
391.10	507.0
392.10	367.0
401.00	351.0
402.00	2335.0
403.00	3208.0
404.10	1637.0
405.00	551.0
414.90	261.0
421.00	2397.0
422.00	2777.0
423.00	18304.0
424.00	7313.0
425.00	1220.0
441.00	42224.0
442.10	278464.0
443.10	55200.0
444.10	5197.0
445.10	356.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

m/z	Abundance
35.00	383.0
37.00	1406.0
38.00	3742.0
39.10	22688.0
40.00	2037.0
41.10	1162.0
42.00	248.0
43.10	563.0
44.00	2990.0
45.00	699.0
45.90	170.0
48.00	233.0
49.00	2913.0
50.10	78480.0
51.10	309056.0
52.10	14868.0
53.00	730.0
55.00	1857.0
56.00	9271.0
57.00	21784.0
58.00	1034.0
59.00	436.0
59.90	270.0
61.00	4544.0
62.10	4890.0
63.00	11587.0
64.00	1747.0
65.00	5510.0
66.20	425.0
67.00	539.0
68.10	5046.0
69.00	301888.0
70.00	1934.0
71.00	353.0
72.00	187.0
73.00	4003.0
74.00	31584.0
75.00	48080.0
76.10	16016.0
77.10	348160.0
78.10	23808.0
79.00	21384.0
80.00	16928.0
81.00	25280.0
82.00	6616.0
83.00	5361.0
84.00	921.0
85.00	4514.0
86.00	7085.0
87.00	3179.0
87.90	1234.0
89.00	706.0
91.00	5641.0
92.00	6427.0
93.00	39728.0
94.00	2487.0
95.00	686.0
96.00	2045.0
97.00	725.0
98.00	29488.0
99.00	25440.0
100.00	2132.0
101.00	13981.0
102.00	799.0
103.00	4436.0
104.00	8289.0
105.00	8790.0
106.00	2374.0
107.00	103632.0
108.00	14608.0
109.00	2998.0
110.00	178240.0
111.00	27112.0
112.00	3326.0
113.00	1140.0
114.10	312.0
115.00	497.0
116.00	4951.0
117.00	92216.0
118.00	6985.0
119.00	596.0
120.00	1337.0
121.00	373.0
122.00	6288.0
123.00	10843.0
124.00	4544.0
125.00	4213.0
126.20	1713.0
127.00	358976.0
128.00	26352.0
129.00	137664.0
130.00	11658.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

131.00	2290.0
132.00	1401.0
133.00	1051.0
134.00	4403.0
135.00	10709.0
136.00	4657.0
137.00	6287.0
138.00	929.0
138.90	847.0
140.00	1731.0
141.00	17968.0
142.00	5557.0
143.00	4112.0
144.00	1218.0
145.00	957.0
146.00	3438.0
147.00	9271.0
148.00	24632.0
149.00	6090.0
150.00	1204.0
151.00	2297.0
151.90	994.0
153.00	5820.0
154.00	4994.0
155.00	11496.0
156.10	14944.0
157.10	2869.0
158.00	2847.0
159.00	3058.0
160.00	6215.0
161.00	9033.0
162.00	2412.0
163.00	857.0
164.00	1064.0
165.00	7107.0
166.00	5718.0
167.00	42152.0
168.00	24920.0
169.00	3874.0
170.00	1108.0
171.00	1342.0
172.00	3322.0
173.00	4852.0
174.00	9168.0
175.00	14728.0
176.00	3791.0
177.00	6712.0
178.00	2182.0
179.00	29880.0
180.00	21440.0
181.00	9667.0
182.10	1110.0
183.10	648.0
184.00	2305.0
185.00	13598.0
186.00	105432.0
187.00	28448.0
188.00	3640.0
189.00	7019.0
190.00	1019.0
191.00	3256.0
192.00	9690.0
193.00	10829.0
194.00	1993.0
195.00	1070.0
196.00	22648.0
197.10	4008.0
198.00	745408.0
199.00	51648.0
200.00	3797.0
201.40	3015.0
203.00	5869.0
204.00	30016.0
205.00	49664.0
206.10	208384.0
207.10	33592.0
208.00	8013.0
209.00	3014.0
210.10	2654.0
211.00	8373.0
212.00	1259.0
212.90	538.0
214.00	196.0
215.00	2181.0
216.00	4636.0
217.00	55776.0
218.00	6934.0
219.00	867.0
221.00	32072.0
221.90	7434.0
223.00	12972.0
224.10	116768.0
225.00	28656.0
226.10	4417.0
227.00	46864.0
228.00	6905.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

229.00	10494.0
230.00	1130.0
231.00	4049.0
232.00	804.0
232.90	903.0
234.00	2968.0
235.00	3335.0
236.00	2298.0
237.00	5610.0
238.00	752.0
238.90	1964.0
240.00	1548.0
241.00	2543.0
242.00	5670.0
243.00	6528.0
244.00	88400.0
245.00	11475.0
246.00	17632.0
247.00	3785.0
248.00	683.0
249.00	2912.0
249.90	791.0
250.90	729.0
252.00	983.0
253.00	2090.0
254.10	3045.0
255.00	456320.0
256.00	68800.0
257.00	5131.0
258.00	26608.0
259.00	5051.0
260.00	825.0
261.00	851.0
263.00	228.0
264.00	718.0
265.00	10776.0
266.00	1670.0
266.90	328.0
270.00	586.0
271.00	990.0
272.00	923.0
273.00	14438.0
274.00	33808.0
275.00	205248.0
276.00	27864.0
277.00	17856.0
278.00	3535.0
279.00	762.0
280.00	183.0
281.10	1381.0
282.00	681.0
283.00	1855.0
284.00	1366.0
285.00	3055.0
286.10	919.0
288.00	154.0
288.90	590.0
289.90	722.0
290.90	437.0
292.00	640.0
293.00	3827.0
294.10	1114.0
295.00	1273.0
296.00	60368.0
297.00	8329.0
298.00	375.0
299.00	163.0
301.10	625.0
302.00	1138.0
303.10	5975.0
304.10	1645.0
305.00	455.0
308.00	830.0
308.80	380.0
310.00	823.0
312.10	171.0
312.90	439.0
314.00	2397.0
315.00	6660.0
316.00	3638.0
316.90	680.0
318.80	196.0
319.90	233.0
321.00	1707.0
322.00	971.0
323.10	17000.0
324.00	3216.0
325.00	515.0
325.90	219.0
327.00	3600.0
328.00	2001.0
329.10	324.0
332.00	1525.0
333.00	1621.0
334.00	12078.0
335.00	3215.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

335.90	279.0
339.10	467.0
339.90	269.0
341.00	2477.0
342.00	529.0
346.00	4248.0
346.90	1035.0
349.90	161.0
350.90	254.0
352.00	5240.0
353.00	3715.0
354.10	6156.0
355.00	1278.0
356.10	235.0
359.00	346.0
361.10	165.0
363.80	180.0
365.00	26800.0
366.00	3816.0
366.90	264.0
370.00	521.0
370.90	1362.0
372.10	8466.0
373.00	2467.0
374.10	168.0
377.00	231.0
383.00	2530.0
384.00	706.0
385.00	336.0
390.00	1064.0
391.00	716.0
392.00	675.0
397.00	215.0
401.00	560.0
402.00	3393.0
403.00	5833.0
404.10	1911.0
405.00	638.0
406.00	172.0
415.00	311.0
421.00	3935.0
422.00	3986.0
423.10	29352.0
424.00	9761.0
425.00	1447.0
439.20	182.0
440.10	296.0
441.00	74376.0
442.10	488576.0
443.10	97808.0
444.00	8411.0
445.00	729.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

m/z	Abundance
36.00	406.0
37.00	998.0
38.00	2858.0
39.10	16912.0
40.00	1711.0
41.00	984.0
42.00	215.0
43.10	504.0
44.00	3472.0
44.90	583.0
47.00	201.0
49.00	2005.0
50.10	58744.0
51.10	220608.0
52.10	11526.0
53.00	538.0
55.00	1352.0
56.00	6416.0
57.00	14740.0
58.00	833.0
58.90	280.0
61.00	2895.0
62.00	3173.0
63.10	8591.0
64.00	1196.0
65.00	4133.0
66.00	510.0
66.90	389.0
68.00	3941.0
69.00	221248.0
70.00	1168.0
71.00	223.0
72.10	253.0
73.10	3226.0
74.00	23776.0
75.00	36864.0
76.10	11648.0
77.10	259392.0
78.10	17680.0
79.00	16816.0
80.00	12844.0
81.00	17904.0
82.00	4357.0
83.10	4155.0
84.00	989.0
85.00	3405.0
86.00	5762.0
87.00	2608.0
88.00	1019.0
88.90	386.0
89.10	375.0
91.00	4478.0
92.00	4630.0
93.00	29944.0
94.00	1744.0
95.10	588.0
96.00	1913.0
97.10	552.0
98.00	23360.0
99.00	19000.0
100.00	1811.0
101.00	9939.0
102.00	430.0
103.00	3331.0
104.00	6638.0
105.00	6704.0
106.00	1821.0
107.00	76616.0
108.00	12151.0
109.10	2523.0
110.00	132864.0
111.00	20984.0
112.00	2233.0
113.00	970.0
113.90	247.0
115.00	619.0
116.00	4166.0
117.00	72880.0
118.00	4696.0
119.00	715.0
120.00	841.0
121.00	296.0
122.00	5574.0
123.00	8239.0
124.00	3507.0
125.00	3386.0
126.20	1208.0
127.00	281600.0
128.00	21328.0
129.00	108104.0
130.00	9683.0
131.10	1774.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

132.00	917.0
133.00	858.0
134.00	3073.0
135.00	9637.0
136.00	4040.0
137.00	5115.0
138.10	1157.0
139.00	447.0
140.00	1585.0
141.00	13941.0
142.00	4728.0
143.00	2948.0
144.00	1080.0
145.00	898.0
146.00	2758.0
147.00	7656.0
148.00	19040.0
149.00	4203.0
150.00	1035.0
151.00	1973.0
151.70	1196.0
153.00	4919.0
154.00	3418.0
155.00	8452.0
156.00	12598.0
157.00	2154.0
158.00	2583.0
158.90	2169.0
160.00	4945.0
161.00	7141.0
162.00	1991.0
163.00	917.0
163.90	789.0
165.00	5809.0
166.00	4471.0
167.00	31904.0
168.00	19464.0
169.00	3080.0
170.00	1045.0
170.90	1505.0
172.00	3159.0
173.10	3728.0
174.00	6991.0
175.00	12178.0
176.00	3120.0
177.00	5747.0
177.90	1649.0
179.00	25616.0
180.00	16584.0
181.00	7805.0
181.90	1112.0
182.90	575.0
184.00	2148.0
185.00	11049.0
186.00	85528.0
187.00	24800.0
188.00	2892.0
189.00	5354.0
190.00	849.0
191.00	3151.0
192.00	8356.0
193.00	9423.0
194.00	2202.0
195.00	1167.0
196.00	18056.0
197.10	3344.0
198.00	617088.0
199.00	40696.0
200.00	3904.0
201.50	2432.0
203.00	4718.0
204.00	23584.0
205.10	40248.0
206.10	169728.0
207.00	27680.0
208.00	7505.0
209.00	2885.0
210.10	2617.0
211.00	6444.0
211.70	1262.0
212.90	616.0
214.00	223.0
215.00	1861.0
216.00	3763.0
217.00	46496.0
218.00	5860.0
218.90	640.0
221.00	24728.0
221.90	5941.0
223.00	11342.0
224.10	98328.0
225.10	24224.0
226.00	2693.0
227.00	41080.0
228.00	6086.0
229.00	8616.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

230.00	1073.0
231.10	3298.0
232.00	582.0
232.90	737.0
234.00	2532.0
235.00	3467.0
236.00	2102.0
237.00	4637.0
238.00	628.0
239.00	1712.0
240.00	1656.0
241.00	2407.0
242.00	5436.0
243.10	5057.0
244.10	77840.0
245.10	9978.0
246.00	14977.0
247.00	3542.0
248.00	647.0
249.00	3006.0
250.00	558.0
251.00	557.0
252.00	774.0
253.00	1300.0
254.10	2370.0
255.00	414592.0
256.00	57672.0
257.00	4977.0
258.00	23968.0
259.00	3848.0
259.90	625.0
261.00	754.0
263.00	159.0
264.10	737.0
265.00	10207.0
266.00	1407.0
267.00	299.0
270.00	563.0
271.00	1127.0
272.00	994.0
273.00	11923.0
274.00	33040.0
275.00	186368.0
276.00	24928.0
277.00	15496.0
278.00	2898.0
279.00	773.0
281.10	1512.0
282.00	778.0
283.00	2220.0
284.00	1326.0
285.10	2547.0
286.00	479.0
287.90	236.0
288.90	687.0
289.90	480.0
291.00	455.0
292.00	772.0
293.00	3484.0
293.90	1284.0
295.00	1227.0
296.00	52680.0
297.00	7268.0
297.90	532.0
301.10	633.0
302.00	1040.0
303.00	6043.0
304.00	1608.0
305.10	247.0
308.00	600.0
309.00	669.0
310.00	764.0
313.00	501.0
314.00	2705.0
315.00	6514.0
316.10	3397.0
316.90	558.0
320.00	174.0
321.00	1557.0
322.00	785.0
323.10	17888.0
324.00	3121.0
325.00	377.0
325.90	335.0
327.00	3355.0
328.00	2041.0
329.10	336.0
332.00	1293.0
333.00	1966.0
334.00	10712.0
335.00	2975.0
336.00	428.0
338.80	333.0
340.00	281.0
341.00	1950.0
342.10	562.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042

342.90	160.0
346.00	3815.0
347.00	754.0
350.90	468.0
352.00	5633.0
353.00	4294.0
354.00	5617.0
355.00	1239.0
359.00	424.0
360.90	160.0
364.10	277.0
365.00	28112.0
365.90	3748.0
367.00	248.0
370.00	592.0
371.00	1371.0
372.10	10309.0
373.10	2285.0
374.00	302.0
377.00	531.0
383.00	2582.0
384.00	971.0
385.00	347.0
390.00	1201.0
391.00	881.0
392.00	426.0
401.00	793.0
402.00	3870.0
403.10	5402.0
404.10	2126.0
404.90	443.0
415.00	440.0
421.00	4366.0
422.00	3986.0
423.00	30624.0
424.00	8093.0
425.00	1012.0
439.20	155.0
439.60	161.0
440.10	243.0
441.10	81712.0
442.10	529920.0
443.10	101184.0
444.10	8918.0
445.00	495.0

Instrument :
BNA_M
ClientSampleId :
DFTPP042