

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061722\
 Data File : BM035486.D
 Acq On : 17 Jun 2022 09:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020127

Quant Time: Jun 17 23:00:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:58:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	186870	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	799550	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	453834	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.203	188	905648	20.000	ng/u1	0.00
79) Chrysene-d12	21.391	240	1052910	20.000	ng/u1	0.00
88) Perylene-d12	23.732	264	1144843	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	37779	8.070	ng/uL	0.00
4) Pyridine-d5	3.687	84	226211	19.620	ng/u1	0.00
7) Phenol-d5	6.998	99	269246	18.800	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.145	67	185721	20.853	ng/u1	0.00
11) 2-Chlorophenol-d4	7.351	132	240368	19.377	ng/u1	0.00
15) 4-Methylphenol-d8	8.539	113	220064	18.355	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	108817	18.283	ng/u1	0.00
24) 2-Nitrophenol-d4	9.698	143	113393	17.669	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	196527	18.036	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	310370	18.103	ng/u1	0.00
46) Dimethylphthalate-d6	13.868	166	574581	18.613	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	721540	18.861	ng/u1	0.00
54) 4-Nitrophenol-d4	14.704	143	65051	13.718	ng/u1	0.00
60) Fluorene-d10	15.445	176	507148	18.973	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.592	200	83019	15.376	ng/u1	0.00
73) Anthracene-d10	17.298	188	761820	18.714	ng/u1	0.00
81) Pyrene-d10	19.597	212	964636	18.635	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.580	264	1054056	19.089	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	38694	8.153	ng/uL	91
5) Pyridine	3.704	79	242733	19.961	ng/u1#	82
6) Benzaldehyde	6.963	77	162383	26.471	ng/u1	92
8) Phenol	7.028	94	295804	19.356	ng/u1	90
10) Bis(2-Chloroethyl)ether	7.239	93	237308	19.882	ng/u1	93
12) 2-Chlorophenol	7.381	128	248512	19.344	ng/u1	95
13) 2-Methylphenol	8.275	108	214018	18.906	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.345	45	420370	23.627	ng/u1	96
16) Acetophenone	8.639	105	353943	19.472	ng/u1#	87
17) N-Nitroso-di-n-propyla...	8.622	70	178250	20.783	ng/u1#	85
18) 4-Methylphenol	8.604	108	225331	18.496	ng/u1	97
19) Hexachloroethane	8.886	117	104729	20.385	ng/u1	86
22) Nitrobenzene	9.016	77	262308	20.763	ng/u1	96
23) Isophorone	9.545	82	468300	19.598	ng/u1	98
25) 2-Nitrophenol	9.733	139	127848	18.302	ng/u1	90
26) 2,4-Dimethylphenol	9.798	107	257142	19.426	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.027	93	304155	19.628	ng/u1	99
29) 2,4-Dichlorophenol	10.280	162	200880	18.169	ng/u1	99
30) Naphthalene	10.657	128	769184	18.831	ng/u1	100
32) 4-Chloroaniline	10.780	127	306092	17.933	ng/u1	99
33) Hexachlorobutadiene	10.945	225	125631	18.059	ng/u1	98
34) Caprolactam	11.574	113	56240	15.107	ng/u1	88
35) 4-Chloro-3-methylphenol	11.927	107	213773	18.444	ng/u1	99
36) 2-Methylnaphthalene	12.274	142	494303	18.611	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061722\
 Data File : BM035486.D
 Acq On : 17 Jun 2022 09:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020127

Quant Time: Jun 17 23:00:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:58:49 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	501742	18.701	ng/ul#	97
39) 1,2,4,5-Tetrachloroben...	12.651	216	228159	17.954	ng/ul	98
40) Hexachlorocyclopentadiene	12.627	237	60058	12.852	ng/ul	99
41) 2,4,6-Trichlorophenol	12.904	196	142263	17.947	ng/ul	98
42) 2,4,5-Trichlorophenol	12.986	196	147784	17.969	ng/ul	97
43) 1,1'-Biphenyl	13.292	154	646305	19.044	ng/ul	98
44) 2-Chloronaphthalene	13.333	162	493431	19.024	ng/ul	98
45) 2-Nitroaniline	13.551	65	140890	21.070	ng/ul	87
47) Dimethylphthalate	13.915	163	586453	18.837	ng/ul	98
48) 2,6-Dinitrotoluene	14.045	165	109586	17.804	ng/ul#	84
50) Acenaphthylene	14.174	152	799640	18.947	ng/ul	99
51) 3-Nitroaniline	14.380	138	109256	17.324	ng/ul	93
52) Acenaphthene	14.515	153	515999	18.430	ng/ul	99
53) 2,4-Dinitrophenol	14.615	184	40948	12.072	ng/ul#	80
55) 4-Nitrophenol	14.721	109	52207	16.940	ng/ul	91
56) Dibenzofuran	14.857	168	730998	19.078	ng/ul	96
57) 2,4-Dinitrotoluene	14.839	165	162257	18.154	ng/ul#	74
58) 2,3,4,6-Tetrachlorophenol	15.092	232	117184	17.252	ng/ul#	82
59) Diethylphthalate	15.280	149	609462	19.319	ng/ul	99
61) Fluorene	15.504	166	584753	18.791	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.498	204	271356	18.579	ng/ul	97
63) 4-Nitroaniline	15.545	138	101778	18.066	ng/ul#	78
66) 4,6-Dinitro-2-methylph...	15.604	198	84518	15.730	ng/ul#	88
67) N-Nitrosodiphenylamine	15.715	169	492010	18.561	ng/ul	100
68) 4-Bromophenyl-phenylether	16.392	248	152612	17.572	ng/ul	92
69) Hexachlorobenzene	16.509	284	174543	17.371	ng/ul#	90
70) Atrazine	16.674	200	163685	17.488	ng/ul	98
71) Pentachlorophenol	16.868	266	76006	14.349	ng/ul	99
72) Phenanthrene	17.245	178	915372	18.772	ng/ul	99
74) Anthracene	17.333	178	918755	18.413	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	239415	17.318	ng/ul	98
76) Pentachlorobenzene	14.780	250	215741	17.552	ng/ul	99
77) Carbazole	17.615	167	830658	18.447	ng/ul	98
78) Di-n-butylphthalate	18.174	149	1076622	17.258	ng/ul	99
80) Fluoranthene	19.262	202	1171872	18.610	ng/ul	99
82) Pyrene	19.627	202	1236223	18.757	ng/ul	99
83) Butylbenzylphthalate	20.527	149	510278	18.009	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.315	252	394875	20.562	ng/ul	98
85) Benzo(a)anthracene	21.374	228	1208530	19.010	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.309	149	763052	18.147	ng/ul	100
87) Chrysene	21.427	228	1214630	19.166	ng/ul	99
89) Di-n-octyl phthalate	22.209	149	1348415	17.946	ng/ul	100
90) Benzo(b)fluoranthene	23.021	252	1294122	19.032	ng/ul	99
91) Benzo(k)fluoranthene	23.068	252	1255853	19.201	ng/ul	99
93) Benzo(a)pyrene	23.627	252	1304441	19.135	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.150	276	1528421	19.150	ng/ul	98
95) Dibenzo(a,h)anthracene	26.162	278	1305635	19.119	ng/ul	99
96) Benzo(g,h,i)perylene	26.891	276	1288313	18.834	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061722\
 Data File : BM035486.D
 Acq On : 17 Jun 2022 09:34
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020127

Quant Time: Jun 17 23:00:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:58:49 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.153 ng/uL
RT: 3.299 min Scan# 30
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion: 88 Resp: 38694
Ion Ratio Lower Upper
88 100
43 40.0 27.8 41.8
58 84.7 61.8 92.8



#5
Pyridine
Concen: 19.961 ng/uL
RT: 3.704 min Scan# 105
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion: 79 Resp: 242733
Ion Ratio Lower Upper
79 100
52 100.7 67.8 101.6
51 50.6 31.4 47.0









#6
Benzaldehyde
Concen: 26.471 ng/ul
RT: 6.963 min Scan# 659
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion: 77 Resp: 162383
Ion Ratio Lower Upper
77 100
105 97.7 84.6 127.0
106 91.8 80.3 120.5



#8
Phenol
Concen: 19.356 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion: 94 Resp: 295804
Ion Ratio Lower Upper
94 100
65 30.7 21.3 31.9
66 40.9 27.4 41.0





#10

Bis(2-Chloroethyl)ether

Concen: 19.882 ng/ul

RT: 7.239 min Scan# 706

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion: 93 Resp: 237308

Ion Ratio Lower Upper

93 100

63 87.9 64.0 96.0

95 31.9 25.8 38.6



#12

2-Chlorophenol

Concen: 19.344 ng/ul

RT: 7.381 min Scan# 730

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34




Tgt Ion: 128 Resp: 248512


Ion Ratio Lower Upper

128 100

64 50.0 36.9 55.3

130 33.6 25.6 38.4









#13
2-Methylphenol
Concen: 18.906 ng/ul
RT: 8.275 min Scan# 882
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:108 Resp: 214018
Ion Ratio Lower Upper
108 100
107 89.5 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.627 ng/ul
RT: 8.345 min Scan# 894
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion: 45 Resp: 420370
Ion Ratio Lower Upper
45 100
77 11.2 10.6 15.8
79 9.0 8.3 12.5









#16
Acetophenone
Concen: 19.472 ng/ul
RT: 8.639 min Scan# 94
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:105 Resp: 353943
Ion Ratio Lower Upper
105 100
77 83.1 59.8 89.6
51 39.2 22.1 33.1#



#17
N-Nitroso-di-n-propylamine
Concen: 20.783 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion: 70 Resp: 178250
Ion Ratio Lower Upper
70 100
42 77.6 49.6 74.4#
101 9.4 8.1 12.1
130 20.3 18.4 27.6









#18
4-Methylphenol
Concen: 18.496 ng/ul
RT: 8.604 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:108 Resp: 225331
Ion Ratio Lower Upper
108 100
107 116.6 90.3 135.5



#19
Hexachloroethane
Concen: 20.385 ng/ul
RT: 8.886 min Scan# 986
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:117 Resp: 104729
Ion Ratio Lower Upper
117 100
201 86.5 81.8 122.6
199 55.4 51.6 77.4









#22
Nitrobenzene
Concen: 20.763 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion: 77 Resp: 262308
Ion Ratio Lower Upper
77 100
123 43.9 37.4 56.0
65 14.5 10.8 16.2



#23
Isophorone
Concen: 19.598 ng/ul
RT: 9.545 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion: 82 Resp: 468300
Ion Ratio Lower Upper
82 100
95 7.2 5.5 8.3
138 17.7 15.3 22.9









#25
2-Nitrophenol
Concen: 18.302 ng/ul
RT: 9.733 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:139 Resp: 127848
Ion Ratio Lower Upper
139 100
65 57.9 39.8 59.6
109 31.0 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 19.426 ng/ul
RT: 9.798 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:107 Resp: 257142
Ion Ratio Lower Upper
107 100
121 51.5 42.7 64.1
122 87.7 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.628 ng/ul
 RT: 10.027 min Scan# 1180
 Delta R.T. -0.000 min
 Lab File: BM035486.D
 Acq: 17 Jun 2022 09:34




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020127


Tgt Ion: 93 Resp: 304155
 Ion Ratio Lower Upper
 93 100
 95 31.6 26.1 39.1
 123 12.0 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 18.169 ng/ul
 RT: 10.280 min Scan# 1223
 Delta R.T. 0.006 min
 Lab File: BM035486.D
 Acq: 17 Jun 2022 09:34

Tgt Ion: 162 Resp: 200880
 Ion Ratio Lower Upper
 162 100
 164 64.5 51.8 77.6
 98 38.0 29.5 44.3









#30
Naphthalene
Concen: 18.831 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:128 Resp: 769184
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.2 10.6 16.0



#32
4-Chloroaniline
Concen: 17.933 ng/ul
RT: 10.780 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:127 Resp: 306092
Ion Ratio Lower Upper
127 100
129 32.1 25.3 37.9









#33
Hexachlorobutadiene
Concen: 18.059 ng/ul
RT: 10.945 min Scan# 1336
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:225 Resp: 125631
Ion Ratio Lower Upper
225 100
223 61.2 50.0 75.0
227 64.7 49.9 74.9



#34
Caprolactam
Concen: 15.107 ng/ul
RT: 11.574 min Scan# 1443
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:113 Resp: 56240
Ion Ratio Lower Upper
113 100
55 200.1 150.2 225.4
56 148.8 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 18.444 ng/ul
RT: 11.927 min Scan# 1503
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:107 Resp: 213773
Ion Ratio Lower Upper
107 100
144 26.9 20.3 30.5
142 80.0 64.4 96.6



#36
2-Methylnaphthalene
Concen: 18.611 ng/ul
RT: 12.274 min Scan# 1562
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:142 Resp: 494303
Ion Ratio Lower Upper
142 100
141 87.8 70.7 106.1









#37
1-Methylnaphthalene
Concen: 18.701 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:142 Resp: 501742
Ion Ratio Lower Upper
142 100
141 91.2 75.2 112.8
116 4.0 2.6 3.8#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 17.954 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:216 Resp: 228159
Ion Ratio Lower Upper
216 100
214 79.5 62.4 93.6
179 22.4 17.3 25.9
108 20.2 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 12.852 ng/ul
RT: 12.627 min Scan# 1622
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:237 Resp: 60058
Ion Ratio Lower Upper
237 100
235 60.1 48.4 72.6
272 12.8 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 17.947 ng/ul
RT: 12.904 min Scan# 1669
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:196 Resp: 142263
Ion Ratio Lower Upper
196 100
198 94.5 77.8 116.6
200 32.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 17.969 ng/ul
RT: 12.986 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:196 Resp: 147784
Ion Ratio Lower Upper
196 100
198 99.0 76.8 115.2
200 31.8 24.9 37.3



#43
1,1'-Biphenyl
Concen: 19.044 ng/ul
RT: 13.292 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:154 Resp: 646305
Ion Ratio Lower Upper
154 100
153 39.4 32.4 48.6
76 13.5 10.1 15.1









#44
2-Chloronaphthalene
Concen: 19.024 ng/ul
RT: 13.333 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:162 Resp: 493431
Ion Ratio Lower Upper
162 100
164 32.5 25.4 38.0
127 41.3 31.6 47.4



#45
2-Nitroaniline
Concen: 21.070 ng/ul
RT: 13.551 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion: 65 Resp: 140890
Ion Ratio Lower Upper
65 100
92 62.4 54.5 81.7
138 96.9 91.2 136.8









#47
Dimethylphthalate
Concen: 18.837 ng/ul
RT: 13.915 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:163 Resp: 586453
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.5 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 17.804 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:165 Resp: 109586
Ion Ratio Lower Upper
165 100
89 66.8 42.5 63.7#
121 24.7 17.2 25.8









#50
Acenaphthylene
Concen: 18.947 ng/ul
RT: 14.174 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:152 Resp: 799640
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1
153 12.7 10.6 15.8



#51
3-Nitroaniline
Concen: 17.324 ng/ul
RT: 14.380 min Scan# 1920
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:138 Resp: 109256
Ion Ratio Lower Upper
138 100
108 11.6 8.1 12.1
92 121.3 90.4 135.6









#52
Acenaphthene
Concen: 18.430 ng/ul
RT: 14.515 min Scan# 1943
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:153 Resp: 515999
Ion Ratio Lower Upper
153 100
152 47.8 37.1 55.7
154 88.8 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 12.072 ng/ul
RT: 14.615 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:184 Resp: 40948
Ion Ratio Lower Upper
184 100
63 88.2 54.2 81.4#
154 73.1 49.3 73.9









#55
4-Nitrophenol
Concen: 16.940 ng/ul
RT: 14.721 min Scan# 1977
Delta R.T. 0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:109 Resp: 52207
Ion Ratio Lower Upper
109 100
139 134.3 124.9 187.3
65 154.9 122.2 183.4



#56
Dibenzofuran
Concen: 19.078 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:168 Resp: 730998
Ion Ratio Lower Upper
168 100
139 39.8 30.1 45.1





#57

2,4-Dinitrotoluene

Concen: 18.154 ng/ul

RT: 14.839 min Scan# 1998

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion:165 Resp: 162257

Ion Ratio Lower Upper

165 100

63 63.0 35.9 53.9#

182 2.6 2.5 3.7



#58

2,3,4,6-Tetrachlorophenol

Concen: 17.252 ng/ul

RT: 15.092 min Scan# 2041

Delta R.T. -0.006 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Tgt Ion:232 Resp: 117184




Ion Ratio Lower Upper


232 100

131 61.9 36.7 55.1#

130 3.3 1.7 2.5#

166 36.3 24.4 36.6









#59
Diethylphthalate
Concen: 19.319 ng/ul
RT: 15.280 min Scan# 2073
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:149 Resp: 609462
Ion Ratio Lower Upper
149 100
177 20.5 16.8 25.2
150 12.5 10.2 15.4



#61
Fluorene
Concen: 18.791 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:166 Resp: 584753
Ion Ratio Lower Upper
166 100
165 96.1 77.2 115.8
167 13.6 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 18.579 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:204 Resp: 271356
Ion Ratio Lower Upper
204 100
206 31.5 25.9 38.9
141 66.6 50.6 75.8



#63
4-Nitroaniline
Concen: 18.066 ng/ul
RT: 15.545 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:138 Resp: 101778
Ion Ratio Lower Upper
138 100
92 54.4 39.2 58.8
108 90.6 51.6 77.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 15.730 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:198 Resp: 84518
Ion Ratio Lower Upper
198 100
182 6.5 4.2 6.4#
77 32.1 19.9 29.9#



#67
N-Nitrosodiphenylamine
Concen: 18.561 ng/ul
RT: 15.715 min Scan# 2147
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:169 Resp: 492010
Ion Ratio Lower Upper
169 100
168 67.3 53.8 80.6
167 36.4 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 17.572 ng/ul
RT: 16.392 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:248 Resp: 152612
Ion Ratio Lower Upper
248 100
250 98.4 77.3 115.9
141 89.2 59.8 89.8



#69
Hexachlorobenzene
Concen: 17.371 ng/ul
RT: 16.509 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:284 Resp: 174543
Ion Ratio Lower Upper
284 100
142 46.0 28.6 42.8#
249 31.7 24.6 37.0





#70

Atrazine

Concen: 17.488 ng/ul

RT: 16.674 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion:200 Resp: 163685

Ion Ratio Lower Upper

200 100

173 26.2 21.3 31.9

215 47.3 39.0 58.6



#71

Pentachlorophenol

Concen: 14.349 ng/ul

RT: 16.868 min Scan# 2343

Delta R.T. -0.006 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34




Tgt Ion:266 Resp: 76006


Ion Ratio Lower Upper

266 100

264 62.5 49.5 74.3

268 62.2 51.0 76.4









#72
Phenanthrene
Concen: 18.772 ng/ul
RT: 17.245 min Scan# 2407
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:178 Resp: 915372
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.2
176 19.5 14.9 22.3



#74
Anthracene
Concen: 18.413 ng/ul
RT: 17.333 min Scan# 2422
Delta R.T. -0.006 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:178 Resp: 918755
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.2
176 18.4 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 17.318 ng/uL
RT: 13.257 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:216 Resp: 239415
Ion Ratio Lower Upper
216 100
214 78.0 60.9 91.3
218 48.1 37.5 56.3



#76
Pentachlorobenzene
Concen: 17.552 ng/uL
RT: 14.780 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:250 Resp: 215741
Ion Ratio Lower Upper
250 100
248 62.6 50.4 75.6
252 63.3 51.1 76.7





#77

Carbazole

Concen: 18.447 ng/ul

RT: 17.615 min Scan# 2470

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion:167 Resp: 830658

Ion Ratio Lower Upper

167 100

166 21.3 17.4 26.0

139 13.7 10.1 15.1



#78

Di-n-butylphthalate

Concen: 17.258 ng/ul

RT: 18.174 min Scan# 2565

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34




Tgt Ion:149 Resp: 1076622


Ion Ratio Lower Upper

149 100

150 8.7 7.3 10.9

104 4.7 3.7 5.5





#80

Fluoranthene

Concen: 18.610 ng/ul

RT: 19.262 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion:202 Resp: 1171872

Ion Ratio Lower Upper

202 100

101 12.9 10.1 15.1

100 10.3 7.8 11.6



#82

Pyrene

Concen: 18.757 ng/ul

RT: 19.627 min Scan# 2812

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34




Tgt Ion:202 Resp: 1236223


Ion Ratio Lower Upper

202 100

101 15.3 11.9 17.9

100 12.7 10.2 15.4









#83
Butylbenzylphthalate
Concen: 18.009 ng/ul
RT: 20.527 min Scan# 2965
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:149 Resp: 510278
Ion Ratio Lower Upper
149 100
91 69.6 53.2 79.8
206 19.6 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 20.562 ng/ul
RT: 21.315 min Scan# 3099
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:252 Resp: 394875
Ion Ratio Lower Upper
252 100
254 63.9 52.4 78.6
126 13.5 10.8 16.2









#85
Benzo(a)anthracene
Concen: 19.010 ng/ul
RT: 21.374 min Scan# 3109
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:228 Resp: 1208530
Ion Ratio Lower Upper
228 100
229 19.8 15.6 23.4
226 27.2 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.147 ng/ul
RT: 21.309 min Scan# 3098
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:149 Resp: 763052
Ion Ratio Lower Upper
149 100
167 26.7 21.5 32.3
279 3.9 3.4 5.2





#87

Chrysene

Concen: 19.166 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion:228 Resp: 1214630

Ion Ratio Lower Upper

228 100

226 29.1 23.7 35.5

229 19.5 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 17.946 ng/ul




RT: 22.209 min Scan# 3251


Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Tgt Ion:149 Resp: 1348415









#90
Benzo(b)fluoranthene
Concen: 19.032 ng/ul
RT: 23.021 min Scan# 3389
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34




Instrument :
BNA_M
ClientSampleId :
SSTD020127


Tgt Ion:252 Resp: 1294122
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 12.7 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 19.201 ng/ul
RT: 23.068 min Scan# 3397
Delta R.T. -0.000 min
Lab File: BM035486.D
Acq: 17 Jun 2022 09:34

Tgt Ion:252 Resp: 1255853
Ion Ratio Lower Upper
252 100
253 22.4 17.4 26.0
125 13.0 10.6 15.8





#93

Benzo(a)pyrene

Concen: 19.135 ng/ul

RT: 23.627 min Scan# 3492

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34

Instrument :

BNA_M

ClientSampleId :

SSTD020127





Tgt Ion:252 Resp: 1304441

Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 14.6 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.150 ng/ul

RT: 26.150 min Scan# 3921

Delta R.T. -0.000 min

Lab File: BM035486.D

Acq: 17 Jun 2022 09:34




Tgt Ion:276 Resp: 1528421


Ion Ratio Lower Upper

276 100

138 30.5 25.6 38.4

277 25.1 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 19.119 ng/ul
 RT: 26.162 min Scan# 3923
 Delta R.T. -0.000 min
 Lab File: BM035486.D
 Acq: 17 Jun 2022 09:34

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020127

Tgt Ion:278 Resp: 1305635
 Ion Ratio Lower Upper
 278 100
 139 19.6 16.2 24.2
 279 23.9 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 18.834 ng/ul
 RT: 26.891 min Scan# 4047
 Delta R.T. -0.006 min
 Lab File: BM035486.D
 Acq: 17 Jun 2022 09:34

Tgt Ion:276 Resp: 1288313
 Ion Ratio Lower Upper
 276 100
 138 26.7 21.3 31.9
 277 23.7 18.6 28.0

