

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061722\
 Data File : BM035493.D
 Acq On : 17 Jun 2022 13:58
 Operator : CG/JU
 Sample : SP5919
 Misc : SFAM SPIKE
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP5919

Quant Time: Jun 17 23:01:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:58:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.816	152	190229	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	802995	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	468525	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.204	188	927893	20.000	ng/u1	0.00
79) Chrysene-d12	21.392	240	928674	20.000	ng/u1	0.00
88) Perylene-d12	23.733	264	1001488	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/uL	
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	0.000	99	0d	0.000	ng/u1	
9) Bis-(2-Chloroethyl)eth...	0.000	67	0d	0.000	ng/u1	
11) 2-Chlorophenol-d4	0.000	132	0d	0.000	ng/u1	
15) 4-Methylphenol-d8	0.000	113	0d	0.000	ng/u1	
21) Nitrobenzene-d5	0.000	128	0d	0.000	ng/u1	
24) 2-Nitrophenol-d4	0.000	143	0	0.000	ng/u1	
28) 2,4-Dichlorophenol-d3	0.000	165	0d	0.000	ng/u1	
31) 4-Chloroaniline-d4	0.000	131	0d	0.000	ng/u1	
46) Dimethylphthalate-d6	0.000	166	0d	0.000	ng/u1	
49) Acenaphthylene-d8	0.000	160	0d	0.000	ng/u1	
54) 4-Nitrophenol-d4	0.000	143	0d	0.000	ng/u1	
60) Fluorene-d10	0.000	176	0d	0.000	ng/u1	
65) 4,6-Dinitro-2-methylph...	0.000	200	0d	0.000	ng/u1	
73) Anthracene-d10	0.000	188	0d	0.000	ng/u1	
81) Pyrene-d10	0.000	212	0d	0.000	ng/u1	
92) Benzo(a)pyrene-d12	0.000	264	0d	0.000	ng/u1	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	144550	29.920	ng/uL	97
5) Pyridine	3.699	79	932238	75.308	ng/u1	94
6) Benzaldehyde	6.958	77	781541	125.153	ng/u1	97
8) Phenol	7.028	94	1231356	79.153	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.240	93	940008	77.364	ng/u1	99
12) 2-Chlorophenol	7.381	128	1055867	80.735	ng/u1	99
13) 2-Methylphenol	8.269	108	921945	80.006	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.352	45	1428155	78.854	ng/u1	100
16) Acetophenone	8.646	105	1462467	79.034	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.628	70	713390	81.709	ng/u1	100
18) 4-Methylphenol	8.605	108	991283	79.931	ng/u1	99
19) Hexachloroethane	8.887	117	416645	79.667	ng/u1	95
22) Nitrobenzene	9.016	77	1037549	81.773	ng/u1	97
23) Isophorone	9.552	82	1972501	82.194	ng/u1	100
25) 2-Nitrophenol	9.734	139	568661	81.059	ng/u1	96
26) 2,4-Dimethylphenol	9.805	107	1076553	80.980	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.028	93	1250345	80.344	ng/u1	100
29) 2,4-Dichlorophenol	10.275	162	906684	81.655	ng/u1	99
30) Naphthalene	10.663	128	3213729	78.339	ng/u1	100
32) 4-Chloroaniline	10.781	127	1283145	74.854	ng/u1	98
33) Hexachlorobutadiene	10.946	225	532387	76.201	ng/u1	99
34) Caprolactam	11.587	113	303921	81.290	ng/u1	96
35) 4-Chloro-3-methylphenol	11.922	107	971388	83.452	ng/u1	97
36) 2-Methylnaphthalene	12.275	142	2092279	78.437	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061722\
 Data File : BM035493.D
 Acq On : 17 Jun 2022 13:58
 Operator : CG/JU
 Sample : SP5919
 Misc : SFAM SPIKE
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP5919

Quant Time: Jun 17 23:01:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:58:49 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.498	142	2101986	78.009	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.651	216	1023246	77.993	ng/ul	99
40) Hexachlorocyclopentadiene	12.628	237	480003	99.494	ng/ul	98
41) 2,4,6-Trichlorophenol	12.898	196	679384	83.021	ng/ul	99
42) 2,4,5-Trichlorophenol	12.981	196	725438	85.441	ng/ul	99
43) 1,1'-Biphenyl	13.293	154	2700249	77.069	ng/ul	99
44) 2-Chloronaphthalene	13.334	162	2122173	79.255	ng/ul	99
45) 2-Nitroaniline	13.545	65	603526	87.426	ng/ul	98
47) Dimethylphthalate	13.922	163	2557659	79.576	ng/ul	100
48) 2,6-Dinitrotoluene	14.045	165	565814	89.042	ng/ul	95
50) Acenaphthylene	14.181	152	3283962	75.370	ng/ul	99
51) 3-Nitroaniline	14.375	138	609304	93.582	ng/ul	98
52) Acenaphthene	14.522	153	2297613	79.493	ng/ul	98
53) 2,4-Dinitrophenol	14.598	184	308456	88.086	ng/ul	92
55) 4-Nitrophenol	14.710	109	256464	80.607	ng/ul	95
56) Dibenzofuran	14.857	168	3031080	76.625	ng/ul	99
57) 2,4-Dinitrotoluene	14.839	165	793040	85.949	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.092	232	605456	86.339	ng/ul	98
59) Diethylphthalate	15.287	149	2563506	78.712	ng/ul	99
61) Fluorene	15.504	166	2478233	77.142	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.498	204	1160134	76.942	ng/ul	99
63) 4-Nitroaniline	15.539	138	598316	102.875	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.604	198	430845	78.264	ng/ul#	94
67) N-Nitrosodiphenylamine	15.716	169	2121067	78.099	ng/ul	99
68) 4-Bromophenyl-phenylether	16.392	248	719743	80.887	ng/ul	98
69) Hexachlorobenzene	16.516	284	807410	78.430	ng/ul	99
70) Atrazine	16.675	200	760989	79.354	ng/ul	98
71) Pentachlorophenol	16.863	266	446949	82.353	ng/ul	99
72) Phenanthrene	17.245	178	3811879	76.297	ng/ul	99
74) Anthracene	17.339	178	3914939	76.581	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	1063296	75.071	ng/ul	98
76) Pentachlorobenzene	14.781	250	986729	78.354	ng/ul	99
77) Carbazole	17.610	167	3606183	78.166	ng/ul	99
78) Di-n-butylphthalate	18.175	149	4383154	68.578	ng/ul	99
80) Fluoranthene	19.263	202	4288649	77.218	ng/ul	97
82) Pyrene	19.627	202	4454695	76.632	ng/ul	99
83) Butylbenzylphthalate	20.527	149	2037068	81.512	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.310	252	1647298	97.253	ng/ul	98
85) Benzo(a)anthracene	21.374	228	4353887	77.646	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.310	149	2942987	79.355	ng/ul	100
87) Chrysene	21.427	228	4162685	74.470	ng/ul	99
89) Di-n-octyl phthalate	22.210	149	5320722	80.948	ng/ul	100
90) Benzo(b)fluoranthene	23.027	252	4793588	80.586	ng/ul	99
91) Benzo(k)fluoranthene	23.074	252	4409677	77.071	ng/ul	99
93) Benzo(a)pyrene	23.633	252	4173032	69.976	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.168	276	5622334	80.527	ng/ul	99
95) Dibenzo(a,h)anthracene	26.174	278	4653868	77.904	ng/ul	99
96) Benzo(g,h,i)perylene	26.909	276	4705037	78.631	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM061722\
 Data File : BM035493.D
 Acq On : 17 Jun 2022 13:58
 Operator : CG/JU
 Sample : SP5919
 Misc : SFAM SPIKE
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SP5919

Quant Time: Jun 17 23:01:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM061622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 17 01:58:49 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 29.920 ng/uL
RT: 3.299 min Scan# 30
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion: 88 Resp: 144550
Ion Ratio Lower Upper
88 100
43 32.4 27.8 41.8
58 79.0 61.8 92.8



#5
Pyridine
Concen: 75.308 ng/uL
RT: 3.699 min Scan# 104
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion: 79 Resp: 932238
Ion Ratio Lower Upper
79 100
52 77.9 67.8 101.6
51 37.5 31.4 47.0









#6
Benzaldehyde
Concen: 125.153 ng/ul
RT: 6.958 min Scan# 658
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion: 77 Resp: 781541
Ion Ratio Lower Upper
77 100
105 102.7 84.6 127.0
106 97.1 80.3 120.5



#8
Phenol
Concen: 79.153 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion: 94 Resp: 1231356
Ion Ratio Lower Upper
94 100
65 26.7 21.3 31.9
66 34.8 27.4 41.0





#10

Bis(2-Chloroethyl)ether

Concen: 77.364 ng/ul

RT: 7.240 min Scan# 706

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58

Instrument :

BNA_M

ClientSampleId :

SP5919





Tgt Ion: 93 Resp: 940008

Ion Ratio Lower Upper

93 100

63 78.6 64.0 96.0

95 32.2 25.8 38.6



#12

2-Chlorophenol

Concen: 80.735 ng/ul

RT: 7.381 min Scan# 730

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58




Tgt Ion:128 Resp: 1055867


Ion Ratio Lower Upper

128 100

64 45.2 36.9 55.3

130 32.6 25.6 38.4









#13
2-Methylphenol
Concen: 80.006 ng/ul
RT: 8.269 min Scan# 881
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:108 Resp: 921945
Ion Ratio Lower Upper
108 100
107 91.3 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 78.854 ng/ul
RT: 8.352 min Scan# 895
Delta R.T. 0.006 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion: 45 Resp: 1428155
Ion Ratio Lower Upper
45 100
77 13.3 10.6 15.8
79 10.2 8.3 12.5









#16
 Acetophenone
 Concen: 79.034 ng/ul
 RT: 8.646 min Scan# 944
 Delta R.T. 0.000 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58




Instrument :
 BNA_M
 ClientSampleId :
 SP5919


Tgt Ion:105 Resp: 1462467
 Ion Ratio Lower Upper
 105 100
 77 75.3 59.8 89.6
 51 27.8 22.1 33.1



#17
 N-Nitroso-di-n-propylamine
 Concen: 81.709 ng/ul
 RT: 8.628 min Scan# 942
 Delta R.T. 0.000 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58

Tgt Ion: 70 Resp: 713390
 Ion Ratio Lower Upper
 70 100
 42 61.7 49.6 74.4
 101 9.9 8.1 12.1
 130 23.3 18.4 27.6









#18
4-Methylphenol
Concen: 79.931 ng/ul
RT: 8.605 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:108 Resp: 991283
Ion Ratio Lower Upper
108 100
107 113.7 90.3 135.5



#19
Hexachloroethane
Concen: 79.667 ng/ul
RT: 8.887 min Scan# 986
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:117 Resp: 416645
Ion Ratio Lower Upper
117 100
201 96.7 81.8 122.6
199 60.3 51.6 77.4









#22
Nitrobenzene
Concen: 81.773 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion: 77 Resp: 1037549
Ion Ratio Lower Upper
77 100
123 49.1 37.4 56.0
65 13.5 10.8 16.2



#23
Isophorone
Concen: 82.194 ng/ul
RT: 9.552 min Scan# 1099
Delta R.T. 0.006 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion: 82 Resp: 1972501
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 19.0 15.3 22.9









#25
2-Nitrophenol
Concen: 81.059 ng/ul
RT: 9.734 min Scan# 1130
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:139 Resp: 568661
Ion Ratio Lower Upper
139 100
65 46.6 39.8 59.6
109 26.4 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 80.980 ng/ul
RT: 9.805 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:107 Resp: 1076553
Ion Ratio Lower Upper
107 100
121 54.1 42.7 64.1
122 88.7 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 80.344 ng/ul
 RT: 10.028 min Scan# 1180
 Delta R.T. 0.000 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58




Instrument :
 BNA_M
 ClientSampleId :
 SP5919


Tgt Ion: 93 Resp: 1250345
 Ion Ratio Lower Upper
 93 100
 95 32.9 26.1 39.1
 123 11.7 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 81.655 ng/ul
 RT: 10.275 min Scan# 1222
 Delta R.T. 0.000 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58

Tgt Ion:162 Resp: 906684
 Ion Ratio Lower Upper
 162 100
 164 63.9 51.8 77.6
 98 35.8 29.5 44.3





#30

Naphthalene

Concen: 78.339 ng/ul

RT: 10.663 min Scan# 1287

Delta R.T. 0.006 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58

Instrument :

BNA_M

ClientSampleId :

SP5919





Tgt Ion:128 Resp: 3213729

Ion Ratio Lower Upper

128 100

129 10.8 8.8 13.2

127 13.1 10.6 16.0



#32

4-Chloroaniline

Concen: 74.854 ng/ul

RT: 10.781 min Scan# 1308

Delta R.T. -0.005 min

Lab File: BM035493.D




Acq: 17 Jun 2022 13:58


Tgt Ion:127 Resp: 1283145

Ion Ratio Lower Upper

127 100

129 32.5 25.3 37.9









#33
Hexachlorobutadiene
Concen: 76.201 ng/ul
RT: 10.946 min Scan# 1336
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:225 Resp: 532387
Ion Ratio Lower Upper
225 100
223 63.1 50.0 75.0
227 63.4 49.9 74.9



#34
Caprolactam
Concen: 81.290 ng/ul
RT: 11.587 min Scan# 1445
Delta R.T. 0.012 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:113 Resp: 303921
Ion Ratio Lower Upper
113 100
55 177.6 150.2 225.4
56 127.2 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 83.452 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:107 Resp: 971388
Ion Ratio Lower Upper
107 100
144 26.5 20.3 30.5
142 83.4 64.4 96.6



#36
2-Methylnaphthalene
Concen: 78.437 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:142 Resp: 2092279
Ion Ratio Lower Upper
142 100
141 89.7 70.7 106.1









#37
1-Methylnaphthalene
Concen: 78.009 ng/ul
RT: 12.498 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:142 Resp: 2101986
Ion Ratio Lower Upper
142 100
141 93.1 75.2 112.8
116 3.4 2.6 3.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 77.993 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:216 Resp: 1023246
Ion Ratio Lower Upper
216 100
214 79.1 62.4 93.6
179 21.3 17.3 25.9
108 17.0 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 99.494 ng/ul
RT: 12.628 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:237 Resp: 480003
Ion Ratio Lower Upper
237 100
235 61.6 48.4 72.6
272 13.0 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 83.021 ng/ul
RT: 12.898 min Scan# 1668
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:196 Resp: 679384
Ion Ratio Lower Upper
196 100
198 96.1 77.8 116.6
200 30.6 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 85.441 ng/ul
RT: 12.981 min Scan# 1682
Delta R.T. -0.011 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:196 Resp: 725438
Ion Ratio Lower Upper
196 100
198 96.8 76.8 115.2
200 30.9 24.9 37.3



#43
1,1'-Biphenyl
Concen: 77.069 ng/ul
RT: 13.293 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:154 Resp: 2700249
Ion Ratio Lower Upper
154 100
153 41.1 32.4 48.6
76 12.8 10.1 15.1









#44
2-Chloronaphthalene
Concen: 79.255 ng/ul
RT: 13.334 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:162 Resp: 2122173
Ion Ratio Lower Upper
162 100
164 33.0 25.4 38.0
127 39.7 31.6 47.4



#45
2-Nitroaniline
Concen: 87.426 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion: 65 Resp: 603526
Ion Ratio Lower Upper
65 100
92 71.0 54.5 81.7
138 112.6 91.2 136.8









#47
Dimethylphthalate
Concen: 79.576 ng/ul
RT: 13.922 min Scan# 1842
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:163 Resp: 2557659
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.1 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 89.042 ng/ul
RT: 14.045 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:165 Resp: 565814
Ion Ratio Lower Upper
165 100
89 49.3 42.5 63.7
121 19.9 17.2 25.8





#50

Acenaphthylene

Concen: 75.370 ng/ul

RT: 14.181 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58

Instrument :

BNA_M

ClientSampleId :

SP5919





Tgt Ion:152 Resp: 3283962

Ion Ratio Lower Upper

152 100

151 19.7 16.1 24.1

153 13.0 10.6 15.8



#51

3-Nitroaniline

Concen: 93.582 ng/ul

RT: 14.375 min Scan# 1919

Delta R.T. -0.005 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58




Tgt Ion:138 Resp: 609304


Ion Ratio Lower Upper

138 100

108 10.9 8.1 12.1

92 110.5 90.4 135.6





#52

Acenaphthene

Concen: 79.493 ng/ul

RT: 14.522 min Scan# 1944

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58

Instrument :

BNA_M

ClientSampleId :

SP5919





Tgt Ion:153 Resp: 2297613

Ion Ratio Lower Upper

153 100

152 46.1 37.1 55.7

154 87.0 71.3 106.9



#53

2,4-Dinitrophenol

Concen: 88.086 ng/ul

RT: 14.598 min Scan# 1957

Delta R.T. -0.011 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58




Tgt Ion:184 Resp: 308456


Ion Ratio Lower Upper

184 100

63 59.7 54.2 81.4

154 56.9 49.3 73.9









#55
4-Nitrophenol
Concen: 80.607 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:109 Resp: 256464
Ion Ratio Lower Upper
109 100
139 160.1 124.9 187.3
65 142.7 122.2 183.4



#56
Dibenzofuran
Concen: 76.625 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:168 Resp: 3031080
Ion Ratio Lower Upper
168 100
139 36.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 85.949 ng/ul
RT: 14.839 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:165 Resp: 793040
Ion Ratio Lower Upper
165 100
63 43.4 35.9 53.9
182 2.8 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 86.339 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:232 Resp: 605456
Ion Ratio Lower Upper
232 100
131 46.5 36.7 55.1
130 2.5 1.7 2.5
166 31.8 24.4 36.6









#59
Diethylphthalate
Concen: 78.712 ng/ul
RT: 15.287 min Scan# 2073
Delta R.T. 0.006 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:149 Resp: 2563506
Ion Ratio Lower Upper
149 100
177 21.3 16.8 25.2
150 12.6 10.2 15.4



#61
Fluorene
Concen: 77.142 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:166 Resp: 2478233
Ion Ratio Lower Upper
166 100
165 97.9 77.2 115.8
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 76.942 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:204 Resp: 1160134
Ion Ratio Lower Upper
204 100
206 32.2 25.9 38.9
141 62.6 50.6 75.8



#63
4-Nitroaniline
Concen: 102.875 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:138 Resp: 598316
Ion Ratio Lower Upper
138 100
92 47.4 39.2 58.8
108 66.1 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 78.264 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:198 Resp: 430845
Ion Ratio Lower Upper
198 100
182 3.6 4.2 6.4#
77 22.0 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 78.099 ng/ul
RT: 15.716 min Scan# 2147
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:169 Resp: 2121067
Ion Ratio Lower Upper
169 100
168 67.7 53.8 80.6
167 35.6 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 80.887 ng/ul
RT: 16.392 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:248 Resp: 719743
Ion Ratio Lower Upper
248 100
250 98.1 77.3 115.9
141 76.4 59.8 89.8



#69
Hexachlorobenzene
Concen: 78.430 ng/ul
RT: 16.516 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:284 Resp: 807410
Ion Ratio Lower Upper
284 100
142 35.3 28.6 42.8
249 30.4 24.6 37.0









#70
 Atrazine
 Concen: 79.354 ng/ul
 RT: 16.675 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58




Instrument :
 BNA_M
 ClientSampleId :
 SP5919


Tgt Ion:200 Resp: 760989
 Ion Ratio Lower Upper
 200 100
 173 27.7 21.3 31.9
 215 47.5 39.0 58.6



#71
 Pentachlorophenol
 Concen: 82.353 ng/ul
 RT: 16.863 min Scan# 2342
 Delta R.T. -0.011 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58

Tgt Ion:266 Resp: 446949
 Ion Ratio Lower Upper
 266 100
 264 63.3 49.5 74.3
 268 64.4 51.0 76.4





#72

Phenanthrene

Concen: 76.297 ng/ul

RT: 17.245 min Scan# 2407

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58

Instrument :

BNA_M

ClientSampleId :

SP5919





Tgt Ion:178 Resp: 3811879

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.2

176 18.9 14.9 22.3



#74

Anthracene

Concen: 76.581 ng/ul

RT: 17.339 min Scan# 2423

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58




Tgt Ion:178 Resp: 3914939


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 18.3 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 75.071 ng/uL

RT: 13.257 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58

Instrument :

BNA_M

ClientSampleId :

SP5919





Tgt Ion:216 Resp: 1063296

Ion Ratio Lower Upper

216 100

214 78.1 60.9 91.3

218 46.7 37.5 56.3



#76

Pentachlorobenzene

Concen: 78.354 ng/uL

RT: 14.781 min Scan# 1988

Delta R.T. 0.000 min

Lab File: BM035493.D

Acq: 17 Jun 2022 13:58




Tgt Ion:250 Resp: 986729


Ion Ratio Lower Upper

250 100

248 63.5 50.4 75.6

252 64.2 51.1 76.7









#77
 Carbazole
 Concen: 78.166 ng/ul
 RT: 17.610 min Scan# 2470
 Delta R.T. -0.005 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58




Instrument :
 BNA_M
 ClientSampleId :
 SP5919


Tgt Ion:167 Resp: 3606183
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.4 26.0
 139 13.1 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 68.578 ng/ul
 RT: 18.175 min Scan# 2565
 Delta R.T. 0.000 min
 Lab File: BM035493.D
 Acq: 17 Jun 2022 13:58

Tgt Ion:149 Resp: 4383154
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.3 10.9
 104 4.8 3.7 5.5









#80
Fluoranthene
Concen: 77.218 ng/u1
RT: 19.263 min Scan# 21
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:202 Resp: 4288649
Ion Ratio Lower Upper
202 100
101 13.8 10.1 15.1
100 11.0 7.8 11.6



#82
Pyrene
Concen: 76.632 ng/u1
RT: 19.627 min Scan# 2812
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:202 Resp: 4454695
Ion Ratio Lower Upper
202 100
101 15.3 11.9 17.9
100 13.0 10.2 15.4









#83
Butylbenzylphthalate
Concen: 81.512 ng/ul
RT: 20.527 min Scan# 2965
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:149 Resp: 2037068
Ion Ratio Lower Upper
149 100
91 67.0 53.2 79.8
206 20.3 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 97.253 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.005 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:252 Resp: 1647298
Ion Ratio Lower Upper
252 100
254 64.1 52.4 78.6
126 13.9 10.8 16.2









#85
Benzo(a)anthracene
Concen: 77.646 ng/ul
RT: 21.374 min Scan# 3109
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:228 Resp: 4353887
Ion Ratio Lower Upper
228 100
229 19.8 15.6 23.4
226 27.3 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 79.355 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:149 Resp: 2942987
Ion Ratio Lower Upper
149 100
167 27.0 21.5 32.3
279 4.3 3.4 5.2









#87
Chrysene
Concen: 74.470 ng/ul
RT: 21.427 min Scan# 3118
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:228 Resp: 4162685
Ion Ratio Lower Upper
228 100
226 30.0 23.7 35.5
229 20.0 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 80.948 ng/ul
RT: 22.210 min Scan# 3251
Delta R.T. 0.000 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:149 Resp: 5320722









#90
Benzo(b)fluoranthene
Concen: 80.586 ng/u1
RT: 23.027 min Scan# 31
Delta R.T. 0.006 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:252 Resp: 4793588
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 12.7 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 77.071 ng/u1
RT: 23.074 min Scan# 3398
Delta R.T. 0.006 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:252 Resp: 4409677
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 12.5 10.6 15.8









#93
Benzo(a)pyrene
Concen: 69.976 ng/ul
RT: 23.633 min Scan# 3493
Delta R.T. 0.006 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58




Instrument :
BNA_M
ClientSampleId :
SP5919


Tgt Ion:252 Resp: 4173032
Ion Ratio Lower Upper
252 100
253 22.2 17.3 25.9
125 14.6 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 80.527 ng/ul
RT: 26.168 min Scan# 3924
Delta R.T. 0.018 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:276 Resp: 5622334
Ion Ratio Lower Upper
276 100
138 32.8 25.6 38.4
277 25.4 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 77.904 ng/ul
RT: 26.174 min Scan# 3923
Delta R.T. 0.012 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Instrument :
BNA_M
ClientSampleId :
SP5919

Tgt Ion:278 Resp: 4653868
Ion Ratio Lower Upper
278 100
139 20.6 16.2 24.2
279 24.1 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 78.631 ng/ul
RT: 26.909 min Scan# 4050
Delta R.T. 0.012 min
Lab File: BM035493.D
Acq: 17 Jun 2022 13:58

Tgt Ion:276 Resp: 4705037
Ion Ratio Lower Upper
276 100
138 27.2 21.3 31.9
277 23.4 18.6 28.0

