

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070621\
 Data File : BM030865.D
 Acq On : 07 Jul 2021 20:14
 Operator : CG/JU
 Sample : M2854-17
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGDW1

Manual Integrations
 APPROVED

mohammad
 7/8/2021 12:38:08 PM

Quant Time: Jul 08 00:47:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM070621.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Jul 06 14:32:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.763	152	82715	20.000	ng/ul	0.00
20) Naphthalene-d8	10.545	136	333345	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.392	164	208545	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.133	188	391399	20.000	ng/ul	0.00
79) Chrysene-d12	21.303	240	292523	20.000	ng/ul	0.00
88) Perylene-d12	23.550	264	302411	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.263	96	4017m	1.898	ng/uL	0.00
4) Pyridine-d5	3.716	84	14447	2.610	ng/ul	0.04
7) Phenol-d5	6.934	99	67744	9.373	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.098	67	49337	10.598	ng/ul	0.00
11) 2-Chlorophenol-d4	7.298	132	56095	10.116	ng/ul	0.00
15) 4-Methylphenol-d8	8.463	113	52216	9.015	ng/ul	0.00
21) Nitrobenzene-d5	8.922	128	24866	10.029	ng/ul	0.00
24) 2-Nitrophenol-d4	9.639	143	24601	9.069	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.175	165	47858	9.167	ng/ul	0.00
31) 4-Chloroaniline-d4	10.692	131	67836	9.150	ng/ul	0.00
46) Dimethylphthalate-d6	13.804	166	172951	11.846	ng/ul	0.00
49) Acenaphthylene-d8	14.086	160	196601	10.660	ng/ul	0.00
54) 4-Nitrophenol-d4	14.621	143	8891m	3.499	ng/ul	0.02
60) Fluorene-d10	15.386	176	148366	11.541	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.521	200	8454	3.368	ng/ul	0.00
73) Anthracene-d10	17.227	188	209065	11.462	ng/ul	0.00
81) Pyrene-d10	19.515	212	180972	11.217	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.403	264	118568	7.596	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070621\
Data File : BM030865.D
Acq On : 07 Jul 2021 20:14
Operator : CG/JU
Sample : M2854-17
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGDW1

Manual Integrations
APPROVED

mohammad
7/8/2021 12:38:08 PM

Quant Time: Jul 08 00:47:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM070621.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Jul 06 14:32:50 2021
Response via : Initial Calibration

