

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071525\
 Data File : BM050450.D
 Acq On : 15 Jul 2025 11:30
 Operator : RC/JU
 Sample : PB168803TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB168803TB

Quant Time: Jul 15 12:21:58 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 08 18:32:25 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	501643	20.000	ng	0.00
21) Naphthalene-d8	10.651	136	1780218	20.000	ng	-0.01
39) Acenaphthene-d10	14.480	164	1111817	20.000	ng	-0.01
64) Phenanthrene-d10	17.215	188	2247046	20.000	ng	-0.01
76) Chrysene-d12	21.439	240	2361654	20.000	ng	-0.02
86) Perylene-d12	24.462	264	2470157	20.000	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.440	112	3793999	130.186	ng	0.00
7) Phenol-d6	7.022	99	4664219	126.959	ng	0.00
23) Nitrobenzene-d5	9.010	82	2747456	78.738	ng	-0.01
42) 2,4,6-Tribromophenol	15.962	330	2007845	148.634	ng	-0.01
45) 2-Fluorobiphenyl	13.110	172	7159065	79.518	ng	-0.01
79) Terphenyl-d14	19.827	244	11514653	85.785	ng	-0.02


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071525\
Data File : BM050450.D
Acq On : 15 Jul 2025 11:30
Operator : RC/JU
Sample : PB168803TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168803TB

Quant Time: Jul 15 12:21:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 08 18:32:25 2025
Response via : Initial Calibration



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM071525\
Data File : BM050450.D
Acq On : 15 Jul 2025 11:30
Operator : RC/JU
Sample : PB168803TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB168803TB

Quant Time: Jul 15 12:21:58 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 08 18:32:25 2025
Response via : Initial Calibration

