

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM072821\
 Data File : BM031205.D
 Acq On : 28 Jul 2021 11:03
 Operator : CG/JU
 Sample : SST0.417
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST0.4017

Quant Time: Jul 28 11:38:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM072821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 28 11:36:18 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	1051	0.40	ng/ul	0.00
4) Naphthalene-d8	10.38	136	3640	0.40	ng/ul	0.00
9) Acenaphthene-d10	14.25	164	2284	0.40	ng/ul	0.00
13) Phenanthrene-d10	16.99	188	4847	0.40	ng/ul	0.00
17) Chrysene-d12	21.18	240	4214	0.40	ng/ul	0.00
23) Perylene-d12	23.35	264	3960	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.20	96	429	0.39	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.98	152	2285	0.36	ng/ul	0.00
18) Fluoranthene-d10	19.02	212	5133	0.40	ng/ul	0.00
Target Compounds				Ovalue		
2) 1,4-Dioxane	3.23	88	466	0.432	ng/ul	100
5) Naphthalene	10.43	128	4086	0.376	ng/ul	100
7) 2-Methylnaphthalene	12.05	142	2797	0.357	ng/ul	100
8) 1-Methylnaphthalene	12.27	142	2729	0.357	ng/ul	100
10) Acenaphthylene	13.97	152	3602	0.386	ng/ul	100
11) Acenaphthene	14.31	153	3038	0.368	ng/ul	100
12) Fluorene	15.30	166	3547	0.365	ng/ul	100
14) Pentachlorophenol	16.65	266	577	0.344	ng/ul	100
15) Phenanthrene	17.03	178	5751	0.374	ng/ul	100
16) Anthracene	17.12	178	5351	0.371	ng/ul	100
19) Fluoranthene	19.05	202	6563	0.387	ng/ul	100
20) Pyrene	19.42	202	6686	0.386	ng/ul	100
21) Benzo(a)anthracene	21.17	228	6274	0.373	ng/ul	100
22) Chrysene	21.22	228	6587	0.362	ng/ul	100
24) Benzo(b)fluoranthene	22.71	252	6627	0.356	ng/ul	100
25) Benzo(k)fluoranthene	22.75	252	6838	0.362	ng/ul	100
26) Benzo(a)pyrene	23.26	252	5736	0.356	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	25.50	276	7644	0.363	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.51	278	6161	0.364	ng/ul	100
29) Benzo(g,h,i)perylene	26.15	276	6564	0.367	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM072821\
Data File : BM031205.D
Acq On : 28 Jul 2021 11:03
Operator : CG/JU
Sample : SST00.417
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4017

Quant Time: Jul 28 11:38:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM072821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 28 11:36:18 2021
Response via : Initial Calibration

